

US010721785B2

(12) **United States Patent**  
**Al-Ali**

(10) **Patent No.:** **US 10,721,785 B2**  
(45) **Date of Patent:** **Jul. 21, 2020**

(54) **PATIENT-WORN WIRELESS  
PHYSIOLOGICAL SENSOR WITH PAIRING  
FUNCTIONALITY**

(71) Applicant: **Masimo Corporation**, Irvine, CA (US)

(72) Inventor: **Ammar Al-Ali**, San Juan Capistrano,  
CA (US)

(73) Assignee: **Masimo Corporation**, Irvine, CA (US)

(\*) Notice: Subject to any disclaimer, the term of this  
patent is extended or adjusted under 35  
U.S.C. 154(b) by 0 days.

(21) Appl. No.: **15/874,652**

(22) Filed: **Jan. 18, 2018**

(65) **Prior Publication Data**

US 2018/0213583 A1 Jul. 26, 2018

**Related U.S. Application Data**

(60) Provisional application No. 62/505,762, filed on May  
12, 2017, provisional application No. 62/447,836,  
filed on Jan. 18, 2017.

(51) **Int. Cl.**

**H04W 76/14** (2018.01)  
**A61B 5/0408** (2006.01)  
**A61B 5/1455** (2006.01)  
**H04L 29/06** (2006.01)  
**A61B 5/00** (2006.01)  
**H04W 12/00** (2009.01)  
**A61B 5/053** (2006.01)  
**A61B 5/01** (2006.01)  
**A61B 5/0476** (2006.01)  
**A61B 5/021** (2006.01)  
**A61B 5/113** (2006.01)  
**H04W 8/00** (2009.01)  
**H04W 4/80** (2018.01)  
**H04W 84/20** (2009.01)

(52) **U.S. Cl.**

CPC ..... **H04W 76/14** (2018.02); **A61B 5/0015**  
(2013.01); **A61B 5/04085** (2013.01); **A61B**  
**5/14551** (2013.01); **A61B 5/6833** (2013.01);

**H04L 63/18** (2013.01); **H04W 12/003**  
(2019.01); **A61B 5/01** (2013.01); **A61B 5/021**  
(2013.01); **A61B 5/0476** (2013.01); **A61B**  
**5/053** (2013.01); **A61B 5/113** (2013.01); **A61B**  
**2562/08** (2013.01); **H04W 4/80** (2018.02);  
**H04W 8/005** (2013.01); **H04W 12/00522**  
(2019.01); **H04W 84/20** (2013.01)

(58) **Field of Classification Search**

CPC ..... H04W 76/14; H04W 8/005; H04W 4/08;  
H04W 84/20; A61B 5/04085; A61B  
5/14551; A61B 5/6833; A61B 5/01; A61B  
5/053; A61B 5/113; A61B 5/0015; A61B  
2562/08  
USPC ..... 370/329  
See application file for complete search history.

(56) **References Cited**

U.S. PATENT DOCUMENTS

4,960,128 A 10/1990 Gordon et al.  
4,964,408 A 10/1990 Hink et al.  
(Continued)

*Primary Examiner* — Yemane Mesfin


*Assistant Examiner* — Intekhaab A Siddiquee

(74) *Attorney, Agent, or Firm* — Knobbe, Martens, Olson  
& Bear, LLP

(57) **ABSTRACT**

Systems and methods described herein use pairing to asso-  
ciate a wireless sensor with a patient monitoring device such  
as a bedside patient monitor or a mobile device. A signal  
emitted by a patient monitoring device can be detected by a  
wireless sensor. The wireless sensor can be associated with  
the detected signal and pair the wireless sensor with the  
patient monitoring device. The wireless sensor can be con-  
figured to enter into a patient parameter sensing mode of  
operation after the association of the wireless sensor with the  
patient monitoring device.

**16 Claims, 12 Drawing Sheets**





(56)

## References Cited

## U.S. PATENT DOCUMENTS

5,041,187	A	8/1991	Hink et al.	6,253,097	B1	6/2001	Aronow et al.
5,069,213	A	12/1991	Polczynski	6,256,523	B1	7/2001	Diab et al.
5,163,438	A	11/1992	Gordon et al.	6,263,222	B1	7/2001	Diab et al.
5,319,355	A	6/1994	Russek	6,278,522	B1	8/2001	Lepper, Jr. et al.
5,337,744	A	8/1994	Branigan	6,280,213	B1	8/2001	Tobler et al.
5,341,805	A	8/1994	Stavriddi et al.	6,285,896	B1	9/2001	Tobler et al.
D353,195	S	12/1994	Savage et al.	6,301,493	B1	10/2001	Marro et al.
D353,196	S	12/1994	Savage et al.	6,308,089	B1	10/2001	von der Ruhr et al.
5,377,676	A	1/1995	Vari et al.	6,317,627	B1	11/2001	Ennen et al.
D359,546	S	6/1995	Savage et al.	6,321,100	B1	11/2001	Parker
5,431,170	A	7/1995	Mathews	6,325,761	B1	12/2001	Jay
D361,840	S	8/1995	Savage et al.	6,334,065	B1	12/2001	Al-Ali et al.
D362,063	S	9/1995	Savage et al.	6,343,224	B1	1/2002	Parker
5,452,717	A	9/1995	Branigan et al.	6,349,228	B1	2/2002	Kiani et al.
D363,120	S	10/1995	Savage et al.	6,360,114	B1	3/2002	Diab et al.
5,456,252	A	10/1995	Vari et al.	6,368,283	B1	4/2002	Xu et al.
5,479,934	A	1/1996	Imran	6,371,921	B1	4/2002	Caro et al.
5,482,036	A	1/1996	Diab et al.	6,377,829	B1	4/2002	Al-Ali
5,490,505	A	2/1996	Diab et al.	6,388,240	B2	5/2002	Schulz et al.
5,494,043	A	2/1996	O'Sullivan et al.	6,397,091	B2	5/2002	Diab et al.
5,533,511	A	7/1996	Kaspari et al.	6,430,437	B1	8/2002	Marro
5,534,851	A	7/1996	Russek	6,430,525	B1	8/2002	Weber et al.
5,561,275	A	10/1996	Savage et al.	6,463,311	B1	10/2002	Diab
5,562,002	A	10/1996	Lalin	6,470,199	B1	10/2002	Kopotic et al.
5,590,649	A	1/1997	Caro et al.	6,501,975	B2	12/2002	Diab et al.
5,602,924	A	2/1997	Durand et al.	6,505,059	B1	1/2003	Kollias et al.
5,632,272	A	5/1997	Diab et al.	6,515,273	B2	2/2003	Al-Ali
5,638,816	A	6/1997	Kiani-Azarbayjany et al.	6,519,487	B1	2/2003	Parker
5,638,818	A	6/1997	Diab et al.	6,525,386	B1	2/2003	Mills et al.
5,645,440	A	7/1997	Tobler et al.	6,526,300	B1	2/2003	Kiani et al.
5,685,299	A	11/1997	Diab et al.	6,541,756	B2	4/2003	Schulz et al.
D393,830	S	4/1998	Tobler et al.	6,542,764	B1	4/2003	Al-Ali et al.
5,743,262	A	4/1998	Lepper, Jr. et al.	6,580,086	B1	6/2003	Schulz et al.
5,758,644	A	6/1998	Diab et al.	6,584,336	B1	6/2003	Ali et al.
5,760,910	A	6/1998	Lepper, Jr. et al.	6,595,316	B2	7/2003	Cybulski et al.
5,769,785	A	6/1998	Diab et al.	6,597,932	B2	7/2003	Tian et al.
5,782,757	A	7/1998	Diab et al.	6,597,933	B2	7/2003	Kiani et al.
5,785,659	A	7/1998	Caro et al.	6,606,511	B1	8/2003	Ali et al.
5,791,347	A	8/1998	Flaherty et al.	6,632,181	B2	10/2003	Flaherty et al.
5,810,734	A	9/1998	Caro et al.	6,639,668	B1	10/2003	Trepagnier
5,823,950	A	10/1998	Diab et al.	6,640,116	B2	10/2003	Diab
5,830,131	A	11/1998	Caro et al.	6,643,530	B2	11/2003	Diab et al.
5,833,618	A	11/1998	Caro et al.	6,650,917	B2	11/2003	Diab et al.
5,860,919	A	1/1999	Kiani-Azarbayjany et al.	6,654,624	B2	11/2003	Diab et al.
5,890,929	A	4/1999	Mills et al.	6,658,276	B2	12/2003	Kiani et al.
5,904,654	A	5/1999	Wohlmann et al.	6,661,161	B1	12/2003	Lanzo et al.
5,919,134	A	7/1999	Diab	6,671,531	B2	12/2003	Al-Ali et al.
5,934,925	A	8/1999	Tobler et al.	6,678,543	B2	1/2004	Diab et al.
5,940,182	A	8/1999	Lepper, Jr. et al.	6,684,090	B2	1/2004	Ali et al.
5,987,343	A	11/1999	Kinast	6,684,091	B2	1/2004	Parker
5,995,855	A	11/1999	Kiani et al.	6,697,656	B1	2/2004	Al-Ali
5,997,343	A	12/1999	Mills et al.	6,697,657	B1	2/2004	Shehada et al.
6,002,952	A	12/1999	Diab et al.	6,697,658	B2	2/2004	Al-Ali
6,011,986	A	1/2000	Diab et al.	RE38,476	E	3/2004	Diab et al.
6,027,452	A	2/2000	Flaherty et al.	6,699,194	B1	3/2004	Diab et al.
6,036,642	A	3/2000	Diab et al.	6,714,804	B2	3/2004	Al-Ali et al.
6,045,509	A	4/2000	Caro et al.	RE38,492	E	4/2004	Diab et al.
6,067,462	A	5/2000	Diab et al.	6,721,582	B2	4/2004	Trepagnier et al.
6,081,735	A	6/2000	Diab et al.	6,721,585	B1	4/2004	Parker
6,088,607	A	7/2000	Diab et al.	6,725,075	B2	4/2004	Al-Ali
6,110,522	A	8/2000	Lepper, Jr. et al.	6,728,560	B2	4/2004	Kollias et al.
6,124,597	A	9/2000	Shehada	6,735,459	B2	5/2004	Parker
6,128,521	A	10/2000	Marro et al.	6,745,060	B2	6/2004	Diab et al.
6,129,675	A	10/2000	Jay	6,760,607	B2	7/2004	Al-Ali
6,144,868	A	11/2000	Parker	6,770,028	B1	8/2004	Ali et al.
6,151,516	A	11/2000	Kiani-Azarbayjany et al.	6,771,994	B2	8/2004	Kiani et al.
6,152,754	A	11/2000	Gerhardt et al.	6,792,300	B1	9/2004	Diab et al.
6,157,850	A	12/2000	Diab et al.	6,813,511	B2	11/2004	Diab et al.
6,165,005	A	12/2000	Mills et al.	6,816,741	B2	11/2004	Diab
6,184,521	B1	2/2001	Coffin, IV et al.	6,822,564	B2	11/2004	Al-Ali
6,206,830	B1	3/2001	Diab et al.	6,826,419	B2	11/2004	Diab et al.
6,229,856	B1	5/2001	Diab et al.	6,830,711	B2	12/2004	Mills et al.
6,232,609	B1	5/2001	Snyder et al.	6,850,787	B2	2/2005	Weber et al.
6,236,872	B1	5/2001	Diab et al.	6,850,788	B2	2/2005	Al-Ali
6,241,683	B1	6/2001	Macklem et al.	6,852,083	B2	2/2005	Caro et al.
				6,861,639	B2	3/2005	Al-Ali
				6,898,452	B2	5/2005	Al-Ali et al.
				6,920,345	B2	7/2005	Al-Ali et al.
				6,931,268	B1	8/2005	Kiani-Azarbayjany et al.



(56)

## References Cited

## U.S. PATENT DOCUMENTS

6,934,570 B2	8/2005	Kiani et al.	7,496,391 B2	2/2009	Diab et al.
6,939,305 B2	9/2005	Flaherty et al.	7,496,393 B2	2/2009	Diab et al.
6,943,348 B1	9/2005	Coffin, IV	D587,657 S	3/2009	Al-Ali et al.
6,950,687 B2	9/2005	Al-Ali	7,499,741 B2	3/2009	Diab et al.
6,961,598 B2	11/2005	Diab	7,499,835 B2	3/2009	Weber et al.
6,970,792 B1	11/2005	Diab	7,500,950 B2	3/2009	Al-Ali et al.
6,979,812 B2	12/2005	Al-Ali	7,509,154 B2	3/2009	Diab et al.
6,985,764 B2	1/2006	Mason et al.	7,509,494 B2	3/2009	Ai-Ali
6,993,371 B2	1/2006	Kiani et al.	7,510,849 B2	3/2009	Schurman et al.
6,996,427 B2	2/2006	Ali et al.	7,526,328 B2	4/2009	Diab et al.
6,999,904 B2	2/2006	Weber et al.	7,530,942 B1	5/2009	Diab
7,003,338 B2	2/2006	Weber et al.	7,530,949 B2	5/2009	Al Ali et al.
7,003,339 B2	2/2006	Diab et al.	7,530,955 B2	5/2009	Diab et al.
7,015,451 B2	3/2006	Dalke et al.	7,563,110 B2	7/2009	Ai-Ali et al.
7,024,233 B2	4/2006	Ali et al.	7,596,398 B2	9/2009	Al-Ali et al.
7,027,849 B2	4/2006	Al-Ali	7,618,375 B2	11/2009	Flaherty
7,030,749 B2	4/2006	Ai-Ali	D606,659 S	12/2009	Kiani et al.
7,039,449 B2	5/2006	Al-Ali	7,647,083 B2	1/2010	Al-Ali et al.
7,041,060 B2	5/2006	Flaherty et al.	D609,193 S	2/2010	Ai-Ali et al.
7,044,918 B2	5/2006	Diab	D614,305 S	4/2010	Al-Ali et al.
7,048,687 B1	5/2006	Reuss et al.	RE41,317 E	5/2010	Parker
7,067,893 B2	6/2006	Mills et al.	7,729,733 B2	6/2010	Al-Ali et al.
7,096,052 B2	8/2006	Mason et al.	7,734,320 B2	6/2010	Ai-Ali
7,096,054 B2	8/2006	Abdul-Hafiz et al.	7,761,127 B2	7/2010	Ai-Ali et al.
7,132,641 B2	11/2006	Schulz et al.	7,761,128 B2	7/2010	Al-Ali et al.
7,142,901 B2	11/2006	Kiani et al.	7,764,982 B2	7/2010	Dalke et al.
7,149,561 B2	12/2006	Diab	D621,516 S	8/2010	Kiani et al.
7,186,966 B2	3/2007	Al-Ali	7,791,155 B2	9/2010	Diab
7,190,261 B2	3/2007	Ai-Ali	7,801,581 B2	9/2010	Diab
7,215,984 B2	5/2007	Diab	7,822,452 B2	10/2010	Schurman et al.
7,215,986 B2	5/2007	Diab	RE41,912 E	11/2010	Parker
7,221,971 B2	5/2007	Diab	7,844,313 B2	11/2010	Kiani et al.
7,225,006 B2	5/2007	Al-Ali et al.	7,844,314 B2	11/2010	Ai-Ali
7,225,007 B2	5/2007	Al-Ali	7,844,315 B2	11/2010	Al-Ali
RE39,672 E	6/2007	Shehada et al.	7,865,222 B2	1/2011	Weber et al.
7,239,905 B2	7/2007	Kiani-Azarbayjany et al.	7,873,497 B2	1/2011	Weber et al.
7,245,953 B1	7/2007	Parker	7,880,606 B2	2/2011	Al-Ali
7,254,429 B2	8/2007	Schurman et al.	7,880,626 B2	2/2011	Ai-Ali et al.
7,254,431 B2	8/2007	Al-Ali	7,891,355 B2	2/2011	Al-Ali et al.
7,254,433 B2	8/2007	Diab et al.	7,894,868 B2	2/2011	Al-Ali et al.
7,254,434 B2	8/2007	Schulz et al.	7,899,507 B2	3/2011	Al-Ali et al.
7,272,425 B2	9/2007	Al-Ali	7,899,518 B2	3/2011	Trepagnier et al.
7,274,955 B2	9/2007	Kiani et al.	7,904,132 B2	3/2011	Weber et al.
D554,263 S	10/2007	Al-Ali	7,909,772 B2	3/2011	Popov et al.
7,280,858 B2	10/2007	Al-Ali et al.	7,910,875 B2	3/2011	Ai-Ali
7,289,835 B2	10/2007	Mansfield et al.	7,919,713 B2	4/2011	Ai-Ali et al.
7,292,883 B2	11/2007	De Felice et al.	7,937,128 B2	5/2011	Al-Ali
7,295,866 B2	11/2007	Al-Ali	7,937,129 B2	5/2011	Mason et al.
7,328,053 B1	2/2008	Diab et al.	7,937,130 B2	5/2011	Diab et al.
7,332,784 B2	2/2008	Mills et al.	7,941,199 B2	5/2011	Kiani
7,340,287 B2	3/2008	Mason et al.	7,951,086 B2	5/2011	Flaherty et al.
7,341,559 B2	3/2008	Schulz et al.	7,957,780 B2	6/2011	Lamego et al.
7,343,186 B2	3/2008	Lamego et al.	7,962,188 B2	6/2011	Kiani et al.
D566,282 S	4/2008	Al-Ali et al.	7,962,190 B1	6/2011	Diab et al.
7,355,512 B1	4/2008	Al-Ali	7,976,472 B2	7/2011	Kiani
7,356,365 B2	4/2008	Schurman	7,988,637 B2	8/2011	Diab
7,371,981 B2	5/2008	Abdul-Hafiz	7,990,382 B2	8/2011	Kiani
7,373,193 B2	5/2008	Al-Ali et al.	7,991,446 B2	8/2011	Al-Ali et al.
7,373,194 B2	5/2008	Weber et al.	8,000,761 B2	8/2011	Al-Ali
7,376,453 B1	5/2008	Diab et al.	8,008,088 B2	8/2011	Bellott et al.
7,377,794 B2	5/2008	Al Ali et al.	RE42,753 E	9/2011	Kiani-Azarbayjany et al.
7,377,899 B2	5/2008	Weber et al.	8,019,400 B2	9/2011	Diab et al.
7,383,070 B2	6/2008	Diab et al.	8,028,701 B2	10/2011	Al-Ali et al.
7,415,297 B2	8/2008	Al-Ali et al.	8,029,765 B2	10/2011	Bellott et al.
7,428,432 B2	9/2008	Ali et al.	8,036,727 B2	10/2011	Schurman et al.
7,438,683 B2	10/2008	Ai-Ali et al.	8,036,728 B2	10/2011	Diab et al.
7,440,787 B2	10/2008	Diab	8,046,040 B2	10/2011	Ali et al.
7,454,240 B2	11/2008	Diab et al.	8,046,041 B2	10/2011	Diab et al.
7,467,002 B2	12/2008	Weber et al.	8,046,042 B2	10/2011	Diab et al.
7,469,157 B2	12/2008	Diab et al.	8,048,040 B2	11/2011	Kiani
7,471,969 B2	12/2008	Diab et al.	8,050,728 B2	11/2011	Al-Ali et al.
7,471,971 B2	12/2008	Diab et al.	RE43,169 E	2/2012	Parker
7,483,729 B2	1/2009	Al-Ali et al.	8,118,620 B2	2/2012	Ai-Ali et al.
7,483,730 B2	1/2009	Diab et al.	8,126,528 B2	2/2012	Diab et al.
7,489,958 B2	2/2009	Diab et al.	8,128,572 B2	3/2012	Diab et al.
			8,130,105 B2	3/2012	Al-Ali et al.
			8,145,287 B2	3/2012	Diab et al.
			8,150,487 B2	4/2012	Diab et al.
			8,175,672 B2	5/2012	Parker



(56)

## References Cited

## U.S. PATENT DOCUMENTS

8,180,420 B2	5/2012	Diab et al.	8,577,431 B2	11/2013	Lamego et al.
8,182,443 B1	5/2012	Kiani	8,581,732 B2	11/2013	Al-Ali et al.
8,185,180 B2	5/2012	Diab et al.	8,584,345 B2	11/2013	Ai-Ali et al.
8,190,223 B2	5/2012	Ai-Ali et al.	8,588,880 B2	11/2013	Abdul-Hafiz et al.
8,190,227 B2	5/2012	Diab et al.	8,600,467 B2	12/2013	Al-Ali et al.
8,203,438 B2	6/2012	Kiani et al.	8,606,342 B2	12/2013	Diab
8,203,704 B2	6/2012	Merritt et al.	8,626,255 B2	1/2014	Al-Ali et al.
8,204,566 B2	6/2012	Schurman et al.	8,630,691 B2	1/2014	Lamego et al.
8,219,172 B2	7/2012	Schurman et al.	8,634,889 B2	1/2014	Al-Ali et al.
8,224,411 B2	7/2012	Al-Ali et al.	8,641,631 B2	2/2014	Sierra et al.
8,228,181 B2	7/2012	Al-Ali	8,652,060 B2	2/2014	Al-Ali
8,229,533 B2	7/2012	Diab et al.	8,663,107 B2	3/2014	Kiani
8,233,955 B2	7/2012	Al-Ali et al.	8,666,468 B1	3/2014	Al-Ali
8,244,325 B2	8/2012	Al-Ali et al.	8,667,967 B2	3/2014	Al-Ali et al.
8,255,026 B1	8/2012	Al-Ali	8,670,811 B2	3/2014	O'Reilly
8,255,027 B2	8/2012	Al-Ali et al.	8,670,814 B2	3/2014	Diab et al.
8,255,028 B2	8/2012	Al-Ali et al.	8,676,286 B2	3/2014	Weber et al.
8,260,577 B2	9/2012	Weber et al.	8,682,407 B2	3/2014	Al-Ali
8,265,723 B1	9/2012	McHale et al.	RE44,823 E	4/2014	Parker
8,274,360 B2	9/2012	Sampath et al.	RE44,875 E	4/2014	Kiani et al.
8,280,473 B2	10/2012	Al-Ali	8,690,799 B2	4/2014	Telfort et al.
8,301,217 B2	10/2012	Al-Ali et al.	8,700,112 B2	4/2014	Kiani
8,306,596 B2	11/2012	Schurman et al.	8,702,627 B2	4/2014	Telfort et al.
8,310,336 B2	11/2012	Muhsin et al.	8,706,179 B2	4/2014	Parker
8,315,683 B2	11/2012	Al-Ali et al.	8,712,494 B1	4/2014	MacNeish, III et al.
RE43,860 E	12/2012	Parker	8,715,206 B2	5/2014	Telfort et al.
8,337,403 B2	12/2012	Al-Ali et al.	8,718,735 B2	5/2014	Lamego et al.
8,346,330 B2	1/2013	Lamego	8,718,737 B2	5/2014	Diab et al.
8,353,842 B2	1/2013	Al-Ali et al.	8,718,738 B2	5/2014	Blank et al.
8,355,766 B2	1/2013	MacNeish, III et al.	8,720,249 B2	5/2014	Al-Ali
8,359,080 B2	1/2013	Diab et al.	8,721,541 B2	5/2014	Al-Ali et al.
8,364,223 B2	1/2013	Al-Ali et al.	8,721,542 B2	5/2014	Al-Ali et al.
8,364,226 B2	1/2013	Diab et al.	8,723,677 B1	5/2014	Kiani
8,374,665 B2	2/2013	Lamego	8,740,792 B1	6/2014	Kiani et al.
8,385,995 B2	2/2013	Al-Ali et al.	8,754,776 B2	6/2014	Poeze et al.
8,385,996 B2	2/2013	Smith et al.	8,755,535 B2	6/2014	Telfort et al.
8,388,353 B2	3/2013	Kiani et al.	8,755,856 B2	6/2014	Diab et al.
8,399,822 B2	3/2013	Al-Ali	8,755,872 B1	6/2014	Marinow
8,401,602 B2	3/2013	Kiani	8,761,850 B2	6/2014	Lamego
8,405,608 B2	3/2013	Ai-Ali et al.	8,764,671 B2	7/2014	Kiani
8,414,499 B2	4/2013	Ai-Ali et al.	8,768,423 B2	7/2014	Shakespeare et al.
8,418,524 B2	4/2013	Ai-Ali	8,771,204 B2	7/2014	Telfort et al.
8,423,106 B2	4/2013	Lamego et al.	8,777,634 B2	7/2014	Kiani et al.
8,428,967 B2	4/2013	Olsen et al.	8,781,543 B2	7/2014	Diab et al.
8,430,817 B1	4/2013	Al-Ali et al.	8,781,544 B2	7/2014	Al-Ali et al.
8,437,825 B2	5/2013	Dalvi et al.	8,781,549 B2	7/2014	Al-Ali et al.
8,455,290 B2	6/2013	Siskavich	8,788,003 B2	7/2014	Schurman et al.
8,457,703 B2	6/2013	Al-Ali	8,790,268 B2	7/2014	Al-Ali
8,457,707 B2	6/2013	Kiani	8,801,613 B2	8/2014	Ai-Ali et al.
8,463,349 B2	6/2013	Diab et al.	8,821,397 B2	9/2014	Ai-Ali et al.
8,466,286 B2	6/2013	Bellot et al.	8,821,415 B2	9/2014	Al-Ali et al.
8,471,713 B2	6/2013	Poeze et al.	8,830,449 B1	9/2014	Lamego et al.
8,473,020 B2	6/2013	Kiani et al.	8,831,700 B2	9/2014	Schurman et al.
8,483,787 B2	7/2013	Al-Ali et al.	8,840,549 B2	9/2014	Al-Ali et al.
8,489,364 B2	7/2013	Weber et al.	8,847,740 B2	9/2014	Kiani et al.
8,498,684 B2	7/2013	Weber et al.	8,849,365 B2	9/2014	Smith et al.
8,504,128 B2	8/2013	Blank et al.	8,852,094 B2	10/2014	Al-Ali et al.
8,509,867 B2	8/2013	Workman et al.	8,852,994 B2	10/2014	Wojtczuk et al.
8,515,509 B2	8/2013	Bruinsma et al.	8,868,147 B2	10/2014	Stippick et al.
8,523,781 B2	9/2013	Al-Ali	8,868,150 B2	10/2014	Al-Ali et al.
8,529,301 B2	9/2013	Ai-Ali et al.	8,870,792 B2	10/2014	Al-Ali et al.
8,532,727 B2	9/2013	Ali et al.	8,886,271 B2	11/2014	Kiani et al.
8,532,728 B2	9/2013	Diab et al.	8,888,539 B2	11/2014	Al-Ali et al.
D692,145 S	10/2013	Al-Ali et al.	8,888,708 B2	11/2014	Diab et al.
8,547,209 B2	10/2013	Kiani et al.	8,892,180 B2	11/2014	Weber et al.
8,548,548 B2	10/2013	Al-Ali	8,897,847 B2	11/2014	Al-Ali
8,548,549 B2	10/2013	Schurman et al.	8,909,310 B2	12/2014	Lamego et al.
8,548,550 B2	10/2013	Al-Ali et al.	8,911,377 B2	12/2014	Al-Ali
8,560,032 B2	10/2013	Al-Ali et al.	8,912,909 B2	12/2014	Al-Ali et al.
8,560,034 B1	10/2013	Diab et al.	8,920,317 B2	12/2014	Al-Ali et al.
8,570,167 B2	10/2013	Ai-Ali	8,921,699 B2	12/2014	Al-Ali et al.
8,570,503 B2	10/2013	Vo et al.	8,922,382 B2	12/2014	Al-Ali et al.
8,571,617 B2	10/2013	Reichgott et al.	8,929,964 B2	1/2015	Al-Ali et al.
8,571,618 B1	10/2013	Lamego et al.	8,942,777 B2	1/2015	Diab et al.
8,571,619 B2	10/2013	Al-Ali et al.	8,948,834 B2	2/2015	Diab et al.
			8,948,835 B2	2/2015	Diab
			8,965,471 B2	2/2015	Lamego
			8,983,564 B2	3/2015	Al-Ali
			8,989,831 B2	3/2015	Al-Ali et al.



(56)

References Cited

U.S. PATENT DOCUMENTS

8,996,085 B2	3/2015	Kiani et al.	9,480,435 B2	11/2016	Olsen
8,998,809 B2	4/2015	Kiani	9,492,110 B2	11/2016	Al-Ali et al.
9,026,053 B2*	5/2015	Molettiere ..... H04B 7/26 455/41.2	9,510,779 B2	12/2016	Poeze et al.
9,028,429 B2	5/2015	Telfort et al.	9,517,024 B2	12/2016	Kiani et al.
9,037,207 B2	5/2015	Al-Ali et al.	9,532,722 B2	1/2017	Lamego et al.
9,060,721 B2	6/2015	Reichgott et al.	9,538,949 B2	1/2017	Al-Ali et al.
9,066,666 B2	6/2015	Kiani	9,538,980 B2	1/2017	Telfort et al.
9,066,680 B1	6/2015	Al-Ali et al.	9,549,696 B2	1/2017	Lamego et al.
9,072,474 B2	7/2015	Al-Ali et al.	9,554,737 B2	1/2017	Schurman et al.
9,078,560 B2	7/2015	Schurman et al.	9,560,996 B2	2/2017	Kiani
9,084,569 B2	7/2015	Weber et al.	9,560,998 B2	2/2017	Al-Ali et al.
9,095,316 B2	8/2015	Welch et al.	9,566,019 B2	2/2017	Al-Ali et al.
9,106,038 B2	8/2015	Telfort et al.	9,579,039 B2	2/2017	Jansen et al.
9,106,307 B2	8/2015	Molettiere et al.	9,591,975 B2	3/2017	Dalvi et al.
9,107,625 B2	8/2015	Telfort et al.	9,622,692 B2	4/2017	Lamego et al.
9,107,626 B2	8/2015	Al-Ali et al.	9,622,693 B2	4/2017	Diab
9,113,831 B2	8/2015	Al-Ali	D788,312 S	5/2017	Al-Ali et al.
9,113,832 B2	8/2015	Al-Ali	9,636,055 B2	5/2017	Al-Ali et al.
9,119,595 B2	9/2015	Lamego	9,636,056 B2	5/2017	Al-Ali
9,131,881 B2	9/2015	Diab et al.	9,649,054 B2	5/2017	Lamego et al.
9,131,882 B2	9/2015	Al-Ali et al.	9,662,052 B2	5/2017	Al-Ali et al.
9,131,883 B2	9/2015	Al-Ali	9,668,679 B2	6/2017	Schurman et al.
9,131,917 B2	9/2015	Telfort et al.	9,668,680 B2	6/2017	Bruinsma et al.
9,138,180 B1	9/2015	Coverston et al.	9,668,703 B2	6/2017	Al-Ali
9,138,182 B2	9/2015	Al-Ali et al.	9,675,286 B2	6/2017	Diab
9,138,192 B2	9/2015	Weber et al.	9,687,160 B2	6/2017	Kiani
9,142,117 B2	9/2015	Muhsin et al.	9,693,719 B2	7/2017	Al-Ali et al.
9,153,112 B1	10/2015	Kiani et al.	9,693,737 B2	7/2017	Al-Ali
9,153,121 B2	10/2015	Kiani et al.	9,697,928 B2	7/2017	Al-Ali et al.
9,161,696 B2	10/2015	Al-Ali et al.	9,717,425 B2	8/2017	Kiani et al.
9,161,713 B2	10/2015	Al-Ali et al.	9,717,458 B2	8/2017	Lamego et al.
9,167,995 B2	10/2015	Lamego et al.	9,724,016 B1	8/2017	Al-Ali et al.
9,176,141 B2	11/2015	Al-Ali et al.	9,724,024 B2	8/2017	Al-Ali
9,186,102 B2	11/2015	Bruinsma et al.	9,724,025 B1	8/2017	Kiani et al.
9,192,312 B2	11/2015	Al-Ali	9,730,640 B2	8/2017	Diab et al.
9,192,329 B2	11/2015	Al-Ali	9,743,887 B2	8/2017	Al-Ali et al.
9,192,351 B1	11/2015	Telfort et al.	9,749,232 B2	8/2017	Sampath et al.
9,195,385 B2	11/2015	Al-Ali et al.	9,750,442 B2	9/2017	Olsen
9,211,072 B2	12/2015	Kiani	9,750,443 B2	9/2017	Smith et al.
9,211,095 B1	12/2015	Al-Ali	9,750,461 B1	9/2017	Telfort
9,218,454 B2	12/2015	Kiani et al.	9,775,545 B2	10/2017	Al-Ali et al.
9,226,696 B2	1/2016	Kiani	9,775,546 B2	10/2017	Diab et al.
9,241,662 B2	1/2016	Al-Ali et al.	9,775,570 B2	10/2017	Al-Ali
9,245,668 B1	1/2016	Vo et al.	9,778,079 B1	10/2017	Al-Ali et al.
9,259,185 B2	2/2016	Abdul-Hafiz et al.	9,782,077 B2	10/2017	Lamego et al.
9,267,572 B2	2/2016	Barker et al.	9,782,110 B2	10/2017	Kiani
9,277,880 B2	3/2016	Poeze et al.	9,787,568 B2	10/2017	Lamego et al.
9,289,167 B2	3/2016	Diab et al.	9,788,735 B2	10/2017	Al-Ali
9,295,421 B2	3/2016	Kiani et al.	9,788,768 B2	10/2017	Al-Ali et al.
9,307,928 B1	4/2016	Al-Ali et al.	9,795,300 B2	10/2017	Al-Ali
9,323,894 B2	4/2016	Kiani	9,795,310 B2	10/2017	Al-Ali
D755,392 S	5/2016	Hwang et al.	9,795,358 B2	10/2017	Telfort et al.
9,326,712 B1	5/2016	Kiani	9,795,739 B2	10/2017	Al-Ali et al.
9,333,316 B2	5/2016	Kiani	9,801,556 B2	10/2017	Kiani
9,339,220 B2	5/2016	Lamego et al.	9,801,588 B2	10/2017	Weber et al.
9,341,565 B2	5/2016	Lamego et al.	9,808,188 B1	11/2017	Perea et al.
9,351,673 B2	5/2016	Diab et al.	9,814,418 B2	11/2017	Weber et al.
9,351,675 B2	5/2016	Al-Ali et al.	9,820,691 B2	11/2017	Kiani
9,364,181 B2	6/2016	Kiani et al.	9,833,152 B2	12/2017	Kiani et al.
9,368,671 B2	6/2016	Wojtczuk et al.	9,833,180 B2	12/2017	Shakespeare et al.
9,370,325 B2	6/2016	Al-Ali et al.	9,839,379 B2	12/2017	Al-Ali et al.
9,370,326 B2	6/2016	McHale et al.	9,839,381 B1	12/2017	Weber et al.
9,370,335 B2	6/2016	Al-Ali et al.	9,847,002 B2	12/2017	Kiani et al.
9,375,185 B2	6/2016	Ali et al.	9,847,749 B2	12/2017	Kiani et al.
9,386,953 B2	7/2016	Al-Ali	9,848,800 B1	12/2017	Lee et al.
9,386,961 B2	7/2016	Al-Ali et al.	9,848,806 B2	12/2017	Al-Ali et al.
9,392,945 B2	7/2016	Al-Ali et al.	9,848,807 B2	12/2017	Lamego
9,397,448 B2	7/2016	Al-Ali et al.	9,861,298 B2	1/2018	Eckerbom et al.
9,408,542 B1	8/2016	Kinast et al.	9,861,304 B2	1/2018	Al-Ali et al.
9,436,645 B2	9/2016	Al-Ali et al.	9,861,305 B1	1/2018	Weber et al.
9,445,759 B1	9/2016	Lamego et al.	9,867,578 B2	1/2018	Al-Ali et al.
9,466,919 B2	10/2016	Kiani et al.	9,872,623 B2	1/2018	Al-Ali
9,474,474 B2	10/2016	Lamego et al.	9,876,320 B2	1/2018	Coverston et al.
9,480,422 B2	11/2016	Al-Ali	9,877,650 B2	1/2018	Muhsin et al.
			9,877,686 B2	1/2018	Al-Ali et al.
			9,891,079 B2	2/2018	Dalvi
			9,895,107 B2	2/2018	Al-Ali et al.
			9,913,617 B2	3/2018	Al-Ali et al.
			9,924,893 B2	3/2018	Schurman et al.



(56)

References Cited

U.S. PATENT DOCUMENTS

9,924,897 B1	3/2018	Abdul-Hafiz	2013/0096936 A1	4/2013	Sampath et al.
9,936,917 B2	4/2018	Poeze et al.	2013/0243021 A1	9/2013	Siskavich
9,943,269 B2	4/2018	Muhsin et al.	2013/0253334 A1	9/2013	Al-Ali et al.
9,949,676 B2	4/2018	Al-Ali	2013/0296672 A1	11/2013	O'Neil et al.
9,955,937 B2	5/2018	Telfort	2013/0296713 A1	11/2013	Al-Ali et al.
9,965,946 B2	5/2018	Al-Ali	2013/0324808 A1	12/2013	Al-Ali et al.
9,980,667 B2	5/2018	Kiani et al.	2013/0331660 A1	12/2013	Al-Ali et al.
9,986,919 B2	6/2018	Lamego et al.	2014/0012100 A1	1/2014	Al-Ali et al.
9,986,952 B2	6/2018	Dalvi et al.	2014/0051953 A1	2/2014	Lamego et al.
9,989,560 B2	6/2018	Poeze et al.	2014/0081175 A1	3/2014	Telfort
9,993,207 B2	6/2018	Al-Ali et al.	2014/0120564 A1	5/2014	Workman et al.
10,007,758 B2	6/2018	Al-Ali et al.	2014/0121482 A1	5/2014	Merritt et al.
10,010,276 B2	7/2018	Al-Ali et al.	2014/0127137 A1	5/2014	Bellott et al.
10,032,002 B2	7/2018	Kiani et al.	2014/0135588 A1	5/2014	Al-Ali et al.
10,039,482 B2	8/2018	Al-Ali et al.	2014/0163344 A1	6/2014	Ai-Ali
10,052,037 B2	8/2018	Kinast et al.	2014/0163402 A1	6/2014	Lamego et al.
10,058,275 B2	8/2018	Al-Ali et al.	2014/0166076 A1	6/2014	Kiani et al.
10,064,562 B2	9/2018	Al-Ali	2014/0171763 A1	6/2014	Diab
10,086,138 B1	10/2018	Novak, Jr.	2014/0180038 A1	6/2014	Kiani
10,092,200 B2	10/2018	Al-Ali et al.	2014/0180154 A1	6/2014	Sierra et al.
10,092,249 B2	10/2018	Kiani et al.	2014/0180160 A1	6/2014	Brown et al.
10,098,550 B2	10/2018	Al-Ali et al.	2014/0187973 A1	7/2014	Brown et al.
10,098,591 B2	10/2018	Al-Ali et al.	2014/0213864 A1	7/2014	Abdul-Hafiz et al.
10,098,610 B2	10/2018	Al-Ali et al.	2014/0266790 A1	9/2014	Ai-Ali et al.
10,123,726 B2	11/2018	Al-Ali et al.	2014/0275808 A1	9/2014	Poeze et al.
10,130,289 B2	11/2018	Al-Ali et al.	2014/0275835 A1	9/2014	Lamego et al.
10,130,291 B2	11/2018	Schurman et al.	2014/0275871 A1	9/2014	Lamego et al.
10,149,616 B2	12/2018	Al-Ali et al.	2014/0275872 A1	9/2014	Merritt et al.
10,154,815 B2	12/2018	Al-Ali et al.	2014/0276115 A1	9/2014	Dalvi et al.
10,159,412 B2	12/2018	Lamego et al.	2014/0288400 A1	9/2014	Diab et al.
10,188,296 B2	1/2019	Al-Ali et al.	2014/0316217 A1	10/2014	Purdon et al.
10,188,331 B1	1/2019	Al-Ali et al.	2014/0316218 A1	10/2014	Purdon et al.
10,188,348 B2	1/2019	Kiani et al.	2014/0316228 A1	10/2014	Blank et al.
10,194,847 B2	2/2019	Al-Ali	2014/0323825 A1	10/2014	Al-Ali et al.
10,194,848 B1	2/2019	Kiani et al.	2014/0323897 A1	10/2014	Brown et al.
10,201,298 B2	2/2019	Al-Ali et al.	2014/0323898 A1	10/2014	Purdon et al.
10,205,272 B2	2/2019	Kiani et al.	2014/0330092 A1	11/2014	Al-Ali et al.
10,205,291 B2	2/2019	Scruggs et al.	2014/0330098 A1	11/2014	Merritt et al.
10,213,108 B2	2/2019	Al-Ali	2014/0357966 A1	12/2014	Al-Ali et al.
10,219,706 B2	3/2019	Al-Ali	2015/0005600 A1	1/2015	Blank et al.
10,219,746 B2	3/2019	McHale et al.	2015/0011907 A1	1/2015	Purdon et al.
10,226,187 B2	3/2019	Al-Ali et al.	2015/0012231 A1	1/2015	Poeze et al.
10,226,576 B2	3/2019	Kiani	2015/0032029 A1	1/2015	Al-Ali et al.
10,231,657 B2	3/2019	Al-Ali et al.	2015/0038859 A1	2/2015	Dalvi et al.
10,231,670 B2	3/2019	Blank et al.	2015/0045637 A1	2/2015	Dalvi
10,231,676 B2	3/2019	Al-Ali et al.	2015/0065046 A1*	3/2015	Wilfred ..... H04W 76/14 455/41.2
10,251,585 B2	4/2019	Al-Ali et al.	2015/0080754 A1	3/2015	Purdon et al.
10,251,586 B2	4/2019	Lamego	2015/0087936 A1	3/2015	Ai-Ali et al.
10,255,994 B2	4/2019	Sampath et al.	2015/0094546 A1	4/2015	Al-Ali
10,258,265 B1	4/2019	Poeze et al.	2015/0097701 A1	4/2015	Ai-Ali et al.
10,258,266 B1	4/2019	Poeze et al.	2015/0099943 A1*	4/2015	Russell ..... A61B 5/0059 600/301
2006/0161054 A1	7/2006	Reuss et al.	2015/0099950 A1	4/2015	Al-Ali et al.
2007/0282478 A1	12/2007	Al-Ali et al.	2015/0099955 A1	4/2015	Ai-Ali et al.
2008/0113618 A1	5/2008	De Leon et al.	2015/0101844 A1	4/2015	Al-Ali et al.
2009/0247984 A1	10/2009	Lamego et al.	2015/0106121 A1	4/2015	Muhsin et al.
2009/0275813 A1	11/2009	Davis	2015/0112151 A1	4/2015	Muhsin et al.
2009/0275844 A1	11/2009	Al-Ali	2015/0116076 A1	4/2015	Al-Ali et al.
2010/0004518 A1	1/2010	Vo et al.	2015/0126830 A1	5/2015	Schurman et al.
2010/0030040 A1	2/2010	Poeze et al.	2015/0165312 A1	6/2015	Kiani
2011/0072263 A1	3/2011	Bishop et al.	2015/0196249 A1	7/2015	Brown et al.
2011/0081860 A1	4/2011	Brown et al.	2015/0216459 A1	8/2015	Al-Ali et al.
2011/0082711 A1	4/2011	Poeze et al.	2015/0230755 A1	8/2015	Al-Ali et al.
2011/0105854 A1	5/2011	Kiani et al.	2015/0238722 A1	8/2015	Al-Ali
2011/0125060 A1	5/2011	Telfort et al.	2015/0245773 A1	9/2015	Lamego et al.
2011/0208015 A1	8/2011	Welch et al.	2015/0245794 A1	9/2015	Al-Ali
2011/0230733 A1	9/2011	Al-Ali	2015/0257689 A1	9/2015	Al-Ali et al.
2012/0165629 A1	6/2012	Merritt et al.	2015/0272514 A1	10/2015	Kiani et al.
2012/0209082 A1	8/2012	Al-Ali	2015/0351697 A1	12/2015	Weber et al.
2012/0209084 A1	8/2012	Olsen et al.	2015/0359429 A1	12/2015	Al-Ali et al.
2012/0283524 A1	11/2012	Kiani et al.	2015/0366507 A1	12/2015	Blank
2012/0319816 A1	12/2012	Al-Ali	2016/0029932 A1	2/2016	Al-Ali
2013/0023775 A1	1/2013	Lamego et al.	2016/0058347 A1	3/2016	Reichgott et al.
2013/0041591 A1	2/2013	Lamego	2016/0066824 A1	3/2016	Al-Ali et al.
2013/0060147 A1	3/2013	Welch et al.	2016/0081552 A1	3/2016	Wojtczuk et al.
2013/0096405 A1	4/2013	Garfio	2016/0095543 A1	4/2016	Telfort et al.
			2016/0095548 A1	4/2016	Al-Ali et al.
			2016/0103598 A1	4/2016	Al-Ali et al.



(56)

References Cited

U.S. PATENT DOCUMENTS

2016/0143548	A1	5/2016	Al-Ali	2018/0103874	A1	4/2018	Lee et al.
2016/0166182	A1	6/2016	Al-Ali et al.	2018/0103905	A1	4/2018	Kiani
2016/0166183	A1	6/2016	Poeze et al.	2018/0110478	A1	4/2018	Al-Ali
2016/0192869	A1	7/2016	Kiani et al.	2018/0116575	A1	5/2018	Perea et al.
2016/0196388	A1	7/2016	Lamego	2018/0125368	A1	5/2018	Lamego et al.
2016/0197436	A1	7/2016	Barker et al.	2018/0125430	A1	5/2018	Al-Ali et al.
2016/0213281	A1	7/2016	Eckerbom et al.	2018/0125445	A1	5/2018	Telfort et al.
2016/0228043	A1	8/2016	O'Neil et al.	2018/0130325	A1	5/2018	Kiani et al.
2016/0233632	A1	8/2016	Scruggs et al.	2018/0132769	A1	5/2018	Weber et al.
2016/0234944	A1	8/2016	Schmidt et al.	2018/0146901	A1	5/2018	Al-Ali et al.
2016/0270735	A1	9/2016	Diab et al.	2018/0146902	A1	5/2018	Kiani et al.
2016/0283665	A1	9/2016	Sampath et al.	2018/0153442	A1	6/2018	Eckerbom et al.
2016/0287090	A1	10/2016	Al-Ali et al.	2018/0153446	A1	6/2018	Kiani
2016/0287786	A1	10/2016	Kiani	2018/0153447	A1	6/2018	Al-Ali et al.
2016/0296169	A1	10/2016	McHale et al.	2018/0153448	A1	6/2018	Weber et al.
2016/0310052	A1	10/2016	Al-Ali et al.	2018/0161499	A1	6/2018	Al-Ali et al.
2016/0314260	A1	10/2016	Kiani	2018/0168491	A1	6/2018	Al-Ali et al.
2016/0324486	A1	11/2016	Al-Ali et al.	2018/0174679	A1	6/2018	Sampath et al.
2016/0324488	A1	11/2016	Olsen	2018/0174680	A1	6/2018	Sampath et al.
2016/0327984	A1	11/2016	Al-Ali et al.	2018/0184917	A1	7/2018	Kiani
2016/0328528	A1	11/2016	Al-Ali et al.	2018/0192924	A1	7/2018	Ai-Ali
2016/0331332	A1	11/2016	Al-Ali	2018/0192953	A1	7/2018	Shreim et al.
2016/0360561	A1	12/2016	Lee et al.	2018/0192955	A1	7/2018	Ai-Ali et al.
2016/0367173	A1	12/2016	Dalvi et al.	2018/0199871	A1	7/2018	Pauley et al.
2017/0000394	A1	1/2017	Al-Ali et al.	2018/0206795	A1	7/2018	Ai-Ali
2017/0007134	A1	1/2017	Al-Ali et al.	2018/0206815	A1	7/2018	Telfort
2017/0007198	A1	1/2017	Al-Ali et al.	2018/0213583	A1	7/2018	Ai-Ali
2017/0014083	A1	1/2017	Diab et al.	2018/0214031	A1	8/2018	Kiani et al.
2017/0014084	A1	1/2017	Al-Ali et al.	2018/0214090	A1	8/2018	Al-Ali et al.
2017/0024748	A1	1/2017	Haider	2018/0218792	A1	8/2018	Muhsin et al.
2017/0027456	A1	2/2017	Kinast et al.	2018/0225960	A1	8/2018	Ai-Ali et al.
2017/0042488	A1	2/2017	Muhsin	2018/0238718	A1	8/2018	Dalvi
2017/0055581	A1	3/2017	Ai-Ali	2018/0242853	A1	8/2018	Ai-Ali
2017/0055582	A1	3/2017	Al-Ali et al.	2018/0242921	A1	8/2018	Muhsin et al.
2017/0055587	A1	3/2017	Al-Ali	2018/0242923	A1	8/2018	Ai-Ali et al.
2017/0055896	A1	3/2017	Ai-Ali et al.	2018/0242924	A1	8/2018	Barker et al.
2017/0079594	A1	3/2017	Telfort et al.	2018/0242926	A1	8/2018	Muhsin et al.
2017/0086723	A1	3/2017	Al-Ali et al.	2018/0247353	A1	8/2018	Ai-Ali et al.
2017/0143281	A1	5/2017	Olsen	2018/0247712	A1	8/2018	Muhsin et al.
2017/0147774	A1	5/2017	Kiani	2018/0249933	A1	9/2018	Schurman et al.
2017/0156620	A1	6/2017	Al-Ali et al.	2018/0253947	A1	9/2018	Muhsin et al.
2017/0173632	A1	6/2017	Al-Ali	2018/0256087	A1	9/2018	Ai-Ali et al.
2017/0187146	A1	6/2017	Kiani et al.	2018/0256113	A1	9/2018	Weber et al.
2017/0188919	A1	7/2017	Al-Ali et al.	2018/0285094	A1	10/2018	Housel et al.
2017/0196464	A1	7/2017	Jansen et al.	2018/0289325	A1	10/2018	Poeze et al.
2017/0196470	A1	7/2017	Lamego et al.	2018/0289337	A1	10/2018	Al-Ali et al.
2017/0202490	A1	7/2017	Al-Ali et al.	2018/0296161	A1	10/2018	Shreim et al.
2017/0214780	A1*	7/2017	Gofman ..... H04W 76/10	2018/0300919	A1	10/2018	Muhsin et al.
2017/0224262	A1	8/2017	Al-Ali	2018/0310822	A1	11/2018	Indorf et al.
2017/0228516	A1	8/2017	Sampath et al.	2018/0310823	A1	11/2018	Al-Ali et al.
2017/0245790	A1	8/2017	Al-Ali et al.	2018/0317826	A1	11/2018	Muhsin
2017/0251974	A1	9/2017	Shreim et al.	2018/0317841	A1	11/2018	Novak, Jr.
2017/0251975	A1	9/2017	Shreim et al.	2018/0333055	A1	11/2018	Lamego et al.
2017/0258403	A1	9/2017	Abdul-Hafiz et al.	2018/0333087	A1	11/2018	Al-Ali
2017/0311891	A1	11/2017	Kiani et al.	2019/0000317	A1	1/2019	Muhsin et al.
2017/0325728	A1	11/2017	Al-Ali et al.	2019/0000362	A1	1/2019	Kiani et al.
2017/0332976	A1	11/2017	Al-Ali et al.	2019/0015023	A1	1/2019	Monfre
2017/0340293	A1	11/2017	Al-Ali et al.	2019/0021638	A1	1/2019	Al-Ali et al.
2017/0360310	A1	12/2017	Kiani et al.	2019/0029574	A1	1/2019	Schurman et al.
2017/0367632	A1	12/2017	Al-Ali et al.	2019/0029578	A1	1/2019	Al-Ali et al.
2018/0008146	A1	1/2018	Al-Ali et al.	2019/0058280	A1	2/2019	Al-Ali et al.
2018/0013562	A1	1/2018	Haider et al.	2019/0058281	A1	2/2019	Al-Ali et al.
2018/0014752	A1	1/2018	Al-Ali et al.	2019/0069813	A1	3/2019	Al-Ali
2018/0028124	A1	2/2018	Al-Ali et al.	2019/0069814	A1	3/2019	Al-Ali
2018/0055385	A1	3/2018	Al-Ali	2019/0076028	A1	3/2019	Al-Ali et al.
2018/0055390	A1	3/2018	Kiani et al.	2019/0082979	A1	3/2019	Al-Ali et al.
2018/0055430	A1	3/2018	Diab et al.	2019/0090748	A1	3/2019	Al-Ali
2018/0064381	A1	3/2018	Shakespeare et al.	2019/0090760	A1	3/2019	Kinast et al.
2018/0069776	A1	3/2018	Lamego et al.	2019/0090764	A1	3/2019	Al-Ali
2018/0070867	A1	3/2018	Smith et al.	2019/0117070	A1	4/2019	Muhsin et al.
2018/0082767	A1	3/2018	Al-Ali et al.	2019/0117139	A1	4/2019	Al-Ali et al.
2018/0085068	A1	3/2018	Telfort	2019/0117140	A1	4/2019	Al-Ali et al.
2018/0087937	A1	3/2018	Al-Ali et al.	2019/0117141	A1	4/2019	Al-Ali
				2019/0117930	A1	4/2019	Al-Ali

\* cited by examiner




FIG. 1A






FIG. 1B






FIG. 2A






FIG. 2B






FIG. 3A




FIG. 3B






FIG. 4




FIG. 5



600




FIG. 6




FIG. 7






FIG. 8




FIG. 9



**PATIENT-WORN WIRELESS  
PHYSIOLOGICAL SENSOR WITH PAIRING  
FUNCTIONALITY**

RELATED APPLICATIONS

This application claims priority benefit under 35 U.S.C. § 119(e) from U.S. Provisional Patent Application Ser. No. 62/505,762, filed May 12, 2017, and also claims priority benefit under 35 U.S.C. § 119(e) from U.S. Provisional Patent Application Ser. No. 62/447,836, filed Jan. 18, 2017, titled "PATIENT-WORN WIRELESS PHYSIOLOGICAL SENSOR WITH PAIRING FUNCTIONALITY", the disclosure of which is hereby incorporated by reference in its entirety.

FIELD OF THE DISCLOSURE

The present disclosure relates to the field of pairing of wireless communication devices. More specifically, the disclosure describes among other things a portable wireless device that communicates with a second device capable of wireless communication with paired electronic devices.

BACKGROUND

In clinical settings, such as hospitals, nursing homes, convalescent homes, skilled nursing facilities, post-surgical recovery centers, and the like, patients are frequently monitored using one more different types of physiological sensors. Various types of sensors include a magnetometer that detects patient movement or orientation to track and prevent patient ulcers, a temperature sensor, an acoustic respiration sensor, an electrocardiogram (ECG) sensor, an electroencephalography (EEG) sensor, one or more pulse oximetry sensors, a moisture sensor, a blood pressure sensor, and an impedance sensor, among other sensors.

Wires leading to and from traditional physiological sensors inhibit patient movement and make it difficult to provide care to a patient. Often sensors are accidentally removed by patient movement. At other times, sensors must be moved or replaced when a patient is moved to a different location or when certain types of care are provided to the patient. Wireless sensors provide a solution to the patient movement and access. However, in busy hospital environments with non-technical staff operating these wireless devices, it can be difficult to correctly configure wireless sensors for communication with the correct monitors.

Similarly, other wireless devices including consumer devices such as, but not limited to, speakers, phones, headphones, watches, keyboards, mice, and so forth, capable of being paired have similar issues. These devices are often used by non-technically oriented users that encounter cumbersome pairing requirements.

SUMMARY

For purposes of summarizing the disclosure, certain aspects, advantages and novel features have been described herein. It is to be understood that not necessarily all such advantages can be achieved in accordance with any particular embodiment disclosed herein. Thus, the embodiments disclosed herein can be embodied or carried out in a manner that achieves or optimizes one advantage or group of advantages as taught or suggested herein without necessarily achieving others.

In certain embodiments, a system for electronically pairing a wireless sensor with a patient monitoring device can include a patient monitoring device, a wireless sensor, and at least one hardware processor. The patient monitoring device can include a first display. The wireless sensor can include a button configured to activate a pairing mode that enables the wireless sensor to electronically pair with the patient monitoring device. The wireless sensor can further comprise an optical detector configured to detect light based signals. The hardware processor can further generate a visual signal from the first display. In some embodiments, the hardware processor can detect the visual signal with the optical detector of the wireless sensor. The hardware processor can associate the wireless sensor with the patient monitoring device based on the detected visual signal, thereby pairing the wireless sensor with the patient monitoring device. The hardware processor can further transmit a confirmation signal from the wireless sensor to indicate that association is complete. The hardware processor can configure the wireless sensor to enter into a patient parameter sensing mode of operation after the association of the wireless sensor with the patient monitoring device.

The system of the preceding paragraph can have any sub-combination of the following features: where the first display is of a first size and the wireless sensor is of a second size, where the second size of the wireless sensor is smaller than the first size of the first display; where the size of the wireless sensor corresponds to a shape of a base of the wireless sensor; where the wireless sensor does not require a separate antenna or any additional components for the pairing with the patient monitoring device; where the wireless sensor does not use a wireless communication protocol for the pairing with the patient monitoring device; where the wireless sensor does not use a wireless communication protocol including a Bluetooth protocol, wifi protocol, or a zigbee protocol; where the one or more hardware processors are configured to detect a shape of the wireless sensor when the wireless sensor is placed directly on the first display and in response to the detected shape, associate the wireless sensor with the patient monitoring device; and where the one or more hardware processors are configured to generate a pattern on the first display and associate the wireless sensor with the patient monitoring device based on a successful placement of the wireless sensor on the first display in relation to the generated pattern.

Additionally, in certain embodiments, a system for electronically pairing a wireless sensor with a patient monitoring device can include a patient monitoring device. The system can include a wireless sensor. The wireless sensor can include a button configured to activate a pairing mode that enables the wireless sensor to electronically pair with the patient monitoring device. The system can include one or more hardware processors. The hardware processor can further generate a signal from the patient monitoring device. The hardware processor can also detect the signal with a detector of the wireless sensor. In some embodiments, the hardware processor can associate the wireless sensor with the patient monitoring device based on the detected signal, thereby pairing the wireless sensor with the patient monitoring device. The hardware processor can also transmit a confirmation signal from the wireless sensor to indicate that association is complete. Moreover, the hardware processor can configure the wireless sensor to enter into a patient parameter sensing mode of operation after the association of the wireless sensor with the patient monitoring device.

The system of the preceding paragraph can have any sub-combination of the following features: where the wire-



less sensor does not require a separate antenna or any additional components for the pairing with the patient monitoring device; where the wireless sensor does not use a wireless communication protocol for the pairing with the patient monitoring device; where the detector comprises a piezoelectric element; where the signal comprises an acoustic signal and where the wireless sensor is configured to detect the acoustic signal with the piezoelectric element; and where the detector comprises an optical detector and where the signal comprises a visual signal and the wireless sensor is configured to detect the visual signal with the optical detector.

In certain embodiments, a method for electronically pairing a wireless sensor with a patient monitoring device can include generating a signal from the patient monitoring device. The method can also include detecting the signal with a detector of the wireless sensor. In some embodiments, the method can include associating the wireless sensor with the patient monitoring device based on the detected signal, thereby pairing the wireless sensor with the patient monitoring device. Furthermore, the method can include transmitting a confirmation signal from the wireless sensor to indicate that association is complete. The method can also include configuring the wireless sensor to enter into a patient parameter sensing mode of operation after the association of the wireless sensor with the patient monitoring device.

The method of the preceding paragraph can have any sub-combination of the following features: where the wireless sensor does not require a separate antenna or any additional components for the pairing with the patient monitoring device; where the wireless sensor does not use a wireless communication protocol for the pairing with the patient monitoring device; where the detector comprises a piezoelectric element; where the detector comprises an optical detector; further generating a pattern on a first display of the patient monitoring device, where the association is further based on a successful placement of the wireless sensor on the first display in relation to the generated pattern.

#### BRIEF DESCRIPTION OF THE DRAWINGS

FIG. 1A is a schematic assembled perspective view of a wireless sensor.

FIG. 1B is a schematic exploded perspective view of a wireless sensor.

FIG. 2A is a schematic perspective view of a wireless sensor.

FIG. 2B is a schematic perspective view of a wireless sensor and a mobile device for pairing.

FIGS. 3A-3B are perspective views of a wireless sensor and a patient monitor pairing.

FIG. 4 is a flow diagram describing a process to pair a wireless sensor with a patient monitor.

FIG. 5 is a flow diagram describing a process to pair a wireless sensor with a patient monitor.

FIG. 6 is a flow diagram describing a process to pair a first wireless device with a second wireless device.

FIG. 7 is a flow diagram describing a process to pair a first wireless device with a second wireless device.

FIG. 8 is a schematic perspective view of the disclosed wireless sensor and a mobile device for pairing.

FIG. 9 is a schematic perspective view of the disclosed wireless sensor and a car console for pairing.

#### DETAILED DESCRIPTION

FIG. 1A is a schematic assembled perspective view of wireless sensor 102. The wireless sensor 102 may also be

referred to herein as “a wireless physiological sensor 102,” “a patient-worn sensor 102,” “a movement sensor 102,” and “a wearable wireless sensor 102.” The wireless sensor 102 includes one or more sensors configured to measure the patient’s position, orientation, and motion. Also illustrated in FIG. 1A is a button or switch 124 located on a top portion of the housing 150. The button or switch 124 can be used to change modes of the wireless sensor 102. For example, pressing and holding the button or switch 124 can cause the wireless sensor 102 to switch into a pairing mode of operation. The pairing mode is used to associate the wireless sensor 102 with a mobile device 210 as shown in FIGS. 2A-B or a bedside patient monitor 310 as shown in FIGS. 3A-B. Wireless sensor 102 may include one or more detecting elements such as: a magnetometer which may also be referred to as a compass, a temperature sensor, an acoustic respiration sensor, an electrocardiogram (ECG) sensor, an electroencephalography (EEG) sensor, one or more pulse oximetry sensors, a moisture sensor, a blood pressure sensor, and an impedance sensor.

The magnetometer may be a three-dimensional magnetometer that provides information indicative of magnetic fields, including the Earth’s magnetic field. A skilled artisan will understand that the accelerometer, gyroscope, and magnetometer can be integrated into a single hardware component such as an inertial measurement unit. The wireless sensor 102 may be configured to calculate the three-dimensional position and orientation of an object derived from inputs from three sensors attached to the object: an accelerometer configured to measure linear acceleration along three axes; a gyroscope configured to measure angular velocity around three axes; and a magnetometer configured to measure the strength of a magnetic field (such as the Earth’s magnetic field) along three axes. The three sensors may attach to the wireless sensor 102 which is affixed to the patient. The sensors may be sampled at a rate between approximately 10 Hz and approximately 100 Hz. One skilled in the art will appreciate that the sensors can be sampled at different rates without deviating from the scope of the present disclosure. The sampled data from the three sensors, which provide nine sensor inputs, are processed to describe the patient’s position and orientation in three-dimensional space. The patient’s position and orientation are described in terms of Euler angles as a set of rotations around a set of X-Y-Z axes of the patient.

An acoustic respiration sensor can be used to sense acoustic and/or vibrational motion from the patient’s body (e.g., the patient’s chest) that are indicative of various physiologic parameters and/or conditions, including without limitation, heart rate, respiration rate, snoring, coughing, choking, wheezing, and respiratory obstruction (e.g., apneic events). The ECG sensor can be used to measure the patient’s cardiac activity. The ECG sensor may include two electrodes and a single lead. The pulse oximetry sensor(s) can be used to monitor the patient’s pulse oximetry, a widely accepted noninvasive procedure for measuring the oxygen saturation level of arterial blood, an indicator of a person’s oxygen supply. A typical pulse oximetry system utilizes an optical sensor clipped onto a portion of the patient’s body (such as, for example, a fingertip, an ear lobe, a nostril, and the like) to measure the relative volume of oxygenated hemoglobin in pulsatile arterial blood flowing within the portion of the body being sensed. Oxygen saturation (SpO<sub>2</sub>), pulse rate, a plethysmograph waveform, perfusion index (PI), pleth variability index (PVI), methemoglobin (MetHb), carboxyhemoglobin (CoHb), total hemoglobin (tHb), glucose, and/or otherwise can be measured and monitored using



5

the pulse oximetry sensor(s). The moisture sensor can be used to determine a moisture content of the patient's skin which is a relevant clinical factor in assessing the patient's risk of forming a pressure ulcer. The impedance sensor can be used to track fluid levels of the patient. For example, the impedance sensor can monitor and detect edema, heart failure progression, and sepsis in the patient.

FIG. 1B is a schematic exploded perspective view of wireless sensor 102 including a bottom base 110, a removable battery isolator 120, a mounting frame 130, a circuit board 140, a housing 150, and a top base 160. The bottom base 110 is a substrate having a top surface on which various components of the wireless sensor 102 are positioned, and a bottom surface that is used to affix the wireless sensor 102 to the patient's body. The bottom base 110 and top base 160 can be made of medical-grade foam material such as white polyethylene, polyurethane, or reticulated polyurethane foams, to name a few. As illustrated in FIG. 1A, the bottom base 110 and the top base 160 are each in a substantially oval shape, with a thickness of approximately 1 mm. The top base 160 includes a cut-out 162 through which the housing 150 fits during assembly. A skilled artisan will understand that there are numerous sizes and shapes suitable for the top and bottom bases 110 and 160 that can be employed without departing from the scope of the present disclosure. The bottom surface of the bottom base 110 is coated with a high tack, medical-grade adhesive, which when applied to the patient's skin, is suitable for long-term monitoring, such as, for example two days or longer. Portions of the top surface of the bottom base 110 are also coated with a medical-grade adhesive, as the bottom base 110 and the top base 160 are adhered together during assembly of the wireless sensor 102. The bottom base 110 may have apertures 112 and 114. These apertures 112 and 114 may permit transmission of thermal energy, electrical energy, light, sound, or any other input to the wireless sensor 102.

The removable battery isolator 120 is a flexible strip made of an electrically insulating material that serves to block electrical communication between the battery 144 and an electrical contact (not shown) on the circuit board 140. The battery isolator 120 is used to preserve battery power until the wireless sensor 102 is ready for use. The battery isolator 120 blocks electrical connection between the battery 144 and the circuit board 140 until the battery isolator 120 is removed from the wireless sensor 102. The battery isolator 120 can be made of any material that possesses adequate flexibility to be slidably removed from its initial position and adequate dielectric properties so as to electrically isolate the battery 144 from the circuit board 140. For example, the battery isolator 120 can be made of plastic, polymer film, paper, foam, combinations of such materials, or the like. The battery isolator 120 includes a pull tab 122 that extends through a slot 152 of the housing 150 when the wireless sensor 102 is assembled. The pull tab 122 can be textured to provide a frictional surface to aid in gripping and sliding the pull tab 122 out of its original assembled position. Once the battery isolator 120 is removed the battery 144 makes an electrical connection with the electrical contact to energize the electronic components of the wireless sensor 102.

The mounting frame 130 is a structural support element that helps secure the battery 144 to the circuit board 140. The mounting frame 130 has wings 132 that, when assembled are slid between battery holder 142 and the battery 144. Additionally, the mounting frame 130 serves to provide rigid structure between the circuit board 140 and the bottom base 110. The rigid structure, which may include an acoustic respiratory sensor, may transmit vibrational motion (vibra-

6

tions) emanating from the patient (such as, for example, vibrational motions related to respiration, heartbeat, snoring, coughing, choking, wheezing, respiratory obstruction, and the like) to the accelerometer 149 positioned on the circuit board 140. The mounting frame 130 may have an aperture 134 that extends through the mounting frame 130. The aperture 134 may be aligned with the aperture 114 in the bottom base 110 as described above. The aperture 134 may permit transmission of thermal energy, electrical energy, light, sound, or any other input to the wireless sensor 102. The aperture 134 may be filled with a thermally conductive material.

A battery holder 142 is attached to two sides of the top portion circuit board 140 and extends (forming a support structure) under the bottom side of the circuit board 140 to hold the battery 144 in position relative to the circuit board 140. An electrical connection between the anode of the battery 144 and the circuit board 140 is made by way of the battery holder which is in electrical contact with the anode of the battery 144 and the circuit board 140. The cathode of the battery 144 is positioned to touch a battery contact (not shown) on the bottom side of the circuit board 140.

The housing 150 is a structural component that serves to contain and protect the components of the wireless sensor 102. The housing 150 can be made of any material that is capable of adequately protecting the electronic components of the wireless sensor 102 such as thermoplastics and thermosetting polymers. The housing 150 includes a slot 152 through which the battery isolator 120 is inserted during assembly. The housing 150 also includes a rim 154 that extends around the outer surface of the housing 150. The rim 154 is used to secure the housing 150 in position relative to the bottom base 110 and the top base 160 when the wireless sensor 102 is assembled.

Assembly of the wireless sensor 102 is as follows: the circuit board 140 and battery holder 142 holding the battery 144 are placed into the housing 150. The wings 132 of the mounting frame 130 are inserted in between the battery 144 and the battery holder 142, so as to align the mounting frame 130 with the circuit board 140. The battery isolator 120 is positioned between the electrical contact and the battery 144. The pull tab 122 of the battery isolator 120 is then fed through the slot 152 in the housing 150. The top base 160 is then positioned over the housing 150. The rim 154 of the housing 150 adheres to the bottom surface of the top base 160, which is coated with high tack, medical-grade adhesive. The resulting partial assembly is positioned centrally onto the top surface of the bottom base 110, aligning the edges of the base top 160 with the edges of the base bottom 110. The bottom surface of the bottom base 110 is then coated with a high tack, medical-grade adhesive, and a release liner (not shown) is placed on the bottom surface of the bottom base 110 to protect the adhesive until it is time for use.

FIG. 2A is a perspective illustration of the patient monitoring system in a clinical setting. The patient monitoring system includes a wireless sensor 102 worn by a patient in proximity to mobile device 210. The mobile device 210 includes screen 220 that may be configured to transmit and receive a visual, optical, and/or light-based pairing signal. The pairing signal may contain a pattern containing a shape, color, or a combination of patterns. The pairing signal may involve displaying instructions on a screen 220 for a user to follow. The instructions may include the user placing a wireless sensor 102 of a particular shape and size in a certain position relative to the screen 220. The instructions may be displayed using an object 214 on the screen 220 to represent



the silhouette of the wireless sensor **102**. To follow the instructions, the user may hold the wireless sensor **102** and position the wireless sensor **102** according to the instructions. The user may place the wireless sensor **102** to resemble the relative position of the object **214** displayed on the screen **220** of the mobile device **210**.

The pairing signal may contain a series of visual, optical, and/or light based signals. The series of signals may utilize variations in color, shade, shape, or visual patterns. The pairing signal may contain a series of flashes, wherein the flashes may vary in intensity or duration. The pairing signal may comprise a combination of visual and audio signals.

For example, the wireless sensor **102** and the mobile device **210** may include a detector that detects visual, optical, or light-based signals and another detector for detecting audio sound or a series of audio sounds. A detector may be configured to detect both visual and audio signals. The pairing signal may include a sequence of visual signals that is synchronized with a sequence of audio signals.

Additionally, the screen **220** may be configured to display motion instructions for a care provider to perform in order to generate motion signals representing a pairing signal. Mobile device **210** may also generate audio-based pairing signals using a speaker (not shown). The sounds may be tuned to a frequency within or outside of the range of human hearing and may comprise various rings or tones. Additionally, the mobile device **210** may have a port for connecting peripheral devices that may generate various signals such as current or voltage based signals. The pairing signal may take the form of an electrical signal.

FIG. **2B** is a perspective illustration of the patient monitoring system in a clinical setting. The patient monitoring system includes a wireless sensor **202** worn by a patient in proximity to mobile device **210**. The mobile device **210** includes screen **220** that may be configured to transmit a visual, optical, and/or light-based pairing signal. The pairing signal may contain a pattern containing a shape, color, or a combination of patterns. The pairing signal may involve displaying instructions on a screen **220** for a user to follow. The instructions may include the user placing a wireless sensor **202** of a particular shape and size in a certain position relative to the screen **220**. The shape and size of the wireless sensor **202** may differ from the shape and size of the wireless sensor **102**. The instructions may be displayed using an object **216** on the screen **220** to represent the silhouette of the wireless sensor **202**. To follow the instructions, the user may hold the wireless sensor **202** and position the wireless sensor **202** according to the instructions. The user may place the wireless sensor **202** to resemble the relative position of the object **216** displayed on the screen **220** of the mobile device **210**.

The pairing signal may contain a series of visual, optical, and/or light based signals. The series of signals may utilize variations in color, shade, shape, or visual patterns. The pairing signal may contain a series of flashes, wherein the flashes may vary in intensity or duration. Mobile device **210** may also generate audio-based pairing signals using a speaker (not shown). The sounds may be tuned to a frequency within or outside of the range of human hearing and may comprise various rings or tones. The pairing signal may comprise a combination of visual and audio signals. For example, the wireless sensor **102** and the mobile device **210** may include a detector that detects visual, optical, or light-based signals and another detector for detecting audio sound or a series of audio sounds. A detector may be configured to detect both visual and audio signals. The pairing signal may

include a sequence of visual signals that is synchronized with a sequence of audio signals.

Additionally, the screen **220** may be configured to display motion instructions for a care provider to perform in order to generate motion signals representing a pairing signal. Additionally, the mobile device **210** may have a port for connecting peripheral devices that may generate various signals such as current or voltage based signals. The pair signal may take the form of an electrical signal.

Wireless sensor **202** may include a probe for taking non-invasive optical measurements. The probe may have an emitter for transmitting an optical signal and a detector for detecting the optical signal transmitted by the emitter. The probe may have a flexible circuit assembly with circuit paths to connect the emitter and the detector. The detector may be further configured to detect the pairing signal emitted by mobile device **210**.

The wireless sensor **202** may also include a button or switch **204**. The button or switch **204** can be used to change modes of the wireless sensor **202**. For example, pressing and holding the button or switch **204** can cause the wireless sensor **202** to switch into a pairing mode of operation. The pairing mode is used to associate the wireless sensor **102** with a mobile device **210** or a bedside patient monitor **310**.

FIG. **3A** is a perspective illustration of the patient monitoring system in a clinical setting. The patient monitoring system includes a wireless sensor **102** worn by a patient in proximity to a bedside patient monitor **310** located at the side of a patient's bed. The bedside patient monitor **310** may include screen **312** that may be configured to transmit a visual, optical, and/or light-based pairing signal. The pairing signal may contain a pattern containing a shape, color, or a combination of patterns. The pairing signal may involve displaying instructions on a screen **312** for a user to follow.

The instructions may include the user placing a wireless sensor **102** of a particular shape and size in a certain position relative to the screen **312**. The instructions may be displayed using an object **214** on the screen **312** to represent the silhouette of the wireless sensor **102**. To follow the instructions, the user may hold the wireless sensor **102** and position the wireless sensor **102** according to the instructions. The user may place the wireless sensor **102** to resemble the relative position of the object **214** displayed on the screen **312** of the bedside patient monitor **310**.

The pairing signal may contain a series of visual, optical, and/or light based signals. The series of signals may utilize variations in color, shade, shape, or visual patterns. Additionally, the pairing signal may contain a series of flashes, wherein the flashes may vary in intensity or duration. Additionally, the screen **312** may be configured to display motion instructions for a care provider to perform in order to generate motion signals representing a pairing signal. Bedside patient monitor **310** may also generate audio-based pairing signals using a speaker (not shown). The sounds may be tuned to a frequency within or outside of the range of human hearing and may comprise various rings or tones. Additionally, the bedside patient monitor **310** may have a port **316** for connecting peripheral devices that may generate various signals such as current or voltage based signals. The pairing signal may take the form of an electrical signal.

The patient monitor **310** may have a button or switch **317** can be used to activate the patient monitor **310** and place it in the pairing mode of operation. Similarly the buttons or switches **318**, **319** can also be used to activate the patient monitor **310** and place it in the pairing mode of operation. This process is discussed more below. When the button/



switch 317 is depressed and/or continuously held down, the patient monitor 310 may enter into the pairing mode of operation.

FIG. 3B is a perspective illustration of the patient monitoring system in a clinical setting. The patient monitoring system includes a wireless sensor 202 worn by a patient in proximity to a bedside patient monitor 310 located at the side of a patient's bed. The bedside patient monitor 310 may include screen 312 that may be configured to transmit a visual, optical, and/or light-based pairing signal. The pairing signal may contain a pattern containing a shape, color, or a combination of patterns. The pairing signal may involve displaying instructions on a screen 312 for a user to follow.

Similar to FIG. 3A, the instructions may include the user placing a wireless sensor 202 of a particular shape and size in a certain position relative to the screen 312. The instructions may be displayed using an object 216 on the screen 312 to represent the silhouette of the wireless sensor 202. To follow the instructions, the user may hold the wireless sensor 202 and position the wireless sensor 202 according to the instructions. The user may place the wireless sensor 202 to resemble the relative position of the object 216 displayed on the screen 312 of the bedside patient monitor 310.

Although described with respect to a bedside patient monitor 310, it is to be understood that mobile device 210 may also perform some or all of the functionality described in relation to bedside patient monitor 310. For example, wireless sensor 102 may pair with either the mobile device 210 or bedside patient monitor 310. Additionally, one skilled in the art will appreciate that numerous types of patient-specific information may be collected and analyzed by either the mobile device 210 or bedside patient monitor 310. Therefore, it is to be understood that patient monitoring system may be implemented with a mobile device 210 or a bedside patient monitor 310.

In some scenarios, it may be desirable to pair, or associate, the wireless sensor 102 with the bedside patient monitor 310 or mobile device 210 to avoid interference from other wireless devices and/or to associate patient-specific information (stored, for example, on the patient monitor 310) with the sensor data that is being collected and transmitted by the wireless sensor 102. Illustratively, such patient-specific information can include, by way of non-limiting example, the patient's name, age, gender, weight, identification number (e.g., social security number, insurance number, hospital identification number, or the like), admission date, length of stay, physician's name and contact information, diagnoses, type of treatment, perfusion rate, hydration, nutrition, pressure ulcer formation risk assessments, patient turn protocol instructions, treatment plans, lab results, health score assessments, and the like. One skilled in the art will appreciate that numerous types of patient-specific information can be associated with the described patient-worn sensor without departing from the scope of the present disclosure. Additionally, pairing the wireless sensor 102 with the mobile device 210 or bedside patient monitor 310 can be performed to provide data security and to protect patient confidentiality. Some wireless systems require the care provider to program the wireless sensor 102 to communicate with the correct mobile device 210 or bedside patient monitor 310. Other wireless systems require a separate token or encryption key and several steps to pair the wireless device 102 with the correct bedside patient monitor 310. Some systems require the token to be connected to the mobile device 210 or the bedside patient monitor 310, then connected to the wireless device 102, and then reconnected to the mobile device 210 or bedside patient monitor

310. In certain scenarios, it may be desirable to share wireless communication information between a wireless sensor 102 and a mobile device 210 or bedside patient monitor 310 without a separate token or encryption key. For security purposes, it may be desirable to use security tokens to ensure that the correct patient monitor 310 receives the correct wirelessly transmitted data. Security tokens prevent the mobile device 210 or bedside patient monitor 310 from accessing the transmitted data unless the wireless sensor 102 and mobile device 210 or patient monitor 310 share the same password. The password may be a word, passphrase, or an array of randomly chosen bytes.

FIG. 4 illustrates a method of associating a wireless sensor 102 with a mobile device 210 or a bedside patient monitor 310, which may be referred to as "pairing." At block 402A, the wireless sensor 102 may be placed near the mobile device 210 or the bedside patient monitor 310 in preparation for receiving a pairing signal. A visual, optical, and/or light-based detector 146 of the wireless sensor 102 may be physically oriented and configured to receive a pairing signal in the form of a visual, optical, and/or light-based pairing signal. The pairing signal may contain a pattern containing a shape, color, or a combination of patterns. The pairing signal may contain a series of visual, optical, and/or light based signals. The series of signals may utilize variations in color, shade, shape, or visual patterns. Additionally, the pairing signal may contain a series of flashes, wherein the flashes may vary in intensity or duration. The signal may be unique to the wireless sensor 102 and the mobile device 210 or the bedside patient monitor 310. By using a pairing signal that uniquely identifies the wireless sensor 102 and the mobile device 210 or the bedside patient monitor 310, a secure connection may be established between the two paired devices.

Various types of sensors can be used with the pairing process of the present disclosure. For example, a pulse oximeter sensor can be paired by facing its light detector toward a bedside patient monitor screen 312 or mobile device screen 220 to receive a visual, optical, and/or light based signal. In another example, an ulcer sensor can be paired by receiving a pairing signal in the form of detected motion. Similarly, other types of wireless sensors can be paired using the included detectors. For example, an acoustic sensor can be paired based on audio signals emitted from the patient monitor 310. EEG and ECG wireless sensors can be paired using small electrical impulses from a special conductor included as part of the patient monitor 310. Other sensors can be paired in a similar fashion depending on the specific detectors included on the wireless sensors.

Returning to block 402A, the wireless sensor 102 may be placed in proximity to the mobile device 210 or the bedside patient monitor 310 such that the visual, optical, and/or light-based detector 146 may receive the visual, optical, and/or light-based pairing signal. The visual, optical, and/or light-based pairing signal may have a pairing signal transmission range of up to approximately three inches. The visual, optical, and/or light-based pairing signal may have a pairing signal transmission range of up to approximately six inches. The visual, optical, and/or light-based pairing signal may have a pairing signal transmission range of up to approximately one foot (i.e., twelve inches) or farther. A skilled artisan will recognize that other ranges can be used for the pairing signal transmission range.

At block 404A the wireless sensor 102 is set to operate in a pairing mode. A user may begin by initiating the pairing mode of operation for the wireless sensor 102. This may include powering on the wireless sensor 102, switching the



## 11

wireless sensor 102 to a special pairing state, and/or the like. For example, the wireless sensor 102 may include a battery isolator 120 which, when removed, activates the wireless sensor 102. Upon activation, the default mode of operation is the pairing mode. The wireless sensor 102 may have a button or switch 124 that can be used to activate the wireless sensor 102 and place it in the pairing mode of operation. For example, a depressible button or switch 124 can be located on the top portion of the housing 150. When the button or switch 124 is depressed and continuously held down, the wireless sensor 102 enters into the pairing mode of operation and remains in the pairing mode of operation for as long as the button or switch 124 is depressed. The wireless sensor 102 enters into the pairing mode by activating the visual, optical, and/or light-based sensor 146. Once activated, the optical sensor 146 may be configured to receive the visual, optical, and/or light-based pairing signal. Similarly, the wireless sensor 202 may have a button or switch 204 that can be used to activate the wireless sensor 202 and place it in the pairing mode of operation.

At block 404B, the mobile device 210 or bedside patient monitor 310 is set to operate in pairing mode. A user may begin by initiating the pairing mode of operation for the mobile device 210 or the bedside patient monitor 310. This may include powering on the mobile device 210 or the bedside patient monitor 310, switching the mobile device 210 or the bedside patient monitor 310 to a special pairing state, and/or the like. The bedside patient monitor 310 may have a button or switch 317 that can be used to activate the patient monitor 310 and place it in the pairing mode of operation. When the button or switch 317 is depressed and/or continuously held down, the patient monitor 310 enters into the pairing mode of operation. Similarly, the mobile device 210 may have a button or switch that can be used to activate the mobile device 210 and place it in the pairing mode of operation. When the button or switch is depressed and/or continuously held down, the mobile device 210 enters into the pairing mode of operation.

As reflected at block 406, the mobile device 210 or the bedside patient monitor 310 transmits a pairing signal to pair, or associate, with wireless sensor 102. The patient monitor screen 312 or mobile device screen 220 may be configured to emit a pairing signal. The mobile device screen 220 or patient monitor screen 312 may be configured to emit a visual, optical, and/or light-based pairing signal. The pairing signal transmission is received by orienting the visual, optical, and/or light-based detector 146 of the wireless sensor 102 toward the mobile device screen 220 or the patient monitor screen 312. The limited range of the visual, optical, and/or light-based pairing signal helps to prevent unintended or incidental association of the wireless sensor 102 with a mobile device 210 or bedside patient monitor 310 that might be nearby but which is not intended to be paired with the wireless sensor 102. Such circumstances can occur in hospitals, healthcare facilities, nursing homes, and the like where wireless sensors 102 mobile devices 210, and bedside patient monitors 310 are located in close physical proximity to one another.

At block 408, the wireless sensor 102 detects the pairing signal from bedside patient monitor 310 or mobile device 210. Upon detection of the pairing signal, at block 410, the wireless sensor 102 associates with the bedside patient monitor 310 thereby configuring the wireless sensor 102 and mobile device 210 or the patient monitor 310 to communicate with each other. Once the pairing is completed, the wireless sensor 102 transmits a confirmation signal confirming that the patient-worn sensor 102 is associated with the

## 12

mobile device 210 or the bedside patient monitor 310, thereby indicating that the pairing process has been successfully completed, as reflected in block 412. At block 414, the mobile device 210 or the bedside patient monitor 310 receives the confirmation signal. And at block 416A, the wireless sensor 102 exits the pairing mode of operation and enters into a patient parameter sensing mode of operation. Similarly, at block 416B, the mobile device 210 or the bedside patient monitor 310 enters a patient parameter sensing mode of operation.

FIG. 5 illustrates a method of associating a wireless sensor 102 with a mobile device 210 or the bedside patient monitor 310, which may be referred to as "pairing." At block 502A, the wireless sensor 102 may be placed near the mobile device 210 or the bedside patient monitor 310 in preparation for receiving a pairing signal. For sensors that utilize an audio sensor, such as respirator sensor, the pairing signal may be an audio sound or a series of audio sounds. A sound or audio-based detector 147 of the wireless sensor 102 may be configured to receive a pairing signal in the form of a sound or audio-based pairing signal. The sounds may be tuned to a frequency within or outside of the range of human hearing and may comprise various rings, chimes, or tones. The series of audio sounds may utilize variations in volume or tone to transmit pairing information. The signal may be unique to the wireless sensor 102 and the mobile device 210 or the patient monitor 310. By using a pairing signal that uniquely identifies the wireless sensor 102 and the mobile device 210 or the patient monitor 310, a secure connection may be established between the two paired devices.

Various types of sensors can be used with the pairing process of the present disclosure. For example, a pulse oximeter sensor can be paired by facing its light detector toward a patient monitor display 312 or mobile device display 220 to receive a visual, optical, and/or light based signal. In another example, an ulcer sensor can be paired by receiving a pairing signal in the form of detected motion. Similarly, other types of wireless sensors can be paired using the included detectors. For example, an acoustic sensor can be paired based on audio signals emitted from the mobile device 210 or the bedside patient monitor 310. EEG and ECG wireless sensors can be paired using small electrical impulses from a special conductor included as part of the patient monitor 310. Other sensors can be paired in a similar fashion depending on the specific detectors included on the wireless sensors.

Returning to block 502A, the wireless sensor 102 may be placed in proximity to the mobile device 210 or the bedside patient monitor 310 such that the sound or audio detector 147 may receive the sound or audio-based pairing signal. The sound or audio-based pairing signal has a pairing signal transmission range of up to approximately three inches. The sound or audio-based pairing signal may have a pairing signal transmission range of up to approximately six inches. The sound or audio-based pairing signal may have a pairing signal transmission range of up to approximately one foot (i.e., twelve inches) or farther. A skilled artisan will recognize that other ranges can be used for the pairing signal transmission range.

At block 504A the wireless sensor 102 is set to operate in a pairing mode. A user may begin by initiating the pairing mode of operation for the wireless sensor 102. This may include powering on the wireless sensor 102, switching the wireless sensor 102 to a special pairing state, and/or the like. For example, the wireless sensor 102 may include a battery isolator 120 which, when removed, activates the wireless sensor 102. Upon activation, the default mode of operation



is the pairing mode. The wireless sensor **102** may have a button or switch **124** that can be used to activate the wireless sensor **102** and place it in the pairing mode of operation. For example, a depressible button or switch **124** can be located on the top portion of the housing **150**. When the button or switch **124** is depressed and continuously held down, the wireless sensor **102** enters into the pairing mode of operation and remains in the pairing mode of operation for as long as the button or switch **124** is depressed. The wireless sensor **102** enters into the pairing mode by activating the sound or audio-based sensor **147**. Once activated, the sound or audio-based sensor **147** may be configured to receive the audio or sound-based pairing signal.

At block **504B**, the mobile device **210** or the bedside patient monitor **310** is set to operate in pairing mode. A user may begin by initiating the pairing mode of operation for the mobile device **210** or the bedside patient monitor **310**. This may include powering on the device, switching the device to a special pairing state, and/or the like. The bedside patient monitor **310** may have a button or switch **317** that can be used to activate the bedside patient monitor **310** and place it in the pairing mode of operation. When the button or switch **317** is depressed and/or continuously held down, the patient monitor **310** enters into the pairing mode of operation.

As reflected at block **506**, the mobile device **210** or the bedside patient monitor **310** transmits a pairing signal to pair, or associate, with wireless sensor **102**. The mobile device **210** or the bedside patient monitor **310** is configured to emit a pairing signal. The speaker of the mobile device **210** or the bedside patient monitor **310** may be configured to emit an audio signal or a series of audio sounds as the pairing signal. The pairing signal transmission is received by orienting the audio or sound-based detector **147** of the wireless sensor **102** toward the bedside patient monitor **310** or mobile device **210**. The limited range of the audio or sound-based pairing signal helps to prevent unintended or incidental association of the wireless sensor **102** with a mobile device or bedside patient monitor **310** that might be nearby but which is not intended to be paired with the wireless sensor **102**. Such circumstances can occur in hospitals, healthcare facilities, nursing homes, and the like where wireless sensors **102** and the mobile device **210** or the bedside patient monitor **310** are located in close physical proximity to one another.

At block **508**, the wireless sensor **102** detects the pairing signal from the mobile device **210** or the patient monitor **310**. Upon detection of the pairing signal, the wireless sensor **102**, at block **510**, associates with the mobile device **210** or the bedside patient monitor **310** thereby configuring the wireless sensor **102** and patient monitor **310** to communicate with each other. Once the pairing is completed, the wireless sensor **102** transmits a confirmation signal confirming that the patient-worn sensor **102** is associated with the mobile device **210** or the bedside patient monitor **310**, thereby indicating that the pairing process has been successfully completed, as reflected in block **512**. At block **514**, the mobile device **210** or the bedside patient monitor **310** receives the confirmation signal. And at block **516A**, the wireless sensor **102** exits the pairing mode of operation and enters into a patient parameter sensing mode of operation. Similarly, at block **516B**, the mobile device **210** or the bedside patient monitor **310** enters a patient parameter sensing mode of operation. In the patient parameter sensing mode of operation, the patient-worn sensor **102** transmits a patient parameter sensing signal having a patient parameter sensing signal transmission range. The wireless sensor **102** increases the power of the patient parameter sensing signal

transmission range to a standard operating range, such as for example, approximately three meters. The patient parameter sensing signal transmission range may be approximately ten feet. Alternatively, the patient parameter sensing signal transmission range may be approximately thirty feet. The pairing signal transmission range may be between approximately three and twelve inches, while the patient parameter sensing signal transmission range may be approximately ten feet. There may be at least an order of magnitude difference between the pairing signal transmission range and the patient parameter sensing signal transmission range. Thus, the pairing signal transmission range is substantially less than the patient parameter sensing transmission range. Once the wireless sensor **102** enters into the patient parameter sensing mode of operation, the wireless sensor **102** is then in condition to be placed on the patient to perform sensing and monitoring functions.

The patient monitor **310** may access the patient's health records and clinician input via a data network. Illustratively, the patients' positional history data, analyzed in view of the patient's health records, may reveal or suggest a treatment protocol) that will likely yield favorable clinical outcomes for the particular patient. Accordingly, the mobile device **210** or the bedside patient monitor **310** analyzes the accessed information in conjunction with the received information from the wireless sensor **102** to determine a recommended treatment protocol for the patient.

In another aspect of the present disclosure, the mobile device **210** or the bedside patient monitor **310** can determine a score that describes the patient's wellness/sickness state, which may also be referred to as a "Halo Index." Illustratively, the patient monitor **310** accesses and analyzes the patient's health records, clinician input, positional history data provided by the wireless sensor, surface structure pressure data, and other physiological parameter data collected and provided by the wireless sensor (such as, by way of non-limiting example, the patient's temperature, respiration rate, heart rate, ECG signal, and the like) to assess the patient's overall health condition.

An extender/repeater may be used to communicate with the wireless sensor **102** instead of the mobile device **210** or the bedside patient monitor **310**. Pairing with the extender/repeater may be performed in the same manner described above with respect to FIG. **4** or FIG. **5**.

In another aspect of the present disclosure, the pairing procedures may be applicable to user products including phones, tablets, headphones, watches, speakers, computer mice, computer keyboards, wearable devices, audio headsets, virtual reality headsets, augmented reality headsets, vehicle consoles, infotainment systems, and any other wireless communication devices known to those of skill in the art.

The wireless communication device may be a mobile phone. The mobile phone may allow voice calls to establish a data connection using a cellular network or Wi-Fi network. The mobile phone may also include features in the operating system or mobile applications that offer various functionalities for the user. The wireless communication device may be configured to collect various data such as the GPS location of the mobile phone. The mobile applications may be configured to receive collected data from wireless devices that are paired with the mobile phone.

The wireless communication device may be a wireless headphone that emits audio signals from an audio source. The audio signal may be transmitted through a wireless data connection established using the disclosed pairing procedures. The wireless communication device may be a watch,



headset, or other wear device. The device may provide functionality such as detecting the location, movement, physical activity, or physiological condition of the user. The detected data may be transmitted to another wireless device for display, storage, analysis, or other uses. The transmission of the detected data may be facilitated by a wireless data connection established using the disclosed pairing procedures. A wireless data connection may be established to facilitate the use of wireless communication device such as a wireless keyboard or wireless computer mouse. The wireless communication device may be connected to another wireless communication device such as a desktop computer or mobile device such as a laptop or tablet. The transmission of data associated with the functionality of the wireless communication device may be facilitated by a wireless data connection established using the disclosed pairing procedures.

The wireless communication device may be a vehicle console or infotainment system. The vehicle console or infotainment system may include a screen for displaying the operation of the vehicle. The screen may be a touch screen that functions as a control interface for the vehicle. The console or infotainment system may receive data such as an audio signal, video signal, GPS location, driving directions, or fare calculations from a paired wireless communication device. A skilled artisan would recognize that other data may be exchanged between the vehicle console or infotainment system and a paired wireless communication device. The transmission of data associated with the functionality of the wireless communication device and vehicle console or infotainment system may be facilitated by a wireless data connection established using the disclosed pairing procedures.

FIG. 6 illustrates a method of associating a first wireless device with a second wireless device, which may be referred to as "pairing." The first wireless device may be a wireless sensor 102 or 202 as described above. The second wireless device may be a mobile device 210 or bedside patient monitor 310 as described above. At block 602A, the first wireless device may be placed near the second wireless device in preparation for receiving a pairing signal. A visual, optical, and/or light-based detector 146 of the first wireless communication device may be physically oriented and configured to receive a pairing signal in the form of a visual, optical, and/or light-based pairing signal. The pairing signal may contain a pattern containing a shape, color, or a combination of patterns. The pairing signal may contain a series of visual, optical, and/or light based signals. The series of signals may utilize variations in color, shade, shape, or visual patterns. The pairing signal may contain a series of flashes, wherein the flashes may vary in intensity or duration. The signal may be unique to the first wireless device and/or the second wireless device. By using a pairing signal that uniquely identifies the first wireless device and/or the second wireless device, a secure connection may be established between the two paired devices.

Various types of sensors can be used with the pairing process of the present disclosure. For example, a pulse oximeter sensor can be paired by facing its light detector toward the display or screen of second wireless communication device to receive a visual, optical, and/or light based signal. In another example, an ulcer sensor can be paired by receiving a pairing signal in the form of detected motion. Similarly, other types of wireless sensors can be paired using the already included detectors that are used for physiological detection or other detection of the surrounding environment during normal use. For example, an acoustic sensor can be

paired based on audio signals emitted from the second wireless communications device. EEG and ECG wireless sensors can be paired using small electrical impulses from a special conductor included as part of the second wireless communications device. Other sensors can be paired in a similar fashion depending on the specific detectors included on the first wireless communication device.

Returning to block 602A, the first wireless device may be placed in proximity to the second wireless device such that the visual, optical, and/or light-based detector 146 may receive the visual, optical, and/or light-based pairing signal. The visual, optical, and/or light-based pairing signal has a pairing signal transmission range of up to approximately three inches. The visual, optical, and/or light-based pairing signal has a pairing signal transmission range of up to approximately six inches. The visual, optical, and/or light-based pairing signal has a pairing signal transmission range of up to approximately one foot (i.e., twelve inches) or farther. A skilled artisan will recognize that other ranges can be used for the pairing signal transmission range.

At block 604A the first wireless device is set to operate in a pairing mode. A user may begin by initiating the pairing mode of operation for the first wireless device. This may include powering on the first wireless device, switching the first wireless device to a special pairing state, and/or the like. For example, the first wireless device may include a battery isolator 120 which, when removed, activates the first wireless device. Upon activation, the default mode of operation is the pairing mode. The first wireless device may have a button/switch 124 that can be used to activate the first wireless device and place it in the pairing mode of operation. For example, a depressible button/switch 124 can be located on the top portion of the housing 150. When the button/switch 124 is depressed and continuously held down, the first wireless device enters into the pairing mode of operation and remains in the pairing mode of operation for as long as the button/switch 124 is depressed. The first wireless device may enter into the pairing mode by activating the sound or audio-based sensor 147. Once activated, the sound or audio-based sensor 147 may be configured to receive the audio or sound-based pairing signal.

At block 604B, the second wireless device is set to operate in pairing mode. A user may begin by initiating the pairing mode of operation for the second wireless device. This may include powering on the device, switching the device to a special pairing state, and/or the like. The second wireless device may have a button or switch that can be used to activate the second wireless device and place it in the pairing mode of operation. When the button or switch is depressed or continuously held down, the second wireless device enters into the pairing mode of operation.

As reflected at block 606, the second wireless device transmits a pairing signal to pair, or associate, with first wireless device. The second wireless communications device may be configured to emit a pairing signal. The display or screen of the second wireless communication device may be configured to emit a visual, optical, and/or light-based pairing signal. The pairing signal transmission is received by orienting the visual, optical, and/or light-based detector 146 of the first wireless communication device toward the display or screen. The limited range of the visual, optical, and/or light-based pairing signal helps to prevent unintended or incidental association of the first wireless communication device with a second wireless communication device that might be nearby but which is not intended to be paired with the first wireless communication device. Such circumstances can occur in residential buildings, office



buildings, commercial facilities, airports, public transportation facilities, hospitals, healthcare facilities, nursing homes, and the like where the first wireless communications device and second communications device are located in close physical proximity to one another.

At block **608**, the first wireless device detects the pairing signal from second wireless device. Upon detection of the pairing signal, the first wireless device, at block **610**, associates with the second wireless device thereby configuring the first wireless device and second wireless device to communicate with each other. Once the pairing is completed, the first wireless device transmits a confirmation signal confirming that the first wireless communication device is associated with the second wireless device, thereby indicating that the pairing process has been successfully completed, as reflected in block **612**. At block **614**, the second wireless device receives the confirmation signal. And at block **616A**, the first wireless device exits the pairing mode of operation and enters into a patient parameter sensing mode of operation. Similarly, at block **616B**, the second wireless device enters a patient parameter sensing mode of operation.

FIG. 7 illustrates another method of associating a first wireless device with a second wireless device, which may be referred to as "pairing." The first wireless device may be a wireless sensor **102** or **202** as described above. The second wireless device may be a mobile device **210** or a bedside patient monitor **310** as described above.

At block **702A**, the first wireless device may be placed near the second wireless device in preparation for receiving a pairing signal. For sensors that utilize an audio sensor, such as respirator sensor, the pairing signal may be an audio sound or a series of audio sounds. A sound or audio-based detector **147** of the first wireless device may be configured to receive a pairing signal in the form of a sound or audio-based pairing signal. The sounds may be tuned to a frequency within or outside of the range of human hearing and may comprise various rings, chimes, or tones. The series of audio sounds may utilize variations in volume or tone to transmit pairing information. The signal may be unique to the first wireless device and/or the second wireless device. By using a pairing signal that uniquely identifies the first wireless device and/or the second wireless device, a secure connection may be established between the two paired devices.

Various types of sensors can be used with the pairing process of the present disclosure. For example, a pulse oximeter sensor can be paired by facing its light detector toward the display or screen of a second wireless communication device to receive a visual, optical, and/or light based signal. In another example, an ulcer sensor can be paired by receiving a pairing signal in the form of detected motion. Similarly, other types of wireless sensors can be paired using the included detectors. For example, an acoustic sensor can be paired based on audio signals emitted from the second wireless communications device. EEG and ECG wireless sensors can be paired using small electrical impulses from a special conductor included as part of the second wireless communications device. Other sensors can be paired in a similar fashion depending on the specific detectors included on the first wireless communication device.

Returning to block **702A**, the first wireless device may be placed in proximity to the second wireless communications device such that the sound or audio detector may receive the sound or audio-based pairing signal. The sound or audio-based pairing signal may have a pairing signal transmission

range of up to approximately three inches. The sound or audio-based pairing signal has a pairing signal transmission range of up to approximately six inches. The sound or audio-based pairing signal may have a pairing signal transmission range of up to approximately one foot (i.e., twelve inches) or farther. A skilled artisan will recognize that other ranges can be used for the pairing signal transmission range.

At block **704A** the first wireless device is set to operate in a pairing mode. A user may begin by initiating the pairing mode of operation for the first wireless device. This may include powering on the first wireless device, switching the first wireless device to a special pairing state, and/or the like. For example, the first wireless device may include a battery isolator **120** which, when removed, activates the first wireless device. Upon activation, the default mode of operation is the pairing mode. The first wireless device may have a button/switch **124** that can be used to activate the first wireless device and place it in the pairing mode of operation. For example, a depressible button/switch **124** can be located on the top portion of the housing **150**. When the button/switch **124** is depressed and continuously held down, the first wireless device enters into the pairing mode of operation and remains in the pairing mode of operation for as long as the button or switch **124** is depressed. The wireless sensor can be placed in pairing mode by being shaken, bounced, or by shining a bright light at the detector. The first wireless device enters into the pairing mode by activating the sound or audio-based sensor **147**. Once activated, the sound or audio-based sensor **147** may be configured to receive the audio or sound-based pairing signal.

At block **704B**, the second wireless device is set to operate in pairing mode. A user may begin by initiating the pairing mode of operation for the second wireless device. This may include powering on the device, switching the device to a special pairing state, and/or the like. The second wireless device may have a button/switch that can be used to activate the second wireless device and place it in the pairing mode of operation. When the button/switch is depressed and/or continuously held down, the second wireless device enters into the pairing mode of operation.

As reflected at block **706**, the second wireless device transmits a pairing signal to pair, or associate, with first wireless device. The second wireless communications device may be configured to emit a pairing signal. The speaker of the second wireless communications device may be configured to emit an audio signal or a series of audio sounds as the pairing signal. The pairing signal transmission is received by orienting the audio or sound-based detector **147** of the first wireless device toward the second wireless communications device. The limited range of the audio or sound-based pairing signal helps to prevent unintended or incidental association of the first wireless device with a second wireless device that might be nearby but which is not intended to be paired with the first wireless device. Such circumstances can occur in residential buildings, office buildings, commercial facilities, airports, public transportation facilities, hospitals, healthcare facilities, nursing homes, and the like where the first wireless communications device and second communications device are located in close physical proximity to one another.

At block **708**, the first wireless device detects the pairing signal from second wireless device. Upon detection of the pairing signal, the first wireless device, at block **710**, associates with the second wireless device thereby configuring the first wireless device and second wireless device to communicate with each other. Once the pairing is completed, the first wireless device transmits a confirmation



signal confirming that the first wireless communication device is associated with the second wireless device, thereby indicating that the pairing process has been successfully completed, as reflected in block 712. At block 714, the second wireless device receives the confirmation signal. And at block 716A, the first wireless device exits the pairing mode of operation and enters into a patient parameter sensing mode of operation. Similarly, at block 716B, the second wireless device enters a patient parameter sensing mode of operation.

FIG. 8 illustrates a method of associating a wireless sensor 102 with a wireless communication device 810, which may be referred to as “pairing.” A wireless sensor 102 may include a motion sensor. The wireless communication device 810 may include a device screen 820. Here, the pairing signal may involve displaying instructions on device screen 820 for a user to follow. The instructions may indicate steps that a user must follow in order to complete the pairing process. In one example, the instructions may be a sequence of motions that the user must perform on the wireless sensor 802. The instructions may be displayed using objects 830, 831, 832 on device screen 820 representing the silhouette of the wireless sensor 102. The objects 830, 831, 832 are displayed on the device screen 820 in sequence, one at a time, along with an arrow representing the direction the wireless sensor 102 should move. To follow the instructions, the user may hold the wireless sensor 102 and move the position of the wireless sensor 102 according to the instructions. For example, to complete STEP 1 of the pairing instruction, the user will move the wireless sensor 102 in an upward motion relative to the device screen 820. The change in position of first wireless communication device 820 will resemble the relative change in position from object 830 to object 831 on the device screen 820. To complete STEP 2 of the pairing instruction, the user will move the wireless sensor 102 in a leftward motion relative to the device screen 820. The change in position of the wireless communication device 820 will resemble the relative change in position from object 831 to object 832 on the device screen 820. The motion detector of the wireless communication device 810 detects each step of the motion. The detected motion is then used as the pairing signal to associate the wireless sensor 102 with wireless communication device 810. One skilled in the art will appreciate that other types and sequences of motion may be used as the pairing signal.

FIG. 9 illustrates a method of associating a first wireless communication device with a second wireless communication device. The first wireless communication device may be a bluetooth headset 902. The wireless communication device may be a vehicle console or infotainment system 910. The vehicle console or infotainment system 910 may include a screen 912 for displaying information related to the operation of the vehicle and control interface. The console or infotainment system 910 may receive data such as an audio signal, video signal, GPS location, driving directions, or fare calculations from a paired wireless communication device 902. A skilled artisan would recognize that other data may be exchanged between the vehicle console or infotainment system and a paired wireless communication device.

The transmission of data associated with the functionality of the wireless communication device 902 and vehicle console or infotainment system 910 may be facilitated by a wireless data connection established using a pairing procedure. The pairing procedure may include using a pairing signal to associate the first wireless communication device 902 with vehicle console or infotainment system 910. The

pairing signal may contain a pattern containing a shape, color, or a combination of patterns. The pairing signal may contain a series of visual, optical, and/or light based signals. The series of signals may utilize variations in color, shade, shape, or visual patterns. The pairing signal may contain a series of flashes, wherein the flashes may vary in intensity or duration. The wireless communication device 902 may have a light 906 that emits various visual or light based signals. Similarly, the wireless communication device 902 may include a speaker 908 that may generate audio-based pairing signals. The sounds may be tuned to a frequency within or outside of the range of human hearing and may comprise various rings or tones. The pairing signal may comprise a combination of visual signals emitted from wireless communication device light 906 and audio signals emitted from wireless communication device speaker 908. The pairing signal may include a sequence of visual signals emitted from console screen 912 that is synchronized with a sequence of audio signals emitted from console speaker. Additionally, the wireless communication device 902 may have a speaker 904 configured to be placed in the user’s ear. The speaker 904 or the speaker 908 may transmit a confirmatory signal that the device has been successfully paired.

The pairing process may be similar to the methods described above. The screen 912 of the vehicle console or infotainment system 910. The pairing signal may involve displaying instructions on the screen 912 for a user to follow. The instructions may include placing a wireless communication device 902 of a particular shape and size in a certain position relative to the screen 912. The user may follow the instructions and hold the wireless communication device 902 in a certain position relative to the screen 912 to pair the wireless communication device 902 to the vehicle console or infotainment system 910.

Additionally, the console screen 912 may be configured to display motion instructions for a user to perform in order to generate motion signals representing a pairing signal. Additionally, the console or infotainment system 910 may have a port for connecting peripheral devices that may generate various signals such as current or voltage based signals. The pair signal may take the form of an electrical signal.

The vehicle console or infotainment system 910 can have a switch or button 916 which, when depressed, places the wireless communication device 902 in a pairing mode of operation, causing the wireless communication device 902 to wait for a pairing signal. Similarly, the vehicle console or infotainment system 910 can have a switch or button 918 on the steering wheel which, when depressed, places the wireless communication device 902 in a pairing mode of operation, causing the wireless communication device 902 to wait for a pairing signal.

Many other variations than those described herein will be apparent from this disclosure. For example, certain acts, events, or functions of any of the algorithms described herein can be performed in a different sequence, can be added, merged, or left out altogether (e.g., not all described acts or events are necessary for the practice of the algorithms). Moreover, acts or events can be performed concurrently, e.g., through multi-threaded processing, interrupt processing, or multiple processors or processor cores or on other parallel architectures, rather than sequentially. In addition, different tasks or processes can be performed by different machines and/or computing systems that can function together.

The first wireless communication device and second wireless communication device may be paired without additional components. The first wireless communication may



be a wireless sensor. The wireless sensor may be a variety of sensors as described herein, such as a magnetometer which may also be referred to as a compass, a temperature sensor, an acoustic respiration sensor, an electrocardiogram (ECG) sensor, an electroencephalography (EEG) sensor, one or more pulse oximetry sensors, a moisture sensor, a blood pressure sensor, and an impedance sensor. The second wireless communication device may be a mobile device **210** or a bedside patient monitor **310**. Similarly, a variety of wireless sensors described herein may be paired to a wireless communication device such as a mobile device or a patient monitoring system. Further, the wireless communication device may also other devices such as phones, tablets, headphones, watches, speakers, computer mice, computer keyboards, wearable devices, audio headsets, virtual reality headsets, augmented reality headsets, vehicle consoles, infotainment systems, and any other wireless communication devices known to those of skill in the art.

The methods and systems described herein can be implemented without any additional hardware components. There is no installation of additional components required. This can provide ease of use and implementation such that users can use devices they are familiar with without additional components. For example, as described above, the shape of the device may be used as a pairing signal. In another example, the vibration of an acoustic sensor may be used as a pairing signal. The properties of the sensors or wireless communication devices can be utilized in the pairing process by the system and methods described herein.

The methods and systems for pairing described herein can also be implemented without the use of wireless communication protocols. For example, the pairing methods and systems can be implemented without the use of protocols such as a Bluetooth protocol, a wifi protocol, or a zigbee protocol.

The various illustrative logical blocks, modules, and algorithm steps described in connection with the disclosure herein can be implemented as electronic hardware, computer software, or combinations of both. To clearly illustrate this interchangeability of hardware and software, various illustrative components, blocks, modules, and steps have been described above generally in terms of their functionality. Whether such functionality is implemented as hardware or software depends upon the particular application and design constraints imposed on the overall system. The described functionality can be implemented in varying ways for each particular application, but such implementation decisions should not be interpreted as causing a departure from the scope of the disclosure.

The various illustrative logical blocks and modules described in connection with the disclosure herein can be implemented or performed by a machine, such as a general purpose processor, a digital signal processor (DSP), an application specific integrated circuit (ASIC), a field programmable gate array (FPGA) or other programmable logic device, discrete gate or transistor logic, discrete hardware components, or any combination thereof designed to perform the functions described herein. A general purpose processor can be a microprocessor, but in the alternative, the processor can be a controller, microcontroller, or state machine, combinations of the same, or the like. A processor can include electrical circuitry configured to process computer-executable instructions. A processor may include an FPGA or other programmable device that performs logic operations without processing computer-executable instructions. A processor can also be implemented as a combination of computing devices, e.g., a combination of a DSP and a

microprocessor, a plurality of microprocessors, one or more microprocessors in conjunction with a DSP core, or any other such configuration. A computing environment can include any type of computer system, including, but not limited to, a computer system based on a microprocessor, a mainframe computer, a digital signal processor, a portable computing device, a device controller, or a computational engine within an appliance, to name a few.

The steps of a method, process, or algorithm described in connection with the disclosure herein can be embodied directly in hardware, in a software module stored in one or more memory devices and executed by one or more processors, or in a combination of the two. A software module can reside in RAM memory, flash memory, ROM memory, EPROM memory, EEPROM memory, registers, hard disk, a removable disk, a CD-ROM, or any other form of non-transitory computer-readable storage medium, media, or physical computer storage known in the art. An example storage medium can be coupled to the processor such that the processor can read information from, and write information to, the storage medium. In the alternative, the storage medium can be integral to the processor. The storage medium can be volatile or nonvolatile. The processor and the storage medium can reside in an ASIC.

Conditional language used herein, such as, among others, “can,” “might,” “may,” “e.g.,” and the like, unless specifically stated otherwise, or otherwise understood within the context as used, is generally intended to convey that certain examples include, while other examples do not include, certain features, elements and/or states. Thus, such conditional language is not generally intended to imply that features, elements and/or states are in any way required for one or more examples or that one or more examples necessarily include logic for deciding, with or without author input or prompting, whether these features, elements and/or states are included or are to be performed in any particular example. The terms “comprising,” “including,” “having,” and the like are synonymous and are used inclusively, in an open-ended fashion, and do not exclude additional elements, features, acts, operations, and so forth. Also, the term “or” is used in its inclusive sense (and not in its exclusive sense) so that when used, for example, to connect a list of elements, the term “or” means one, some, or all of the elements in the list. Further, the term “each,” as used herein, in addition to having its ordinary meaning, can mean any subset of a set of elements to which the term “each” is applied.

While the above detailed description has shown, described, and pointed out novel features as applied to various examples, it will be understood that various omissions, substitutions, and changes in the form and details of the systems, devices or methods illustrated can be made without departing from the spirit of the disclosure. As will be recognized, certain examples described herein can be embodied within a form that does not provide all of the features and benefits set forth herein, as some features can be used or practiced separately from others.

The term “and/or” herein has its broadest, least limiting meaning which is the disclosure includes A alone, B alone, both A and B together, or A or B alternatively, but does not require both A and B or require one of A or one of B. As used herein, the phrase “at least one of” A, B, “and” C should be construed to mean a logical A or B or C, using a non-exclusive logical or.

The apparatuses and methods described herein may be implemented by one or more computer programs executed by one or more processors. The computer programs include processor-executable instructions that are stored on a non-



transitory tangible computer readable medium. The computer programs may also include stored data. Non-limiting examples of the non-transitory tangible computer readable medium are nonvolatile memory, magnetic storage, and optical storage.

Although the foregoing disclosure has been described in terms of certain preferred examples, other examples will be apparent to those of ordinary skill in the art from the disclosure herein. Additionally, other combinations, omissions, substitutions and modifications will be apparent to the skilled artisan in view of the disclosure herein. Accordingly, the present invention is not intended to be limited by the description of the preferred examples, but is to be defined by reference to claims.

What is claimed is:

**1.** A system for electronically pairing a wireless physiological sensor with a patient monitoring device, the system comprising:

a patient monitoring device comprising a first display;  
a wireless physiological sensor, the wireless physiological sensor comprising a button configured to activate a pairing mode that enables the wireless physiological sensor to electronically pair with the patient monitoring device, the wireless physiological sensor further comprising an optical detector configured to detect light based signals associated with physiological parameters of a patient; and

one or more hardware processors configured to:

generate a visual signal from the first display;  
detect the visual signal with the same optical detector of the same wireless physiological sensor configured to detect light based signals associated with physiological parameters of the patient;

associate the wireless physiological sensor with the patient monitoring device based on the detected visual signal, thereby pairing the wireless physiological sensor with the patient monitoring device;

transmit a confirmation signal from the wireless physiological sensor to indicate that association is complete; and

configure the wireless physiological sensor to enter into a patient parameter sensing mode of operation after the association of the wireless physiological sensor with the patient monitoring device.

**2.** The system of claim **1**, wherein the first display is of a first size and the wireless physiological sensor is of a second size, wherein the second size of the wireless sensor is smaller than the first size of the first display.

**3.** The system of claim **2**, wherein the size of the wireless physiological sensor corresponds to a shape of a base of the wireless physiological sensor.

**4.** The system of claim **1**, wherein the wireless physiological sensor does not require a separate antenna or any additional components for the pairing with the patient monitoring device.

**5.** The system of claim **1**, wherein the wireless physiological sensor does not use a wireless communication protocol for the pairing with the patient monitoring device.

**6.** The system of claim **5**, wherein the wireless communication protocol comprises at least one of: a Bluetooth protocol, wifi protocol, or a zigbee protocol.

**7.** The system of claim **1**, wherein the one or more hardware processors are configured to detect a shape of the wireless physiological sensor when the wireless sensor is placed directly on the first display and in response to the detected shape, associate the wireless physiological sensor with the patient monitoring device.

**8.** The system of claim **1**, wherein the one or more hardware processors are configured to generate a pattern on the first display and associate the wireless physiological sensor with the patient monitoring device based on a successful placement of the wireless physiological sensor on the first display in relation to the generated pattern.

**9.** A system for electronically pairing a wireless physiological sensor with a patient monitoring device, the system comprising:

a patient monitoring device comprising a first display;  
a wireless physiological sensor comprising a detector configured to detect light based signals associated with a physiological parameter of a patient, the wireless physiological sensor comprising a button configured to activate a pairing mode that enables the wireless physiological sensor to electronically pair with the patient monitoring device; and

one or more hardware processors configured to:

generate a signal from the patient monitoring device;  
detect the signal with the same detector of the same wireless physiological sensor configured to detect light based signals associated with a physiological parameter of a patient;

associate the wireless sensor with the patient monitoring device based on the detected signal, thereby pairing the wireless physiological sensor with the patient monitoring device;

transmit a confirmation signal from the wireless physiological sensor to indicate that association is complete; and

configure the wireless physiological sensor to enter into a patient parameter sensing mode of operation after the association of the wireless physiological sensor with the patient monitoring device.

**10.** The system of claim **9**, wherein the wireless physiological sensor does not require a separate antenna or any additional components for the pairing with the patient monitoring device.

**11.** The system of claim **9**, wherein the wireless physiological sensor does not use a wireless communication protocol for the pairing with the patient monitoring device.

**12.** A method for electronically pairing a wireless physiological sensor with a patient monitoring device, the method comprising:

generating a signal from the patient monitoring device;  
detecting the signal with a detector of the wireless physiological sensor configured to detect light-based signals associated with a physiological parameter of a patient;

associating the wireless physiological sensor with the patient monitoring device based on the detected signal, thereby pairing the wireless physiological sensor with the patient monitoring device;

transmitting a confirmation signal from the wireless physiological sensor to indicate that association is complete; and

configuring the wireless physiological sensor to enter into a patient parameter sensing mode of operation after the association of the wireless physiological sensor with the patient monitoring device.

**13.** The method of claim **12**, wherein the wireless physiological sensor does not require a separate antenna or any additional components for the pairing with the patient monitoring device.

**14.** The method of claim **12**, wherein the wireless physiological sensor does not use a wireless communication protocol for the pairing with the patient monitoring device.



15. The method of claim 12, wherein the detector comprises an optical detector.

16. The method of claim 12, further comprising generating a pattern on a first display of the patient monitoring device, wherein the association is further based on a successful placement of the wireless physiological sensor on the first display in relation to the generated pattern.

\* \* \* \* \*