

(12) **United States Patent**
Aboytes et al.

(10) **Patent No.:** **US 10,898,200 B2**
(45) **Date of Patent:** **Jan. 26, 2021**

(54) **DEVICES AND METHODS FOR THE TREATMENT OF VASCULAR DEFECTS**

(71) Applicant: **Covidien LP**, Mansfield, MA (US)

(72) Inventors: **Maria G. Aboytes**, Palo Alto, CA (US);
Arturo S. Rosqueta, San Jose, CA (US)

(73) Assignee: **Covidien LP**, Mansfield, MA (US)

(*) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 148 days.

(21) Appl. No.: **16/147,883**

(22) Filed: **Oct. 1, 2018**

(65) **Prior Publication Data**

US 2019/0105054 A1 Apr. 11, 2019

Related U.S. Application Data

(63) Continuation of application No. 15/846,964, filed on Dec. 19, 2017, which is a continuation of application (Continued)

(51) **Int. Cl.**

A61B 17/12 (2006.01)
A61M 29/00 (2006.01)
A61B 17/00 (2006.01)

(52) **U.S. Cl.**

CPC .. **A61B 17/12113** (2013.01); **A61B 17/12163** (2013.01); **A61B 17/12168** (2013.01); (Continued)

(58) **Field of Classification Search**

CPC **A61B 17/12031**; **A61B 17/12113**; **A61B 17/1214**; **A61B 17/12145**; (Continued)

(56) **References Cited**

U.S. PATENT DOCUMENTS

5,250,071 A 10/1993 Palermo
5,354,295 A 10/1994 Guglielmi et al.
(Continued)

FOREIGN PATENT DOCUMENTS

CN 200991273 Y 12/2007
DE 102011102933 12/2012
(Continued)

OTHER PUBLICATIONS

Extended European Search Report dated Sep. 19, 2019; European Patent Application No. 19183387.0; 6 pages.

(Continued)


Primary Examiner — Robert A Lynch

(74) *Attorney, Agent, or Firm* — Fortem IP LLP; Mary L. Fox

(57) **ABSTRACT**

Devices and methods for treating vascular defects, such as, for example, balloon-type aneurysms, are described herein. In one embodiment, an apparatus includes an insertion portion and an expandable implant. The expandable implant is configured to be deployed in an aneurysm and is coupled to the insertion portion. The expandable implant has a first portion and a second portion coupled to the first portion. The expandable implant is movable between a first configuration in which the first portion and the second portion are substantially linearly aligned and a second configuration in which the second portion at least partially overlaps the first portion.

22 Claims, 42 Drawing Sheets



Related U.S. Application Data

No. 15/162,073, filed on May 23, 2016, now Pat. No. 10,064,627, which is a continuation of application No. 14/661,233, filed on Mar. 18, 2015, now Pat. No. 9,855,051, which is a continuation of application No. 13/727,029, filed on Dec. 26, 2012, now Pat. No. 8,998,947, which is a continuation-in-part of application No. 13/421,122, filed on Mar. 15, 2012, now abandoned, which is a continuation-in-part of application No. 13/230,628, filed on Sep. 12, 2011, now Pat. No. 8,974,512.

- (60) Provisional application No. 61/381,770, filed on Sep. 10, 2010.
- (52) **U.S. Cl.**
CPC *A61B 17/12172* (2013.01); *A61M 29/00* (2013.01); *H05K 999/99* (2013.01); *A61B 17/1214* (2013.01); *A61B 2017/00867* (2013.01); *A61B 2017/12054* (2013.01)
- (58) **Field of Classification Search**
CPC A61B 17/12168; A61B 17/12172; A61B 2017/1205; A61B 2017/12054
See application file for complete search history.

(56) **References Cited**

U.S. PATENT DOCUMENTS

5,645,558 A 7/1997 Horton
 5,669,931 A 9/1997 Kupiecki et al.
 5,725,552 A 3/1998 Kotula et al.
 5,733,294 A 3/1998 Forber et al.
 5,741,333 A 4/1998 Frid
 5,749,891 A 5/1998 Ken et al.
 5,749,919 A 5/1998 Blanc
 5,814,062 A 9/1998 Sepetka et al.
 5,911,731 A 6/1999 Pham et al.
 5,916,235 A 6/1999 Guglielmi
 5,925,060 A 7/1999 Forber
 5,928,260 A 7/1999 Chin et al.
 5,935,148 A 8/1999 Villar et al.
 5,944,738 A 8/1999 Amplatz et al.
 5,951,599 A 9/1999 McCrory
 5,964,797 A 10/1999 Ho
 5,976,169 A 11/1999 Imran
 5,980,554 A 11/1999 Lenker et al.
 6,022,374 A 2/2000 Imran
 6,033,423 A 3/2000 Ken et al.
 6,036,720 A 3/2000 Abrams et al.
 6,063,070 A 5/2000 Eder
 6,063,104 A 5/2000 Villar et al.
 6,086,577 A 7/2000 Ken et al.
 6,090,125 A 7/2000 Horton
 6,093,199 A 7/2000 Brown et al.
 6,096,021 A 8/2000 Helm et al.
 6,123,715 A 9/2000 Amplatz
 6,139,564 A 10/2000 Teoh
 6,152,144 A 11/2000 Lesh et al.
 6,159,531 A 12/2000 Dang et al.
 6,168,615 B1 1/2001 Ken et al.
 6,168,622 B1 1/2001 Mazzocchi
 6,309,367 B1 10/2001 Boock
 6,344,048 B1 2/2002 Chin et al.
 6,346,117 B1 2/2002 Greenhalgh
 6,350,270 B1 2/2002 Roue
 6,368,339 B1 4/2002 Amplatz
 6,371,980 B1 4/2002 Rudakov et al.
 6,375,668 B1 4/2002 Gifford et al.
 6,383,174 B1 5/2002 Eder
 6,391,037 B1 5/2002 Greenhalgh
 6,428,558 B1 8/2002 Jones et al.
 6,447,531 B1 9/2002 Amplatz

6,451,050 B1 9/2002 Rudakov et al.
 6,454,780 B1 9/2002 Wallace
 6,494,884 B2 12/2002 Gifford, III et al.
 6,506,204 B2 1/2003 Mazzocchi
 6,511,468 B1 1/2003 Cragg et al.
 6,547,804 B2 4/2003 Porter et al.
 6,551,303 B1 4/2003 Van Tassel et al.
 6,579,303 B2 6/2003 Amplatz
 6,585,748 B1 7/2003 Jeffree
 6,589,256 B2 7/2003 Forber
 6,589,265 B1 7/2003 Palmer et al.
 6,599,308 B2 7/2003 Amplatz
 6,602,261 B2 8/2003 Greene, Jr. et al.
 6,605,101 B1 8/2003 Schaefer et al.
 6,605,102 B1 8/2003 Mazzocchi et al.
 6,605,111 B2 8/2003 Bose et al.
 6,613,074 B1 9/2003 Mittelberq et al.
 6,626,939 B1 9/2003 Burnside et al.
 6,635,068 B1 10/2003 Dubrul et al.
 6,652,556 B1 11/2003 VanTassel et al.
 6,666,882 B1 12/2003 Bose et al.
 6,669,721 B1 12/2003 Bose et al.
 6,682,546 B2 1/2004 Amplatz
 6,689,486 B2 2/2004 Ho et al.
 6,730,108 B2 5/2004 Van Tassel et al.
 6,746,468 B1 6/2004 Sepetka et al.
 6,746,890 B2 6/2004 Gupta et al.
 6,780,196 B2 8/2004 Chin et al.
 6,802,851 B2 10/2004 Jones et al.
 6,811,560 B2 11/2004 Jones et al.
 6,855,153 B2 2/2005 Saadat
 6,855,154 B2 2/2005 Abdel-Gawwad
 6,878,384 B2 4/2005 Cruise et al.
 6,905,503 B2 6/2005 Gifford, III et al.
 6,936,055 B1 8/2005 Ken et al.
 6,991,617 B2 1/2006 Hektner et al.
 6,994,092 B2 2/2006 van der Burg et al.
 6,994,717 B2 2/2006 Konya et al.
 7,011,671 B2 3/2006 Welch
 7,029,487 B2 4/2006 Greene, Jr. et al.
 7,083,632 B2 8/2006 Avellanet et al.
 7,128,073 B1 10/2006 van der Burg et al.
 7,128,736 B1 10/2006 Abrams et al.
 7,169,177 B2 1/2007 Obara
 7,195,636 B2 3/2007 Avellanet et al.
 7,229,461 B2 6/2007 Chin et al.
 7,232,461 B2 6/2007 Ramer
 7,326,225 B2 2/2008 Ferrera et al.
 7,331,980 B2 2/2008 Dubrul et al.
 7,419,503 B2 9/2008 Pulnev et al.
 7,597,704 B2 10/2009 Frazier et al.
 7,601,160 B2 10/2009 Richter
 7,708,754 B2 5/2010 Balgobin et al.
 7,879,064 B2 2/2011 Monstadt et al.
 RE42,625 E 8/2011 Guglielmi
 8,043,326 B2 10/2011 Hancock et al.
 8,142,456 B2 3/2012 Rosqueta et al.
 8,211,160 B2 7/2012 Garrison et al.
 8,333,783 B2 12/2012 Braun et al.
 8,343,167 B2 1/2013 Henson
 8,351,104 B2 1/2013 Jones et al.
 8,351,138 B2 1/2013 Adams
 8,361,104 B2 1/2013 Jones et al.
 8,361,138 B2 1/2013 Adams
 8,425,541 B2 4/2013 Masters et al.
 8,470,013 B2 6/2013 Duggal et al.
 8,715,317 B1 5/2014 Janardhan et al.
 8,834,515 B2 9/2014 Nin et al.
 8,906,057 B2 12/2014 Connor et al.
 8,974,512 B2 3/2015 Aboytes et al.
 8,998,947 B2 4/2015 Aboytes et al.
 9,211,202 B2 12/2015 Strother et al.
 9,486,224 B2 11/2016 Riina et al.
 9,833,309 B2 12/2017 Levi et al.
 9,844,380 B2 12/2017 Furey
 9,844,382 B2 12/2017 Aboytes et al.
 9,855,051 B2 1/2018 Aboytes et al.
 9,907,684 B2 3/2018 Connor et al.
 9,962,146 B2 5/2018 Hebert et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

			2009/0287291	A1	11/2009	Becking et al.
			2009/0287292	A1	11/2009	Becking et al.
			2009/0287294	A1*	11/2009	Rosqueta A61B 17/12022 623/1.15
			2009/0297582	A1	12/2009	Meyer et al.
			2009/0318892	A1	12/2009	Aboytes et al.
			2009/0318941	A1	12/2009	Sepetka et al.
			2009/0319023	A1	12/2009	Hildebrand et al.
			2010/0030200	A1	2/2010	Strauss et al.
			2010/0036410	A1	2/2010	Krolik et al.
			2010/0139465	A1	6/2010	Christian et al.
			2010/0144895	A1	6/2010	Porter
			2010/0174269	A1	7/2010	Tompkins et al.
			2010/0185271	A1	7/2010	Zhang
			2010/0256527	A1	10/2010	Lippert et al.
			2010/0256528	A1	10/2010	Lippert et al.
			2010/0256601	A1	10/2010	Lippert et al.
			2010/0256602	A1	10/2010	Lippert et al.
			2010/0256603	A1	10/2010	Lippert et al.
			2010/0256604	A1	10/2010	Lippert et al.
			2010/0256605	A1	10/2010	Lippert et al.
			2010/0256606	A1	10/2010	Lippert et al.
			2010/0262014	A1	10/2010	Huang
			2010/0268201	A1	10/2010	Tieu et al.
			2011/0022149	A1	1/2011	Cox et al.
			2011/0046658	A1	2/2011	Connor et al.
			2011/0077620	A1	3/2011	deBeer
			2011/0137332	A1	6/2011	Sepetka et al.
			2011/0137405	A1	6/2011	Wilson et al.
			2011/0144669	A1	6/2011	Becking et al.
			2011/0152993	A1	6/2011	Marchand et al.
			2011/0202085	A1	8/2011	Loganathan et al.
			2011/0208227	A1	8/2011	Becking
			2011/0224776	A1	9/2011	Sepetka et al.
			2011/0265943	A1	11/2011	Rosqueta et al.
			2011/0319926	A1	12/2011	Becking et al.
			2012/0022572	A1	1/2012	Braun et al.
			2012/0041472	A1	2/2012	Tan et al.
			2012/0101510	A1	4/2012	Lenker et al.
			2012/0116350	A1	5/2012	Strauss et al.
			2012/0165919	A1	6/2012	Cox et al.
			2012/0197283	A1	8/2012	Marchand et al.
			2012/0239074	A1	9/2012	Aboytes et al.
			2012/0283769	A1	11/2012	Cruise et al.
			2012/0316598	A1	12/2012	Becking et al.
			2012/0316632	A1	12/2012	Gao
			2012/0330341	A1	12/2012	Becking et al.
			2012/0330347	A1	12/2012	Becking et al.
			2012/0330348	A1	12/2012	Strauss et al.
			2013/0066357	A1	3/2013	Aboytes et al.
			2013/0066360	A1	3/2013	Becking et al.
			2013/0085522	A1	4/2013	Becking et al.
			2013/0116722	A1	5/2013	Aboytes et al.
			2013/0253572	A1	9/2013	Molaei et al.
			2013/0274866	A1	10/2013	Cox et al.
			2014/0012307	A1	1/2014	Franano et al.
			2014/0058420	A1	2/2014	Hannes et al.
			2014/0316012	A1	10/2014	Freyman et al.
			2014/0371734	A1	12/2014	Truckai
			2015/0216684	A1	8/2015	Enzmann et al.
			2015/0250628	A1	9/2015	Monstadt et al.
			2015/0272590	A1	10/2015	Aboytes et al.
			2015/0313737	A1	11/2015	Tippett et al.
			2015/0327843	A1	11/2015	Garrison
			2015/0342613	A1	12/2015	Aboytes et al.
			2016/0066921	A1	3/2016	Seifert et al.
			2016/0135984	A1	5/2016	Rudakov et al.
			2016/0206320	A1	7/2016	Connor
			2016/0206321	A1	7/2016	Connor
			2016/0262766	A1	9/2016	Aboytes et al.
			2017/0150971	A1	6/2017	Hines
			2017/0156903	A1	6/2017	Shobayashi
			2017/0189035	A1	7/2017	Porter
			2017/0266023	A1	9/2017	Thomas
			2017/0340333	A1	11/2017	Badruddin et al.
			2017/0367708	A1	12/2017	Mayer et al.
			2018/0049859	A1	2/2018	Stoppenhagen et al.
			2018/0125686	A1	5/2018	Lu
			2018/0140305	A1	5/2018	Connor
10,028,745	B2	7/2018	Morsi			
2001/0000797	A1	5/2001	Mazzocchi			
2001/0001835	A1	5/2001	Greene et al.			
2002/0062145	A1	5/2002	Rudakov et al.			
2002/0169473	A1	11/2002	Sepetka et al.			
2002/0193812	A1	12/2002	Patel et al.			
2002/0193813	A1	12/2002	Helkowski et al.			
2003/0004533	A1	1/2003	Dieck et al.			
2003/0004568	A1	1/2003	Ken et al.			
2003/0018294	A1	1/2003	Cox			
2003/0028209	A1	2/2003	Teoh et al.			
2003/0040772	A1	2/2003	Hyodoh et al.			
2003/0055440	A1	3/2003	Jones et al.			
2003/0093111	A1	5/2003	Ken et al.			
2003/0113478	A1	6/2003	Dana et al.			
2003/0114918	A1	6/2003	Garrison et al.			
2004/0064093	A1	4/2004	Hektner et al.			
2004/0115164	A1	6/2004	Pierce et al.			
2004/0161451	A1	8/2004	Pierce et al.			
2004/0172056	A1	9/2004	Guterman et al.			
2005/0085836	A1	4/2005	Raymond			
2005/0222580	A1	10/2005	Gifford et al.			
2005/0267511	A1	12/2005	Marks et al.			
2005/0277978	A1	12/2005	Greenhalgh			
2006/0034883	A1	2/2006	Dang et al.			
2006/0052816	A1	3/2006	Bates et al.			
2006/0064151	A1	3/2006	Guterman et al.			
2006/0116709	A1	6/2006	Sepetka et al.			
2006/0116712	A1	6/2006	Sepetka et al.			
2006/0116713	A1	6/2006	Sepetka et al.			
2006/0116714	A1*	6/2006	Sepetka A61B 17/12022 606/200			
2006/0155323	A1	7/2006	Porter et al.			
2006/0190025	A1	8/2006	Lehe et al.			
2006/0190070	A1	8/2006	Dieck et al.			
2006/0200234	A1	9/2006	Hines			
2006/0206140	A1	9/2006	Shaolian et al.			
2006/0206198	A1	9/2006	Churchwell et al.			
2006/0206199	A1	9/2006	Churchwell et al.			
2006/0241686	A1	10/2006	Ferrera et al.			
2006/0247680	A1	11/2006	Amplatz et al.			
2006/0271162	A1	11/2006	Vito et al.			
2007/0010850	A1	1/2007	Balgobin et al.			
2007/0083226	A1	4/2007	Buiser et al.			
2007/0100426	A1	5/2007	Rudakov et al.			
2007/0167876	A1	7/2007	Euteneuer et al.			
2007/0167877	A1	7/2007	Euteneuer et al.			
2007/0167972	A1	7/2007	Euteneuer et al.			
2007/0175536	A1	8/2007	Monetti et al.			
2007/0179520	A1	8/2007	West			
2007/0185442	A1	8/2007	Euteneuer et al.			
2007/0185443	A1	8/2007	Euteneuer et al.			
2007/0185444	A1	8/2007	Euteneuer et al.			
2007/0185457	A1	8/2007	Euteneuer et al.			
2007/0191924	A1	8/2007	Rudakov			
2007/0219619	A1	9/2007	Dieck et al.			
2007/0276426	A1	11/2007	Euteneuer			
2007/0276427	A1	11/2007	Euteneuer			
2007/0282373	A1	12/2007	Ashby et al.			
2007/0288083	A1	12/2007	Hines			
2008/0082176	A1	4/2008	Slazas			
2008/0114391	A1	5/2008	Dieck et al.			
2008/0114436	A1	5/2008	Dieck et al.			
2008/0125852	A1	5/2008	Garrison et al.			
2008/0200945	A1	8/2008	Amplatz et al.			
2008/0200979	A1	8/2008	Dieck et al.			
2008/0221600	A1	9/2008	Dieck et al.			
2008/0281350	A1	11/2008	Sepetka et al.			
2009/0043375	A1	2/2009	Rudakov et al.			
2009/0062899	A1	3/2009	Dang et al.			
2009/0112251	A1*	4/2009	Qian A61B 17/12163 606/194			
2009/0187214	A1	7/2009	Amplatz et al.			
2009/0264978	A1	10/2009	Dieck et al.			
2009/0275974	A1	11/2009	Marchand et al.			

(56)

References Cited

U.S. PATENT DOCUMENTS

2018/0161185 A1 6/2018 Kresslein et al.
 2018/0193025 A1 7/2018 Walzman
 2018/0193026 A1 7/2018 Yang et al.
 2018/0206852 A1 7/2018 Moeller
 2019/0053811 A1 2/2019 Garza et al.

FOREIGN PATENT DOCUMENTS

EP 0717969 6/1996
 EP 1295563 A1 3/2003
 EP 1813213 8/2007
 EP 2208483 7/2010
 EP 2609888 7/2013
 FR 2890306 3/2007
 FR 2890306 A1 3/2007
 JP 2005261951 9/2005
 JP 2008521492 6/2008
 JP 2010523260 T 7/2010
 WO WO9406502 3/1994
 WO WO9409705 5/1994
 WO WO9907294 2/1999
 WO WO9929260 6/1999
 WO 0164112 A1 9/2001
 WO WO02054980 7/2002
 WO WO02089863 11/2002
 WO WO2005099634 10/2005
 WO WO2006034149 3/2006
 WO WO2007121405 10/2007
 WO 2008036156 A1 3/2008
 WO WO2008074027 6/2008
 WO WO2009/014528 1/2009
 WO WO2010009019 1/2010
 WO WO2010027363 3/2010
 WO WO2010/077599 7/2010
 WO 2011066962 A1 6/2011
 WO WO2011095966 8/2011
 WO WO2012/034135 3/2012

WO WO2013112944 8/2013
 WO WO2013/138615 9/2013
 WO 2017074411 A1 5/2017
 WO 2018051187 A1 3/2018

OTHER PUBLICATIONS

Extended European Search Report dated Jul. 1, 2016 in European Patent Application No. 13867906.3; 9 pages.
 Extended European Search Report dated Mar. 2, 2016 in European Patent Application No. 137616710.0; 13 pages.
 Extended European Search Report dated Jan. 30, 2017 in European Patent Application No. 16190494.1; 9 pages.
 Extended European Search Report dated Jan. 7, 2015 in European Patent Application No. 11824250.2, 7 pages.
 International Search Report and Written Opinion for International Application No. PCT/US2013/077767, dated Mar. 19, 2014, 9 pages.
 International Search Report and Written Opinion for International Application No. PCT/US2013/031466, dated Jun. 25, 2014, 11 pages.
 International Search Report and Written Opinion for International Application No. PCT/US11/51268, dated Jan. 2, 2012, 10 pages.
 Prosecution History for U.S. Appl. No. 13/230,628.
 Prosecution History for U.S. Appl. No. 14/603,998.
 Prosecution History for U.S. Appl. No. 14/693,417.
 Prosecution History for U.S. Appl. No. 15/831,974.
 Prosecution History for U.S. Appl. No. 15/832,394.
 Prosecution History for U.S. Appl. No. 13/421,122.
 Prosecution History for U.S. Appl. No. 14/661,233.
 Prosecution History for U.S. Appl. No. 15/162,073.
 Prosecution History for U.S. Appl. No. 15/228,278.
 Prosecution History for U.S. Appl. No. 15/683,627.
 Prosecution History for U.S. Appl. No. 16/156,536.
 Extended European Search Report dated Jun. 17, 2020; European Application No. 19207991.1; 14 pages.

* cited by examiner




FIG. 1




FIG. 2




FIG. 3




FIG. 4




FIG. 5A




FIG. 5B







FIG. 5C






FIG. 9




FIG. 10




FIG. 11




FIG. 12




FIG. 13




FIG. 14




FIG. 15




FIG. 16




FIG. 17A




FIG. 17B




FIG. 18




FIG. 19A




FIG. 19B




FIG. 20





FIG. 22




FIG. 23




FIG. 24




FIG. 25




FIG. 26




FIG. 27




FIG. 28




FIG. 29




FIG. 30




FIG. 31




FIG. 32




FIG. 33




FIG. 34




FIG. 35




FIG. 36




FIG. 37




FIG. 38




FIG. 39




FIG. 40








FIG. 41








FIG. 47




FIG. 48




FIG. 49




FIG. 50




FIG. 51




FIG. 52




FIG. 53




FIG. 54




FIG. 55




FIG. 56




FIG. 57

FIG. 58





FIG. 59




FIG. 60

DEVICES AND METHODS FOR THE TREATMENT OF VASCULAR DEFECTS

CROSS-REFERENCE TO RELATED APPLICATIONS

This application is a continuation of U.S. patent application Ser. No. 15/846,964, filed Dec. 19, 2017, now U.S. Pat. No. 10,675,037, which is a continuation of U.S. patent application Ser. No. 15/162,073, filed May 23, 2016, now U.S. Pat. No. 10,064,627, and of U.S. patent application Ser. No. 14/661,233, filed Mar. 18, 2015, now U.S. Pat. No. 9,855,051, both of which are continuations of U.S. patent application Ser. No. 13/727,029, filed Dec. 26, 2012, now U.S. Pat. No. 8,998,947, which is a continuation-in-part of U.S. patent application Ser. No. 13/421,122, filed Mar. 15, 2012, now abandoned, which is a continuation-in-part of U.S. patent application Ser. No. 13/230,628, filed Sep. 12, 2011, now U.S. Pat. No. 8,974,512, which claims priority to and the benefit of U.S. Provisional Patent Application No. 61/381,770, filed Sep. 10, 2010. All of the foregoing disclosures are hereby incorporated by reference herein in their entirety.

BACKGROUND

The invention relates generally to medical devices and more particularly to expandable medical devices and methods for treating vascular defects. For example, the invention can relate to expandable medical devices and methods for treating an aneurysm. Aneurysms are dilations in a blood vessel caused from weakening of a blood vessel wall. The dilation is produced by the pressure exerted by normal blood flow, which can cause the weakened segment of the blood vessel to swell. In some cases, this swelling results in a sac, or balloon-like polyp protruding from the main or parent vessel. Continued growth and/or eventual rupture of the ballooned arterial wall can have devastating results for a patient. As such, unruptured aneurysms should be treated to prevent hemorrhage. Additionally, ruptured aneurysms can be treated to avert a subsequent rupture and/or additional damage.

Some known medical devices and treatment methods used for treating an aneurysm include delivering a platinum coil to the sac of the aneurysm. The platinum coil is electrolytically separated from a delivery wire, thus inducing a charge in the coil which can cause a thrombotic effect in the aneurysm. In known procedures, about 30% of the volume of the aneurysm is packed with coils. Such known devices and methods, however, often have an about 30% recanalization rate, meaning blood flow returns to the aneurysm again and can cause the coil-packed aneurysm to swell further. Additionally, such known devices and methods require prolonged procedure times for the patient and correspondingly increased exposure to radiation for the patient. Moreover, such devices and methods do not treat the neck of the aneurysm, which is the area between the parent blood vessel and the sac of the aneurysm.

Another known treatment method includes the use of both a coil and a stent. The coil is delivered to the sac of the aneurysm as described above, and the stent is positioned within the parent blood vessel such that a portion of the stent is disposed over the neck of the aneurysm. Such procedures have several drawbacks. For one, delivery of two separate types of devices (i.e., coil(s) and a stent) is a more complex procedure, often resulting in a longer procedure time for the patient. The stent may lead to intra-stent stenosis of the

blood vessel. Additionally, a patient would likely be required to take a blood thinner indefinitely following the procedure. Moreover, such devices and methods are not suitable for treatment of aneurysms positioned at a bifurcation of the blood vessel (i.e., between adjacent branches of a vessel).

Another known device and treatment method includes the use of a flow diverter delivered to the parent blood vessel adjacent the neck of the aneurysm. Generally, the flow diverter is positioned within the parent blood vessel over the neck of the aneurysm to prevent additional blood flow into the aneurysm from the vessel. In current procedures, more than one flow diverter is required per aneurysm to ensure blood flow is appropriately diverted from the aneurysm. Such a device and treatment method has similar drawbacks to the use of a stent, described above. Specifically, the flow diverter may lead to stenosis of the blood vessel and the patient would likely be required to take a blood thinner indefinitely following the procedure. Additionally, known flow diverters are not suitable for treating an aneurysm positioned at a bifurcation of the blood vessel. Moreover, long term follow-up of patients treated using a flow diverter is showing an increased rate of recanalization to the aneurysm.

Thus, there is a need for improved systems, devices and methods for treating vascular defects, such as balloon-type aneurysms, as described herein.

SUMMARY

Devices and methods for treating vascular defects, such as, for example, balloon-type aneurysms, are described herein. In one embodiment, an apparatus includes an insertion portion and an expandable implant. The expandable implant is configured to be deployed in an aneurysm and is coupled to the insertion portion. The expandable implant has a first portion and a second portion coupled to the first portion. The expandable implant is movable between a first configuration in which the first portion and the second portion are substantially linearly aligned and a second configuration in which the second portion at least partially overlaps the first portion.

BRIEF DESCRIPTION OF THE DRAWINGS

FIG. 1 is a schematic illustration of a medical device according to an embodiment in a first configuration.

FIG. 2 is a schematic illustration of a medical device according to an embodiment in a second configuration.

FIG. 3 is a side view of a medical device according to an embodiment in a first configuration.

FIG. 4 is a side view of a medical device according to an embodiment in a second configuration.

FIG. 5A is a view of the medical device of FIG. 3 in a first configuration during insertion into an aneurysm.

FIG. 5B is a view of the medical device of FIG. 3 in a second configuration during insertion into an aneurysm.

FIG. 5C is a view of the medical device of FIG. 3 in a third configuration during insertion into an aneurysm.

FIG. 6 is a view of a portion of a medical device in an expanded configuration, according to an embodiment.

FIGS. 7-13 are views of a medical device in an expanded configuration, according to embodiments.

FIG. 14 is a view of a medical device in a partially collapsed configuration, according to an embodiment.

FIG. 15 is a view of the medical device of FIG. 14 in an expanded configuration, according to an embodiment.

FIG. 16 is a view of a portion of a medical device in an expanded configuration according to an embodiment, with a first portion spaced apart from a second portion.

FIG. 17A is a view of a portion of a medical device in a collapsed configuration according to an embodiment.

FIG. 17B is a view of a portion of a medical device in an expanded configuration according to an embodiment.

FIG. 18 is a flowchart of a method according to an embodiment.

FIG. 19A is a view of a portion of a medical device in an expanded configuration, according to an embodiment.

FIG. 19B is a schematic illustration of the medical device of FIG. 19A.

FIG. 20 is a view of a portion of a medical device in an expanded configuration, according to an embodiment.

FIG. 21 is a view of a portion of a medical device in an expanded configuration, according to an embodiment.

FIG. 22 is a view of a portion of the medical device of FIG. 21 in a collapsed configuration.

FIG. 23 is a view of a portion of a medical device in a collapsed configuration, according to another embodiment.

FIG. 24 is a view of the portion of the medical device of FIG. 23 in an expanded configuration.

FIG. 25 is a view of a portion of a medical device in a collapsed configuration, according to an embodiment.

FIG. 26 is a view of the portion of the medical device of FIG. 25 in a partially expanded configuration.

FIG. 27 is a view of a portion of the medical device of FIG. 25 in an expanded configuration.

FIGS. 28 and 29 are each a different view of a portion of a medical device in an expanded configuration, according to an embodiment.

FIGS. 30 and 31 are each a view of a portion of a medical device in an expanded configuration, according to different embodiments.

FIG. 32 is a view of a portion of a medical device in a collapsed configuration, according to an embodiment.

FIG. 33 is a view of the portion of the medical device of FIG. 32, shown in an expanded configuration.

FIG. 34 is a schematic illustration of the portion of the medical device of FIG. 33.

FIG. 35 is a schematic illustration of a portion of an insertion device, according to an embodiment, shown in a first configuration and coupled to a schematic illustration of a portion of an expandable implant.

FIG. 36 is a schematic illustration of the portion of the insertion device and expandable implant of FIG. 35, shown in a second configuration.

FIG. 37 is a schematic illustration of the portion of the insertion device of FIG. 35 shown removed from the expandable implant.

FIG. 38 is a schematic illustration of a portion of an insertion device, according to another embodiment.

FIG. 39 is a view of a portion of an insertion device, according to another embodiment.

FIG. 40 is a schematic illustration of a portion of an insertion device coupled to an expandable implant, according to another embodiment.

FIG. 41 is a flowchart illustrating a method of deploying an expandable implant, according to an embodiment.

FIG. 42 is a view of a portion of a medical device in an expanded configuration, according to an embodiment.

FIG. 43 is a view of a portion of the medical device of FIG. 42 in a collapsed configuration.

FIG. 44 is a view of a portion of a medical device in an expanded configuration, according to an embodiment.

FIG. 45 is a view of a portion of the medical device of FIG. 44 in a collapsed configuration.

FIG. 46 is a view of a portion of the medical device of FIG. 44 shown partially deployed within an aneurysm.

FIGS. 47-54 are each a schematic illustration of a portion of an insertion device, according to a different embodiment.

FIG. 55 is a side view of a portion of an insertion device according to an embodiment.

FIG. 56 is a side view of a portion of the insertion device of FIG. 55 shown coupled to an expandable implant.

FIG. 57 is a view of a portion of a medical device, shown partially deployed, according to another embodiment.

FIG. 58 is a schematic illustration of a portion of a medical device shown in a collapsed configuration, according to another embodiment.

FIG. 59 is a view of the portion of the medical device of FIG. 58, shown in an expanded configuration.

FIG. 60 is a schematic illustration of a portion of an expandable implant, according to another embodiment, shown in a collapsed configuration.

DETAILED DESCRIPTION

Medical devices and methods of treatment are described herein to treat patients experiencing a vascular defect, such as an aneurysm, in a circulatory blood vessel and the effects of that defect, including hemorrhagic stroke. For example, the devices and methods described herein can be useful for treating vascular defects present in vasculature that is tortuous, of small-diameter, and/or that is otherwise difficult to access. More specifically, the devices and methods described herein can be useful for treating saccular (also referred to as balloon-type or berry) aneurysms, bifurcation aneurysms, fistulas, and other defects in vasculature, including defects in neurovasculature. The medical devices and methods of treatment described herein can reduce hemorrhagic events while promoting endothelialization of an opening between an aneurysm and a parent blood vessel from which the aneurysm bulge formed (e.g., at a neck of the aneurysm).

Various embodiments of a medical device for occupying all or substantially all of the volume of an aneurysm and/or promoting endothelialization at or proximate to the aneurysm are described herein. In some embodiments, the medical device includes an expandable implant including an electropositive woven or braided material. The filaments or strands forming the braid or weave are configured to encourage recruitment and/or retention of endothelial cells to the device and therefore within the defect. The expandable implant is configured to assume a non-linear predetermined three-dimensional shape within a sac of the aneurysm upon release from a tubular or other delivery constraint (e.g., a catheter or cannula). The electropositive woven or braided material has a particular porosity and includes multiple openings between the filaments or strands when the expandable implant is in the expanded configuration. Such openings are ideal in the blood environment for harboring endothelial cells recruited to the site. The electropositivity of the material encourages endothelialization in the presence of the electronegative charges of the blood and body tissues. Said another way, the electropositivity of the expandable implant in relation to a charge of blood and tissue (which is electronegative in comparison) provides an environment in the defect that promotes endothelialization. Endothelialization within the defect can ultimately result in the defect walling-off from the parent vessel. For example, the growth and development of an endothelial layer over a neck of an aneurysm can wall off the aneurysm from the parent vessel

and allow flow dynamics to equilibrate at the defect. As such, the device can be configured to facilitate healing the defect and preventing recanalization because tissue is created from within the body that resists aberrant blood flow and redistributes the flow pressure that may have created the defect. Upon healing with endothelialization, the pressure is evenly distributed along the parent vessel in a manner that precludes recanalization at the defect post-treatment. Furthermore, blood from within the parent vessel no longer has access to the walled off defect once the endothelialization process is complete. Additionally, at least a portion of the expandable implant can be positioned over the neck of the aneurysm once the implant is deployed within the aneurysm such that the portion disrupts the flow of blood from the parent vessel into the aneurysm. As such, the expandable implant provides blood flow disruption in advance of and in addition to growth and development of the endothelial layer over the neck of the aneurysm.

A medical device described herein can include an insertion portion (e.g., a guide wire) and an expandable implant formed with, for example, woven or braided filaments in a mesh-like configuration. The terms mesh and braid can each refer herein to a fabric or material of woven or braided filaments or strands of wire or polymer. The expandable implant of the medical device can be configured to compress or collapse for delivery into a blood vessel. In some embodiments, the medical device can be inserted while in a collapsed or compressed configuration through a delivery device, such as, for example, a microcatheter, cannula, delivery tube or sheath. In some embodiments, the medical device can be deployed without the use of such a delivery device.

The expandable implant of the medical device can have a collapsed or compressed configuration such that the expandable implant has a diameter that can fit within the narrow constraints of the neurovasculature and/or within a lumen of a delivery catheter. The expandable implant of the medical device can be formed with, for example, an arrangement of strands (e.g., a mesh or braid arrangement of strands or filaments) that can compress and expand. Such materials include Nitinol, MP35N, stainless steel, cobalt chromium, titanium, platinum, tantalum, tungsten, or alloys thereof, or polyester, polyethylene (PET), Dacron, PEEK, vectron, and suture materials, and are available from Fort Wayne Metals of Fort Wayne, Ind., California Fine Wire Company of Grover Beach, Calif., other metal manufacturers, Ethicon Inc. of Somerville, N.J., Genzyme of Cambridge, Mass., Poly-Med, Inc. of Anderson, S.C., and/or other medical grade suture and fiber manufacturers. The expandable implant can be compressed over and/or along the insertion portion of the medical device. The insertion portion can be, for example, a wire. In some embodiments, a medical device includes an insertion portion movably disposable within a lumen of a delivery device. A distal portion of the insertion portion can be coupled to the expandable implant. The expandable implant can be moved from a collapsed configuration to an expanded configuration while disposed within, or as it is being inserted into, a defect (e.g., an aneurysm).

In some embodiments, the expandable implant can be formed with filaments of superelastic or shape memory material (such as, e.g., nitinol) and the braid or mesh can be set in a predefined shape prior to attaching the expandable implant to the insertion portion of the medical device. In such an embodiment, when the expandable implant is deployed and expands, it assumes a biased predetermined shape. The predetermined shape can be a generic shape, such

as that of a sphere, or can be a custom-made shape based on a shape of a target aneurysm within a patient. Suitable materials are described in more detail herein.

The medical devices described herein can include one or more expandable implants formed with a woven mesh or braid that has variably sized apertures (also referred to herein as "openings" or "pores"). Said another way, the devices are formed with a material that has a particular porosity or pore density. In some embodiments, an expandable implant can have sections of mesh or braid having variation in density of the filaments and may include portions or bands of densely spaced filaments (i.e., lower porosity) spaced by portions or bands that are less dense (i.e., higher porosity). The less dense braid portion can have larger openings in the braid, while the more dense braid portion can have smaller openings in the braid. Material (e.g., bodily tissue such as endothelial cells) can be encouraged to enter and/or attach to interstices of the mesh of the expandable implant. For example, the more dense braid portion can be used to encourage greater endothelial cell attachment and the less dense braid portion can be used to reduce the overall weight and or material to be implanted in the patient. The less dense sections can also direct the final shape of the expandable implant. For example, sections of less dense (more open) mesh or braid can direct the effects of expansion of the implant.

In some embodiments, a medical device can be delivered to a desired treatment site within a vasculature by inserting the medical device through a lumen of a delivery catheter (e.g., a microcatheter). The expandable medical device can be inserted through the delivery catheter in a collapsed or compressed configuration. The expandable implant of the expandable medical device can be moved out through a distal end of the delivery catheter at the treatment site (e.g., into a sac of an aneurysm) and moved to an expanded configuration. In some embodiments, the delivery catheter is used to compress or collapse the expandable implant. For example, the expandable implant can be formed with a biased expanded configuration and when it is placed within a lumen of a catheter it is compressed. When the expandable implant is moved outside of the catheter, it can assume its biased expanded configuration. In the expanded configuration, a first portion of the expandable implant substantially overlaps a second portion of the expandable implant. The first and second portions of the expandable implant can be discrete structures or can be portions of a unitary or monolithically constructed device.

A medical device, such as an expandable implant, described herein can include a first porous member and a second porous member coupled to the first porous member. Each of the first and second porous members includes a first end and a second end. The first and second porous members each have a collapsed configuration for insertion through a blood vessel and an expanded configuration for occupying at least a portion of the volume defined by the sac of an aneurysm. In some embodiments, the first porous member is substantially elongate and has a greater width in its expanded configuration than in its collapsed configuration. The second porous member is substantially elongate and has a greater width in its expanded configuration than in its collapsed configuration. In some embodiments, the width of the first porous member is greater than the width of the second porous member, for example, when each of the first and second porous members are in their expanded configurations.

In some embodiments, the first porous member is configured to occupy a first volume in its collapsed configura-

tion and a second, greater, volume in its expanded configuration. For example, the first porous member can have a substantially spherical, oblong, or other suitable shape in its expanded configuration that occupies a greater volume than the substantially elongate shape of the first porous member in its collapsed configuration. The second porous member can be configured to move or curve into a three dimensional configuration in the expanded configuration such that a first segment of the second porous member overlaps with a second segment of the second porous member. In its expanded configuration, the second porous member can define an interior region configured to receive the first porous member in its expanded configuration. For example, in some embodiments, the second porous member has a substantially spherical shape with an open interior region configured to receive the first porous member.

In some embodiments, a medical device, such as an expandable implant, described herein can include a first porous member and a second porous member. Each of the first and second porous members includes a first end and a second end. The first and second porous members each have a collapsed configuration for insertion through a blood vessel and an expanded configuration for occupying at least a portion of the volume defined by a sac of an aneurysm. The first and second porous members are each substantially elongate in the collapsed configuration. In its expanded configuration, the first porous member has a three-dimensional shape including a first segment configured to overlap with a second segment and defining an interior region. The second porous member is configured to be disposed in the interior region of the first porous member when each of the first and second porous members is in their respective expanded configurations. In some embodiments, the second porous member can be formed integrally or monolithically with the first porous member. In some embodiments, the second porous member can be woven or braided using the same filaments that form the first porous member.

In some embodiments, the expandable implant is in the form of a braided tube that includes fibers of a superelastic shape memory alloy, or polymeric fibers. In some embodiments, the expandable implant can effect a shape deformation inducing a substantially spherical contour. In some embodiments, the expandable implant can effect a shape deformation inducing a helical contour. In some embodiments, the shape deformation can include inducing radial expansion and/or axial shortening.

The medical devices described herein can be used to occupy at least a portion of the volume defined by a sac of an aneurysm and/or to promote endothelialization of the neck of the aneurysm to inhibit or stop blood flow into the aneurysm, which can lead to, for example, hemorrhagic stroke. In some embodiments, wire or polymer filaments can be used to form a woven mesh or braided strands that can be expandable, and have apertures sized to promote endothelial cell attachment at the aneurysm.

It is noted that, as used in this written description and the appended claims, the singular forms “a,” “an” and “the” include plural referents unless the context clearly dictates otherwise. Thus, for example, the term “a lumen” is intended to mean a single lumen or a combination of lumens. Furthermore, the words “proximal” and “distal” refer to direction closer to and away from, respectively, an operator (e.g., surgeon, physician, nurse, technician, etc.) who would insert the medical device into the patient, with the tip-end (i.e., distal end) of the device inserted first inside a patient’s body. Thus, for example, the end first inserted inside a patient’s body would be the distal end of the medical device, while the

end outside of or inserted later into a patient’s body would be the proximal end of the medical device. Additionally, the terms “first,” “second,” “third,” and so on, used to describe similarly identified elements are for purposes of clarity only, and are not meant to imply a priority or that such numerical identifier must be associated with that particular element in the claims.

FIGS. 1 and 2 are schematic illustrations of a vascular medical device **100** according to an embodiment in a first configuration and a second configuration, respectively. The medical device is configured to promote healing of an aneurysm. More specifically, at least a portion of the medical device is configured to occupy at least a portion of the volume defined by a sac of the aneurysm and, in some embodiments, at least a portion of the medical device is configured to promote endothelial cell attachment over a neck of the aneurysm. Once endothelialization over the aneurysm neck is complete, blood flow into the aneurysm sac from a parent blood vessel (i.e., the vessel on which the aneurysm formed) is prevented.

The medical device **100** can include an insertion portion **102** and an expandable implant **110**. The insertion portion **102** is coupled to the expandable implant **110**, such as, for example, at a proximal portion **112** of the expandable implant **110**. In some embodiments, the insertion portion **102** is removably coupled to the expandable implant **110**. In this manner, the insertion portion **102** can be separated from the expandable implant **110** following delivery of the expandable implant to the aneurysm and removed from a patient’s vasculature. The insertion portion **102** can be, for example, a guide wire or a distal end portion of a wire. The medical device **100** can be used with a cannula or catheter **104** (shown in dashed lines in FIGS. 1 and 2) to, for example, deliver the expandable implant **110** to the aneurysm.

The expandable implant **110** is configured to be deployed in the aneurysm (e.g., in a sac of an aneurysm). The expandable implant **110** has a first portion **120** and a second portion **130**. As shown in FIG. 1, the expandable implant **110** has a first configuration in which the first portion **120** and the second portion **130** are substantially linearly aligned. In its first configuration, the expandable implant **110** is configured for insertion through a blood vessel. The expandable implant **110** is also configured for insertion through a neck of the aneurysm when in its first configuration.

The expandable implant **110** is movable between its first configuration and a second configuration in which the second portion **130** at least partially overlaps the first portion **120**, as shown in FIG. 2. For example, the second portion **130** can be configured to bend, curve and/or twist in multiple turns such that multiple segments of the first portion **120** and the second portion **130** are overlapped. Additionally, at least one of the first portion **120** and the second portion **130** can be configured to bend or curve in multiple turns such that the respective first or second portion is overlapped with itself. In some embodiments, the expandable implant **110** can be understood to have multiple first portions and multiple second portions. In other words, the expandable implant can continually overlap itself in its deployed configuration to occupy all or substantially all of the volume of the aneurysm.

In its second configuration, the expandable implant **110** is configured to occupy at least a portion of the volume defined by the sac of the aneurysm. In some embodiments, when the expandable implant **110** is in its second configuration, at least a portion of the expandable implant is configured to be positioned over the neck of the aneurysm. For example, the

portion of the expandable implant **110** at which the second portion **130** overlaps the first portion **120** can be configured to be positioned over the neck of the aneurysm. As such, the portion of the expandable implant **110** disposed over the aneurysm neck has an increased density (e.g., a dual density compared to the first portion **120** or the second portion **130** individually), which helps to limit or prevent blood flow from entering the sac of the aneurysm. The portion of the expandable implant **110** positioned over the aneurysm neck can be a scaffold for endothelial cell attachment at the aneurysm neck. For example, the portion of the expandable implant **110** positionable over the aneurysm neck can be porous, such as by including a porous mesh, as described in more detail herein. In some embodiments, the first portion **120** and the second portion **130** of the expandable implant **110** are biased to the second configuration.

As noted above, in some embodiments, at least a portion of the expandable implant **110** is porous. For example, in some embodiments, at least a portion of the expandable implant **110** can include and/or be constructed of a mesh (e.g., woven, braided, or laser-cut) material such that a wall or layer of the expandable implant **110** defines multiple openings or interstices **118**. More specifically, in some embodiments, at least one of or both the first portion **120** and the second portion **130** of the expandable implant **110** can include the porous mesh. The porous mesh can have a first porosity when the expandable implant **110** is in its first configuration and a second porosity when the expandable implant is in its second configuration. More specifically, in some embodiments, the porous mesh can have a greater porosity when the expandable implant **110** is in its second configuration than when the expandable implant is in its first configuration. The porosity of the porous mesh can be increased, for example, because one or more individual pores or openings are larger when in the second configuration than in the first configuration. For example, the porous mesh can be expanded in the second configuration, thereby increasing the space between filaments of the mesh (and thus the size of one or more openings of the mesh). In other words, an overall volume of pore openings can be increased. In another example, the porosity of the porous mesh can be increased because one or more openings that were closed off when the expandable implant **110** was collapsed into its first configuration are reopened when the expandable implant is moved to its second configuration. In other words, a number of open pores can be increased.

In some embodiments, the first portion **120** and the second portion **130** can have one of the same or different porosities. For example, the first portion **120** can have a porosity greater than a porosity of the second portion **130**. In another example, the second portion **130** can have a porosity greater than the porosity of the first portion **120**. In still another example, the first and second portions **120**, **130** can have substantially equivalent porosities in the expanded configuration.

In some embodiments, at least one of the first portion **120** and the second portion **130** includes one, two, three, or more layers. For example, in some embodiments, the first portion **120** of the expandable implant **110** includes a first layer (not shown in FIG. 1 or 2) of porous mesh and a second layer (not shown in FIG. 1 or 2) of porous mesh. The first layer and the second layer can have the same or different porosities. In some embodiments, the first layer is offset from the second layer. As such, the porosity of the first portion is determined by the porosities of the first and second layers and the manner in which the first layer is offset from the second layer.

In some embodiments, at least a portion of the expandable implant **110**, such as at least one of the first portion **120** or the second portion **130** can include a shape-memory material, such as, for example, nitinol, and can be preformed to assume a desired shape. Thus, in such an embodiment, the portion of the expandable implant **110** (e.g., the first portion **120** and/or the second portion **130**) can be biased into an expanded second configuration and moved to a collapsed first configuration by restraining or compressing the portion of the expandable implant.

In some embodiments, at least a portion of the expandable implant **110**, such as at least one of the first portion **120** or the second portion **130** can include an electropositive material, described in more detail below.

The expandable implant **110** when in the expanded configuration can have a variety of different shapes, sizes and configurations. For example, in some embodiments, when in the expanded configuration the expandable implant **110** can be substantially spherical. In some embodiments, the expandable implant **110** can be substantially helical. In some embodiments, the expandable implant **110** can be substantially circular, disc-shaped, or ring-shaped. In some embodiments, the expandable implant **110** can be a custom-made shape based on a shape of a target aneurysm within a patient: for example, a shape modeled after the shape of the target aneurysm as detected by an imaging device. For example, an image of the aneurysm shape can be acquired using an angiogram, and the expandable implant **110** can be modeled after the shape of the aneurysm shown in the angiogram. In some embodiments, the expandable implant **110** can include multiple portions having varying outer perimeters or outer diameters. For example, in some embodiments, when in the expanded configuration the expandable implant **110** can include a first portion having a first outer perimeter, a second portion having a second outer perimeter and a third portion having a third outer perimeter. In such an embodiment, the second outer perimeter can be smaller than each of the first outer perimeter and the third outer perimeter.

In one example use of the medical device **100**, a catheter **104** can be inserted into a blood vessel and directed to a desired treatment site near a vascular defect, such as the aneurysm. The expandable implant **110** is inserted into an elongate lumen of the catheter **104** for delivery to the treatment site. A distal portion of the catheter **104** is positioned adjacent the aneurysm within the blood vessel. The expandable implant **110** is moved from a first position inside the catheter to a second position outside the catheter. When the expandable implant **110** is in its first position, each of the first portion **120** and the second portion **130** are in a first configuration. For example, in the first configuration, each of the first and second portions **120**, **130** can be compressed or collapsed within the lumen of the catheter **104** and are substantially linear in configuration.

The expandable implant **110** can be oriented with respect to an opening in the vessel wall in fluid communication with the aneurysm such that the expandable implant can enter a sac of the aneurysm when the expandable implant **110** is moved to its second position. The expandable implant **110** can be moved from its first position to its second position with the assistance of the insertion portion **102** such that the expandable implant **110** is directed into and positioned within a sac of the aneurysm. When the expandable implant **110** is in its second position, the first and second portions each have a second configuration. For example, in the second configuration, each of the first and second portions **120**, **130** can be expanded into a three-dimensional shape. The three-dimensional shape of the first portion **120** in the

11

second configuration can be similar to or different from the three-dimensional shape of the second portion 130. In the second configuration, the first portion 120 of the expandable implant 110 substantially overlaps the second portion 130. In some embodiments, the second portion 130 is disposed in an interior region defined by the first portion when each of the first portion and the second portion are in their respective second configurations.

The first and second portions 120, 130 can be moved to their respective second configurations concurrently or sequentially. For example, in some embodiments, the second portion 130 is moved to its second configuration before the first portion 120 is moved to its second configuration. The expandable implant 110 can assume a biased expandable configuration such that the walls of the expandable implant 110 contact at least a portion of the wall of the aneurysm and/or such that a portion of the expandable implant is disposed over the neck of the aneurysm. The presence of the expandable implant 110 over the neck of the aneurysm can substantially reduce and/or prevent further blood flow from the parent vessel into the aneurysm sac because the expandable implant can act as a physical flow disruptor for blood flowing from the parent vessel and as a scaffold for endothelial cell attachment at the aneurysm neck to promote endothelialization of the neck/vessel wall. The insertion portion 102 can then be disconnected from a proximal end of the expandable implant 110 and removed through the catheter 104.

FIGS. 3, 4, 5A, 5B and 5C illustrate a medical device according to an embodiment. The medical device 200 can include all or some of the same features and functions as described above for medical device 100. The medical device 200 includes an insertion portion 202 and an expandable implant 210. The expandable implant 210 is removably coupled at its proximal end to a distal end of the insertion portion 202.

The expandable implant 210 includes a first portion 220 and a second portion 230. As shown in FIGS. 3 and 5A, the expandable implant 210 has a first, or collapsed, configuration in which the first and second portions 220, 230 are substantially linearly aligned. In this manner, the expandable implant 210 can be disposed within a lumen of a catheter 204 for delivery through a blood vessel V to a treatment site, such as to an aneurysm A. In its first configuration, the expandable implant 210 has a first width W_1 , as shown in FIG. 2. As shown in FIGS. 4 and 5B-5C, the expandable implant 210 is moveable to a second, or expanded or deployed, configuration. The insertion portion 202 is configured to move the expandable implant 210 from the first configuration to the second configuration. The insertion portion 202 can be disconnected from the expandable implant 210 when the expandable implant 210 is in its second configuration.

In its second configuration, the expandable implant 210 is configured to occupy at least a portion of the volume defined by a sac of the aneurysm A. As such, the expandable implant 210 has a second width W_2 in the second, expanded, configuration greater than its first width W_1 . For example, the expandable implant 210 can be substantially narrow and elongate in its first configuration and can assume a three-dimensional shape in its second configuration. In the embodiments illustrated in FIGS. 3-5C, the expandable implant 210 has a substantially spherical shape in its second configuration. The expandable implant 210 can be compliant such that its three-dimensional shape can accommodate any irregularities in the shape of the aneurysm. In the second configuration, the second portion 230 of the expandable

12

implant 210 at least partially overlaps the first portion 220. At least a portion of the expandable implant 210 is configured to be positioned over a neck N of the aneurysm A when the expandable implant is in its second configuration within the sac of aneurysm A. The expandable implant 210 is configured to facilitate endothelial cell attachment at the neck N of the aneurysm A, as described in more detail herein.

In the embodiment illustrated in FIG. 3, the first portion (or member) 220 is a first ribbon-like strand and the second portion (or member) 230 is a second ribbon-like strand discrete from the first portion. In other embodiments, an expandable implant can include a first portion and a second portion from a single ribbon-like strand (e.g., integrally or monolithically constructed), instead of discrete portions. A first end 222 of the first portion 220 is coupled to a first end 232 of the second portion 230. Any suitable mechanism for coupling the first end 222 of the first portion 220 to the first end 232 of the second portion 230 can be used, such as an adhesive, a mechanical coupler, a weld, or the like, or any combination of the foregoing. For example, the first ends 222, 232 can be coupled by a band 240. The band 240 can also be configured to help couple the insertion portion 202 to the expandable implant 210. The band 240 can be or can include, for example, a radiopaque marker.

A second end 224 of the first portion 220 and a second end 234 of the second portion 230 each have a radiopaque marker 242, 244, respectively, coupled thereto. The radiopaque markers 242, 244 are configured to facilitate imaging of the expandable implant 210 during delivery to the treatment site and/or subsequent to implantation. The markers 242, 244 are configured to be wholly disposed within the sac of the aneurysm A when the expandable implant 210 is in its second configuration. As such, the markers 242, 244 will not puncture the wall of the aneurysm A or the vessel V, and the markers 242, 244 will not interfere with endothelial cell attachment at the aneurysm neck. This is also beneficial because if the markers 242, 244 were positioned at or proximate to the neck of the aneurysm, blood from a parent blood vessel could have a tendency to clot around the marker.

When the expandable member 210 is moved between its first configuration and its second configuration, at least one of the first portion 220 and the second portion 230 is also moveable between a first configuration and a second configuration. The first portion or member 220 has a first, collapsed, configuration in which the first portion 220 is substantially elongate and has a first width. The first portion 220 has a second, expanded, configuration, in which the first portion 220 has a second width greater than the first width. For example, the first portion 220 can be moveable from a substantially linear, elongate collapsed configuration to a multi-dimensional (e.g., three-dimensional) shape in the expanded or deployed configuration. As shown in FIGS. 4 and 5C, the first portion 220 can have a three-dimensional shape in the expanded configuration that lends an overall spherical shape to the expandable implant 210. The first portion 220 can be biased to its second, expanded, configuration.

The first portion or member 220 is porous and, for example, can include or be constructed of a porous mesh. The porous mesh can be formed using filaments that are woven or braided together in a manner that openings or interstices are present between portions of the filaments at least when the expandable implant 210 is in its second configuration. For example, the porous mesh can include a plurality of braided wires. Suitable mesh material is

described in more detail herein. The porous mesh can have a first porosity when the first portion **220** is in the first configuration and a second porosity when the first portion **220** is in the second configuration. For example, when the first portion **220** is moved from its first, collapsed, configuration to its second, expanded, configuration, the mesh can be expanded such that the size of the openings of the mesh is increased, thus increasing the porosity of the mesh. The porous mesh is configured to act as a scaffold that promotes clot formation and endothelium cell attachment when the mesh is disposed within the aneurysm A. Specifically, endothelial cells will migrate to the openings of the mesh.

The first portion **220** of the expandable implant **210** includes a first layer of porous mesh and a second layer of porous mesh. In this manner, the density of the first portion **220** is greater than the density of either the first or second layers individually. Such a dual-density structure can help to limit or prevent blood flow into the aneurysm A, for example when the first and second layers of the first portion **220** are disposed over the neck N of the aneurysm A. The first layer of porous mesh and the second layer of porous mesh can have the same porosities, or different porosities. The first layer of porous mesh can be offset from the second layer of porous mesh. In this manner, the overall porosity of the first portion **220** is greater than the porosity of either the first or second layers individually. The first and second layers of porous mesh can be coupled together in any suitable manner. For example, the first portion **220** can be formed using an elongate tubular mesh having an elongate lumen there-through. In such an embodiment, the elongate mesh can be flattened from a tubular structure to a ribbon-like structure such that a first side, or layer, of the mesh is disposed on or proximate to a second side, or layer, of the mesh, thus forming a dual density, or dual-layered, mesh structure.

The second portion, or member, **230** of the expandable implant **210** can be configured the same as or similar to, and can be used in the same or similar manner, as the first portion **220**. When the expandable member **210** is moved between its first configuration and its second configuration, the second portion **230** is also moveable between a first, collapsed, configuration in which the second portion is substantially elongate and has a third width, and a second, expanded, configuration, in which the second member has a fourth width greater than the third width. For example, the second portion **230** can be moveable from a substantially linear, elongate collapsed configuration to a multi-dimensional (e.g., three-dimensional) shape in the expanded configuration. As shown in FIGS. 4 and 5C, the second portion **230** can have a three-dimensional shape in the expanded configuration that lends an overall spherical shape to the expandable implant **210**. The second portion **230** can be biased to its second, expanded, configuration.

The second portion **230** is porous and can include or be constructed of a porous mesh. The porous mesh can be configured the same as or similar to, and can be used in the same or similar manner, as the porous mesh described above with respect to the first portion **220** of the expandable implant **210**. For example, the porous mesh can include a weave or braid of filaments that is porous at least when the expandable implant **210** is in its second configuration. Additionally, the porous mesh of the second portion **230** can have a first porosity when the second portion **230** is in the first configuration and a second porosity when the second portion **230** is in the second configuration. In some embodiments, the second portion **230** of the expandable implant **210** includes a first layer of porous mesh and a second layer of porous mesh, which can be of the same or different porosi-

ties. In this manner, the total density of the second portion **230** is greater than the density of either the first or second layers individually. The first layer of porous mesh can be offset from the second layer of porous mesh such that the overall porosity of the second portion **230** is greater than the porosity of either the first or second layers individually. Similarly as described above with respect to the first portion **220**, the first and second layers of porous mesh of the second portion **230** can be formed from a monolithically constructed elongate tubular mesh that is flattened into a ribbon-like structure.

The first portion **220** and the second portion **230** of the expandable implant **210** can be the same or different sizes. For example, as shown in FIG. 5A, the first portion **220** can have a length in its first, collapsed, configuration, that is less than a length of the second portion **230** in its first, collapsed, configuration. In this manner, the markers **242**, **244** will be sequentially introduced through the neck N of the aneurysm A, which permits the expandable implant **210** to be introduced through a narrower neck N. In another example, the first portion **220** and the second portion **230** can have the same or different widths. In some embodiments, for example, the first width of the first portion **220** in its first configuration is wider than the third width of the second portion **230** in its first configuration. The second width of the first portion **220** in its second configuration can also be wider than the fourth width of the second portion **230** in its second configuration. In another example, the fourth, expanded, width of the second portion **230** can be greater than the second, expanded, width of the first portion **220**. In some embodiments, the porous mesh of the first portion **220** can have a multi-dimensional shape with a first width when the expandable implant **210** is in its second configuration, and the porous mesh of the second portion **230** can have a multi-dimensional shape with a second width less than the first width when the expandable implant is in its second configuration.

In some embodiments, for example, the first portion **220** (or the porous mesh of the first portion) can have a width of about 8 mm when the expandable implant is expanded in its second configuration, and the second portion **230** (or the porous mesh of the second portion) can have a width of about 9.5 mm when the expandable implant is expanded in its second configuration. As such, in an embodiment in which the first portion **220** has a smaller overall size in the expanded configuration than the second portion **230**, the first portion **220** can be configured to be disposed within an open interior region formed by the second portion **230** in its second configuration.

In some embodiments, a variation of medical device **200** is contemplated. For example, in such an embodiment, the first portion of the expandable implant can include a first tubular mesh that defines a lumen therethrough, and the second portion of the expandable implant can include a second tubular mesh disposed within the lumen of the first tubular mesh. The first and second tubular mesh structures can be formed into a substantially ribbon-like strand. As such, the expandable implant has a four-layer density. The expandable implant can include additional ribbon-like strands in addition to the strand formed by the first and second portions. For example, the expandable implant can include one, two, three, four, five, six, seven, eight, or nine strands, with each of the strands having a desired number of layers (e.g., two, four, or more layers). As such, an expandable implant can be formed that has a desired amount of density. As noted above, a highly dense structure helps to prevent blood flow from the parent blood vessel into the

aneurysm. Each layer or portion of the expandable implant can have the same or different density as the other layers or portions. Furthermore, each layer or portion of the expandable implant can have the same or different porosity as the other layers or portions.

FIG. 6 illustrates a portion of another embodiment of a medical device. The medical device 300 can include the same or similar features and functions as described above for previous embodiments. For example, the medical device 300 includes an expandable implant 310 and an insertion portion or member (not shown in FIG. 6). The expandable implant 310 is shown in an expanded configuration and can be moved between a compressed or collapsed configuration in which the expandable implant is substantially elongate and the expanded configuration in the same or similar manner as described above for expandable implant 210. In the expanded configuration, a first portion 320 of the expandable implant 310 is overlapped by a second portion 330 of the expandable implant. Additionally, at least a portion of the first portion 320 is disposed within an open interior region 336 defined by the second portion 330 when the expandable implant 310 is in its expanded configuration.

The expandable implant 310 includes a ribbon-like strand of porous mesh. At least a portion of the porous mesh is configured to be positioned over a neck of an aneurysm when the expandable implant 310 is in the expanded configuration. The porous mesh is configured to bend, curve, and/or twist at multiple turns into a substantially spherical shape when the expandable implant 310 is in the expanded configuration. The porous mesh can be a ribbon-like structure that is wider than the porous mesh of expandable implant 210. In this manner, the porous mesh of expandable implant 310 can be a shorter length than that of expandable implant 210 and still provide a similar amount of coverage within the aneurysm (and over the neck of the aneurysm) as expandable implant 210. The porous mesh can include one, two, or more layers depending on the desired density and porosity of the expandable implant 310. In some embodiments, a first radiopaque marker 342 is coupled to a first end 312 of the expandable implant 310 and a second radiopaque marker 344 is coupled to a second end 314 of the expandable implant. The expandable implant 310 is configured to be wholly disposed within the aneurysm such that the radiopaque markers 342, 344 are wholly disposed within the aneurysm sac and the porous mesh is disposed over the neck of the aneurysm. In some embodiments, the radiopaque markers are configured to be positioned at a side of the aneurysm (i.e., disposed away from the neck of the aneurysm).

FIG. 7 illustrates another embodiment of a medical device. The medical device 400 can include the same or similar features and functions as described above for previous embodiments. For example, the medical device 400 includes an expandable implant 410 and an insertion portion or member 402. The expandable implant 410 is sized to occupy the sac of an aneurysm, and the insertion member 402 is configured to facilitate delivery of the expandable implant into the sac of the aneurysm. The expandable implant 410 is shown in an expanded configuration and can be moved between a compressed or collapsed configuration and the expanded configuration in the same or similar manner as described above for previous embodiments.

The expandable implant 410 includes at least one ribbon-like strand of porous mesh configured to be expanded within the aneurysm as a 360 degree spiral or ring-shaped structure. In the expanded configuration, a first portion 420 of the expandable implant 410 is overlapped by a second portion

(not shown in FIG. 7) of the expandable implant, which is overlapped by a third portion 450 of the expandable implant. In this manner, at least a portion of the expandable implant 410 includes two, three, four, or more layers of implant material (e.g., porous mesh, as described above in previous embodiments), which can be positioned over the neck of the aneurysm from within the aneurysm to function as a dense flow disruptor. In some embodiments, a radiopaque marker 442 is coupled to the expandable implant 410.

FIG. 8 illustrates another embodiment of a medical device. The medical device 500 can include the same or similar features and functions as described above for medical device 400. For example, the medical device 500 includes an expandable implant 510 and an insertion portion or member 502. The medical device 500 can be delivered to an aneurysm or other vascular defect using a microcatheter 504. The expandable implant 510 is sized to occupy at least a portion of the volume defined by the sac of the aneurysm, and the insertion member 502 is configured to facilitate delivery of the expandable implant into the sac of the aneurysm. The expandable implant 510 is shown in an expanded configuration and can be moved between a compressed or collapsed configuration and the expanded configuration in the same or similar manner as described above for previous embodiments.

The expandable implant 510 includes a porous mesh configured to be expanded within the aneurysm as a substantially circular or disc-shaped structure, as shown in FIG. 8. In the expanded configuration, a first end portion 512 of the expandable implant 510 is engaged with and/or overlapped with a second end portion 514 of the expandable implant. The expandable implant 510 includes a first portion 520 having a first density of porous mesh and a second portion 530 having a second, higher, density of porous mesh. More specifically, a weave or braid of the porous mesh has a higher density in the second portion 530 than in the first portion 520 of the expandable implant. The expandable implant 510 is configured to be disposed within the aneurysm (or other vascular defect) such that at least a portion of the second portion 530 is disposed over the neck of the aneurysm, because the higher density promotes endothelial cell attachment to the expandable implant. The expandable implant 510 includes at least one radiopaque marker 542, which can be disposed on one of the first end portion 512 (as shown in FIG. 8) and/or the second end portion 514. When the expandable implant 510 is disposed within the aneurysm in its expanded configuration such that the higher density second portion 530 is disposed over the neck of the aneurysm, the at least one radiopaque marker 542 is disposed within the sac of the aneurysm away from the neck of the aneurysm.

FIG. 9 illustrates another embodiment of a medical device. The medical device 600 can include the same or similar features and functions as described above for previous embodiments. For example, the medical device 600 includes an expandable implant 610 and an insertion portion or member 602. The expandable implant 610 is sized to occupy at least a portion of a volume defined by the sac of the aneurysm, and the insertion member 602 is configured to facilitate delivery of the expandable implant into the sac of the aneurysm. The expandable implant 610 is shown in an expanded configuration and can be moved between a compressed or collapsed configuration and the expanded configuration in the same or similar manner as described above for previous embodiments.

The expandable implant 610 includes a ribbon-like strand of porous mesh having at least two layers of mesh. The

expandable implant **610** is configured to be expanded within the aneurysm as a substantially helical or coil shaped structure, as shown in FIG. **9**. The expandable implant **610** can be disposed within the aneurysm (or other vascular defect) such that at least a portion of the implant is disposed over the neck of the aneurysm to facilitate endothelial cell attachment at the neck. The expandable implant **610** includes at least one radiopaque marker **642**, which can be disposed on an end of the expandable implant **610**, as shown in FIG. **9**. The insertion member **602** can be removably coupled to the expandable implant at the radiopaque marker.

FIG. **10** illustrates another embodiment of a medical device. A medical device **700** includes all the same or similar features and functions as described above for medical device **600**. For example, the medical device **700** includes an expandable implant **710**, an insertion portion or member **702**, and a radiopaque marker **742** coupled to an end of the expandable implant. The expandable implant **710** includes a porous mesh formed of a tubular or rounded braid structure. The rounded braid structure can lend more softness to the expandable implant **710** than, for example, the flattened ribbon-like structure previously described.

FIG. **11** illustrates another embodiment of a medical device. The medical device **800** can include the same or similar features and functions as described above for previous embodiments. For example, the medical device **800** includes an expandable implant **810** and an insertion portion or member **802**. The medical device **800** can be delivered to an aneurysm or other vascular defect using a microcatheter **804**. The expandable implant **810** is sized to occupy at least a portion of the volume of the sac of the aneurysm, and the insertion member **802** is configured to facilitate delivery of the expandable implant from the microcatheter **804** into the sac of the aneurysm. The expandable implant **810** is shown in an expanded configuration and can be moved between a compressed or collapsed configuration and the expanded configuration in the same or similar manner as described above for previous embodiments.

The expandable implant **810** includes a first member **820** and a second member **830**. The first and second members **820**, **830** are coupled at a first end **812** of the expandable implant **810** and a second end **814** of the expandable implant. The first and second members **820**, **830** are also coupled together at at least one middle portion of the expandable implant **810** between the first end **812** and the second end **814**. The first and second members **820**, **830** can be coupled, for example, using radiopaque markers **842**, **844**, **846**. Each site of coupling is configured to be a folding point of the expandable implant **810** when the expandable implant is delivered into the aneurysm and is expanded within the aneurysm to comply with the shape of the aneurysm. As such, the expandable implant **810** can be more densely packed into the aneurysm, for example, as compared to an implant that cannot bend or fold in response to the shape of the aneurysm. At least one of the first member **820** and the second member **830** of the expandable implant **810** includes a porous mesh formed of a tubular or rounded braid structure.

FIG. **12** illustrates another embodiment of a medical device. The medical device **900** can include the same or similar features and functions as described above for previous embodiments. For example, the medical device **900** includes an expandable implant **910** and an insertion portion or member **902**. The expandable implant **910** is sized to occupy the sac of the aneurysm, and the insertion member **902** is configured to facilitate delivery of the expandable implant from a microcatheter (not shown in FIG. **12**) into the

sac of the aneurysm. The expandable implant **910** is shown in an expanded configuration and can be moved between a compressed or collapsed configuration and the expanded configuration in the same or similar manner as described above for previous embodiments.

The expandable implant **910** includes a series of expandable portions **920**, **922**, **924**, **926**, **928** separated by a series of constricted portions **930**, **932**, **934**, **936**. The expandable portions **920**, **922**, **924**, **926**, **928** can be configured to expand to any suitable multi-dimensional shape, including, for example, that resembling a sphere, a disc, a parabola, or the like. Additionally, each expandable portion **920**, **922**, **924**, **926**, **928** can have an expanded shape distinct from an expanded shape of another expandable portion.

When the expandable implant **910** is in its expanded configuration, as shown in FIG. **12**, the expandable portions **920**, **922**, **924**, **926**, **928** are more porous and less dense than the constricted portions **930**, **932**, **934**, **936**. The density and/or porosity of each expandable portion **920**, **922**, **924**, **926**, **928** can be varied from the other expandable portions **920**, **922**, **924**, **926**, **928**, and the density and/or porosity of each expandable portion **920**, **922**, **924**, **926**, **928** can be varied along a length and/or width of the respective expandable portion. For example, a first expandable portion **920** can be more dense and/or less porous proximate to a first constriction portion **930** and less dense and/or more porous at a middle, wider portion of the first expandable portion **920**. Additionally, the expandable portions **920**, **922**, **924**, **926**, **928** are each configured to have a width greater than when the expandable implant **910** is in its collapsed configuration, and the constricted portions **930**, **932**, **934**, **936** are each configured to have a width narrower than a width of the expandable portions **920**, **922**, **924**, **926**, **928**. As such, the expandable implant **910** is configured to bend, curve, and/or fold at the constricted portions **930**, **932**, **934**, **936** to help comply with the shape of the aneurysm.

When the expandable implant **910** is in its expanded configuration, the first expandable portion **920** is configured to have a width greater than the width of the other expandable portions **922**, **924**, **926**, **928**. The first expandable portion **920** can be, as illustrated in FIG. **12**, the most proximal of the expandable portions **920**, **922**, **924**, **926**, **928**. The first expandable portion **920** is configured to be positioned over a neck of the aneurysm when the expandable implant **910** is disposed within the aneurysm in its expanded configuration. In this manner, the first expandable portion **920** is configured to act as a flow disruptor at the neck of the aneurysm to help limit the flow of blood into the aneurysm from the parent blood vessel. The remaining, more distal, expandable portions **922**, **924**, **926**, **928** are configured to be packed into the aneurysm to embolize the aneurysm.

The expandable implant **910** includes a first radiopaque marker **942** coupled to a first end **912** of the implant and a second radiopaque marker **944** coupled to a second end **914** of the implant. The radiopaque markers **942**, **944** are configured to be wholly disposed within the sac of the aneurysm when the expandable implant **910** is disposed in the aneurysm in its expanded configuration.

FIG. **13** illustrates another embodiment of a medical device. The medical device **1000** can include the same or similar features and functions as described above for previous embodiments. For example, the medical device **1000** includes an expandable implant **1010** and an insertion portion or member **1002**. The expandable implant **1010** is sized to occupy the sac of the aneurysm, and the insertion member **1002** is configured to facilitate delivery of the expandable implant into the sac of the aneurysm. The expandable

implant **1010** is shown in an expanded configuration and can be moved between a compressed or collapsed configuration and the expanded configuration in the same or similar manner as described above for previous embodiments.

The expandable implant **1010** includes a first porous member **1020** and a second porous member **1030**. The first porous member **1020** includes a porous mesh configured to have a multi-dimensional shape when the expandable implant **1010** is in its expanded configuration. As such, the first porous member **1020** has a second width in the expanded configuration that is greater than a first width of the first porous member in the collapsed configuration. The first porous member **1020** can be configured to expand to any suitable multi-dimensional shape, including, for example, that resembling a parabola, as shown in FIG. **13**, a sphere, a disc, or the like. The first porous member **1020** is configured to be positioned over a neck of the aneurysm when the expandable member **1010** is disposed within the sac of the aneurysm to disrupt and/or stop the flow of blood into the aneurysm from the parent blood vessel. Additionally, the porous mesh of the first porous member **1020** is configured to promote endothelial cell attachment at the neck of the aneurysm, which can help to heal over the neck of the aneurysm.

The second porous member **1030** includes a porous mesh configured to have a multi-dimensional shape when the expandable implant **1010** is in its expanded configuration. As such, the second porous member **1030** has a fourth width in the expanded configuration greater than a third width of the second porous member in the collapsed configuration. The second porous member **1030** can be configured to expand to any suitable multi-dimensional shape, including, for example, that resembling a tube, as shown in FIG. **13**, a sphere, a disc, a parabola, or the like. In the embodiment illustrated in FIG. **13**, the second width of the first porous member **1020** is greater than the fourth width of the second porous member **1030**. The second porous member **1030** is configured to be disposed within the sac of the aneurysm such that the first porous member **1020** is disposed between the second porous member **1030** and the neck of the aneurysm. The second porous member **1030** is configured to be packed into the aneurysm to embolize the aneurysm.

A radiopaque marker **1044** is disposed between the first porous member **1020** and the second porous member **1030**, and can be used to couple the first and second porous members. The expandable implant **1010** is configured to bend, curve, and/or fold at the radiopaque marker **1044**, which can help the expandable implant **1010** comply with the shape of the sac of the aneurysm. Another radiopaque marker **1042** can be disposed on a proximate end of the expandable implant **1010**, and can be used to couple the insertion portion **1002** to the expandable implant. The radiopaque markers **1042**, **1044** are configured to be wholly disposed within the sac of the aneurysm when the expandable implant **1010** is disposed in the aneurysm in its expanded configuration.

FIGS. **14-15** illustrate another embodiment of a medical device. The medical device **1100** can include the same or similar features and functions as described above for previous embodiments. For example, the medical device **1100** includes a first porous member **1120**, a second porous member **1130**, and an insertion portion or member **1102** removably couplable to the first and second porous members **1120**, **1130**.

The first porous member **1120** has a first end **1122** and a second end **1124**. As shown in FIG. **14**, the first porous member **1120** has a collapsed configuration for insertion

through a blood vessel. In its collapsed configuration, the first porous member **1120** is substantially elongate with a first length. As shown in FIG. **15**, the first porous member **1120** has an expanded configuration for occupying a sac of an aneurysm. When the first porous member **1120** is in its expanded configuration, it has a three-dimensional shape and defines an open interior region **1126**. The first porous member **1120** can have any suitable three-dimensional shape. For example, the first porous member **1120** can be configured to curve into a substantially spherical shape, as shown in FIG. **15**. Additionally, in its expanded configuration, the first porous member **1120** includes a first segment configured to overlap with a second segment, which can be similar in many respects as described above with respect to expandable implants **210** and **310**, for example. For example, the first porous member **1120** can include a mesh having a first segment configured to overlap with a second segment of the porous mesh to form a higher density portion of the first porous member **1120**.

The second porous member **1130** has a first end **1132** and a second end **1134**. The second porous member **1130** has a collapsed, first, configuration (not shown in FIG. **14** or **15**) for insertion through a blood vessel. In its collapsed configuration, the second porous member **1130** is substantially elongate with a second length less than the first length of the first porous member, and is configured to occupy a first volume. As shown in FIGS. **14** and **15**, the second porous member **1130** has an expanded, second, configuration for occupying at least a portion of the volume of the sac of the aneurysm. When the second porous member **1130** is in its expanded configuration, it has a three-dimensional shape and is configured to occupy a second volume greater than the first volume. The second porous member **1130** can have any suitable three-dimensional shape. For example, the second porous member **1130** can be configured to expand into a substantially ball (e.g., spherical, round, oblong, or the like) shape, as shown in FIGS. **14** and **15**. In the expanded configuration, the second porous member **1130** can have a porosity the same as, or different than, a porosity of the first porous member **1120**. The second porous member **1130** is configured to be disposed in the interior region **1126** of the first porous member **1120** when each of the first porous member and the second porous member are in the deployed or expanded configurations.

In the embodiment illustrated in FIGS. **14** and **15**, the second porous member **1130** is coupled to the first porous member **1120**. Specifically, the first end **1122** of the first porous member **1120** is coupled to the first end **1132** of the second porous member **1130**. At least one of the first porous member **1120** and the second porous member **1130** includes a radiopaque marker. As shown in FIG. **14**, a first radiopaque marker **1142** can be disposed on the first ends **1122**, **1132** of the first and second porous members **1120**, **1130** to couple the first and second porous members together. A second radiopaque marker **1144** can be disposed on the second end **1134** of the second porous member **1130**. When the first and second porous members **1120**, **1130** are in their respective expanded configurations, the second radiopaque marker **1144** is disposed within the interior region defined by the first porous member **1120**.

In use, the first and second porous members **1120**, **1130**, and the first and second radiopaque markers **1142**, **1194**, are wholly disposed within the aneurysm. The second porous member **1130** can be inserted into the aneurysm first and assume its expanded configuration therein. The first porous member **1120** can then be inserted into the aneurysm such that the first porous member curves, coils, or otherwise

wraps around the second porous member **1130** as the first porous member moves to its expanded configuration. The first porous member **1120** is configured to be disposed within the aneurysm such that a portion of the first porous member is disposed over the neck of the aneurysm. For example, the higher density portion of the first porous member **1120** at which the first segment overlaps the second segment can be positioned over the neck of the aneurysm to promote endothelial cell attachment at the aneurysm neck. The second porous member **1130** can help to embolize the aneurysm by providing additional porous mesh within the sac of the aneurysm for cell attachment and/or clot formation. As such, the second porous member occupies a portion of the volume of the sac of the aneurysm such that blood flow through the aneurysm is further inhibited.

Although the medical device **1100** includes discrete first and second porous members **1120**, **1130**, respectively, in other embodiments, the first and second porous members can be differently constructed. For example, referring to FIG. **16**, an embodiment of a medical device **1200** is illustrated. The medical device **1200** can include the same or similar features and functions as described above for medical device **1100**, or other previous embodiments. For example, the medical device **1200** includes a first porous member **1220**, a second porous member **1230**, and an insertion portion or member (not shown in FIG. **16**) removably coupleable to the first and second porous members. Each of the first porous member **1220** and the second porous member **1230** can be similar in form and function as the first porous member **1120** and the second porous member **1130**, respectively, described above.

In the embodiment illustrated in FIG. **16**, however, the second porous member **1230** is monolithically constructed with the first porous member **1220**. It should be noted that in FIG. **16**, the first and second porous members **1220**, **1230**, are shown in an expanded configuration but the second porous member **1230** is shown spaced apart from the first porous member **1220** for illustration purposes only. In use, in their respective deployed or expanded configurations, the second porous member **1230** is disposed within an interior region **1226** defined by the first porous member **1220** in a similar manner as that illustrated in FIG. **15** with respect to medical device **1100**. Additionally, the medical device **1200** includes two radiopaque markers **1242**, **1244**. A first radiopaque marker **1242** is disposed at an end of a porous mesh of the first porous member **1220**, and the second radiopaque marker **1244** is disposed at an opposing end of porous mesh of the second porous member **1230**.

In some embodiments, a medical device includes an expandable implant that has a substantially continuous outer surface when in an expanded configuration. Referring to FIGS. **17A** and **17B**, a portion of a medical device **1300** according to an embodiment is illustrated in a collapsed configuration and an expanded configuration, respectively. The medical device **1300** can include the same or similar features and functions as described herein for other embodiments. For example, the medical device **1300** can include an expandable implant **1310** configured to move from the collapsed configuration (e.g., for delivery through a blood vessel) to the expanded configuration (e.g., for deployment within an aneurysm). The expandable implant **1310** includes at least a first portion **1320** and a second portion **1330**, and can include additional portions **1340**, **1350**, **1360**. When the expandable implant **1310** is in its expanded configuration, the expandable implant **1310** has a three-dimensional shape (e.g., a substantially spherical shape) with a substantially continuous outer surface such that edges of at least two of

the portions **1320**, **1330**, **1340**, **1350**, **1360** overlap. For example, edges of the first portion **1320** and the second portion **1330** can overlap, as shown in FIG. **17B**. In other words, the expandable implant **1310** moves into the expanded configuration such that few or no openings or spaces remain between edges of the portions **1320**, **1330**, **1340**, **1350**, **1360** of the expandable implant **1310**.

FIG. **18** is a flowchart illustrating a method **80** of using a medical device to disrupt blood flow into an aneurysm and to promote healing of the aneurysm, as described herein, according to an embodiment. The method **80** includes at **82**, positioning a catheter adjacent to an aneurysm of a blood vessel. For example, a distal portion of the catheter can be positioned adjacent an opening from the blood vessel into the aneurysm. The catheter defines an elongate lumen, which can be configured to receive at least a portion of the medical device for delivery to the aneurysm.

At **84**, optionally, an expandable implant of the medical device is inserted into the catheter. The expandable implant includes a first portion and a second portion, each of which has a first (e.g., insertion or collapsed) configuration and a second (e.g., deployed or expanded) configuration. In the second configuration, the first portion substantially overlaps the second portion. Each of the first portion and the second portion also include a porous mesh. The porous mesh has a first porosity when in the first configuration and a second porosity when in the second configuration. The second porosity can be, for example, greater than the first porosity. The expandable implant can be biased in its second configuration before being inserted into the catheter. The expandable implant is in its first configuration when the expandable implant is disposed in the lumen of the catheter. The expandable implant can be inserted into the catheter after the catheter is positioned within the blood vessel, before the catheter is introduced into the blood vessel, or any time therebetween.

At **86**, the expandable implant is optionally oriented to the opening in the vessel wall in fluid communication with the aneurysm. In this manner, the expandable implant is oriented to enter a sac of the aneurysm when the expandable implant is moved out of the catheter, as described in more detail herein.

At **88**, the expandable implant is moved from a first position inside the catheter to a second position outside the catheter. For example, the expandable implant can be moved from a first position inside the lumen of the catheter to a second position in at least one of the blood vessel or the aneurysm outside of the catheter. As noted above, the expandable implant is in its first configuration when in its first position inside the catheter. The expandable implant is moved to its second configuration when in its second position outside of the constraint of the catheter. The second portion of the expandable implant can be moved to its second configuration before the first portion is moved to its second configuration. In their respective second configurations, the second portion can be disposed in an interior region defined by the first portion. For example, the second portion can be moved to its second configuration in which it has a multi-dimensional expanded shape, and then the first portion can be moved to its second configuration in which it curves into a multi-dimensional expanded shape around the second portion.

The medical device can include an insertion portion configured to move the expandable implant from its first position to its second position. The insertion portion can be, for example, a wire coupled to one of the first portion or the second portion of the expandable implant. At **90**, the inser-

tion portion is optionally disconnected from the expandable implant. For example, the insertion portion can be disconnected from a proximal end of the expandable implant, such as after the expandable implant has been inserted into the aneurysm. At 92, the insertion portion is optionally removed from the blood vessel through the catheter.

After the expandable implant is disposed within the aneurysm, or other target vascular defect, the portion of a patient's body including the aneurysm can be imaged (e.g., using X-ray or other suitable imaging techniques) to determine whether the expandable implant is properly positioned within the aneurysm. For example, the expandable implant can include one or more radiopaque markers that are visible using X-ray. In another example, the patient can be injected intravenously with a radiopaque dye at a desired time following implantation of the expandable implant to determine the success of endothelial cell attachment and/or healing over of the neck of the aneurysm following the procedure. If radiopaque dye is visible within the parent blood vessel adjacent the aneurysm, but not within the aneurysm itself, the expandable implant has operated to successfully prevent further blood flow into the aneurysm. If radiopaque dye is visible within the aneurysm, blood flow from the parent blood vessel has not been completely prevented and additional treatment options may be considered by the health care practitioner.

FIG. 19A illustrates a portion of another embodiment of a medical device. The medical device 1400 can include the same or similar features and functions as described above for previous embodiments. For example, the medical device 1400 includes an expandable implant 1410 and an insertion portion or member (not shown in FIG. 19A). The expandable implant 1410 is shown in an expanded configuration and can be moved between a compressed or collapsed configuration in which the expandable implant 1410 is substantially elongate and the expanded configuration in the same or similar manner as described above for previous embodiments.

The expandable implant 1410 includes a ribbon-like strand of porous mesh and includes petal-like portions or sections 1425 and 1427 along its length. At least a portion of the porous mesh is configured to be positioned over a neck of an aneurysm when the expandable implant 1410 is in the expanded configuration. The expandable implant 1410 includes a first portion 1420 that includes the petal-like portions 1427 and a second portion 1430 that includes the petal-like portions 1427. The petal-like portions 1425 of the second portion 1430 are larger than the petal-like portions 1427 of the first portion 1420 such that when the expandable implant 1410 is moved to its expanded configuration, the petal-like portions 1425 of the second portion at least partially overlap the petal-like portions 1427 of the first portion 1420. During deployment of the expandable implant 1410 (e.g., when moved from its collapsed configuration to its expanded configuration) the petal-like portions 1425 of the second portion 1430 will deploy first, and then the petal-like portions 1427 of the first portion 1420 will deploy at least partially within an interior region defined by the second portion 1430. The petal-like portions 1425 of the second portion 1430 can be sized and configured to be disposed at a neck of an aneurysm when the expandable implant 1410 is in the expanded configuration. The petal-like portions 1427 of the first portion 1420 can be formed in a smaller diameter fixture than the petal-like portions 1425, and can be sized and configured to substantially fill the aneurysm and to hold the second portion 1430 in place at the neck of the aneurysm when the expandable implant 1410 is

in the expanded configuration. For example, the petal-like portions 1427 of the first portion 1420 can have a diameter of about 2 mm-12 mm, and the petal-like portions 1425 of the second portion 1430 can have a corresponding diameter of about 1 mm larger than the petal-like portions 1427 of the first portion 1420. For example, the petal-like portions 1425 of the second portion 1430 can be about 3 mm-13 mm. FIG. 19B is a schematic illustration of the expandable implant 1410 in its expanded configuration showing the positional relationship of the first portion 1420 to the second portion 1430.

As described for previous embodiments, a first radiopaque marker 1442 is coupled to a first end of the expandable implant 1410 and a second radiopaque marker (not shown) is coupled to a second end of the expandable implant 1410. The expandable implant 1410 is configured to be wholly disposed within the aneurysm such that the radiopaque markers are wholly disposed within the aneurysm sac and the porous mesh is disposed over the neck of the aneurysm. In some embodiments, the radiopaque markers are configured to be positioned at a side of the aneurysm (i.e., disposed away from the neck of the aneurysm).

FIG. 20 illustrates a portion of another embodiment of a medical device. The medical device 1500 can include the same or similar features and functions as described above for previous embodiments. For example, the medical device 1500 includes an expandable implant 1510 and an insertion portion or member (not shown in FIG. 20). The expandable implant 1510 is shown in an expanded configuration and can be moved between a compressed or collapsed configuration in which the expandable implant 1510 is substantially elongate and the expanded configuration in the same or similar manner as described above for previous embodiments.

As with the previous embodiment, the expandable implant 1510 includes a ribbon-like strand of porous mesh. At least a portion of the porous mesh is configured to be positioned over a neck of an aneurysm and at least another portion of the porous mesh substantially fills the volume of the aneurysm when the expandable implant 1510 is in the expanded configuration. The expandable implant 1510 includes a first portion 1520 and a second portion 1530. In this embodiment, each of the first portion 1520 and the second portion 1530 form a sphere when the expandable implant 1510 is in its expanded configuration. One of the first portion 1520 or the second portion 1530 can be configured to be disposed at a neck of the aneurysm and the other of the first portion 1520 or the second portion 1530 can substantially fill the volume of the aneurysm. For example, in this embodiment, the first portion 1520 can be configured to be deployed at the dome of an aneurysm and serve as an anchor for the second portion 1530 and the second portion 1530 can be disposed across the neck of the aneurysm when the expandable implant 1510 is in the expanded configuration. The expandable implant 1510 can also include radiopaque markers (not shown) as described above for previous embodiments.

FIGS. 21 and 22 illustrate another embodiment of a medical device. The medical device 1600 can include the same or similar features and functions as described above for previous embodiments. For example, the medical device 1600 includes an expandable implant 1610 and an insertion portion or member (not shown). The expandable implant 1610 is shown in an expanded configuration and can be moved between a compressed or collapsed configuration as shown in FIG. 22 and the expanded configuration as shown in FIG. 21 in the same or similar manner as described above for previous embodiments.

As with the previous embodiment, the expandable implant **1610** includes a ribbon-like strand of porous mesh that includes a first portion **1620** in the form of a disc-shaped structure and a second portion **1630** that includes petal-like portions or sections along its length (similar to the embodiment of FIG. **19A**). The disc or spherical shaped structure of the first portion **1620** can be disposed at various locations along the length (e.g., middle, end, etc.) of the expandable implant **1610**. At least a portion of the porous mesh is configured to be positioned over a neck of an aneurysm when the expandable implant **1610** is in the expanded configuration. In this embodiment, when the expandable implant **1610** is in the expanded configuration, the petal-like portions of the second portion **1630** at least partially overlap the disc-shaped structure of the first portion **1620**. For example, when the expandable implant **1610** is in its expanded configuration, the petal-like portions of the second portion **1630** can define a diameter greater than a diameter defined by the disc or spherical shaped structure of the first portion **1620**. The expandable implant **1610** can also include a first radiopaque marker **1642** coupled to a first end **1612** of the expandable implant **1610** and a second radiopaque marker (not shown) coupled to a second end (not shown) of the expandable implant **1610**. The expandable implant **1610** can also include a connector **1652** coupled to a first end **1612** of the expandable implant **1610**.

When the expandable implant **1610** is in its expanded configuration, the expandable implant **1610** has a three-dimensional shape (e.g., a substantially spherical shape) with a substantially continuous outer surface such that edges of at least two of the petal-like portions **1625** overlap each other (in a similar manner as the embodiment of FIGS. **17A** and **17B**), and at least partially overlap the disc-shaped portion **1620**. The expandable implant **1610** can move into the expanded configuration such that few or no openings or spaces remain between petal-like portions **1625** of the expandable implant **1610**.

FIGS. **23** and **24** illustrate a portion of another embodiment of a medical device. The medical device **1800** can include the same or similar features and functions as described above for previous embodiments. For example, the medical device **1800** includes an expandable implant **1810** and an insertion portion or member (not shown in FIGS. **23** and **24**). The expandable implant **1810** can be moved between a collapsed configuration as shown in FIG. **23** and an expanded configuration as shown in FIG. **24**.

Similar to the embodiment of FIG. **19A**, the expandable implant **1810** includes a ribbon-like strand of porous mesh that includes petal-like portions or sections **1825** along its length. At least a portion of the porous mesh is configured to be positioned over a neck of an aneurysm when the expandable implant **1810** is in the expanded configuration. When the expandable implant **1810** is in its expanded configuration, the expandable implant **1810** has a three-dimensional shape (e.g., a substantially spherical shape) with a substantially continuous outer surface such that edges of at least two of the petal-like portions **1825** overlap each other as shown in FIG. **24**.

In this embodiment, when the implantable implant **1810** is formed, the ribbon-like strand of porous mesh is wrapped around the forming fixture in a multi-directional fashion. For example, a portion of the mesh can be wrapped in a continuous manner around the fixture as indicated at C in FIG. **23**, and a portion of the mesh can be wrapped in an s-shape manner as indicated at S in FIG. **23**. With such forming, when the expandable implant **1810** is moved to its expanded configuration, the petal-like portions **1825** that

have been formed by wrapping in a continuous manner will follow each other (each petal-like portion **1825** will cause the adjacent petal-like portion **1825** to collapse), and the petal-like portions **1825** that have been formed in a s-shape manner will individually self-deploy or collapse. The multi-directional heat forming of the expandable implant **1810** can allow the expandable implant **1810** to deploy fragmented within an aneurysm.

In this embodiment, the medical device **1800** also includes a PT coil or PT strand **1835** disposed along the length of the expandable implant **1810** to provide for a portion of the expandable implant **1810** to be radiopaque. As shown in FIG. **23**, the PT strand **1835** is disposed along a length of the expandable implant **1810** and across or within the petal-like portions **1825**. The PT strand **1835** can be coupled to, for example, marker bands (not shown) disposed on a proximal end and a distal end of the expandable implant **1810**. In some embodiments, a PT strand **1835** can be braided within the mesh of the expandable implant **1810**.

In some embodiments, the PT strand **1835** can also be used to prevent over-stretching of the expandable implant **1810** when being delivered to a treatment site. For example, as described above, the PT strand **1835** can be coupled to the proximal end and the distal end of the expandable implant **1810**. Thus, the PT strand **1835** can define a maximum length in which the expandable implant **1810** can be stretched or extended lengthwise during insertion and prevent overstretching. In alternative embodiments, a separate component can be used to limit the length of the expandable implant **1810**. For example, in some embodiments, a separate wire member in addition to a PT strand can be used. In some embodiments, an expandable implant may not include a PT strand, such as PT strand **1835**. In such embodiments, a separate wire member can be coupled to the proximal end and distal end of the expandable member and used to limit the length or amount of stretch of the expandable implant in a similar manner.

In some embodiments, a medical device can include a strand formed with, for example, a suture that extends along or within the medical device. The suture strand can reinforce the medical device along its length. In some embodiments, a radiopaque coil can be placed over the suture strand to enhance visibility of the medical device under fluoroscopy.

FIGS. **25-27** illustrate a portion of another embodiment of a medical device. The medical device **1900** can include the same or similar features and functions as described above for previous embodiments. For example, the medical device **1900** includes an expandable implant **1910** and an insertion portion or member (not shown in FIGS. **25-27**). The expandable implant **1910** can be moved between a collapsed configuration (as shown in FIG. **25**, a partially expanded configuration as shown in FIG. **26**, and an expanded configuration as shown in FIG. **27**).

The expandable implant **1910** includes a ribbon-like strand of porous mesh that includes a first portion **1920** (see FIGS. **25-27**) and a second portion **1930** (shown only in FIG. **27**). In this embodiment, the first portion **1920** and the second portion **1930** are separate components that can be deployed together. The first portion **1920** includes disc-shaped portions **1945** along its length, and the second portion **1930** includes petal-like portions **1925**, as described above for previous embodiments. When the expandable implant **1910** is in its expanded configuration, the expandable implant **1910** has a three-dimensional shape (e.g., a substantially spherical shape) as shown in FIG. **27**.

During deployment of the medical device **1900**, the second portion **1930** can be deployed first such that the

petal-like portions **1925** are moved to an expanded configuration and define an interior region **1936**. The first portion **1920** can then be deployed such that the disc-shape portions **1945** will collapse upon each other (as shown in FIGS. **26** and **27**) within the interior region **1936** of the second portion **1930**, as shown in FIG. **27**. In other words, when the expandable implant **1910** is in the expanded configuration, the second portion **1930** at least partially overlaps the first portion **1920**, as shown in FIG. **27**. At least a portion of the porous mesh is configured to be positioned over a neck of an aneurysm when the expandable implant **1910** is in the expanded configuration. For example, when the expandable implant **1910** is in its expanded configuration, the second portion **1930** can be disposed at the neck of the aneurysm to disrupt blood flow, and the first portion **1920** can help occlude the aneurysm at a relatively fast rate. Although this embodiment illustrates the first portion **1920** and the second portion **1930** as separate components, in an alternative embodiment, the first portion **1920** and the second portion **1930** can be formed with a single mesh component.

In this embodiment, the medical device **1900** can also include a PT coil or PT strand (not shown) disposed along the length of first portion **1920** and/or the second portion **1930** of the expandable implant **1910** in a similar manner as described above for medical device **1800**. The PT strand can be coupled to a first marker band **1942** disposed at a first end **1912** of the expandable implant **1910** and a second marker band **1944** disposed on a second end of the expandable implant **1910** as shown in FIG. **27**. As described above, the PT strand can be braided within the mesh of the expandable implant **1910**. As shown in FIGS. **26** and **27**, the expandable member **1910** also includes a connector member **1952** that can be used to couple the expandable member **1910** to a detachment device as described in more detail below (see e.g., discussion of FIG. **40**).

FIGS. **28** and **29** illustrate another embodiment of a medical device. A medical device **2000** can include all the same or similar features and functions as described above for previous embodiments. For example, the medical device **2000** includes an expandable implant **2010**, an insertion portion or member **2002**, a first radiopaque marker **2042** coupled to a first end **2012** of the expandable implant **2010** and a second radiopaque marker **2044** coupled to a second end **2014** of the expandable implant **2010**. The expandable implant **2010** can be moved between a collapsed configuration (not shown) and an expanded configuration as shown in FIGS. **28** and **29**.

In this embodiment, the expandable implant **2010** includes three tubular or rounded strands **2020**, **2030** and **2015** formed of a porous mesh similar to the tubular structures described above, for example, with respect to FIGS. **10** and **11**. In some embodiments, the strands **2020**, **2030** and **2015** can be braided. In alternative embodiments, the strands **2020**, **2030** and **2015** can be formed with ribbon-like strands of porous mesh rather than tubular strands. When the expandable implant **2010** is in its expanded configuration, at least a portion of the tubular strands **2020**, **2030** and **2015** can overlap each other as shown in FIG. **29**. The expandable implant **2010** can be used to fill a volume of an aneurysm and can be used alone or in conjunction with another expandable implant to fill the volume of an aneurysm.

The tubular mesh can be, for example, 1 mm tubular mesh. In this embodiment, the tubular strands **2020**, **2030**, **2015** can be heat-shaped such that the expandable implant **2010** has a 2D configuration when the expandable implant **2010** is in its expanded configuration. In this embodiment, three tubular strands are included, but in alternative embodi-

ments a different number of tubular strands can be included. For example, an expandable implant can be formed with 1-10 tubular strands. The tubular strands **2020**, **2030** and **2015** can be coupled together at various locations along their lengths with marker bands, such as marker band **2046** shown in FIG. **29**. In alternative embodiments, the tubular strands can be twisted together, or braided together rather than using marker bands. In some embodiments, the strands are not coupled together.

FIG. **30** illustrates another embodiment of a medical device including tubular structures. A medical device **2100** can include all the same or similar features and functions as described above for previous embodiments. For example, the medical device **2100** includes an expandable implant **2110** and an insertion portion or member **2102**. Although not shown in FIG. **30**, the medical device **2100** can also include radiopaque markers coupled to end portions to the expandable implant **2110**. The expandable implant **2110** can be moved between a collapsed configuration (not shown) and an expanded configuration as shown in FIG. **30**.

The expandable implant **2110** includes three tubular or rounded strands **2120**, **2130** and **2115** formed of a porous mesh similar to the tubular strands described above for medical device **2000**. When the expandable implant **2110** is in its expanded configuration, at least a portion of the tubular strands **2120**, **2130** and **2115** can overlap each other as shown in FIG. **30**. In this embodiment, the tubular strands **2120**, **2130**, **2115** can be heat-shaped to have a 3D configuration when the expandable implant **2110** is in the expanded configuration. In this embodiment, three tubular strands are included, but in alternative embodiments a different number of tubular strands can be included. For example, an expandable implant can be formed with 1-10 tubular strands. The tubular strands **2120**, **2130** and **2115** can be coupled together at various locations along their lengths with marker bands (not shown) as described above for medical device **2000**, or can be coupled using other coupling methods, such as being twisted together, or braided together. In some embodiments, the tubular strands are not coupled together.

FIG. **31** illustrates another embodiment of a medical device including tubular structures. A medical device **2200** can include all the same or similar features and functions as described above for previous embodiments. For example, the medical device **2200** includes an expandable implant **2210** and an insertion portion or member **2202**. Although not shown in FIG. **31**, the medical device **2200** can also include radiopaque markers coupled to end portions to the expandable implant **2210**, such as radiopaque marker **2242** coupled to an end **2212** shown in FIG. **31**. The expandable implant **2210** can be moved between a collapsed configuration (not shown) and an expanded configuration as shown in FIG. **31**.

In this embodiment, the expandable implant **2210** includes a single tubular or rounded braid structure **2215** formed of a porous mesh similar to the tubular structures described above for medical devices **2000** and **2100**. When the expandable implant **2210** is in its expanded configuration, at least a first portion of the tubular structure **2215** can overlap a second portion of the tubular structure **2215**, as shown in FIG. **31**. In this embodiment, the tubular structure **2215** is formed in a 2D shape configuration and the tubular structure is formed with a larger porosity mesh than medical devices **2000** and **2100**. For example, the tubular structure **2215** can be formed with a 3 mm mesh.

FIGS. **32-33** illustrate a portion of another embodiment of a medical device. The medical device **2400** can include the same or similar features and functions as described above for previous embodiments. The medical device **2400**

includes an expandable implant **2410** and can include an insertion portion or member (not shown in FIGS. **32-33**). The expandable implant **2410** can be moved between a collapsed configuration as shown in FIG. **32** and an expanded configuration as shown in FIG. **33**.

In this embodiment, the expandable implant **2410** includes a first portion **2420** formed with a ribbon-like strand of porous mesh and includes petal-like portions **2425**, and a second portion **2430** in the form of a tubular or rounded strand **2415** formed of a porous mesh similar to the tubular strands described above, for example, with respect to FIGS. **28-30**. The tubular strand **2415** can be heat formed as either a 2D or 3D configuration. In some embodiments, the tubular strand **2415** can be braided.

When the expandable implant **2410** is in its expanded configuration, at least a portion of the first portion **2420** (e.g., petal-like portions **2425**) can overlap the tubular strand **2415** of the second portion **2430**. At least a portion of the expandable implant **2410** is configured to be positioned over a neck of an aneurysm when the expandable implant **2410** is in the expanded configuration. The petal-like portions **2425** and the tubular strand **2415** can each be a variety of different sizes (e.g., diameters), such that when the expandable implant **2410** is moved to its expanded configuration, the petal-like portions **2425** of the second portion **2410** define an interior region and the tubular strand **2415** of the first portion **2420** substantially fills the interior region of the second portion **2430**. Thus, the tubular strand **2415** can be used as a filler to substantially fill a volume of an aneurysm as described above for expandable implants **2010** and **2110**.

The first portion **2420** and the second portion **2430** can be coupled together, for example, with marker bands at end portions of the first portion **2420** and the second portion **2430** and/or at other locations along a length of each of the first portion **2420** and the second portion **2430**. The first portion **2420** and the second portion **2430** can have the same or substantially the same length or can have different lengths. For example, in some embodiments, the second portion **2430** can be longer than the first portion and vice versa.

The expandable implant **2410** also includes a first radiopaque marker band **2442** disposed at a first end **2412** of the expandable member and a second radiopaque marker band **2444** disposed at a second end **2414** of the expandable implant **2410** as shown in FIG. **35**, which is a schematic illustration of the expandable implant **2410**. As shown in FIG. **34**, which is a schematic illustration of the expandable implant **2410**, the expandable member **2410** also includes a connector member **2452** that can be used to couple the expandable member to a detachment device as described in more detail below.

FIGS. **35-37** are each a schematic illustration of an insertion device that can be used to insert and deploy an implant, such as an expandable implant as described herein, at a desired location within a patient's body (e.g., within an aneurysm). An insertion device **2554** can be used in conjunction with a cannula, such as, for example, the cannula **104** described herein. For example, the insertion device **2254** can be used instead of the insertion portion **102** described herein and can be releasably or removably coupled to an implant as described in more detail below.

The insertion device **2554** includes a first elongate member **2556** defining a lumen **2557** through which a second elongate member **2558** can be movably disposed. A marker band **2564** is coupled to a distal end portion of the first elongate member **2556**. An expandable coupling member **2562** is also coupled to the distal end portion of the first

elongate member **2556**, for example, by adhesively coupling a portion of the expandable coupling member **2562** between the marker band **2564** and an outer wall of the first elongate member **2556**. The expandable coupling member **2562** can be various lengths and can in some embodiments have a length, for example, of about 1-2 mm. The expandable coupling member **2562** can be formed, for example, with a mesh material and/or a braided material.

The second elongate member **2558** can be, for example, a core wire and includes a ball member **2560** (also referred to as a "coupling member") disposed at a distal end of the second elongate member **2558**. The second elongate member **2558** can be moved between a first position in which the ball member **2560** is disposed outside of the expandable coupling member **2562** as shown in FIGS. **35** and **37**, and a second position in which the ball member **2560** is disposed within an interior region defined by the expandable coupling member **2562** as shown in FIG. **36**. Although the ball member **2560** is shown circular shaped, in alternative embodiments, the ball member **2560** can be other shapes, such as, for example, oval, elliptical, square, rectangular, triangular or other desired shapes (as shown in a side view).

To insert and deploy an expandable implant (e.g., an expandable implant as described herein) within a patient's body, a proximal end portion of the expandable implant can be coupled to a distal end portion of the insertion device **2554**. Specifically, as shown in FIG. **35**, an expandable implant **2510** (also referred to as "implant") can include an outer marker band **2543** and an inner marker band **2541** each coupled to a proximal end portion **2512** of the implant **2510**. The outer marker band **2543** can be used to hold the implant **2510**, and the inner marker band **2541** can be disposed within the outer marker band **2543**. The inner marker band **2541** can provide a channel through which a distal end portion of the insertion device **2554**, including the expandable coupling member **2562** and the ball member **2560**, can be inserted. The second elongate member **2558** is then pulled proximally (in a direction of arrow A in FIG. **36**) causing the ball member **2560** to become wedged within the expandable coupling member **2562** as shown in FIG. **36**. For example, the expandable coupling member **2562** can be moved between collapsed or relaxed configuration as shown in FIG. **35** to an expanded configuration as shown in FIG. **36** in which the expandable coupling member **2562** flexes outward or expands as the ball member **2560** is moved proximally within the expandable coupling member **2562**. A locking mechanism (not shown) can be used to lock the second elongate member **2558** in position relative to the first elongate member **2556**. For example, a handle (not shown) can be coupled to the second elongate member **2558** and can include a locking mechanism that can lock the second elongate member **2558** in the position shown in FIG. **36**. With the expandable coupling member **2562** expanded as shown in FIG. **36**, the implant **2510** is maintained coupled to the insertion device **2554**.

With the insertion device **2554** coupled to the implant **2510**, a distal end portion of the implant **2510** can be inserted into, for example, an insertion cannula or catheter (not shown) (e.g., cannula **102** described above), and the insertion cannula can be used to insert the implant **2510** into a blood vessel in a similar manner as described above with respect to FIGS. **1** and **2**. For example, the implant **2510** with the insertion device **2554** coupled thereto can be inserted into the insertion cannula such that the implant **2510** is moved into a collapsed configuration. The insertion cannula can then be inserted into a blood vessel of the patient to deliver the implant **2510** to a desired location (e.g., an

aneurysm) within the patient. At the desired location, the implant **2510** can be moved out of a distal end of the cannula and moved to its expanded configuration as described above. After the implant **2510** has been deployed, the implant **2510** can be detached from the insertion device **2554**. Specifically, to detach the insertion device **2554** from the implant **2510**, the second elongate member **2558** is unlocked and moved distally (in a direction of arrow B shown in FIG. 37) such that the ball member **2560** is moved distally outside the expandable coupling member **2562** allowing the expandable coupling member **2562** to move back to its collapsed or relaxed configuration as shown in FIG. 37. The insertion device **2554** can then be removed by pulling the insertion device **2554** proximally (in a direction of arrow A in FIG. 37).

FIG. 38 is a schematic illustration of another embodiment of an insertion device that can be used to insert and deploy an implant, such as an expandable implant as described herein. An insertion device **2654** can be used in conjunction with a cannula and can be releasably or removably coupled to an implant, as described above for insertion device **2554**.

The insertion device **2654** includes a first elongate member **2656** defining a lumen **2657** through which a second elongate member **2658** can be movably disposed. A coupling element **2666** is coupled to a distal end portion of the first elongate member **2656**, for example, by gluing a portion of the coupling element **2666** to an interior wall of the first elongate member **2656**. The coupling element **2666** can include, for example, a length of suture material, and can be various lengths. For example, the coupling element **2666** can in some embodiments have a length of about 1-2 mm. The first elongate member **2656** can also include a marker band **2664** similar to the marker band **2564** that can be coupled to a distal end portion of the first elongate member **2656**.

The second elongate member **2658** includes a ball member **2660** disposed at a distal end of the second elongate member **2658** and can be moved between a first position in which the ball member **2660** is disposed at a distance from the coupling element **2666** (e.g., at a position distal of the coupling element **2666**), and a second position in which the ball member **2660** is disposed in contact with the coupling element **2666**. For example, when the second elongate member **2658** is in its second position, the ball member **2660** is disposed at a location along a length of the coupling element **2666** and contacting the coupling element **2666** such that an interference fit is created between the ball member **2660** and the coupling element **2666** as shown in FIG. 38.

To insert and deploy an expandable implant, such as the expandable implants described herein, within a patient's body, a proximal end portion of the expandable implant (also referred to as "implant") can be coupled to a distal end portion of the insertion device **2654**. Specifically, the implant can include an outer marker band **2643** and an inner marker band **2641** each coupled to a proximal end portion of the implant. As with the previous embodiment, the outer marker band **2643** can be used to hold the implant and the inner marker band **2641** can be disposed within the outer marker band **2643** and provide a channel **2647** through which the distal end portion of the insertion device **2654** can be inserted.

With the second elongate member **2658** in its first position (i.e., with the ball member **2660** disposed at a distance from the coupling element **2666**) and the coupling element **2666** in its first configuration, the ball member **2660** and the coupling element **2666** are inserted through the inner marker band **2641** and disposed within the implant. The second

elongate member **2658** is then pulled proximally (in the direction of arrow A in FIG. 38) such that the second elongate member **2658** is moved to its second position (with the ball member **2660** contacting the coupling element **2666**) and the coupling element **2666** is moved to a second configuration as shown in FIG. 38. When the second elongate member **2658** is in its second position and the coupling element **2666** is in its second configuration an interference fit is created between the ball member **2660** and the coupling element **2666**. This interference fit holds the implant to the insertion device **2654**. As described above for the previous embodiment, a locking mechanism (not shown) can be used to lock the second elongate member **2658** in position relative to the first elongate member **2656**. For example, a handle **2655** is coupled to the second elongate member **2658** and can include a locking mechanism (not shown) that can lock the second elongate member **2658** in its second position, as shown in FIG. 38.

With the insertion device **2654** coupled to the implant, a distal end portion (not shown) of the implant can be inserted into, for example, an insertion cannula (not shown) (e.g., cannula **102** described above), and the insertion cannula can be used to insert the implant into a blood vessel in a similar manner as described above with respect to FIGS. 1 and 2 and FIGS. 35-37. For example, the implant with the insertion device **2654** coupled thereto can be pushed distally within the cannula to move the implant to a collapsed configuration. The cannula can then be inserted into a blood vessel of the patient to deliver the implant to a desired location within the patient, such as, for example, within an aneurysm, as described above. After the implant has been deployed (e.g., moved out of a distal end of the cannula), the insertion device **2654** can be detached from the implant in a similar manner as described above for the previous embodiment. Specifically, to detach the insertion device **2654** from the implant, the second elongate member **2658** is unlocked and moved distally (in the direction of arrow B in FIG. 38) such that the ball member **2660** is moved away (e.g., distally) from the coupling element **2666**, eliminating the interference fit between the ball member **2660** and the coupling element **2666**. The insertion device **2654** can then be removed by pulling the insertion device **2654** proximally (in a direction of arrow A in FIG. 38).

FIG. 39 illustrates an embodiment of an insertion device **2754** that is similar to the insertion device **2654**. The insertion device **2754** can include the same as or similar features and function the same as or similar to the insertion device **2654**. For example the insertion device **2754** can be used in the deployment of an implant as described above. The insertion device **2754** includes a first elongate member **2756** defining a lumen (not shown) through which a second elongate member **2758** (e.g., a core wire) can be movably disposed. A coupling element **2766** is coupled to a distal end portion of the first elongate member **2756**, for example, by gluing a portion of the coupling element **2766** to an interior wall of the first elongate member **2756**. The coupling element **2766** can be various lengths and can in some embodiments have a length, for example, of about 1-2 mm. The first elongate member **2756** can also include a marker band (not shown) coupled to a distal end portion of the first elongate member **2756**.

A ball member **2760** is disposed at a distal end of the second elongate member **2758** and the second elongate member **2758** can be moved between a first position in which the ball member **2760** is disposed at a distance from the coupling element **2766** (e.g., distal of the coupling element **2766**) as shown in FIG. 40 and a second position in

which the ball member 2760 is disposed in contact with the coupling element 2766 at a location along a length of the coupling element 2762 such that an interference fit is created between the ball member 2760 and the coupling element 2766. The insertion device 2754 can be used to insert and deploy an implant and be detached from the implant in the same or similar manner as described above for insertion device 2654.

FIG. 40 is a schematic illustration of another embodiment of an insertion device that can be used to insert and deploy an implant, such as an expandable implant as described herein. An insertion device 2854 can be used in conjunction with a cannula and can be releasably or removably coupled to an implant, as described above for example, for insertion device 2554.

The insertion device 2854 includes a first elongate member 2856 defining a lumen 2857 through which a second elongate member 2858 can be movably disposed. The first elongate member 2856 can also include a marker band (not shown) coupled to a distal end portion of the first elongate member 2756. An insertion ball member 2860 is disposed at a distal end of the second elongate member 2858. The insertion device 2854 can be coupled to an expandable implant 2810 similar to or the same as the expandable implants described herein. The expandable implant 2810 includes a marker band 2842 and a connector member 2852 coupled to the marker band 2842. The connector member 2852 includes a wire 2868 coupled to the marker band 2842 and/or the implant 2810 and an implant ball member 2870 coupled to (or formed integrally or monolithically with) the wire 2868. The wire 2868 and implant ball member 2870 collectively can have a length L that in some embodiments can be, for example, 1.5 mm. Although not discussed in detail above, the connector members 1652, 1952, 2352, and 2452 described above for previous embodiments of an expandable implant can include the same or similar features and functions as the connector 2852.

To insert and deploy the expandable implant 2810 within a patient's body, the expandable implant 2810 is first coupled to the insertion device 2854. Specifically, the second elongate member 2858 is moved distally (in a direction of arrow B in FIG. 40) such that the insertion ball member 2860 is disposed outside a distal end of the first elongate member 2856. The implant ball member 2870 is then inserted into the distal end of the first elongate member 2856 as shown in FIG. 40. The second elongate member 2858 is then moved proximally (in the direction of arrow A in FIG. 40) such that the insertion ball member 2860 locks or traps the implant ball member 2870 within the lumen 2857 of the first elongate member 2856 as shown in FIG. 40. For example, each of the insertion ball member 2860 and the implant ball member 2870 can have a diameter greater than half the diameter of the lumen 2857 of the first elongate member 2856 such that when the implant ball member 2870 is disposed within the lumen 2857 and the insertion ball member 2860 is moved proximally into the lumen 2857, the implant ball member 2860 cannot be pulled back out of the lumen 2857.

With the implant ball member 2870 trapped within the lumen 2857 of the first elongate member 2856, the expandable implant 2810 will be held to the insertion device 2854. As described above for the previous embodiment, a locking mechanism (not shown) can be used to lock the second elongate member 2658 in this position relative to the first elongate member 2856. With the insertion device 2854 coupled to the expandable implant 2810, a distal end portion of the expandable implant 2810 can be inserted into, for

example, an insertion cannula (not shown) (e.g., cannula 102 described above), and the insertion cannula can be used to insert the implant 2810 into a blood vessel in a similar manner as described above with respect to previous embodiments. After the expandable implant 2810 has been deployed, the insertion device 2854 can be detached from the expandable implant 2810 in a similar manner as described above for the previous embodiment. Specifically, to detach the insertion device 2854 from the expandable implant 2810, the second elongate member 2858 is unlocked and moved distally (in the direction of arrow B) such that the insertion ball member 2860 is moved distally outside of the first elongate member 2856, un-trapping the implant ball member 2870. The insertion device 2854 can then be removed by pulling the first elongate member 2856 and the second elongate member 2858 proximally.

FIG. 41 is a flowchart illustrating a method of deploying an expandable implant within an aneurysm using an insertion device as described herein. The method includes, at 2982, coupling a distal end portion of an insertion device to a proximal end portion of an expandable implant. For example, the insertion device can be an insertion device as described herein and the expandable implant can be an expandable implant as described herein. The insertion device can include a first elongate member that defines a lumen and a second elongate member movably disposed at least partially within the lumen of the first elongate member. The coupling can include moving the second elongate member proximally relative to the first elongate member such that a first coupling member on a distal end of the second elongate member engages a second coupling member on at least one of the first elongate member or the expandable implant and secures a portion of the expandable implant to the insertion device. In some embodiments, the second coupling member can be disposed on the first elongate member, and the moving the second elongate member proximally relative to the first elongate member causes the second coupling member to be moved from a collapsed configuration to an expanded configuration. In some embodiments, the second coupling member is disposed on the expandable implant, and prior to moving the second elongate member proximally, the second coupling member is inserted through a distal end of the first elongate member such that the second coupling member is disposed within the lumen of the first elongate member.

At 2984, the expandable implant can be inserted within a blood vessel of a patient while the expandable implant is in a collapsed configuration and coupled to the insertion device. For example, the expandable implant can be moved to a collapsed configuration using a cannula as described herein. At 2986, the expandable implant can be deployed within an aneurysm such that the expandable implant moves to an expanded configuration within the aneurysm. For example, the expandable implant can be moved outside the cannula such that it can move to its expanded configuration. At 2988, the insertion device can be decoupled from the expandable implant, and at 2990, the insertion device can be removed from the blood vessel of the patient.

FIGS. 42 and 43 illustrate a portion of another embodiment of a medical device. The medical device 3000 can include the same or similar features and functions as described above for previous embodiments. For example, the medical device 3000 includes an expandable implant 3010 and an insertion portion or member 3002. The expandable implant 3010 can be moved between a collapsed configuration, as shown in FIG. 43 and an expanded configuration, as shown in FIG. 42.

35

The expandable implant **3010** includes a ribbon-like strand of porous mesh that includes a first portion **3020** and a second portion **3030** formed as a single component. In this embodiment, when the expandable implant **3010** is in the expanded configuration, the second portion **3030** forms a ball-like structure that defines an interior region **3036** and the first portion **3020** can be deployed within the interior region **3036**. Specifically, during deployment of the medical device **3000**, the second portion **3030** can be deployed first such that it can be expanded to the ball-shaped structure within an aneurysm, and then the first portion **3020** can be deployed within the interior region **3036** to substantially fill the second portion **3030** as shown in FIG. **42**.

FIGS. **44-46** illustrate a portion of another embodiment of a medical device. The medical device **3100** can include the same or similar features and functions as described above for previous embodiments. For example, the medical device **3100** includes an expandable implant **3110** and an insertion portion or member **3102**. The expandable implant **3110** can be moved between a collapsed configuration, as shown in FIG. **45** and an expanded configuration, as shown in FIG. **44**.

The expandable implant **3110** is an example of a multi-layer implant that includes a ribbon-like strand of porous mesh that includes a first portion **3130**, a second portion **3120** and a third portion **3115** formed with a single mesh component. Such an embodiment may be desirable in that the implant can fit in a small delivery catheter, but can have high flow disruption by having more than two layers of material, and forming the layers in-vivo. For example, in this embodiment, when the expandable implant **3110** is in the expanded configuration, the second portion **3120** can be expanded within the third portion **3115** and the first portion **3130** can be expanded within the second portion **3120**. Specifically, during deployment within an aneurysm **A**, as shown in FIG. **46**, the medical device **3100** can first be inserted into a delivery catheter **3104** such that the expandable implant **3110** is moved to its collapsed configuration. At the deployment site, the expandable implant **3110** can be moved outside the delivery catheter **3104** and deployed within an aneurysm. During deployment, the third portion **3115** can be deployed first, then the second portion **3120** can be deployed within an interior region defined by the third portion **3115**, and then the first portion **3130** can be deployed within an interior region defined by the second portion **3120**. FIG. **46** illustrates the expandable implant **3110** with the third portion **3115** and the second portion **3120** deployed and the first portion **3130** still within the catheter **3104**. In some embodiments, the insertion portion **3102** can be coupled to the second portion **3120**, such that during detachment of the insertion portion **3102** (e.g., after the expandable implant **3110** has been deployed within an aneurysm), the detachment can occur inside the second portion to avoid any part of the implant from extending or hanging within the blood vessel **V**.

FIG. **47** is a schematic illustration of another embodiment of an insertion device that can be used to insert and deploy an implant, such as an expandable implant as described herein. An insertion device **3254** can be used in conjunction with a cannula or catheter, and can be releasably or removably coupled to an implant, as described for previous embodiments.

The insertion device **3254** includes a first elongate member **3256** defining a lumen **3257** through which a second elongate member **3258** can be movably disposed. The first elongate member **3256** includes an inner marker band **3265** coupled to a distal end portion of the first elongate member **3256**. In this embodiment, a distal end portion **3267** of the

36

second elongate member **3258** is tapered as shown in FIG. **47**. The insertion device **3254** also includes a handle **3255** disposed at a proximal end portion of the insertion device **3254**.

The insertion device **3254** can be coupled to an expandable implant **3210** similar to, or the same as, the expandable implants described herein. The expandable implant **3210** includes a marker band **3242** and a connector member **3252** coupled to the marker band **3242**. The connector member **3252** can be similar to or the same as, for example, the connector member **2852** described above. For example, the connector member **3252** includes a wire **3268** coupled to the marker band **3242** and an implant ball member **3270** coupled to (or formed monolithically or integrally with) the wire **3268**.

To insert and deploy the expandable implant **3210** within a patient's body, the expandable implant **3210** is first coupled to the insertion device **3254**. Specifically, in this embodiment, the second elongate member **3258** is moved proximally (in a direction of arrow **A** in FIG. **47**) such that the tapered distal end portion **3267** is moved proximally within the lumen **3257**. This allows the implant ball member **3270** to be inserted into the lumen **3257** of the first elongate member **3256**. The second elongate member **3258** is then moved distally (in the direction of arrow **B** in FIG. **47**) such that the tapered distal end portion **3267** of the second elongate member **3256** engages the implant ball member **3270** and traps or wedges the implant ball member **3270** within the lumen **3257** of the first elongate member **3256** between the tapered distal end portion **3267** and the inner marker band **3265**.

With the implant ball member **3270** locked or wedged within the lumen **3257** of the first elongate member **3256**, the expandable implant **3210** will be held to the insertion device **3254**. As described above for previous embodiments, a locking mechanism (not shown) coupled to the handle **3255** can be used to lock the second elongate member **3258** in this position relative to the first elongate member **3256**. With the insertion device **3254** coupled to the expandable implant **3210**, the expandable implant **3210** can be inserted into, for example, an insertion cannula (not shown) (e.g., cannula **102** described above) to move the expandable implant **3210** to a collapsed configuration, and the insertion cannula can be used to insert the implant into a blood vessel in a similar manner as described above with respect to previous embodiments.

After the expandable implant **3210** has been deployed within, for example, an aneurysm, the insertion device **3254** can be detached from the expandable implant **3210** and removed from the patient's body. Specifically, to detach the insertion device **3254** from the expandable implant **3210**, in this embodiment, the second elongate member **3258** is unlocked from the handle **3255** and moved proximally (in the direction of arrow **A**) such that the tapered distal end portion **3267** is moved proximally and disengages the implant ball member **3270**. With the tapered distal end portion **3267** moved proximally, the implant ball member **3260** will be free to move outside of the lumen **3257** of the first elongate member **3256**. The insertion device **3254** can then be removed by pulling the insertion device **3254** proximally.

FIG. **48** is a schematic illustration of another embodiment of an insertion device that can be used to insert and deploy an implant, such as an expandable implant as described herein. An insertion device **3354** can be used in conjunction

with a cannula or catheter, and can be releasably or removably coupled to an implant, as described for previous embodiments.

The insertion device **3354** includes a first elongate member **3356** defining a lumen **3357** through which a second elongate member **3358** can be movably disposed. The first elongate member **3356** includes a tapered distal end portion **3392** as shown in FIG. **48**. In alternative embodiments, the first elongate member **3356** can have a constant diameter as with previous embodiments. The first elongate member **3356** also includes an outer marker band **3364** coupled to the tapered distal end portion **3392**. An insertion ball member **3360** is disposed at a distal end of the second elongate member **3358** as shown in FIG. **48**. The insertion device **3354** can also include a handle (not shown) disposed at a proximal end portion of the insertion device **3354** as described above for previous embodiments.

The insertion device **3354** can be coupled to an expandable implant **3310** similar to, or the same as, the expandable implants described herein. The expandable implant **3310** includes a marker band **3342** at a proximal end portion of the expandable implant **3310**, and a connector member **3352** coupled to the marker band **3342**. The connector member **3352** can be similar to, or the same as, for example, the connector member **2852** described above. For example, the connector member **3352** includes a wire **3368** coupled to the marker band **3342** and an implant ball member **3370** coupled to (or formed monolithically or integrally with) the wire **3368**. In this embodiment, as shown in FIG. **48**, the insertion ball member **3360** is larger than the implant ball member **3370** and defines a slot **3371** on a side portion thereof through which the wire **3368** of the connector member **3352** can be disposed when the implant **3310** is coupled to the insertion device **3354**.

To insert and deploy the expandable implant **3310** within a patient's body, the expandable implant **3310** is first coupled to the insertion device **3354**. Specifically, in this embodiment, the second elongate member **3358** is moved distally (in a direction of arrow B in FIG. **48**) such that insertion ball member **3360** is moved distally outside of the lumen **3357** of the first elongate member **3356**. The implant ball member **3370** can be inserted into the lumen **3357** of the first elongate member **3356** and the wire **3368** can be placed or disposed within the slot **3371** of the insertion ball member **3360**. The second elongate member **3358** is then moved proximally (in the direction of arrow A in FIG. **48**) such that the insertion ball member **3360** and the implant ball member **3370** are moved into the lumen **3357** of the first elongate member **3356** and the insertion ball member **3360** locks or traps the implant ball member **3370** within the lumen **3357** of the first elongate member **3356** as shown in FIG. **48**.

With the implant ball member **3370** trapped within the lumen **3357** of the first elongate member **3356**, the expandable implant **3310** will be coupled to the insertion device **3354**. As described above for previous embodiments, a locking mechanism (not shown) coupled to the handle (not shown) can be used to lock the second elongate member **3358** in this position relative to the first elongate member **3356**. With the insertion device **3354** coupled to the expandable implant **3310**, the expandable implant **3310** can be inserted into, for example, an insertion cannula (not shown) to move the expandable implant **3310** to a collapsed configuration. The insertion cannula can be used to insert the implant **3310** into a blood vessel in a similar manner as described above with respect to previous embodiments.

After the expandable implant **3310** has been deployed within, for example, an aneurysm, the insertion device **3354**

can be detached from the expandable implant **3310** and removed from the patient's body. Specifically, to detach the insertion device **3354** from the expandable implant **3310**, in this embodiment, the second elongate member **3358** is unlocked from the handle **3355** and moved distally (in the direction of arrow B in FIG. **48**) such that the insertion ball member **3360** is moved distally allowing the implant ball member **3370** to be free to be moved outside of the lumen **3357** of the first elongate member **3356**. The insertion device **3354** can then be removed by pulling the insertion device **3354** proximally.

FIG. **49** is a schematic illustration of another embodiment of an insertion device that can be used to insert and deploy an implant, such as an expandable implant as described herein. An insertion device **3454** can be used in conjunction with a cannula or catheter, and can be releasably or removably coupled to an implant, as described for previous embodiments.

The insertion device **3454** includes a first elongate member **3456** defining a lumen **3457** through which a second elongate member **3458** can be movably disposed. The first elongate member **3458** includes a tapered distal end portion **3492** as shown in FIG. **49**, but can in alternative embodiments have a constant diameter. The first elongate member **3456** also includes an outer marker band **3464** coupled to the tapered distal end portion **3492**. A plunger or bumper member **3494** is disposed at a distal end of the second elongate member **3458**, as shown in FIG. **49**. The insertion device **3454** can also include a handle (not shown) disposed at a proximal end portion of the insertion device **3454** as described above for previous embodiments.

The insertion device **3454** can be coupled to an expandable implant **3410** similar to, or the same as, the expandable implants described herein. The expandable implant **3410** includes a marker band **3442** at a proximal end portion of the expandable implant **3410**, and a connector member **3452** coupled to the marker band **3442**. The connector member **3452** can be similar to, or the same as, for example, the connector member **2852** described above and includes a wire **3468** coupled to the marker band **3442** and an implant ball member **3470** coupled to (or formed monolithically or integrally with) the wire **3468**.

To insert and deploy the expandable implant **3410** within a patient's body, the expandable implant **3410** is first coupled to the insertion device **3454**. Specifically, in this embodiment, the second elongate member **3458** is moved distally (in a direction of arrow B in FIG. **49**) such that insertion plunger member **3494** is moved distally outside of the lumen **3457** of the first elongate member **3456**. The implant ball member **3470** can then be inserted into the lumen **3457** of the first elongate member **3456**. The second elongate member **3458** is then moved distally (in the direction of arrow B in FIG. **49**) such that the plunger member **3494** locks or traps the insertion ball member **3470** within the lumen **3457** of the first elongate member **3456** as shown in FIG. **49**.

With the implant ball member **3470** trapped within the lumen **3457** of the first elongate member **3456**, a locking mechanism (not shown) coupled to the handle (not shown) can be used to lock the second elongate member **3458** in this position relative to the first elongate member **3456**. With the insertion device **3454** coupled to the expandable implant **3410**, the expandable implant **3410** can be inserted into, for example, an insertion cannula (not shown) to move the expandable implant **3410** to a collapsed configuration. The insertion cannula can be used to insert the implant **3410** into

a blood vessel in a similar manner as described above with respect to previous embodiments.

After the expandable implant **3410** has been deployed within, for example, an aneurysm, the insertion device **3454** can be detached from the expandable implant **3410** and removed from the patient's body. Specifically, to detach the insertion device **3454** from the expandable implant **3410**, in this embodiment, the second elongate member **3458** is unlocked from the handle and moved distally (in the direction of arrow B in FIG. 49) such that the plunger member **3494** is moved distally allowing the implant ball member **3470** to be free to be moved outside of the lumen **3457** of the first elongate member **3456**. The insertion device **3454** can then be removed by pulling the insertion device **3454** proximally.

FIG. 50 is a schematic illustration of another embodiment of an insertion device that can be used to insert and deploy an implant, such as an expandable implant as described herein. An insertion device **3554** can be used in conjunction with a cannula or catheter, and can be releasably or removably coupled to an implant, as described for previous embodiments.

The insertion device **3554** includes a first elongate member **3556** defining a lumen **3557** through which a second elongate member **3558** can be movably disposed. The first elongate member **3558** includes a tapered distal end portion **3592** as shown in FIG. 50, but can in alternative embodiments, have a constant diameter. The first elongate member **3556** also includes an outer marker band **3564** coupled to the tapered distal end portion **3592**. An insertion ball member **3560** is disposed at a distal end of the second elongate member **3558**, as shown in FIG. 50. In this embodiment, the insertion device **3554** also includes an elongate locking member **3596**. The locking member **3596** can have a constant diameter or outer perimeter along its length or can be tapered. For example a distal end portion of the locking member **3596** can have a smaller diameter than a proximal end portion of the locking member **3596**. The locking member **3596** is used in conjunction with the insertion ball member **3560** to lock the implant ball member **3570** to the insertion device **3554** as described in more detail below. The insertion device **3554** can also include a handle (not shown) disposed at a proximal end portion of the insertion device **3554** as described above for previous embodiments.

As with previous embodiments, the insertion device **3554** can be coupled to an expandable implant **3510** similar to, or the same as, the expandable implants described herein. The expandable implant **3510** includes a marker band **3542** at a proximal end portion, and a connector member **3552** coupled to the marker band **3542**. The connector member **3552** can be similar to, or the same as, for example, the connector members described above and includes a wire **3568** coupled to the marker band **3542** and an implant ball member **3570** coupled to (or formed monolithically or integrally with) the wire **3568**.

To insert and deploy the expandable implant **3510** within a patient's body, the expandable implant **3510** is coupled to the insertion device **3554**. Specifically, in this embodiment, the locking member **3596** is moved proximally (in the direction of arrow A in FIG. 50) such that a distal end portion of the locking member **3596** is disposed proximally of the insertion ball member **3560**. This allows the implant ball member **3570** to be inserted into the lumen **3557** of the first elongate member **3556**. In other words, the insertion ball member **3560** and the implant ball member **3570** can each be sized (e.g., each can have a diameter) such that when the locking member **3596** is moved proximally, disengaging the

insertion ball member **3560**, the implant ball member **3570** can be moved in and out of the lumen **3557** while the implant ball member **3570** is disposed within the lumen **3557**. After the implant ball member **3570** is placed within the lumen **3557** of the first elongate member **3556**, the locking member **3596** can be moved distally (in a direction of arrow B in FIG. 50) such that the distal end portion of the locking member **3596** is wedged between an inner wall of the first elongate member **3556** and the insertion ball member **3560**. With the locking member **3596** in this position, the implant ball member **3570** will be held or trapped within the lumen **3557** of the first elongate member **3556** as shown in FIG. 50.

With the implant ball member **3570** trapped within the lumen **3557** of the first elongate member **3556**, a locking mechanism (not shown) coupled to the handle (not shown) can be used to lock the locking member **3596** in position relative to the first elongate member **3556**. With the insertion device **3554** coupled to the expandable implant **3510**, the expandable implant **3510** can be inserted into, for example, an insertion cannula (not shown) to move the expandable implant **3510** to a collapsed configuration. The insertion cannula can be used to insert the implant **3510** into a blood vessel in a similar manner as described above with respect to previous embodiments.

After the expandable implant **3510** has been deployed within, for example, an aneurysm, the insertion device **3554** can be detached from the expandable implant **3510** and removed from the patient's body. Specifically, to detach the insertion device **3554** from the expandable implant **3510**, in this embodiment, the locking member **3596** is unlocked from the handle and moved proximally (in the direction of arrow A in FIG. 50) such that the distal end portion of the locking member **3596** is moved proximally away from the insertion ball member **3560** allowing the implant ball member **3570** to be free to be moved outside of the lumen **3557** of the first elongate member **3556**. The insertion device **3554** can then be removed by pulling the insertion device **3554** proximally.

FIG. 51 is a schematic illustration of another embodiment of an insertion device that can be used to insert and deploy an implant, such as an expandable implant as described herein. An insertion device **3654** can be used in conjunction with a cannula or catheter, and can be releasably or removably coupled to an expandable implant, as described for previous embodiments.

The insertion device **3654** includes a first elongate member **3656** defining a lumen **3657** through which a second elongate member **3658** can be movably disposed. The first elongate member **3656** also includes an extension member **3672** and a stopper **3674** disposed within the lumen **3657**. The extension member **3672** defines an opening or window **3673**. An insertion ball member **3660** is disposed at a distal end of the second elongate member **3658** and a bumper member **3694** disposed at a spaced distance proximally of the insertion ball member **3660**, as shown in FIG. 51. The insertion device **3654** can also include a handle **3655** disposed at a proximal end portion of the insertion device **3654** as described above for previous embodiments.

As with previous embodiments, the insertion device **3654** can be coupled to an expandable implant **3610** similar to, or the same as, the expandable implants described herein. In this embodiment, the expandable implant **3610** includes a marker band **3642** at a proximal end portion, and a connector member **3652** coupled to the marker band **3642**. The connector member **3652** can be similar to, or the same as, for example, the connector members described above and includes a wire **3668** coupled to the marker band **3642** and

an implant ball member **3670** coupled to (or formed monolithically or integrally with) the wire **3668**.

The expandable implant **3610** can also include a lead-in portion **3676** (also referred to herein as “lead-in member”) disposed at a distal end portion of the expandable implant **3610**. The lead-in portion **3676** can be formed with, for example, a shape memory material such as nitinol, such that the lead-in portion **3676** has a biased curved shape when not constrained within, for example a cannula. The curved shape of the lead-in portion **3676** can reduce or eliminate possible sharp edges when inserting the expandable implant **3610** within a vasculature of a patient. The lead-in portion **3676** can be a separate component coupled to the expandable implant **3610** or can be formed integrally or monolithically with the expandable implant **3610**. In some embodiments, the lead-in portion **3676** can be crimped to the distal end portion of the expandable implant **3610**. In some embodiments, the lead-in portion **3676** can be formed integrally or monolithically with a wire member or radiopaque wire (as described for example with respect to FIGS. **23** and **24**) that extends through the expandable implant **3610**. For example, such a wire member can extend beyond the distal end portion of the expandable implant **3610** and form the lead-in portion **3676**.

To insert and deploy the expandable implant **3610** within a patient’s body, the expandable implant **3610** is coupled to the insertion device **3654**. Specifically, in this embodiment, the second elongate member **3658** is moved proximally (in a direction of arrow A in FIG. **51**) such that the insertion ball member **3660** is disposed proximally of the window **3673** defined in the extension member **3672**. This allows the implant ball member **3670** to be inserted through an opening (not shown) defined at a distal end of the first elongate member **3656** and into the lumen **3657** of the first elongate member **3656**. For example, the insertion ball member **3660** and the implant ball member **3670** can each be sized (e.g., each can have a diameter) such that collectively the insertion ball member **3660** and the implant ball member **3670** have a size (e.g., a diameter) greater than a diameter of the lumen **3657**. Thus, the insertion ball member **3660** is moved to a position to provide clearance or space for the implant ball member **3670** to be inserted into the lumen **3657** and disposed near or adjacent to the window **3673**. The second elongate member **3658** can then be moved distally (in a direction of arrow B in FIG. **51**) such that the insertion ball member **3660** is moved distally and contacts the implant ball member **3670** and moves or pushes the implant ball member **3670** at least partially through the window **3673**. The second elongate member **3658** is moved distally until the insertion ball member **3660** is moved to a position distal of the implant ball member **3670**, allowing the implant ball member **3670** to move back into the lumen **3657**. Further, as the second elongate member **3658** is moved distally, the bumper member **3694** on the second elongate member **3658** can contact the stopper **3674** to limit the movement of the second elongate member **3658** in the distal direction. With the insertion ball member **3660** and the implant ball member **3670** interlocked within the lumen **3657** and the implant ball member **3670** now positioned proximal of the insertion ball member **3660**, the implant **3610** is maintained coupled to the insertion device **3654**.

With the implant ball member **3670** held or trapped within the lumen **3657** of the first elongate member **3656**, a locking mechanism (not shown) coupled to the handle **3655** can be used to lock the second elongate member **3658** in this position relative to the first elongate member **3656**. With the insertion device **3654** coupled to the expandable implant

3610, the expandable implant **3610** can be inserted into the lumen of an insertion cannula **3604** to move the expandable implant **3610** to a collapsed configuration. The insertion cannula **3604** can be used to insert the implant **3610** into a blood vessel in a similar manner as described above with respect to previous embodiments.

After the expandable implant **3610** has been deployed within, for example, an aneurysm, the insertion device **3654** can be detached from the expandable implant **3610** and removed from the patient’s body. Specifically, to detach the insertion device **3654** from the expandable implant **3610**, the second elongate member **3558** is moved proximally such that the insertion ball member **3660** contacts the implant ball member **3670** and moves the implant ball member **3670** at least partially within the window **3673**. The second elongate member **3658** is moved proximally until the insertion ball member **3660** is disposed proximal of the window **3673** such that the implant ball member **3670** can move back within the lumen **3657** of the first elongate member **3656**. The stopper **3674** can limit the movement of the second elongate member **3658** by engaging the insertion ball member **3660**. With the implant ball member **3670** disposed distal of the insertion ball member **3660**, the implant **3610** can be released from the insertion device **3654**. The insertion device **3654** can then be removed by pulling the insertion device **3654** proximally.

FIG. **52** is a schematic illustration of another embodiment of an insertion device that can be used to insert and deploy an implant, such as an expandable implant as described herein. An insertion device **3754** can be used in conjunction with a cannula or catheter, and can be releasably or removably coupled to an implant, as described for previous embodiments.

The insertion device **3754** includes a first elongate member **3756** defining a lumen **3757** through which a second elongate member **3758** can be movably disposed. The first elongate member **3756** includes an inner stopper **3774** coupled to a distal end portion of the first elongate member **3756** within the lumen **3757**. The inner stopper **3774** defines a channel **3775** that can be used to trap or hold an implant **3710** to the insertion device **3654** as described in more detail below.

The second elongate member **3758** includes a distal end portion **3767** that can be smaller in size (e.g., diameter) than a remaining portion **3777** of the second elongate member **3758**. In some embodiments, the distal end portion **3767** can be tapered. In some embodiments, the distal end portion **3767** can be a separate component coupled to the remaining portion **3777** of the second elongate member **3758**. The second elongate member **3758** can be formed for example, with a shape-memory material and define a bend or bends along its length. The insertion device **3754** also includes a handle **3755** disposed at a proximal end portion of the insertion device **3754**.

The insertion device **3754** can be coupled to an expandable implant **3710** similar to, or the same as, the expandable implants described herein. The expandable implant **3710** includes a marker band **3742** and a connector member **3752** coupled to the marker band **3742**. The connector member **3752** can be similar to or the same as, for example, the connector members described above for previous embodiments. For example, the connector member **3752** includes a wire **3768** coupled to the marker band **3742** and an implant ball member **3770** coupled to (or formed monolithically or integrally with) the wire **3768**.

To insert and deploy the expandable implant **3710** within a patient’s body, the expandable implant **3710** is first

coupled to the insertion device **3754**. Specifically, in this embodiment, the second elongate member **3758** is moved proximally (in a direction of arrow A in FIG. **52**) such that the distal end portion **3767** is moved proximally within the lumen **3757**. This allows the implant ball member **3770** to be inserted into the lumen **3757** of the first elongate member **3756**. The second elongate member **3758** is then moved distally (in the direction of arrow B in FIG. **52**) such that the distal end portion **3767** of the second elongate member **3756** engages and urges the implant ball member **3770** at least partially within the channel **3775**, trapping or wedging the implant ball member **3770** within the lumen **3757** of the first elongate member **3756** between the tapered distal end portion **3767** and the inner stopper **3774**.

With the implant ball member **3770** wedged or trapped within the lumen **3757** of the first elongate member **3756**, the expandable implant **3710** will be held to the insertion device **3754**. As described above for previous embodiments, a locking mechanism (not shown) coupled to the handle **3755** can be used to lock the second elongate member **3758** in this position relative to the first elongate member **3756**. With the insertion device **3754** coupled to the expandable implant **3710**, the expandable implant **3710** can be inserted into, for example, an insertion cannula (not shown) to move the expandable implant **3710** to a collapsed configuration, and the insertion cannula can be used to insert the implant **3710** into a blood vessel in a similar manner as described above with respect to previous embodiments.

After the expandable implant **3710** has been deployed within, for example, an aneurysm, the insertion device **3754** can be detached from the expandable implant **3710** and removed from the patient's body. Specifically, to detach the insertion device **3754** from the expandable implant **3710**, the second elongate member **3758** is unlocked from the handle **3755** and moved proximally (in the direction of arrow A) such that the tapered distal end portion **3767** is moved proximally and disengages the implant ball member **3770**. With the tapered distal end portion **3767** moved proximally, the implant ball member **3770** will be free to move outside of the lumen **3757** of the first elongate member **3756**. The insertion device **3754** can then be removed by pulling the insertion device **3754** proximally.

FIG. **53** is a schematic illustration of another embodiment of an insertion device that can be used to insert and deploy an implant, such as an expandable implant as described herein. An insertion device **3854** can be used in conjunction with a cannula or catheter, and can be releasably or removably coupled to an implant, as described for previous embodiments.

The insertion device **3854** includes a first elongate member **3856** defining a lumen **3857** through which a second elongate member **3858** can be movably disposed. The insertion device **3854** can also include a handle (not shown) disposed at a proximal end portion of the insertion device **3854** as described above for previous embodiments. The first elongate member **3856** includes an inner stopper **3874** coupled to a distal end portion of the first elongate member **3856**. The inner stopper **3874** can be, for example, an inner marker band as described above for previous embodiments.

The second elongate member **3858** includes a distal end portion **3867** that can engage a portion of an expandable implant as described in more detail below. In some embodiments, the distal end portion **3867** can be tapered. The second elongate member **3858** also includes a bumper member **3894** and a coil member **3876**. In some embodiments, the distal end portion **3867** can be a separate component coupled to the bumper member **3894**. In some

embodiments, the distal end portion **3867** is formed integral or monolithically with a remaining portion **3877** of the second elongate member **3858**. For example, the distal end portion and/or the remaining portion **3877** can extend through a lumen (not shown) of the bumper member **3894** and a lumen (not shown) of the coil member **3876**, and extend to a proximal end of the insertion device **3854**.

The insertion device **3854** can be coupled to an expandable implant **3810** similar to, or the same as, the expandable implants described herein. The expandable implant **3810** includes a marker band **3842** and a connector member **3852** coupled to the marker band **3842**. The connector member **3852** can be similar to or the same as, for example, the connector members described above. For example, the connector member **3852** includes a wire **3868** coupled to the marker band **3842** and an implant ball member **3870** coupled to (or formed monolithically or integrally with) the wire **3868**.

To insert and deploy the expandable implant **3810** within a patient's body, the expandable implant **3810** is first coupled to the insertion device **3854**. Specifically, in this embodiment, the second elongate member **3858** is moved proximally (in a direction of arrow A in FIG. **53**) such that the distal end portion **3867** is moved proximally to a position proximal of the inner stopper **3874**. This allows the implant ball member **3870** to be inserted through an opening in a distal end of the first elongate member **3856** and into the lumen **3857** of the first elongate member **3856**. The second elongate member **3858** is then moved distally (in the direction of arrow B in FIG. **53**) such that the distal end portion **3867** of the second elongate member **3856** engages the implant ball member **3870** and traps or wedges the implant ball member **3870** within the lumen **3857** of the first elongate member **3856** between the distal end portion **3867** and the inner stopper **3874**.

With the implant ball member **3870** locked or wedged within the lumen **3857** of the first elongate member **3856**, the expandable implant **3810** will be held to the insertion device **3854**. As described above for previous embodiments, a locking mechanism (not shown) can be coupled to the handle and can be used to lock the second elongate member **3858** in this position relative to the first elongate member **3856**. With the insertion device **3854** coupled to the expandable implant **3810**, the expandable implant **3810** can be inserted into, for example, an insertion cannula (not shown) (e.g., cannula **102** described above) to move the expandable implant **3810** to a collapsed configuration, and the insertion cannula can be used to insert the implant **3810** into a blood vessel in a similar manner as described above with respect to previous embodiments.

After the expandable implant **3810** has been deployed within, for example, an aneurysm, the insertion device **3854** can be detached from the expandable implant **3810** and removed from the patient's body. Specifically, to detach the insertion device **3854** from the expandable implant **3810**, the second elongate member **3858** is unlocked from the handle and moved proximally (in the direction of arrow A) such that the distal end portion **3867** is moved proximally and disengages the implant ball member **3870**. With the distal end portion **3867** moved proximally, the implant ball member **3870** will be free to move outside of the lumen **3857** of the first elongate member **3856**. The insertion device **3854** can then be removed by pulling the insertion device **3854** proximally.

FIG. **54** is a schematic illustration of another embodiment of an insertion device that can be used to insert and deploy an implant, such as an expandable implant as described

herein. An insertion device **3954** can be used in conjunction with a cannula or catheter, and can be releasably or removably coupled to an implant, as described for previous embodiments.

The insertion device **3954** includes a first elongate member **3956**, a second elongate member **3958** and an outer shrink tube **3938**. The first elongate member **3956** defines a lumen **3957** through which the second elongate member **3958** can be movably disposed and the shrink tube **3938** and the first elongate member **3956** collectively define a lumen **3939** that the second elongate member **3958** can also be movably disposed within.

The first elongate member **3956** defines a skived or cutout portion **3953** that extends between an intermediate portion **3959** of the first elongate member **3956** and a distal end portion **3972** of the first elongate member **3956**. The outer shrink tube **3938** can be coupled to the first elongate member **3956** at least along a portion of the first elongate member **3956** that defines the skived portion **3953**. The skived portion **3953** can reduce the mass of the first elongate member **3956** and allow the first elongate member **3956** to be more flexible along the skived portion. The outer shrink tube **3938** can be, for example, a material that is heat shrunk to the outer surface of the first elongate member **3956** to provide an outer boundary or perimeter of the insertion device **3954** along the skived portion **3953** of the first elongate member **3956**. The outer shrink tube **3938** can be formed with a flexible material such that the portion of the first elongate member **3956** including the skived portion **3953** and the outer shrink tube **3938** is flexible and can be maneuvered through tortuous vasculature.

The distal end portion **3972** of the first elongate member **3956** defines a side window **3973** in fluid communication with a lumen **3937** defined by the distal end portion **3972**. One or more tab members **3951** (three shown in FIG. **54**) are disposed on the first elongate member **3956** at spaced locations along a length of the first elongate member **3956**. The tab members **3951** can be, for example, semi-circular or c-shaped defining an open portion or can be circular or ring shaped forming a closed loop. The tab members **3951** can be separate components coupled to the first elongate member **3956** or formed integrally or monolithically with the first elongate member **3956**. A sleeve member **3949** is coupled to one of the tab members **3951** and/or to the first elongate member **3956**. The sleeve member **3949** can be, for example, welded to the tab member **3951** and/or the first elongate member **3956**. The sleeve member **3949** defines a lumen **3933** through which the second elongate member **3958** can be movably disposed.

As shown in FIG. **54**, the second elongate member **3958** can be movably disposed through the lumen **3957**, the lumen **3939**, a lumen **3933** of the sleeve member **3949** and the lumen **3937** of the distal end portion **3972**. An insertion ball member **3960** is disposed at a distal end of the second elongate member **3958** and a bumper **3994** is coupled to the second elongate member **3958** proximally of the insertion ball member **3960**. In addition, a stopper **3948** is coupled to the second elongate member **3958** at a spaced distance proximal of the bumper **3994** and a radiopaque marker **3961** is coupled to the second elongate member **3958** proximal of the stopper **3948**. The bumper **3994**, the stopper **3948** and the radiopaque marker **3961** can each be, for example, welded to the second elongate member **3958**. The insertion device **3954** can also include a handle (not shown) disposed at a proximal end portion of the insertion device **3954** as described above for previous embodiments.

As with previous embodiments, the insertion device **3954** can be coupled to an expandable implant **3910** similar to, or the same as, the expandable implants described herein. In this embodiment, the expandable implant **3910** includes a marker band **3942** at a proximal end portion, and a connector member **3952** coupled to the marker band **3942**. The connector member **3952** can be similar to, or the same as, for example, the connector members described above and includes a wire **3968** coupled to the marker band **3942** and an implant ball member **3970** coupled to (or formed monolithically or integrally with) the wire **3968**.

In use, to insert and deploy the expandable implant **3910** within a patient's body, the expandable implant **3910** is first coupled to the insertion device **3954**. Specifically, in this embodiment, the second elongate member **3958** is moved proximally (in a direction of arrow A in FIG. **54**) such that the insertion ball member **3960** is disposed proximally of the window **3973** defined by the distal end portion **3972** of the first elongate member **3956**. This allows the implant ball member **3970** to be inserted through an opening **3963** defined at a distal end of the first elongate member **3956** and into the lumen **3937** of the distal end portion **3972** of the first elongate member **3956**. For example, the insertion ball member **3960** and the implant ball member **3970** can each be sized (e.g., each can have a diameter) such that collectively the insertion ball member **3960** and the implant ball member **3970** have a size (e.g., a diameter) greater than a diameter of the lumen **3937**. Thus, the insertion ball member **3960** is moved to a position proximal of the window **3973** to provide clearance or space for the implant ball member **3970** to be inserted into the lumen **3937** and disposed near or adjacent to the window **3973**. With the implant ball member **3970** disposed near the window **3973**, the second elongate member **3958** can then be moved distally (in a direction of arrow B in FIG. **54**) such that the insertion ball member **3960** is moved distally and contacts the implant ball member **3970**, and moves or pushes the implant ball member **3970** at least partially through the window **3973**. The bumper **3994** provides rigidity to the distal portion of the second elongate member **3958** as the insertion ball member **3960** is moved distally. The second elongate member **3958** is moved distally until the insertion ball member **3960** is moved to a position distal of the implant ball member **3970**, allowing the implant ball member **3970** to move back at least partially within the lumen **3937**. Further, as the second elongate member **3958** is moved distally, the stopper **3948** on the second elongate member **3958** can contact the sleeve member **3949** to limit the movement of the second elongate member **3958** in the distal direction. With the insertion ball member **3960** and the implant ball member **3970** interlocked within the lumen **3937** and the implant ball member **3970** now positioned proximal of the insertion ball member **3960**, the implant **3910** is maintained coupled to the insertion device **3954**.

With the implant ball member **3970** held or trapped within the lumen **3937** of the distal end portion **3972**, a locking mechanism (not shown) can be used to lock the second elongate member **3958** in this position relative to the first elongate member **3956**. For example, a locking mechanism can be coupled to a handle (not shown) as described above for previous embodiments. With the insertion device **3954** coupled to the expandable implant **3910**, the expandable implant **3910** can be inserted into the lumen of an insertion cannula (not shown) to move the expandable implant **3910** to a collapsed configuration. The insertion cannula can be

used to insert the implant **3910** into a blood vessel in a similar manner as described above with respect to previous embodiments.

After the expandable implant **3910** has been deployed within, for example, an aneurysm, the insertion device **3954** can be detached from the expandable implant **3910** and removed from the patient's body. Specifically, to detach the insertion device **3954** from the expandable implant **3910**, the second elongate member **3958** is moved proximally (in the direction of arrow A) such that the insertion ball member **3960** contacts the implant ball member **3970** and moves the implant ball member **3970** at least partially through the window **3973**. The second elongate member **3958** is moved proximally until the insertion ball member **3960** is disposed proximal of the window **3973** such that the implant ball member **3970** can move back within the lumen **3937** of the distal end portion **3972**. The sleeve member **3949** can limit the proximal movement of the second elongate member **3958** by engaging the insertion ball member **3960**. For example, the insertion ball member **3960** can have a larger diameter than an inner diameter of the sleeve member **3949**. With the implant ball member **3970** disposed distal of the insertion ball member **3960**, the implant **3910** can be released from the insertion device **3954**. For example, the insertion device **3954** can be removed by pulling the insertion device **3954** proximally and as the insertion device **3954** is moved proximally, the implant ball member **3970** can move through the distal opening **3963** leaving the implant **3910** implanted within the patient's body.

FIGS. **55** and **56** illustrate another embodiment of an insertion device. An insertion device **4054** includes a first elongate member **4056**, a second elongate member **4058** and an outer shrink tube **4038**. The first elongate member **4056** defines a lumen (not shown) through which the second elongate member **4058** can be movably disposed and the shrink tube **4038** and the first elongate member **4056** collectively define a lumen (not shown) that the second elongate member **4058** can also be movably disposed within. The insertion device **4054** can be used in conjunction with a cannula or catheter, and can be releasably or removably coupled to an implant, as described for previous embodiments.

The first elongate member **4056** defines a skived or cutout portion **4053** that extends between an intermediate portion **4059** of the first elongate member **4056** and a distal end portion **4072** of the first elongate member **4056**. The first elongate member **4056** can also include additional skived or cutout portions (not shown). The outer shrink tube **4038** can be coupled to the first elongate member **4056** at least along a portion of the first elongate member **4056** that defines the skived portion **4053**. The skived portion **4053** can reduce the mass of the first elongate member **4056** and allow the first elongate member **4056** to be more flexible along the skived portion. The outer shrink tube **4038** can be, for example, a material that is heat shrunk to the outer surface of the first elongate member **4056** to provide an outer boundary or perimeter of the insertion device **4054** along the skived portion **4053** of the first elongate member **4056**. The outer shrink tube **4038** can be formed with a flexible material such that the portion of the first elongate member **4056** including the skived portion **4053** and the outer shrink tube **4038** is flexible and can be maneuvered through tortuous vasculature.

The distal end portion **4072** of the first elongate member **4056** defines a side window **4073** and a lumen (not shown) in fluid communication with the side window **4073**. One or more tab members **4051** (only one tab member is shown in

FIGS. **55** and **56**) are disposed on the first elongate member **4056** at spaced locations along a length of the first elongate member **4056**. As shown in FIGS. **55** and **56**, in this embodiment, the tab member **4051** is substantially c-shaped defining an open portion. The tab member **4051** can be, for example welded to the first elongate member **4056**.

As shown in FIGS. **55** and **56**, the second elongate member **4058** can be movably disposed through the lumen of the first elongate member **4056**, the lumen defined collectively by the first elongate member **4056** and the outer shrink tube **4038**, and the lumen of the distal end portion **4072**. An insertion ball member (not shown) is disposed at a distal end of the second elongate member **4058** and a bumper **4094** is coupled to the second elongate member **4058** proximally of the insertion ball member. As shown in FIGS. **55** and **56**, in this embodiment, the bumper **4094** includes a spring. In addition, a stopper (not shown) and a radiopaque marker (not shown) can be coupled to the second elongate member **4058** at a spaced distance proximal of the bumper **4094** as with the previous embodiments, and can provide the same function as described above for insertion device **3954**. The insertion device **4054** can also include a handle (not shown) disposed at a proximal end portion of the insertion device **4054** as described above for previous embodiments. The bumper **4094**, the stopper and the radiopaque marker can each be, for example, welded to the second elongate member **4058**.

As with previous embodiments, the insertion device **4054** can be coupled to an expandable implant **4010** similar to, or the same as, the expandable implants described herein. The expandable implant **4010** includes a marker band **4042** at a proximal end portion, and a connector member **4052** coupled to the marker band **4042**. The connector member **4052** includes a wire **4068** coupled to the marker band **4042** and an implant ball member **4070** (see FIG. **56**) coupled to (or formed monolithically or integrally with) the wire **4068**. FIG. **56** illustrates the implant ball member **4070** inserted into the distal end portion **4072** of the first elongate member **4056** and disposed near the window **4073**.

In use, the insertion device **4054** can function the same as or similar to the insertion device **3954** described above. For example, the implant **4010** can be coupled to the insertion device **4054** and locked in position by the insertion ball member in the same or similar manner as described above for insertion device **3954**. Likewise, the implant **4010** can be released from the insertion device **4054** in the same or similar manner as described above for insertion device **3954**.

FIG. **57** illustrates another embodiment of a medical device that includes a lead-in portion or member disposed at a distal end portion of an expandable implant. The medical device **4100** includes an expandable implant **4110** that can be configured the same as or similar to any of the embodiments of an expandable implant described herein. For example, the expandable implant **4110** can be deployed within an aneurysm of a patient as described herein. As shown in FIG. **57**, a lead-in member **4176** is coupled to a distal end portion of the expandable implant **4110**. In this embodiment, the lead-in member **4176** is coupled to the distal end portion of the expandable implant **4110** with a crimp **4178**. The lead-in member **4176** can be formed with, for example a shape memory material such, as nitinol, such that the lead-in member **4176** has a biased curved shape when not constrained within, for example a cannula **4105**. Thus, the lead-in member **4176** can provide a smooth surface free of sharp edges when inserting the expandable implant **4110** within a vasculature of a patient. The lead-in member **4176** can have a substantially linear configuration

when constrained within the cannula **4105**, or a delivery device as described herein. Although not shown in FIG. **57**, the medical device **4100** can be delivered within a vasculature of a patient using a delivery device as described herein for other embodiments.

FIGS. **58-60** illustrate a portion of a medical device **4200** according to an embodiment. The medical device **4200** can include the same or similar features and functions as described herein for other embodiments. For example, the medical device **4200** can include an expandable implant **4210** configured to move from the collapsed configuration (e.g., for delivery through a blood vessel) to the expanded configuration (e.g., for deployment within an aneurysm) and an insertion member or device **4254** (shown in FIG. **59**) as described herein.

Similar to the expandable implant **1810**, the expandable implant **4210** includes a ribbon-like strand of porous mesh that includes one or more petal-like portions or sections **4225** along its length. In this embodiment, there are four petal-like portions **4225** included within an outer petal segment **4291** of the expandable implant **4210** and three petal-like portions **4225** included within an inner petal segment **4290** of the expandable implant **4210**.

At least a portion of the porous mesh can be configured to be positioned over a neck of an aneurysm when the expandable implant **4210** is in the expanded configuration. When the expandable implant **4210** is in its expanded configuration, the expandable implant **4210** has a three-dimensional shape (e.g., a substantially spherical shape) with a substantially continuous outer surface such that a portion (e.g., edges) of at least two of the petal-like portions **4225** overlap each other as shown in FIG. **59**. For example, as the expandable implant **4210** is being deployed within an aneurysm, the petal-like portions **4225** of the outer petal segment **4191** expands first and forms an outer layer that covers the aneurysm. The petal-like portions **4225** of the inner petal segment **4290** then form a second spherical layer of material inside the petal-like portions **4225** of the outer petal portion **4291** to provide greater surface area to further promote thrombosis.

In this embodiment, a suture strand **4235** extends along the length of the expandable implant **4210** to provide reinforcement to the expandable implant **4210** and can also provide for a radiopaque coil to be disposed over at least a portion of the suture strand **4235** to provide visibility of the expandable implant **4210** during, for example, fluoroscopy. As shown in FIGS. **58** and **59**, the suture strand **4235** is disposed along a length of the expandable implant **4210** and across or within the petal-like portions **4225**. The suture strand **4235** can be coupled to, for example, marker bands **4242** and **4244** disposed on a proximal end and a distal end, respectively, of the expandable implant **4210**.

In this embodiment, the outer petal segment **4291** and the inner petal segment **4290** can be formed as separate components and coupled together by the suture strand **4235**. This creates an articulation point or joint **4279** between the outer petal segment **4291** and the inner petal segment **4290**. For example, the inner petal segment **4290** can include the marker band **4242** at a proximal end and a marker band **4294** at a distal end. The outer petal segment **4291** can include the marker band **4244** at a distal end and a marker band **4295** at a proximal end. The articulation joint **4279** is defined where the marker band **4294** and the marker band **4295** are coupled to the suture strand **4235**.

The articulation joint **4279** can provides greater freedom of motion of the petal-like portions **4225**, which can allow more uniform expansion of the petal-like portions **4225**. In

4290, the separate construction of the outer petal segment **4291** and the inner petal segment **4290** can allow for one spherical layer of the expandable implant to be formed at a time, which may be advantageous and/or easier to manufacture. The ability to manufacture the expandable implant **4210** in multiple segments can also allow for the addition to, or removal of, segments of an expandable implant to provide a selected length or size of the expandable implant to meet a particular need.

As shown in FIG. **58**, the expandable implant **4210** can also include a lead-in member **4276** coupled to a distal end portion of the expandable implant **4210** with the marker band **4244**. The lead-in member **4276** can be formed with, for example a shape memory material such, as nitinol, such that the lead-in member **4276** has a biased curved shape when not constrained within, for example a cannula (not shown) as described above for expandable implant **4110**. In some embodiments, the lead-in member **4276** can be coupled to the distal end portion of the expandable implant **4210** with a crimp similar to the implant **4110**. Although not shown, the expandable implant **4210** can also include a coupling member to releasably couple the expandable implant **4210** to the delivery device **4254** as described above for previous embodiments.

FIG. **60** illustrates another embodiment of a medical device **4300** that includes an expandable implant **4310** that has multiple articulation joints **4379**. The medical device **4300** can include the same or similar features and functions as described herein for other embodiments. For example, the medical device **4300** can be configured to move from a collapsed configuration a shown in FIG. **60** (e.g., for delivery through a blood vessel) to an expanded configuration (not shown) (e.g., for deployment within an aneurysm). The medical device **4300** can also include an insertion member or device (not shown in FIG. **60**) to which the expandable implant **4310** can be releasably coupled, as described above for previous embodiments.

The expandable implant **4310** includes a ribbon-like strand of porous mesh that includes one or more petal-like portions or sections **4325** along its length. In this embodiment, there are three petal-like portions **4325** included within a first petal segment **4392** of the expandable implant **4310**, four petal-like portions **4325** included within a second petal segment **4391**, and three petal-like portions **4325** included within a third petal segment **4390** of the expandable implant **4310**.

As with the previous embodiment, at least a portion of the porous mesh can be configured to be positioned over a neck of an aneurysm when the expandable implant **4310** is in the expanded configuration. When the expandable implant **4310** is in its expanded configuration, the expandable implant **4310** can have a three-dimensional shape (e.g., a substantially spherical shape) with a substantially continuous outer surface as described above for previous embodiments.

A suture strand **4335** extends along the length of the expandable implant **4310** to provide reinforcement to the expandable implant **4310** and can also provide for a radiopaque coil to be disposed over at least a portion of the suture strand **4335** to provide visibility of the expandable implant **4310** during, for example, fluoroscopy. The suture strand **4335** can be coupled to, for example, marker bands **4342** and **4344** disposed on a proximal end and a distal end, respectively, of the expandable implant **4310**.

As shown in FIG. **60**, the expandable implant **4310** can also include a lead-in member **4376** coupled to a distal end portion of the expandable implant **4310** with the marker band **4344**. The lead-in member **4376** can be formed the

same as or similar to the lead-in members described above. Although not shown, the expandable implant **4310** can also include a coupling member to releasably couple the expandable implant **4310** to a delivery device as described above for previous embodiments.

In this embodiment, the first petal segment **4392**, the second petal segment **4391** and the third petal segment **4390** can be formed as separate components and coupled together by the suture strand **4335**. This creates a first articulation point or joint **4379** between the first petal segment **4392** and the second petal segment **4391**, and a second articulation point or joint **4379'** between the second petal segment **4391** and the third petal segment **4390**. In this embodiment, the first petal segment **4392** includes the marker band **4344** on a distal end and a marker band **4397** on a proximal end, the second petal segment **4391** includes a marker band **4396** on a distal end and a marker band **4395** on a proximal end, and the third petal segment **4390** includes the marker band **4342** at a proximal end and a marker band **4394** at a distal end. The first articulation joint **4379** is defined where the marker band **4397** and the marker band **4396** are coupled to the suture strand **4335**, and the second articulation joint **4379'** is defined where the marker band **4395** and the marker band **4394** are couple to the suture strand **4335**.

As discussed above for expandable implant **4210**, the articulation joints **4379**, **4379'** can provide greater freedom of motion of the petal-like portions **4325** of the expandable implant **4310**, which can allow more uniform expansion of the petal-like portions **4325** within an aneurysm. In addition, with three petal segments **4392**, **4391**, **4390**, the expandable implant **4310** can have a greater density when deployed within an aneurysm which can further enhance thrombosis.

In alternative embodiments, an expandable implant can have a different number of articulation joints and a different number of petal segments than described above for expandable implants **4210** and **4310**. In some embodiments, it may be desirable to have at least two petal-like portions (e.g., **4225**, **4325**) between the articulation joints. In other words it may be desirable for each petal segment to have at least two petal-like portions. A greater number of articulation points or joints can provide increased freedom of motion of the petal-like portions, which can lead to a more uniform expansion of the expandable implant. The petal segments or layers can also have variable stiffness. For example, in an expandable implant, such as, expandable implant **4310**, it may be desirable for the first petal segment to have a greater stiffness such that the first petal segment (e.g., petal segment **4392**) can frame the aneurysm as the expandable implant is being deployed within the aneurysm. In this example it may be desirable for the second petal segment (e.g., petal layer **4391**) to have a medium stiffness (e.g., stiffness less than the first petal segment and greater than the third petal segment) to fill the aneurysm, and the third petal segment (e.g., petal segment **4390**) to be the softest segment to pack the aneurysm.

The petal width can also be varied between segments. For example, it may be desirable for the distal segment (e.g., first petal segment **4392**) to have a greater width than the remaining segments and the proximal petal segments (e.g., the second petal segment **4391** and/or the third petal segment **4390**) to be shorter and narrower to fit inside the distal segment (e.g., the first petal segment **4392**).

The insertion devices (e.g., **2554**, **2654**, **2754**, **2854**, **3254**, **3354**, **3454**, **3554**, **3654**, **3754**, **3854**, **3954**, **4054**) described herein can be used to deliver an expandable implant as described herein. For example, any of the expandable implants described herein can include an outer marker band

and an inner marker band coupled to a proximal end portion of the expandable implant that can be used to couple the expandable implant to an insertion device, such as, for example, the insertion devices **2554**, **2654** and **2754**. In addition, any of the expandable implants described herein can include a connector member (e.g., **1652**, **1952**, **2452**, **2852**, **3252**, **3352**, **3452**, **3552**, **3652**, **3752**, **3852**, **3952**, **4052**) as described above, including a wire and ball member configured to be coupled to an insertion device, such as, for example, insertion devices **2854**, **3254**, **3354**, **3454**, **3554**, **3654**, **3754**, **3854**, **3954** and **4054**. Further, although the ball members (insertion or implant ball members) are shown as circular, any of the ball members described herein can be other shapes, such as, for example, oval, elliptical, square, rectangular, triangular or other desired shape (as shown in a side view).

The various devices described herein can be made of any material suitable for the defined purpose, including, for example, drawn filled tube DFT®. DFT is available as wire, cable or ribbon. DFT is a metal-to-metal composite developed to combine the desired physical and mechanical attributes of two or more materials into a single wire or ribbon system, which can be used for the expandable implant.

Filaments or wires for the braid or mesh (e.g., the expandable implants) can include, for example, filaments of materials such as MP35N, stainless steel, nitinol, cobalt chromium, titanium, platinum, tantalum, tungsten, or alloys thereof, or polyester, polyethylene (PET), Dacron, PEEK, vectron, and suture materials. Each strand may have a diameter between 0.0005"-0.010", e.g., about 0.002". In some embodiments, an outer material of the mesh or braid can be formed with nitinol that is superelastic at body temperature, and an inner material can be radiopaque, or alternatively platinum wires may be included in the braid to provide additional radiopacity. For example, in some embodiments, an expandable implant can include radiopaque material(s) woven within the mesh material such that the expandable implant can be highly visible without the use of a radioactive die.

Suitable materials can be chosen based on their electropositivity. For example, an expandable implant can include titanium, tungsten, or another material listed below in Table 1, or any combination thereof. In use, the electropositive material of the expanded expandable implant creates an electrically favorable region within the vascular defect and through the blood, and the region in the defect containing blood, fluid or tissue is then predisposed for endothelialization to occur.

TABLE 1

PERIODIC TABLE ELEMENT	ABBREVIATION	FULL NAME	COMPOSITE CHARGE VALUE
22	Ti	titanium	1.36
23	V	vanadium	1.53
40	Zr	zirconium	1.22
41	Nb	niobium or columbium	1.33
42	Mo	molybdenum	1.47
72	Hf	hafnium	1.16
73	Ta	tantalum	1.30
74	W	tungsten	1.47

In some embodiments, the expandable implants described herein can be formed with tubular braid, or sheets of woven filaments (forming a mesh, weave or fabric). The filaments can be wire or polymer or other suitable material. The

expandable implants can be braided wire (e.g. NiTi wire), and can include a mixture of wire types and wire sizes (e.g. NiTi and Platinum wire, and e.g. 0.001" wire braided with 0.00125" wire). The expandable implants can also be made with polymer fibers, or polymer fibers and metal wire mixed together. In some embodiments, the filaments or wires for the braid or mesh can be formed with a radiopaque material. In some embodiments, the filaments or wires for the braid or mesh can include, for example, a wire coextruded with a platinum core surrounded by nitinol (NiTi). In other words, the wire includes two concentric circles when viewed in a cross-sectional view, with the center or core wire being platinum, and the outer wire being nitinol. The percentage of platinum can be, for example, between 5% platinum to 50% platinum and several variations in between (e.g., 10%, 15%, 20%, 25%, 30%, 35%, 40%, 45%). Said another way, a percentage of a diameter of the wire can be, for example 5% to 50% platinum. In some embodiments, the percentage of platinum to nitinol is 30% platinum and 70% nitinol. In some embodiments, the expandable implants can be formed with one or more bioabsorbable materials. In some embodiments, after the expandable implant is formed, the mesh of the implant can be etched to remove an outer oxide layer. This can provide corrosion reduction and/or help thrombosis form faster.

The expandable implants described herein can be formed with one or more soft pliable materials such that the expandable implant can be deployed, for example, in a ruptured or unruptured aneurysm. In some embodiments an expandable implant as described herein can be formed with one or more materials such that the expandable implant has variable stiffness. For example, a first portion of the expandable implant can be formed with a first material and a second portion of the expandable implant can be formed with a second material different than the first material, or the first material can have a different thickness than the second material. For example, in some embodiments, a distal end portion of the expandable implant can be formed with a first material and a proximal end portion of the expandable implant can be formed with a second material different than the first material. In some embodiments, a proximal end portion of an expandable implant can be formed with a first material that provides for greater stiffness than a second material with which a distal end portion of the expandable implant is formed. Such an embodiment may be desirable such that the softer distal end portion of the implant can be deployed within an aneurysm and the stiffer proximal end portion can provide more structure to help support the implant at, for example, a neck of the aneurysm.

The mesh of the expandable implants can be made by a variety of different forms, including, but not limited to, braiding, weaving, welding, or laser cutting. The mesh can have an operating length, for example, in a range of about 0.5 cm to about 70 cm. In some embodiments, the mesh can have a length of 30 cm. In some embodiments, the mesh can have a diameter in a range of about 0.5-60 mm. In some embodiments, the mesh can have a diameter of up to about 10 mm when expanded (e.g., about 9.5 mm for an outer porous member or portion, about 8 mm for an inner porous member or portion). The mesh can have a single density or can have two or more densities. For example, in some embodiments, the number of variable densities can be in a range of about 2 to about 10. For example, a first density can be about 100 PPI and a second density can be about 40 PPI (PPI=pies per inch). The braid pattern can be any pattern suitable, for example, a one-over-one configuration, or two-over-one configuration, etc. Strand count for the mesh can be

in a range of about 4 strands to about 288 strands. In some embodiments, the strand count is about 48 strands. Common multiples of 4, 8, 16, 24, 32, 64, 72, 96, 128, 144, 192 and 288 strands for braid are available using commercial braid-ers.

A single expandable implant can include wires of the same size or a combination of 2 different wire sizes. For example, the expandable implant can have 24 wires of 0.001" and 24 wires of 0.0005". The thicker wires can impart additional strength to the expandable implant and the thinner wire can provide density. In addition, any combination of wire count, wire diameter, braid angle or pics per inch can be used to make the mesh of the expandable implant.

CONCLUSION

While various embodiments of the invention have been described above, it should be understood that they have been presented by way of example only, and not limitation. Where methods and steps described above indicate certain events occurring in certain order, those of ordinary skill in the art having the benefit of this disclosure would recognize that the ordering of certain steps may be modified and that such modifications are in accordance with the variations of the invention. Additionally, certain of the steps may be performed concurrently in a parallel process when possible, as well as performed sequentially as described above. For example, the expandable implant can be inserted into the catheter concurrently with positioning of the expandable catheter adjacent the aneurysm.

The embodiments have been particularly shown and described, but it will be understood that various changes in form and details may be made. For example, although various embodiments have been described as having particular features and/or combinations of components, other embodiments are possible having any combination or sub-combination of any features and/or components from any of the embodiments described herein. The specific configurations of the various components can also be varied.

For example, although the embodiments (e.g., medical device **1010**) illustrated and described herein include one or two porous members or portions (e.g., porous members **1020**, **1030**), in other embodiments, any suitable number of porous members or portions can be included. For example, in some embodiments, the medical device **1010** can also include a third porous member (not shown) having a first end and a second end and coupled to at least one of the first porous member **1020** and the second porous member **1030**. Like the first and second porous members **1020**, **1030**, the third porous member can have a collapsed configuration for insertion through the blood vessel and an expanded configuration for occupying the sac of the aneurysm. The third porous member can be substantially elongate and have a width in its expanded configuration that is greater than its width in its collapsed configuration.

In another example, a radiopaque marker of a medical device illustrated and described can be differently positioned on an expandable implant of the medical device. Moreover, the size and specific shape of the various components can be different than the embodiments shown, while still providing the functions as described herein.

We claim:

1. An occlusive device configured to be positioned within an aneurysm, the device comprising:
 - a coupler;
 - a mesh having a length, a first end of the length, and a second end of the length opposite the first end, wherein

55

the coupler is disposed at the first end, and wherein the mesh has a preset shape such that, at least when the mesh is in a deployed state, the mesh includes a flattened broad portion comprising a longitudinally flattened tubular braid having a width that tapers in the direction of the first end until terminating at the coupler; and

a coil comprising an elongated member wound around an axis such that the wound elongated member forms a tubular sidewall defining a lumen therethrough, the coil having a length, a first end of the length, and a second end of the length opposite the first end, wherein the first end of the coil is coupled to the first end of the mesh at the coupler.

2. The occlusive device of claim 1, wherein the mesh is a braid.

3. The occlusive device of claim 1, wherein the mesh is a braid comprising superelastic materials.

4. The occlusive device of claim 1, wherein the mesh is a braid, and wherein the braid is formed of a plurality of drawn-filled tube ("DFT") wires.

5. The occlusive device of claim 4, wherein the DFT wires comprise a platinum core surrounded by an outer nitinol layer.

6. The occlusive device of claim 1, further comprising a suture strand, wherein the coil is disposed over at least a portion of the suture strand.

7. The occlusive device of claim 1, wherein the coil is radiopaque.

8. The occlusive device of claim 1, wherein the mesh is configured to be positioned across a neck of the aneurysm to cover the neck of the aneurysm and substantially reduce or prevent blood flow into the aneurysm.

9. The occlusive device of claim 1, wherein the mesh is configured to be positioned across a neck of the aneurysm such that the mesh serves as a scaffold for endothelial cell attachment to promote endothelialization of the neck.

10. The occlusive device of claim 1, wherein the coupler is a radiopaque marker band.

11. The occlusive device of claim 1, wherein the mesh is a flattened tubular braid.

12. An occlusive device configured to be positioned within an aneurysm, the device comprising:

a first coupler;

a second coupler;

a mesh having a length, a first end of the length, and a second end of the length opposite the first end, wherein the first coupler is disposed at the first end and the

56

second coupler is disposed at the second end, and wherein the mesh has a preset shape such that, at least when the mesh is in a deployed state, the mesh includes a flattened broad portion comprising a longitudinally flattened tubular braid having a width that tapers in the direction of the first end until terminating at the first coupler;

a coil comprising an elongated member wound around an axis such that the wound elongated member forms a tubular sidewall defining a lumen therethrough, the coil having a length, a first end of the length, and a second end of the length opposite the first end, wherein the first end of the coil is coupled to the first end of the mesh at the first coupler; and

a lead-in member coupled to the second end of the mesh at the second coupler.

13. The occlusive device of claim 12, wherein the mesh is a braid.

14. The occlusive device of claim 12, wherein the mesh is a braid comprising superelastic materials.

15. The occlusive device of claim 12, wherein the mesh is a braid, and wherein the braid is formed of a plurality of drawn-filled tube ("DFT") wires.

16. The occlusive device of claim 15, wherein the DFT wires comprise a platinum core surrounded by an outer nitinol layer.

17. The occlusive device of claim 12, further comprising a suture strand, wherein the coil is disposed over at least a portion of the suture strand.

18. The occlusive device of claim 12, wherein the coil is radiopaque.

19. The occlusive device of claim 12, wherein the mesh is configured to be positioned across a neck of the aneurysm to cover the neck of the aneurysm and substantially reduce or prevent blood flow into the aneurysm.

20. The occlusive device of claim 12, wherein the mesh is configured to be positioned across a neck of the aneurysm such that the mesh serves as a scaffold for endothelial cell attachment to promote endothelialization of the neck.

21. The occlusive device of claim 12, wherein the first coupler is a first marker band and the second coupler is a second marker band.

22. The occlusive device of claim 12, wherein the mesh is a flattened tubular braid.

* * * * *

UNITED STATES PATENT AND TRADEMARK OFFICE
CERTIFICATE OF CORRECTION

PATENT NO. : 10,898,200 B2
APPLICATION NO. : 16/147883
DATED : January 26, 2021
INVENTOR(S) : Aboytes et al.

Page 1 of 1

It is certified that error appears in the above-identified patent and that said Letters Patent is hereby corrected as shown below:

On the Title Page

Item “(63)” in Column 1, under “Related U.S. Application Data”, Line 2, after “2017,” insert
-- now Pat. No. 10,675,037, --.

Signed and Sealed this
Second Day of March, 2021



Drew Hirshfeld
*Performing the Functions and Duties of the
Under Secretary of Commerce for Intellectual Property and
Director of the United States Patent and Trademark Office*