



US011819246B2

(12) **United States Patent**
Edelhauser et al.

(10) **Patent No.:** **US 11,819,246 B2**
(45) **Date of Patent:** ***Nov. 21, 2023**

(54) **SOFTWARE FOR USE WITH DEFORMITY CORRECTION**

(71) Applicant: **Stryker European Operations Holdings LLC**, Kalamazoo, MI (US)

(72) Inventors: **Adam John Edelhauser**, Nyack, NY (US); **Ashish Gangwar**, Gurgaon (IN); **Sandeep Menon**, 1st Cross Cooke Town (IN)

(73) Assignee: **Stryker European Operations Holdings LLC**, Kalamazoo, MI (US)

(*) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 394 days.

This patent is subject to a terminal disclaimer.

(21) Appl. No.: **17/109,222**

(22) Filed: **Dec. 2, 2020**

(65) **Prior Publication Data**
US 2021/0077149 A1 Mar. 18, 2021

Related U.S. Application Data

(63) Continuation of application No. 16/218,778, filed on Dec. 13, 2018, now Pat. No. 10,881,433, which is a (Continued)

(51) **Int. Cl.**
A61B 17/62 (2006.01)
A61B 17/66 (2006.01)
(Continued)

(52) **U.S. Cl.**
CPC **A61B 17/62** (2013.01); **A61B 17/66** (2013.01); **A61B 34/10** (2016.02); **A61B 90/37** (2016.02);
(Continued)

(58) **Field of Classification Search**
CPC A61B 17/60; A61B 17/62; A61B 17/64; A61B 17/6416-6491; A61B 34/00-10; A61B 2034/101
See application file for complete search history.

(56) **References Cited**
U.S. PATENT DOCUMENTS

5,546,942 A 8/1996 Zhang
5,681,309 A 10/1997 Ross, Jr. et al.
(Continued)

FOREIGN PATENT DOCUMENTS

DE 102006048451 A1 4/2008
EP 2767252 A1 8/2014
(Continued)

OTHER PUBLICATIONS

Extended European Search Report for Application No. 14154820.6 dated Jun. 16, 2014.


(Continued)

Primary Examiner — Eduardo C Robert
Assistant Examiner — Michelle C Eckman
(74) *Attorney, Agent, or Firm* — SLEMAN & LUND LLP

(57) **ABSTRACT**

The present disclosure relates to software used in planning the correction of bone deformities preoperatively or post-operatively, and in particular relates to virtually manipulating rings and struts of an external fixation frame in order to plan the steps for making a desired correction to two or more bone portions of a patient. The software can be used prior to surgery, allowing a user to virtually define a bone deformity, and virtually add and manipulate fixation rings and struts to the bone deformity. Based on the virtual manipulations, a correction plan can be generated that describes length adjustments that should be made to the plurality of model struts over a period of time to correct the bone deformity.

(Continued)



The software can also be used after surgical fixation of the fixation frame and struts to the deformed bone.

17 Claims, 18 Drawing Sheets

Related U.S. Application Data

continuation of application No. 15/627,900, filed on Jun. 20, 2017, now Pat. No. 10,194,944, which is a continuation of application No. 14/926,576, filed on Oct. 29, 2015, now Pat. No. 9,724,129, which is a continuation of application No. 13/770,056, filed on Feb. 19, 2013, now Pat. No. 9,204,937.

- (51) **Int. Cl.**
A61B 34/10 (2016.01)
A61B 90/00 (2016.01)
- (52) **U.S. Cl.**
 CPC ... *A61B 2034/101* (2016.02); *A61B 2034/104* (2016.02); *A61B 2034/105* (2016.02); *A61B 2090/364* (2016.02); *A61B 2090/376* (2016.02)

(56) **References Cited**

U.S. PATENT DOCUMENTS

5,682,886 A 11/1997 Delp et al.
 5,702,389 A 12/1997 Taylor et al.
 5,728,095 A 3/1998 Taylor et al.
 5,769,092 A 6/1998 Williamson, Jr.
 5,824,085 A 10/1998 Sahay et al.
 5,880,976 A 3/1999 DiGioia, III et al.
 5,891,143 A 4/1999 Taylor et al.
 5,971,984 A 10/1999 Taylor et al.
 6,030,386 A 2/2000 Taylor et al.
 6,112,109 A 8/2000 D'Urso
 6,129,727 A 10/2000 Austin et al.
 6,205,411 B1 3/2001 DiGioia, III et al.
 6,701,174 B1 3/2004 Krause et al.
 6,711,432 B1 3/2004 Krause et al.
 7,039,225 B2 5/2006 Tanaka et al.
 7,280,683 B2 10/2007 Bi et al.
 7,388,972 B2 6/2008 Kitson
 7,394,946 B2 7/2008 Dewaele
 7,547,307 B2 6/2009 Carson et al.
 RE40,914 E 9/2009 Taylor et al.
 7,837,621 B2 11/2010 Krause et al.
 7,967,868 B2 6/2011 White et al.
 8,055,487 B2 11/2011 James
 8,157,800 B2 4/2012 Vvedensky et al.
 8,257,353 B2 9/2012 Wong et al.
 8,265,949 B2 9/2012 Haddad
 8,296,094 B2 10/2012 Harrison et al.
 8,311,306 B2 11/2012 Pavlovskaja et al.
 8,333,766 B2 12/2012 Edelhauser et al.
 8,419,732 B2 4/2013 Mullaney
 8,439,914 B2 5/2013 Ross et al.
 8,484,001 B2 7/2013 Glozman et al.
 8,617,171 B2 12/2013 Park et al.
 8,654,150 B2 2/2014 Haskell
 8,715,291 B2 5/2014 Park et al.
 8,731,885 B2 5/2014 Iannotti et al.
 8,737,700 B2 5/2014 Park et al.
 8,777,946 B2 7/2014 Lindahl et al.
 8,860,753 B2 10/2014 Bhandarkar et al.
 8,864,750 B2 10/2014 Ross et al.
 8,864,763 B2 10/2014 Murray et al.
 8,923,590 B2 12/2014 Chen et al.
 8,945,128 B2 2/2015 Singh et al.
 8,952,986 B2 2/2015 Haskell
 9,101,398 B2 8/2015 Singh et al.
 9,204,937 B2 12/2015 Edelhauser et al.
 9,524,581 B2 12/2016 Haskell

9,724,129 B2 8/2017 Edelhauser et al.
 9,788,861 B2 10/2017 Murray et al.
 9,895,167 B2 2/2018 Edelhauser et al.
 10,082,384 B1 9/2018 Singh
 10,154,884 B2 12/2018 Kumar et al.
 10,194,944 B2* 2/2019 Edelhauser A61B 34/10
 10,251,705 B2 4/2019 Kumar et al.
 10,258,377 B1 4/2019 Lavi et al.
 10,603,112 B2 3/2020 Kumar et al.
 2002/0010465 A1 1/2002 Koo et al.
 2003/0191466 A1 10/2003 Austin et al.
 2004/0039259 A1* 2/2004 Krause G06T 15/20
 600/300
 2004/0068187 A1 4/2004 Krause et al.
 2004/0073211 A1 4/2004 Austin et al.
 2004/0073212 A1 4/2004 Kim
 2005/0004451 A1 1/2005 Vilsmeier et al.
 2005/0054917 A1 3/2005 Kitson
 2005/0215997 A1 9/2005 Austin et al.
 2005/0267360 A1 12/2005 Birkenbach et al.
 2006/0015120 A1 1/2006 Richard et al.
 2006/0079745 A1 4/2006 Viswanathan
 2006/0161052 A1 7/2006 Colombet et al.
 2006/0189842 A1 8/2006 Hoeg et al.
 2006/0276786 A1 12/2006 Brinker
 2007/0055234 A1 3/2007 McGrath et al.
 2007/0078678 A1 4/2007 DiSilvestro et al.
 2007/0133845 A1 6/2007 Fradkin et al.
 2007/0173815 A1 7/2007 Murase
 2007/0219561 A1 9/2007 Lavallee et al.
 2008/0051779 A1 2/2008 Mackenzie et al.
 2008/0108912 A1 5/2008 Node-Langlois
 2008/0119719 A1 5/2008 Ascenzi et al.
 2008/0137923 A1 6/2008 Spahn
 2008/0177203 A1 7/2008 von Jako
 2008/0198966 A1 8/2008 Hjarn et al.
 2008/0234554 A1 9/2008 Vvedensky et al.
 2008/0243127 A1 10/2008 Lang et al.
 2008/0269741 A1 10/2008 Karidis
 2008/0275467 A1 11/2008 Liao et al.
 2008/0319448 A1 12/2008 Lavallee et al.
 2009/0054887 A1 2/2009 Podhajsky
 2010/0036393 A1 2/2010 Unsworth
 2010/0087819 A1 4/2010 Mullaney
 2010/0130858 A1 5/2010 Arai et al.
 2010/0286995 A1 11/2010 Pekar et al.
 2011/0004199 A1 1/2011 Ross et al.
 2011/0009868 A1 1/2011 Sato
 2011/0029116 A1 2/2011 Jordan et al.
 2011/0103556 A1 5/2011 Carn
 2011/0103676 A1 5/2011 Mullaney
 2011/0116041 A1 5/2011 Hartung et al.
 2011/0188726 A1 8/2011 Nathaniel et al.
 2011/0304332 A1 12/2011 Mahfouz
 2011/0313418 A1 12/2011 Nikonovas
 2011/0313419 A1 12/2011 Mullaney
 2011/0313424 A1 12/2011 Bono et al.
 2012/0071802 A1 3/2012 Reiley et al.
 2012/0130687 A1 5/2012 Otto et al.
 2012/0155732 A1 6/2012 Finkelstein et al.
 2012/0214121 A1 8/2012 Greenberg
 2012/0328174 A1 12/2012 Jadiyahappa
 2012/0330312 A1* 12/2012 Burgherr A61B 90/98
 606/54
 2013/0018666 A1 1/2013 Murphy et al.
 2013/0089253 A1 4/2013 Chabanas et al.
 2013/0096373 A1 4/2013 Chabanas et al.
 2013/0121612 A1 5/2013 Falco, Jr. et al.
 2013/0172783 A1 7/2013 Ikits et al.
 2013/0201212 A1* 8/2013 Haskell A61B 34/10
 345/632
 2013/0211792 A1 8/2013 Kang et al.
 2013/0215114 A1 8/2013 Cherkashin et al.
 2013/0253511 A1 9/2013 Cheng et al.
 2013/0289575 A1 10/2013 Edelhauser et al.
 2014/0039663 A1 2/2014 Boyer et al.
 2014/0073907 A1 3/2014 Kumar et al.
 2014/0189508 A1 7/2014 Granchi et al.
 2014/0236153 A1 8/2014 Edelhauser

(56)

References Cited

U.S. PATENT DOCUMENTS

2014/0270437 A1 9/2014 Shreiber et al.
 2014/0276821 A1 9/2014 Murray et al.
 2014/0303486 A1 10/2014 Baumgartner et al.
 2014/0324403 A1 10/2014 Gotte
 2014/0328460 A1 11/2014 Egli et al.
 2014/0343586 A1 11/2014 Sakuragi
 2014/0350389 A1 11/2014 Powell et al.
 2014/0357984 A1 12/2014 Wallace et al.
 2014/0379356 A1 12/2014 Sachdeva et al.
 2015/0049083 A1 2/2015 Bidne et al.
 2015/0087965 A1 3/2015 Tokuda et al.
 2015/0227679 A1 8/2015 Kamer et al.
 2015/0238271 A1 8/2015 Wollowick et al.
 2016/0042571 A1 2/2016 Mikheev et al.
 2016/0045225 A1 2/2016 Edelhauser et al.
 2016/0092651 A1* 3/2016 Austin G06F 3/04842
 705/2
 2016/0331463 A1 11/2016 Notzli et al.
 2017/0281233 A1 10/2017 Edelhauser et al.
 2017/0303966 A1 10/2017 Edelhauser et al.
 2017/0348054 A1 12/2017 Kumar et al.
 2018/0055569 A1 3/2018 Wahl et al.
 2018/0368928 A1 12/2018 Abedinnasab
 2019/0282276 A1 9/2019 Burgherr et al.

FOREIGN PATENT DOCUMENTS

RU 2448663 C1 4/2012
 RU 2471447 C1 1/2013

RU 2489106 C2 8/2013
 WO 2009076296 A2 6/2009
 WO 2010104567 A1 9/2010
 WO 2013013170 A1 1/2013

OTHER PUBLICATIONS

European Patent Office (ISA), International Search Report and Written Opinion dated Jun. 25, 2013 for International Application No. PCT/US2013/024548, International filing date Feb. 3, 2013.
 LITOS GmbH, "Ilizarov Hexapod System," available from <http://d3llyibkg2zj6z.cloudfront.net/ImageAnexo/Ilozarov-Hexapod-System.-PDF>, dated May 23, 2007.
 Response Ortho LLC, Smart Correction Computer Assisted Circular Hexapod System Brochure, date not known.
 Vreden Russian Research Institute of Traumatology and Orthopedics Ortho-SUV Ltd., Deformity Correction and Fracture Treatment by Software-based Ortho-SUV Frame, Saint-Petersburg, 2013.
 MED Surgical, Adam Frame with Paley's Method, Workshop, Oct. 2010.
 Craveiro-Lopes, MD, Software Assisted "Ortho-SUV Frame", Int'l Congress on External Fixation & Bone Reconstruction, Oct. 22, 2010.
 Smart Correction, Computer-Assisted Circular External Fixator System, website printout, Feb. 2, 2011.
 Extended European Search Report for Application No. EP17173905, dated Oct. 30, 2017.

* cited by examiner

102 →

<p>LOGIN</p> <p>USERNAME <input type="text" value="HLRFDEMO"/> 104</p> <p>PASSWORD <input type="password" value="*****"/> 106</p> <p><input type="button" value="LOGIN"/></p> <p><u>CAN'T ACCESS YOUR ACCOUNT ?</u></p>	<p>CREATE NEW USER ACCOUNT</p> <p>NEW USER? PLEASE CLICK THE BUTTON TO REQUEST FOR NEW USER ACCOUNT</p> <p><input type="button" value="REQUEST NEW ACCOUNT"/></p>
--	--

FIG. 1

110

[HOME](#) [WELCOME USER](#) [SETTINGS](#) | [HELP](#) | [LOGOUT](#)

[START NEW CASE](#) [EXPRESS SEARCH](#) SHOW CASES CREATED:

MY CASES CASES FROM OTHERS

CASE ID	PATIENT NAME	ANATOMY	CREATION DATE	MODIFIED DATE	DELETE	SHARE
CS0000001	JOHN A	LEFT FEMUR	MM/DD/YYYY 10:16:40 PM	MM/DD/YYYY 9:34:30 PM	DELETE	SHARE
CS0000002	JOHN B	LEFT FEMUR	MM/DD/YYYY 10:11:00 PM	MM/DD/YYYY 10:13:30 PM	DELETE	SHARE
CS0000003	JOHN C	LEFT FEMUR	MM/DD/YYYY 10:12:10 PM	MM/DD/YYYY 10:13:11 PM	DELETE	SHARE
CS0000004	JOHN D	LEFT FEMUR	MM/DD/YYYY 10:08:15 PM	MM/DD/YYYY 10:09:14 PM	DELETE	SHARE
CS0000005	JOHN E	LEFT FEMUR	MM/DD/YYYY 9:57:51 PM	MM/DD/YYYY 10:08:02 PM	DELETE	SHARE
CS0000006	JOHN F	LEFT FEMUR	MM/DD/YYYY 8:43:11 PM	MM/DD/YYYY 9:57:25 PM	DELETE	SHARE
CS0000007	JOHN G	LEFT FEMUR	MM/DD/YYYY 8:47:24 PM	MM/DD/YYYY 8:48:18 PM	DELETE	SHARE
CS0000008	JOHN H	LEFT FEMUR	MM/DD/YYYY 8:41:53 PM	MM/DD/YYYY 8:42:03 PM	DELETE	SHARE
CS0000009	JOHN I	LEFT FEMUR	MM/DD/YYYY 8:40:41 PM	MM/DD/YYYY 8:40:51 PM	DELETE	SHARE
CS0000010	JOHN J	LEFT FEMUR	MM/DD/YYYY 8:39:41 PM	MM/DD/YYYY 8:39:50 PM	DELETE	SHARE

[NEXT >](#)

114

FIG. 2



120

WELCOME USER [SETTINGS](#) | [HELP](#) | [LOGOUT](#)

[HOME](#) |
 [CASE DETAILS](#) |
 [DEFORMITY DEFINITION](#) |
 [RING CONFIGURATION](#) |
 [STRUT CONFIGURATION](#) |
 [LAF INPUT](#) |
 [CORRECTION PLAN](#) |
 [STUDY REPORT](#) |
 [SMART TOOL](#)

CASE ID CS0000001

IMAGE PREVIEW

 <p>201 AP XRAY</p>	<p>MMDD/YYYY 6:07:25 PM</p> <p>VIEW DELETE</p>
<p>LATERAL XRAY</p> 	<p>MMDD/YYYY 6:07:54 PM</p> <p>VIEW DELETE</p>

UPLOAD IMAGES

NO FILE CHOSEN

DETAILS
 LATERAL ▾

PATIENT DETAILS

FIRST NAME*
LAST NAME*
MIDDLE NAME*
D.O.B (MMDD/YYYY)*

GENDER* ▾
ANATOMY* ▾ ▾
RACE ▾

STUDY NOTES

ENTER NOTES HERE

* MANDATORY FIELDS

FIG. 3

200





FIG. 4

200

CASE DETAILS DEFORMITY DEFINITION RING CONFIGURATION STRUT CONFIGURATION LAF INPUT CORRECTION PLAN STUDY REPORT SMART TOOL



204

206

IMAGE MANIPULATION

DEFORMITY DEFINITION

REFERENCE FRAGMENT: PROXIMAL

ANTERIOR POSTERIOR PLANE

TRANSLATION(MM)*: 3

ANGULATION(DEG)*: -16

LATERAL PLANE

TRANSLATION(MM)*: 0

ANGULATION(DEG)*: 0

AXIAL PLANE

TRANSLATION(MM)*: -2

ANGULATION(DEG)*: 0

DEFORMITY APEX

OSTEOTOMY PLANE

PREVIOUS NEXT

FIG. 5

200

WELCOME USER [SETTINGS](#) | [HELP](#) | [LOGOUT](#)

[HOME](#) |
 [CASE DETAILS](#) |
 [DEFORMITY DEFINITION](#) |
 [RING CONFIGURATION](#) |
 [STRUT CONFIGURATION](#) |
 [LAF INPUT](#) |
 [CORRECTION PLAN](#) |
 [STUDY REPORT](#) |
 [SMART TOOL](#)

REFERENCE FRAGMENT*

ANTERIOR POSTERIOR PLANE

TRANSLATION(MM)*

ANGULATION(DEG)*

LATERAL PLANE

TRANSLATION(MM)*

ANGULATION(DEG)*

AXIAL PLANE

DISTANCE(MM)*

ROTATION(DEG)*

OSTECTOMY PLANE

* MANDATORY FIELDS

FIG. 6

250

HOME WELCOME USER SETTINGS | HELP | LOGOUT

CASE DETAILS DEFORMITY DEFINITION RING CONFIGURATION STRUT CONFIGURATION LAF INPUT CORRECTION PLAN STUDY REPORT SMART TOOL

AP		LATERAL		POSTERIOR	
MEDIAL		LATERAL		ANTERIOR	
AXIAL		PERSPECTIVE			
MEDIAL		LATERAL			

SELECT MODE OF FRAME APPLICATION*

PRE-OP POST-OP

* MANDATORY FIELDS

SAVE CASE DATA << PREVIOUS NEXT >>

FIG. 7

300

HOME WELCOME USER SETTINGS | HELP | LOGOUT

CASE DETAILS DEFORMITY DEFINITION RING CONFIGURATION STRUT CONFIGURATION LAF INPUT CORRECTION PLAN STUDY REPORT SMART TOOL

SELECT MODE OF FRAME APPLICATION *

PRE-OP POST-OP

RING SELECTION

REFERENCE RING (MM) MOVING RING (MM)

REFERENCE RING PLACEMENT

AP OFFSET (MM)

LAT OFFSET (MM)

AXIAL OFFSET (MM)

AXIAL ANGLE (DEG)

MOVING RING PLACEMENT

POSITION (MM)

ORIENTATION (DEG)

AP: LAT: AXIAL:

* MANDATORY FIELDS

SAVE CASE DATA << PREVIOUS NEXT >>

FIG. 8

350





FIG. 9


400

HOME WELCOME USER SETTINGS | HELP | LOGOUT


CASE DETAILS DEFORMITY DEFINITION RING CONFIGURATION STRUT CONFIGURATION LAF INPUT CORRECTION PLAN STUDY REPORT SMART TOOL




AP



LATERAL



MEDIAL



PERSPECTIVE

POSSIBLE STRUT COMBINATIONS

COMPUTE POSSIBLE STRUT COMBINATIONS

STRUT1 TYPE	STRUT2 TYPE	STRUT3 TYPE
<input checked="" type="radio"/> SMALL	LARGE	STANDARD
<input type="radio"/> STANDARD	LARGE	STANDARD

*OPTIMAL STRUT COMBINATION IS HIGHLIGHTED

STRUTS POSITION

STRUT	STRUT LENGTH (MM)	STRUT ANGLE (DEG)	BASE ANGLE (DEG)
STRUT1	154.91	43.03	80.5
STRUT2	208.21	53.82	103.57
STRUT3	187.57	70.83	81.12

*MANDATORY FIELDS

SAVE CASE DATA

<< PREVIOUS NEXT >>

FIG. 10

450

HOME WELCOME USER SETTINGS | HELP | LOGOUT

CASE DETAILS DEFORMITY DEFINITION RING CONFIGURATION STRUT CONFIGURATION LAF INPUT CORRECTION PLAN STUDY REPORT SMART TOOL

AP		LATERAL	
MEDIAL		ANTERIOR	
AXIAL		POSTERIOR	
MEDIAL		PERSPECTIVE	

LIMITING ANATOMICAL FACTOR

AP OFFSET(MM) LATERAL OFFSET(MM) AXIAL OFFSET(MM)

MAX. DISTRACTION RATE(MM/DAY)

MINIMUM CORRECTION TIME * OVERRIDE

*MANDATORY FIELDS

FIG. 11

500

HOME WELCOME USER SETTINGS | HELP | LOGOUT

CASE DETAILS DEFORMITY DEFINITION RING CONFIGURATION STRUT CONFIGURATION LAF INPUT CORRECTION PLAN STUDY REPORT SMART TOOL

CS0000001

START DATE MMDDYYYY

GENERATE CORRECTION SCHEDULE

0000000000

RUN SIMULATION

SAVE CASE DATA

<< PREVIOUS

NEXT >>

FIG. 12

500

HOME
WELCOME USER
SETTINGS
HELP
LOGOUT

CASE DETAILS
DEFORMITY DEFINITION
RING CONFIGURATION
STRUT CONFIGURATION
LAF INPUT
CORRECTION PLAN
STUDY REPORT
SMART TOOL

CS9001101


START DATE:

CORRECTION SCHEDULE

DAY	DATE	STRUT1 (DEG)		STRUT2 (DEG)		STRUT3 (DEG)	
		POSITION	ANGLE	POSITION	ANGLE	POSITION	ANGLE
1	MM/DD/YYYY	277.2	26.4	18.4	43.6	178.9	12.2
2	MM/DD/YYYY	279.2	26.9	16	42.5	178.5	13.6
3	MM/DD/YYYY	281.2	27.4	14.1	41.3	178.2	14.9
4	MM/DD/YYYY	283.3	27.8	12.5	40	177.8	16.3
5	MM/DD/YYYY	285.5	28.2	11.1	38.6	177.4	17.7
6	MM/DD/YYYY	287.7	28.6	10	37.2	177	19.1
7	MM/DD/YYYY	289.9	28.9	9.1	35.8	176.7	20.5
12							

DAY NO = 1

520



510

FIG. 13

600

WELCOME USER SETTINGS | HELP | LOGOUT

CASE DETAILS DEFORMITY DEFINITION RING CONFIGURATION STRUT CONFIGURATION LAF INPUT CORRECTION PLAN STUDY REPORT SMART TOOL

PRINT REPORT

STUDY REPORT

CASE DETAILS

CASE ID	CS000001	CREATED DATE	MM/DD/YYYY
PATIENT NAME	JOHN A	DATE OF BIRTH	MM/DD/YYYY
SEX	MALE	ANATOMY	LEFT TIBIA

DEFORMITY DEFINITION

	AP	LATERAL	AXIAL
TRANSLATION(MM)	0 MM - MEDIAL	0 MM - ANTERIOR	23.9 MM - LONG
ANGULATION(DEG)	13 DEG - APEXLATERAL	15.6 DEG - APEXPOSTERIOR	4.5 DEG - EXTERNAL

LAF INPUT

	AP OFFSET (MM)	LATERAL OFFSET (MM)	AXIAL OFFSET (MM)
LAF POSITIN	0	0	0

SAVE CASE DATA << PREVIOUS NEXT >>

FIG. 14

700

WELCOME USER SETTINGS | HELP | LOGOUT

CORRECTION PLAN STUDY REPORT SMART TOOL

HOME CASE DETAILS DEFORMITY DEFINITION RING CONFIGURATION STRUT CONFIGURATION LAF INPUT

LAUNCH SMART CLIENT

CS9000001

SMART TOOL PLANE

DAY	DATE	STRUT1		STRUT2		STRUT3		
		SCREW A (DEG)	SCREW B (DEG)	SCREW A (DEG)	SCREW B (DEG)	SCREW A (DEG)	SCREW B (DEG)	
1	MM/DD/YYYY	0	0	0	0	0	0	
2	MM/DD/YYYY	163	-50	-1376	-18	-45	-35	
3	MM/DD/YYYY	172	-42	-1385	-19	-50	-35	
4	MM/DD/YYYY	179	-34	-1395	-19	-54	-35	
5	MM/DD/YYYY	184	-26	-1404	-20	-59	-34	
6	MM/DD/YYYY	187	-18	-1412	-21	-64	-34	
7	MM/DD/YYYY	187	-10	-1420	-22	-69	-34	
8	MM/DD/YYYY	186	-3	-1428	-23	-74	-34	
9	MM/DD/YYYY	184	5	-1435	-24	-78	-34	
10	MM/DD/YYYY	179	12	-1442	-25	-83	-33	
							12	

SAVE CASE DATA


<< PREVIOUS

FIG. 15

800

HOME WELCOME USER SETTINGS | HELP | LOGOUT

CASE DETAILS DEFORMITY DEFINITION RING CONFIGURATION STRUT CONFIGURATION LAF INPUT CORRECTION PLAN STUDY REPORT SMART TOOL



SELECT MODE OF FRAME APPLICATION *

PRE-OP POST-OP

RING SELECTION

REFERENCE RING (MM) MOVING RING (MM)

REFERENCE RING PLACEMENT

AP OFFSET (MM)

LAT OFFSET (MM)

AXIAL OFFSET (MM)

AXIAL ANGLE (DEG)

MOVING RING POSITION WILL BE CALCULATED AUTOMATICALLY.

* MANDATORY FIELDS

SAVE CASE DATA << PREVIOUS NEXT >>

FIG. 16

900

HOME WELCOME USER SETTINGS HELP LOGOUT

CASE DETAILS DEFORMITY DEFINITION RING CONFIGURATION STRUT CONFIGURATION LAF INPUT CORRECTION PLAN STUDY REPORT SMART TOOL

STRUT SELECTION

STRUT 1* STRUT 2* STRUT 3*

STRUTS POSITION

STRUT	STRUT LENGTH (MM)	STRUT ANGLE (DEG)
STRUT 1	160.2	52
STRUT 2	206.5	60
STRUT 3	184.3	54

MOVING RING POSITION (MM) MOVING RING ORIENTATION (DEG)

	MOVING RING POSITION (MM)	MOVING RING ORIENTATION (DEG)
AP	-12.7184	2.32
LAT	-33.9753	28.78
AXIAL	-124.8888	-3.6

*MANDATORY FIELDS

FIG. 17

1000

WELCOME USER [SETTINGS](#) | [HELP](#) | [LOGOUT](#)

[HOME](#)

CS0000001

ANATOMY

FEMUR

PATIENT DETAILS

PATIENT NAME

JOHN A

SEX

MALE

CREATED DATE

MMDDYYYY

DATE OF BIRTH

MMDDYYYY

STUDIES

[CREATE NEW](#)

[CASE SUMMARY](#)

S NO	STUDY DATE	STUDY NOTES	
1	MMDDYYYY	ENTER NOTES HERE	OPEN DELETE

FIG. 18

SOFTWARE FOR USE WITH DEFORMITY CORRECTION

CROSS-REFERENCE TO RELATED APPLICATIONS

This application is a continuation of U.S. patent application Ser. No. 16/218,778, filed Dec. 13, 2018, which is a continuation of U.S. patent application Ser. No. 15/627,900, filed Jun. 20, 2017 and issued as U.S. Pat. No. 10,194,944, which is a continuation of U.S. patent application Ser. No. 14/926,576, filed Oct. 29, 2015 and issued as U.S. Pat. No. 9,724,129, which is a continuation of U.S. patent application Ser. No. 13/770,056, filed Feb. 19, 2013 and issued as U.S. Pat. No. 9,204,937, the disclosures of which are incorporated by reference herein.

FIELD OF THE INVENTION

The present disclosure relates to software used in planning the correction of bone deformities preoperatively or postoperatively, and in particular relates to virtually manipulating rings and struts of an external fixation frame in order to plan the steps for making a desired correction to two or more bone portions of a patient.

BACKGROUND OF THE INVENTION

Currently, external fixation systems may be used to correct skeletal deformities using the distraction osteogenesis process, for example. The Ilizarov external fixation device (or similar system) may be used for such a purpose. The Ilizarov-type devices generally translate bone segments by manipulating the position of rings connected to each bone segment.

These external fixation devices generally utilize threaded rods fixated to through-holes in the rings to build the frame. In order to build the desired frame, these rods generally have to have different lengths.

Once the frame is installed, the patient or surgeon moves the rings or percutaneous fixation components manually or mechanically by adjusting a series of nuts.

As fixation devices become more complex, the task of determining the optimal lengths and positions of the struts with respect to rings of the fixation frame, as well as creating a correction plan for manipulating the struts to correct the bone deformity, becomes more difficult.

The increasing difficulty of these determinations decreases the attractiveness of using complex fixation frames. It would be advantageous to have an at least partially automated method for determining the optimal configuration of a fixation frame in reference to a deformed bone, as well as a correction plan for manipulating the fixation frame to correct the bone deformity.

BRIEF SUMMARY OF THE INVENTION

In one embodiment of the invention, a method of generating a correction plan for correcting a deformed bone includes displaying on a visual medium a first model of the deformed bone in a first plane, the first model of the deformed bone having a position and orientation on the visual medium. A model bone having a first configuration is overlaid on the first model of the deformed bone in the first plane and a position and orientation of the model bone is manipulated into a second configuration being substantially similar to the position and orientation of the first model of

the deformed bone in the first plane. A second model of the deformed bone is displayed on the visual medium in a second plane, the second model of the deformed bone having a position and orientation on the visual medium. The model bone in the second configuration is overlaid on the second model of the deformed bone in the second plane and the position and orientation of the model bone is manipulated into a third configuration being substantially similar to the position and orientation of the second model of the deformed bone in the second plane. The model bone in the third configuration is projected onto a three dimensional axis. A model of first and second fixation rings is and positional data corresponding to a position and orientation of the models of the first and second fixation rings with respect to the three dimensional axis is displayed. The positional data corresponding to the models of the first and second fixation rings is manipulated until the first and second model fixation rings are each in a desired position relative to the model bone in the third configuration.

The first model of the deformed bone may be an x-ray image displayed on the visual medium in an anterior-posterior view, while the second model of the deformed bone may be an x-ray image displayed on the visual medium in a lateral view.

Positional data corresponding to the position and orientation of the model bone with respect to the three dimensional axis may be displayed on the visual medium. The model bone may have a plurality of portions and the positional data corresponding to the model bone may include coordinate locations and angular orientations of at least one of the plurality of portions of the model bone on the three dimensional axis. The step of manipulating the position and orientation of the model bone may include one of entering numerical values into an input box and moving a slide-bar corresponding to the numerical values.

The step of manipulating the position and orientation of the model bone may include changing the position and orientation of the model bone in an anterior-posterior plane, a lateral plane, and an axial plane. Changing the position and orientation of the model bone in the anterior-posterior plane, the lateral plane, and the axial plane may each include changing at least one of a translation or angulation value.

Combinations of sizes of a plurality of model struts to connect the models of the first and second fixation rings may be determined with an algorithm using the positional data corresponding to the desired position of the models of the first and second fixation rings. One of the combinations of sizes of the plurality of model struts may be selected. The correction plan may be determined with an algorithm using the manipulated position and orientation of the model bone in the third configuration, the positional data corresponding to the desired position of the models of the first and second fixation rings, and the selected combinations of sizes of the plurality of model struts. The correction plan describes length adjustments that should be made to the plurality of model struts over a period of time. The correction plan results in length adjustments made to the plurality of model struts such that the models of the first and second fixation rings are in corrected positions.

In another embodiment of the invention, a method of generating a correction plan for correcting a deformed bone includes the step of displaying on a visual medium a first model of the deformed bone in a first plane, the first model of the deformed bone having a position and orientation on the visual medium. A model bone having a first configuration is overlaid on the first model of the deformed bone in the first plane and a position and orientation of the model

bone is manipulated into a second configuration being substantially similar to the position and orientation of the first model of the deformed bone in the first plane. A second model of the deformed bone is displayed on the visual medium in a second plane, the second model of the deformed bone having a position and orientation on the visual medium. The model bone in the second configuration is overlaid on the second model of the deformed bone in the second plane and the position and orientation of the model bone is manipulated into a third configuration being substantially similar to the position and orientation of the second model of the deformed bone in the second plane. The model bone is projected in the third configuration onto a three dimensional axis. A model of a first fixation ring having a first configuration and positional data corresponding to a position and orientation of the model of the first fixation ring with respect to the three dimensional axis is displayed. The positional data corresponding to the position and orientation of the model of the first fixation ring is manipulated into a second configuration relative to the model bone being substantially similar to a position and orientation of a first fixation ring relative to the deformed bone.

The first model of the deformed bone may be an x-ray image displayed on the visual medium in an anterior-posterior view, and the second model of the deformed bone may be an x-ray image displayed on the visual medium in a lateral view.

Positional data corresponding to the position and orientation of the model bone with respect to the three dimensional axis is displayed on the visual medium. The model bone has a plurality of portions and the positional data corresponding to the model bone includes coordinate locations and angular orientations of at least one of the plurality of portions of the model bone on the three dimensional axis.

The step of manipulating the position and orientation of the model bone includes one of entering numerical values into an input box and moving a slide-bar corresponding to the numerical values. The step of manipulating the position and orientation of the model bone includes changing the position and orientation of the model bone in an anterior-posterior plane, a lateral plane, and an axial plane. Changing the position and orientation of the model bone in the anterior-posterior plane, the lateral plane, and the axial plane each includes changing at least one of a translation or angulation value.

A plurality of model struts having a first configuration and positional data corresponding to a position and orientation of the plurality of model struts in relation to the model of the first fixation ring are displayed on the visual medium. The positional data corresponding to the position and orientation of the plurality of model struts is manipulated into a second configuration relative to the model of the first fixation ring being substantially similar to a position and orientation of a plurality of struts relative to the first fixation ring. A position and orientation of a second model fixation ring is determined, wherein the positional data of the second model fixation ring relative to the model bone is substantially similar to a position and orientation of a second fixation ring relative to the deformed bone.

The correction plan is determined with an algorithm using the manipulated position and orientation of the model bone in the third configuration, the manipulated positional data corresponding to the first model fixation ring, and the manipulated positional data corresponding to the plurality of

model struts. The correction plan describes length adjustments that should be made to the plurality of model struts over a period of time.

BRIEF DESCRIPTION OF THE DRAWINGS

FIG. 1 illustrates a login screen of a deformity correction application.

FIG. 2 illustrates a home page screen of a deformity correction application.

FIG. 3 illustrates a case details screen of a deformity correction application.

FIG. 4 illustrates an image manipulation function of a deformity correction application.

FIG. 5 illustrates a deformity definition function of a deformity correction application.

FIG. 6 illustrates an alternate deformity definition function of a deformity correction application.

FIG. 7 illustrates a first ring configuration screen of a deformity correction application.

FIG. 8 illustrates a second ring configuration screen of a pre-operative mode of a deformity correction application.

FIG. 9 illustrates a first strut configuration screen of a pre-operative mode of a deformity correction application.

FIG. 10 illustrates a second strut configuration screen of a pre-operative mode of a deformity correction application.

FIG. 11 illustrates a limiting anatomical factor input screen of a deformity correction application.

FIG. 12 illustrates a correction plan generation screen of a deformity correction application during the generation of a correction plan.

FIG. 13 illustrates a correction plan generation screen of a deformity correction application displaying the correction plan.

FIG. 14 illustrates a study report screen of a deformity correction application.

FIG. 15 illustrates a smart tool screen of a deformity correction application.

FIG. 16 illustrates a second ring configuration screen of a post-operative mode of a deformity correction application.

FIG. 17 illustrates a strut configuration screen of a post-operative mode of a deformity correction application.

FIG. 18 illustrates an existing case screen of a deformity correction application.

DETAILED DESCRIPTION

In one embodiment of the invention, software aids a user, such as a physician, surgeon, or other medical personnel, in planning and carrying out the correction of a bone deformity using a limb reconstruction frame using a web application, for example.

As shown in FIG. 1, upon starting the application, the user is presented with a login screen 102. The login screen 102 preferably includes a username field 104 and password field 106 in which the user enters, respectively, a username and password to gain further access to the application. This step of authentication may, for example, help maintain compliance with patient privacy regulations. In cases where a first time user tries to gain further access to the application, a new user account will have to be created.

As shown in FIG. 2, upon logging in, the user is taken to the home page 110 (FIG. 2). From the home page 110, the user has the option of starting a new case by choosing the new case option 112, or choosing an existing case by, for example, searching for an existing case using keywords or case number identifiers or choosing from a list of existing

5

cases **114**. Each listed case includes summary information of the case, such as case numbers, patient names, relevant anatomy to be corrected, and dates that the case was created and/or last modified. The initial list of displayed existing cases **114** can, for example, display the most recent cases viewed by the user. For existing cases, the user also has the option to share the case with another user. For example, a physician can choose the case sharing option to send the case information to another physician using the software. This might be useful, for example, if the first physician wants advice from a second physician on the case, or if the first physician believes his case will be helpful to a second physician handling a similar case.

A user chooses the new case option for a patient whose information has not yet been entered into the software. When a user selects the new case option **112**, the user is brought to a case details screen **120** (FIG. 3). At the case details screen **120**, the user has the option of entering, viewing, or modifying patient details such as the patient's name, gender, race, date of birth, anatomy relevant to the case, and notes as the user sees fit. The user is presented with a preview of uploaded X-ray images **201** if any have been uploaded. If no X-ray images **201** have been uploaded, the user can upload images, for example one X-ray image of the deformed bone in the anterior-posterior ("AP") plane and one X-ray image of the deformed bone in the lateral plane. These X-ray images **201**, if used, may help the user to define the bone deformity, described more fully below. To upload an X-ray image **201**, the user chooses the "choose file" option and is able to import the desired image file from a memory device, such as a flash drive. Prior to uploading, the user also provides details relating to the image **201**, such as the view (e.g. lateral plane) in which the image was taken. Upon selecting the anatomy within the patient details (e.g. right or left side, femur, tibia, humerus, radius), the user can proceed to define the deformity.

A deformity definition screen **200** (FIGS. 4-5) within the application presents a model bone **202**. The deformity definition screen **200** includes an image manipulation function and a deformity definition function. In the image manipulation function, shown in FIG. 4, the application presents a model bone **202** overlaid on an X-ray image **201** previously uploaded on the case details screen **120**. The model bone **202** is chosen based on input of the relevant anatomy (e.g. left femur) from the case details screen **120**, and is initially presented in a non-deformed state. The model bone **202** is presented in the same plane which the user defined for the X-ray image **201**. The image manipulation function allows the user to manipulate the X-ray image **201**, for example by rotating, zooming, or repositioning the image. The user can also manipulate the transparency of the X-ray image **201** as desired for the best view of the model bone **202** in relation to the X-ray image. The user manipulates the X-ray image **201** with the goal of generally matching the size and position of the bone shown in the X-ray image with the model bone **202** displayed on the deformity definition screen **200**. The image **201** will often require manipulation because X-ray images, along with the actual bones being imaged, may vary in size and the initial position of the uploaded image may not align with the position of the model bone **202**. Once the user is satisfied that the model bone **202** generally corresponds to the size and position of the X-ray image **201**, the user can choose to begin the deformity definition function. This step can be repeated for each X-ray image **201**. For example, if a user uploads one X-ray image **201** in each of the AP and lateral planes, two separate image manipulations could be performed.

6

The deformity definition function of the deformity definition screen **200** is illustrated in FIG. 5. The manipulated X-ray image **201** is displayed in the background with the model bone **202** in the foreground. The model bone **202** includes a plurality of portions, including a proximal portion **204** and a distal portion **206**. A panel, to the right of the model bone **202** as illustrated in FIG. 5, allows the user to enter parameters relating to geometry of the deformed bone to modify the bone model **202** to approximate the actual bone deformity.

To start, the user designates a reference bone fragment, for example either the proximal portion **204** or distal portion **206**. The reference fragment, as depicted in FIG. 5, remains aligned with the vertical axis during manipulation of the bone model **202**, with the non-reference fragment changing position, with respect to the reference fragment, based on user input. The user inputs values for translation and angulation of the non-reference fragment with respect to the reference fragment in one or more of the AP plane, the lateral plane, and the axial plane. The user can input this data by, for example, entering the numerical value of translation (for example in millimeters) and/or angulation (for example in degrees) of the non-reference fragment with respect to the reference fragment in any of the desired planes. The user may alternately or additionally change the translation and/or angulation values using a slide-bar function in which sliding a bar in one direction decreases the value of the relevant parameter while sliding the bar in the other direction increases the value.

The user can also input a value for the deformity apex and the position of the osteotomy along the length of the model bone **202**. In FIG. 5, the deformity apex would be the point at which a line through the center of the proximal portion **204** intersects with the distal portion **206**. A change to the deformity apex value would be visualized in FIG. 5 as a change in the location at which the angled deformity of the proximal portion **204** is seen. The osteotomy plane, as seen in FIG. 5, is the location of the separation between the proximal portion **204** and the distal portion **206**.

As the user enters or changes the above-mentioned values, the graphical representation of the non-reference bone portion will change to reflect the new values. The change in the position of the model bone **202** against the backdrop of the X-ray image **201** allows the user to obtain visual confirmation that the parameters applied to the model bone accurately represent the parameters of the deformed bone. This step may be performed using more than one X-ray image **201**, or without any X-ray images, as described below. When using multiple X-ray images **201**, each can be viewed while setting values for the deformity. For example, a user may view the model bone **202** against an X-ray image **201** in the AP plane while setting values for the deformity in the AP plane, and then switch to a view of the model bone **202** against an X-ray image **201** in the lateral plane while setting values for the deformity in the lateral plane.

X-ray images **201** are not a necessary part of the deformity definition step. For example, in addition or as an alternative to using X-ray images **201**, the user can perform the same model bone **202** manipulations without the backdrop of an X-ray image. As seen in FIG. 6, a deformity definition screen **200** presents a model bone **202** simultaneously in the AP, lateral, and axial planes, as well as a perspective view. One or more combinations of views, including additional views not specifically identified herein, may be used and is largely a matter of design choice. Similar to the process described with reference to FIG. 5, the user enters values corresponding to the position and orientation

of the non-reference fragment (distal portion **206** as illustrated in FIG. **6**) in the AP plane, the lateral plane, and the axial plane. The user may also enter or change a value for the osteotomy plane and the deformity apex (deformity apex option not illustrated in FIG. **6**), largely in the same manner as described with reference to FIG. **5**.

For each value entered, the user can select a corresponding direction from a drop down menu. For example, the user can enter a 20 mm translation in the AP plane, and assign that value either a medial or lateral direction. Alternatively, the value can be a positive or negative value, with one direction assigned to positive values and the opposite direction assigned to negative values (e.g. 20 mm corresponding to a 20 mm medial translation and -20 mm corresponding to a 20 mm lateral translation). Similar to the method described with reference to FIG. **5**, the user can directly enter numerical values based on experience or previous measurements of the bone deformity. In addition or as an alternative, the user can change the position of the slide bars corresponding to the non-reference bone parameters to reach a final value that accurately represents the geometry of the bone deformity.

Once the user is satisfied that the model bone **202** is an accurate representation of the deformed bone, the user can proceed to the first ring configuration screen **250** (FIG. **7**). The first ring configuration screen **250** allows the user to choose to continue in either a pre-operative (“pre-op”) or post-operative (“post-op”) mode by selecting the respective radio button. Generally speaking, the pre-op mode is used prior to the surgical fixation of the limb reconstruction device to the deformed bone. The post-op mode is to be used after the limb reconstruction device, with associated rings and struts, has already been affixed to the patient. In a single case, the pre-op mode can be used alone, the post-op mode can be used alone, or each mode can be used prior to and following surgery, respectively.

If the pre-op mode is selected, the user can continue to a pre-op ring configuration screen **300** (FIG. **8**). At this point, the user inputs the size of the desired rings, including a reference ring **305** and a moving ring **310**. For example, a user may be able to choose between a 155 mm, 180 mm, or 210 mm ring. The user may also be able to choose the type of ring, such as a full ring or partial ring. Different types of rings are known in the art and the inclusion of different rings as options in the software is largely a matter of design choice.

The rings **305**, **310** are displayed along with the model bone **202** on the screen, preferably in an AP view, a lateral view, and an axial view. Additional views, such as a perspective view, may be included. The position and orientation of the proximal portion **204** and distal portion **206** of the model bone **202** are based on the input received during the deformity definition stage.

Once a size and/or type of ring is selected for the reference ring, it is displayed perpendicular to the reference fragment with a longitudinal axis of the reference fragment extending through the center of the reference ring. Similarly, once a size and/or type of ring is selected for the moving ring, it is displayed perpendicular to the non-reference fragment with a longitudinal axis of the non-reference fragment extending through the center of the moving ring. Similar to the deformity definition screen **200**, the user enters position and orientation values for the reference ring **305** and the moving ring **310**. For the rings, the user can directly enter the values, or move a slide-bar corresponding to the values to scroll through a range of values. Because this is the pre-op mode and no fixation device has yet been attached to the patient, the user chooses the ring sizes,

positions and orientations that he believes will be effective for the correction based, for example, on his experience and knowledge.

As the values are entered, or as the slide-bar is moved, the graphical representation of the rings changes to reflect the new values. For the reference ring **305**, the position values include an AP offset, a lateral offset, an axial offset, and an axial angle. The moving ring **310** includes these values, and additionally includes an AP angle and a lateral angle. Once the user is satisfied that the reference ring **305** and moving ring **310** are at locations on the model bone **202** representative of where the actual rings should be located on the patient’s deformed bone, the user can proceed to the first strut configuration screen **350**. The software may also provide maximum and minimum values for the placement and orientation of the reference ring **305** and moving ring **310**, based, for example, on the feasibility of actually achieving those values in the operating room. These limits would help ensure that the planned values for the ring positions input by the user can likely be achieved during surgery on the patient.

The first strut configuration screen **350** allows the user to initiate an automatic calculation of possible strut combinations to connect the reference ring **305** to the moving ring **310** (FIG. **9**). Once the calculation is complete, the user is presented with the second strut configuration screen **400** that displays at least one possible strut combination, and preferably more than one (FIG. **10**). The possible strut combinations **410** are presented in a table with a description of each strut in a particular combination. The model bone **202** and rings **305**, **310** are displayed in an AP view, a lateral view and an axial view. Additional views such as a perspective view can be included. This display is based on values entered during the deformity definition and ring configuration steps. Additionally, the struts of a particular strut combination **410** are displayed with the model bone and rings. Each strut combination **410** includes a radio button.

When the user selects a particular strut combination **410** by selecting the corresponding radio button, the views update to show that particular strut combination. An optimal strut combination is highlighted among all the strut combinations **410** to suggest to the user a particularly desired combination to select. Among other factors, the optimal strut combination is based on the combination that will require the least amount of strut change-outs during the correction procedure.

In the pre-op mode, when a particular strut combination **410** is selected, the orientation of each strut, including strut length, strut angle, and base angle are displayed. After selecting a desired strut combination **410**, the user may optionally choose to over-constrain the frame with an additional strut (not illustrated). For example, in a fixation frame that uses three struts, a fourth strut could be added after the three-strut combination is chosen. The fourth strut over-constrains the fixation frame by increasing stiffness and reducing play in the frame. The length, angle, and position of the additional strut is provided by the software once the option for the additional strut is chosen. Once the desired strut combination is chosen, and any additional desired struts are chosen, the user proceeds to a limiting anatomical factor (“LAF”) input screen **450**.

The LAF input screen **450** (FIG. **11**) allows a user to input a limiting anatomical factor (“LAF”) and to calculate a minimum amount of time for the correction process. The LAF is a reference point to which a user may want to set a maximum distraction rate or velocity. For example, if a segment of bone is to be moved a certain distance per day, surrounding tissues will also be moved at the same rate. A

user may want to limit the maximum rate of movement of the bone or surrounding structures. Also of note is that certain portions of bone or tissue may move at different rates for a given distraction rate. For example, as the distal portion of the bone illustrated in FIG. 11 is moved to align with the proximal portion, a point on the distal portion of the bone near the osteotomy plane will move at a lower rate than a point on the distal portion of the bone farther away from the osteotomy plane because of the angle.

The LAF may default to the center of a moving fragment, one of the ends of the moving fragment, or anywhere else desired. If a user desires a LAF location other than the default location, he may change the location by changing the AP, lateral, and axial offset values. By setting the LAF location, a user may ensure that the maximum distraction rate applies to the LAF location. This ensures that a user may limit the maximum rate of movement of a particular portion of the bone or surrounding tissue.

The views of the second strut configuration screen 400 are displayed again on the LAF input screen 450. The user inputs LAF values for the AP offset, lateral offset, and axial offset. The user also enters a maximum distraction rate. The maximum distraction rate may represent, for example, the maximum, safest, or optimal amount of millimeters that the length of a strut can increase or decrease in a day, for example 1 mm/day. Based on the LAF input values and the maximum distraction rate, the user can initiate a calculation of the minimum amount of time it will take for the correction of the patient's deformed bone using the ring and strut configurations chosen in the previous steps. If the user is satisfied with the minimum correction time, he can move to the next step of generating the correction plan. If he is not satisfied, he can override the minimum correction time and enter a different value, and then continue to the step of generating the correction plan. For example, if the minimum correction time is initially output as 10 days, but the user (or a separate physician) will not be able to see the patient again for 14 days, the user may override the correction time to a value of 14 days. This potentially would allow for a more gradual correction plan, which may be of benefit.

On the correction plan generation screen 500, the user enters the date on which the user or patient will begin adjusting the fixation mechanism according to the correction plan. Once entered, the user commands the computer to generate a correction plan (FIG. 12). Once calculated, the correction plan 510 is displayed on the correction screen 500 (FIG. 13). The correction plan 510 can include, for example, the position and angle of each strut of the limb reconstruction frame for each day of the correction, along with the date and day number (e.g. first day, second day) of the correction plan 510. When the software is determining the deformity correction plan 510, an angle of the proximal portion of the bone 204 in relation to the distal portion of the bone 206 is utilized. In one embodiment, the software determines this angle based on the midlines of the two bone portions 204, 206. In another embodiment, the user may input the diameter of the bone. With the value of diameter of the bone, the software can base the value of the angle on the edges of the two bone portions 204, 206. Utilizing the outer edges of the bone portions 204, 206 provides for a more accurate correction plan 510.

The correction plan 510 may call for changing out one strut for a strut of a larger or smaller size during the period of correction. For example, a relatively small strut may be initially utilized between the reference and moving rings 305, 310. The angle and length of that small strut will be adjusted over time. At a certain point, the correction plan

510 may require to replace the relatively small strut with a relatively large strut because, for example, the relatively small strut may be close to reaching its maximum length. The software may provide an option to the user of changing out the relatively small strut or, instead, changing the point of connection of the small strut. By changing the point or points of connection of a strut, which may for example be a hole in the reference ring 305 and a hole in the moving ring 310, the length and position limitations of the strut are overcome. This option may be provided alternatively to changing out struts, or in addition, and applies to struts of all sizes.

The correction plan 510 may also show a relationship between positions of the struts and discrete user or patient actions. For example, if the correction plan 510 calls for a strut to be lengthened by 1 millimeter on the first day, the correction plan may indicate that the user or patient should increase the length of that strut four separate times, for example by 0.25 millimeters in the morning, 0.25 millimeters at noon, 0.25 millimeters in the evening and another 0.25 millimeters at night. Besides use as an instructional tool, the correction plan 510 also aids a physician or surgeon in monitoring the progress of the correction of the bone deformity, for example by checking at periodic intervals that the struts of the fixation frame are in the proper position as called for by the correction plan. In addition to the correction plan 510, the correction screen 500 may also include a simulation 520 of the correction. The user can view the simulation 520 to see what the progress of the correction should look like. The simulation 520 allows the user to see what the model bone, rings and struts will look like in one or more views on each day of the correction plan 510 or with each discrete correction made as called for by the correction plan. This helps the user ensure that the correction plan 510 is appropriate for the given case, and further aids the user in determining whether the correction is progressing according to the plan, as the user can compare the model of what the bone and fixation frame should look like to what the actual patient's bone and fixation frame look like on a given day.

The user has the option of viewing a report of all the pertinent details of a specific case in the study report screen 600, as seen in FIG. 14. The study report screen 600 is a comprehensive report of the case and can include information such as case details, parameter of the deformity definition, parameters for LAF input, parameters for the ring and strut configurations, and the complete correction plan, including the schedule of when one or more struts should be changed out for different struts, if necessary.

The user also has access to a smart tool screen 700 (FIG. 15). The smart tool screen 700 displays pertinent data for a smart tool plan and provides an option to launch a smart tool client to interface with the smart tool. Generally, a smart tool may be used in conjunction with the fixation frame and software described herein. The smart tool is an automated tool that a patient uses to perform the corrective fixation frame alterations described by the correction plan. For example, the smart tool may include a motor that rotates screws of the struts to lengthen the struts based on instructions downloaded from the software, which instructions correspond to the correction plan. The smart tool may limit the patient from changing the lengths of the struts in a manner other than described in the correction plan, and may communicate with the software and the case physician to reflect whether, or to what degree, the patient is following the correction plan. Such a smart tool is described more fully in U.S. application Ser. No. 13/167,101, the entire contents of which are hereby incorporated by reference herein.

11

As mentioned above, the application can be used in a post-op mode in addition or as an alternative to the pre-op mode. This mode can be used once the patient has already undergone surgery to attach the fixation frame to the deformed bone. The post-op mode can be used as an alternative to the pre-op mode, for example in cases in which time is limited and surgery must be performed without the benefit of the planning provided in the pre-op mode described above. Contrariwise, the pre-op mode may be especially useful in cases of a congenital deformity or in cases in which a deformity is stable, where planning time is available without risking the health of the patient.

In practice, the post-op mode should always be used, regardless of whether the pre-op mode is used. Even if the pre-op mode is used, a surgeon is likely to desire to confirm that the placement of rings and struts in surgery actually matches the pre-op plan.

In the post-op mode, the steps described above with reference to the login screen 102, home page 110, case details screen 120, and deformity definition screen 200 are the same (FIGS. 1-6). At the first ring configuration screen 250, the user would choose the post-op mode instead of the pre-op mode to continue in the post-op mode.

After selecting the post-op mode, the user proceeds to a post-op ring configuration screen 800 (FIG. 16). The user enters the sizes of both the reference ring 305 and the moving ring 310 as well as the position and orientation of the reference ring. These values are based on the sizes of the actual rings attached to the patient. For the reference ring, the user follows the same procedure as described for the pre-op ring configuration screen 300, either entering ring position parameters directly into the corresponding fields or by using a slide-bar to set the values. The user does not need to enter the positions of the moving ring 310, however, as this will be automatically determined during the next stages based on the positions of the struts.

The user proceeds to the post-op strut configuration screen 900 to enter the sizes and positions of the struts of the fixation frame attached to the patient during surgery (FIG. 17). The input values include the size of each strut, for example small, standard or large, the length of each strut, and the angle of each strut. The user may also have the option to enter the position along each ring that each strut is attached. For example, each ring may contain a plurality of through-holes for attachment of the struts, and each strut may be attached at recommended, predetermined through holes. These recommended, predetermined attachment through-holes are set as default attachment points in the software. However, if the physician is unable or unwilling to use the recommended, predetermined locations for strut attachment, he has the option to attach the struts at any other through-hole and the software can accommodate this modification. Again, these values are based on the actual fixation frame attached to the patient during surgery. Once the values are entered, the user can command the application to calculate the position of the moving ring 310 based on the position of the reference ring 305 and the struts. Once the moving ring position is calculated, the position and orientation fields for the AP, lateral, and axial axes populate with the calculated values. Once calculated, views of the fixation frame and model bone, for example AP, lateral, axial and perspective views, also update to reflect the calculated position and orientation of the moving ring. This provides additional confirmation to the surgeon that the model accurately represents the fixation frame of the case.

The remainder of the process from this point is the same as described with reference to the pre-op mode (FIGS.

12

12-15). The user advances through the LAF input screen 450 by entering LAF values and a minimum distraction rate to calculate a minimum correction time. The user then proceeds to the correction plan generation screen 500, enters the start date for the correction, and initiates the generation of a correction plan. As described with reference to the pre-op mode, the correction plan 510 and a simulation are then displayed on the correction screen 500. Again, the user can view a study report in the study report screen 600 and interface with a smart tool using the smart tool screen 700.

While the above description relates to creating a new case, the user may choose to open an existing case in the software if one exists. After logging in, the user can choose an existing case by, for example, searching for an existing case or choosing from a list of existing cases 114 (FIG. 2). After opening the existing case, the user is taken to an existing case screen 1000 (FIG. 18) that shows patient details and the existing cases for the patient. From this screen, the user can choose to open a particular study of interest, create a new case for the patient, or view a report of the case summary. If the user opens the existing case, he is able to review or edit the case using the procedures described above.

Although the invention herein has been described with reference to particular embodiments, it is to be understood that these embodiments are merely illustrative of the principles and applications of the present invention. It is therefore to be understood that numerous modifications may be made to the illustrative embodiments and that other arrangements may be devised without departing from the spirit and scope of the present invention as defined by the appended claims.

The invention claimed is:

1. A method of generating a correction plan for correcting a deformed bone of a patient comprising the steps of:

prior to generating the correction plan, coupling an external fixation frame to the patient so that a proximal fixation ring is coupled to a proximal fragment of the deformed bone, a distal fixation ring is coupled to a distal fragment of the deformed bone, and a plurality of struts couple the proximal fixation ring to the distal fixation ring;

displaying on a display device a first x-ray image of the deformed bone in a first plane;

overlying a first proximal reference axis with a first central axis of the proximal fragment of the deformed bone in the first x-ray image on the display device so that the first proximal reference axis is substantially aligned with the first central axis of the proximal fragment in the first x-ray image;

overlying a first distal reference axis with a first central axis of the distal fragment of the deformed bone in the first x-ray image on the display device so that the first distal reference axis is substantially aligned with the first central axis of the distal fragment in the first x-ray image;

displaying on the display device a second x-ray image of the deformed bone in a second plane;

overlying a second proximal reference axis with a second central axis of the proximal fragment of the deformed bone in the second x-ray image on the display device so that the second proximal reference axis is substantially aligned with the second central axis of the proximal fragment in the second x-ray image;

overlying a second distal reference axis with a second central axis of the distal fragment of the deformed bone in the second x-ray image on the display device so that

13

the second distal reference axis is substantially aligned with the second central axis of the distal fragment in the second x-ray image;
 displaying a model of the proximal fixation ring on the display device;
 manipulating the model of the proximal fixation ring so that a position and orientation of the model of the proximal fixation ring corresponds to a position and orientation of the proximal fixation ring coupled to the proximal fragment of the deformed bone; and
 generating the correction plan based at least in part on the first and second proximal reference axes, the first and second distal reference axes, and the position and orientation of the model proximal fixation ring,
 wherein the first proximal reference axis and the second proximal reference axis are initially presented together as part of a model bone in a non-deformed state which is later deformed by repositioning a distal portion of the model bone relative to a proximal portion of the model bone.

2. The method of claim 1, wherein the first proximal reference axis is a longitudinal axis of the proximal portion of the model bone in the first plane.

3. The method of claim 2, wherein the second proximal reference axis is the longitudinal axis of the proximal portion of the model bone in the second plane.

4. The method of claim 1, wherein the first distal reference axis is a longitudinal axis of the distal portion of the model bone in the first plane.

5. The method of claim 4, wherein the second distal reference axis is the longitudinal axis of the distal portion of the model bone in the second plane.

6. The method of claim 1, wherein overlying the first proximal reference axis with the first central axis of the proximal fragment of the deformed bone in the first x-ray image on the display device is performed prior to overlying the second proximal reference axis with the second central axis of the proximal fragment of the deformed bone in the second x-ray image on the display device.

7. The method of claim 1, wherein overlying the first distal reference axis with the first central axis of the distal fragment of the deformed bone in the first x-ray image on the display device is performed prior to overlying the second distal reference axis with the second central axis of the distal fragment of the deformed bone in the second x-ray image on the display device.

8. The method of claim 1, wherein the first central axis of the proximal fragment in the first x-ray image on the display device is maintained parallel to a vertical axis of the display device.

14

9. The method of claim 8, wherein the second central axis of the proximal fragment in the second x-ray image on the display device is maintained parallel to the vertical axis of the display device.

10. The method of claim 1, wherein the step of overlying the proximal reference axis with the axis of the proximal fragment of the deformed bone in the first image on the display device so that the proximal reference axis is substantially aligned with the axis of the proximal fragment includes at least one of entering numerical values into an input box and moving a slide-bar corresponding to numerical values.

11. The method of claim 1, wherein overlying the first proximal reference axis with the first central axis of the proximal fragment of the deformed bone in the first image on the display device includes adjusting at least one of a translation and an angulation of the first proximal reference axis with respect to the first distal reference axis.

12. The method of claim 1, wherein the step of manipulating the model of the proximal fixation ring further includes changing at least one of the position and orientation of the model of the proximal fixation ring with respect to a model of the deformed bone displayed on the display device.

13. The method of claim 1, further comprising:
 displaying a model of a distal fixation ring on the display device; and
 manipulating the model of the distal fixation ring so that a position and orientation of the model of the distal fixation ring corresponds to a position and orientation of the distal fixation ring coupled to the distal fragment of the deformed bone.

14. The method of claim 13, further comprising:
 displaying on the display device a plurality of model struts, each of the model struts having at least one of a size, position, and orientation corresponding to one of the plurality of struts coupling the proximal fixation ring to the distal fixation ring.

15. The method of claim 1, wherein overlying the first proximal reference axis with the first central axis of the proximal fragment of the deformed bone in the first image on the display device is performed manually by a user via an input device.

16. The method of claim 1, wherein the first plane of the first x-ray image is an anterior-posterior (“AP”) plane.

17. The method of claim 16, wherein the second plane of the second x-ray image is a lateral plane.

* * * * *