



(19) **United States**

(12) **Patent Application Publication**  
**Lee et al.**

(10) **Pub. No.: US 2017/0014646 A1**  
(43) **Pub. Date: Jan. 19, 2017**

(54) **GUIDED CHARGED PARTICLE  
IMAGING/TREATMENT APPARATUS AND  
METHOD OF USE THEREOF**

(71) Applicants: **W. Davis Lee**, Newburyport, MA (US);  
**Mark R. Amato**, South Hamilton, MA  
(US); **Susan L. Michaud**, Brewster,  
MA (US)

(72) Inventors: **W. Davis Lee**, Newburyport, MA (US);  
**Mark R. Amato**, South Hamilton, MA  
(US); **Susan L. Michaud**, Brewster,  
MA (US)

(21) Appl. No.: **15/281,031**

(22) Filed: **Sep. 29, 2016**

**Related U.S. Application Data**

(63) Continuation-in-part of application No. 15/167,617, filed on May 27, 2016, which is a continuation-in-part of application No. 15/152,479, filed on May 11, 2016, which is a continuation-in-part of application No. 14/216,788, filed on Mar. 17, 2014, which is a continuation-in-part of application No. 13/087,096, filed on Apr. 14, 2011, now Pat. No. 9,044,600, said application No. 14/216,788 is a continuation-in-part of application No. 13/788,890, filed on Mar. 7, 2013, now Pat. No. 8,907,309, said application No. 15/152,479 is a continuation-in-part of application No. 14/952,817, filed on Nov. 25, 2015, which is a continuation-in-part of application No. 14/293,861, filed on Jun. 2, 2014, now Pat. No. 9,498,649, which is a continuation-in-part of application No. 12/985,039, filed on Jan. 5, 2011, now Pat. No. 8,598,543, said application No. 15/152,479 is a continuation-in-part of application No. 14/860,577, filed on Sep. 21, 2015, which is a continuation of application No. 14/223,289, filed on Mar. 24, 2014, now Pat. No. 9,177,751, which is a continuation-in-part of application No. 14/216,788, filed on Mar. 17, 2014, which is a continuation-in-part of application No. 12/985,039, filed

on Jan. 5, 2011, now Pat. No. 8,598,543, said application No. 15/152,479 is a continuation-in-part of application No. 15/073,471, filed on Mar. 17, 2016, said application No. 15/167,617 is a continuation-in-part of application No. 14/860,577, filed on Sep. 21, 2015, which is a continuation of application No. 14/223,289, filed on Mar. 24, 2014, now Pat. No. 9,177,751, which is a continuation-in-part of application No. 14/216,788, filed on Mar. 17, 2014, which is a continuation-in-part of application No. 13/572,542, filed on Aug. 10, 2012, now Pat. No. 9,056,199, which is a continuation-in-part of application No. 12/425,683, filed on Apr. 17, 2009, now Pat. No. 7,939,809.


(60) Provisional application No. 61/324,776, filed on Apr. 16, 2010, provisional application No. 61/324,776, filed on Apr. 16, 2010, provisional application No. 61/324,776, filed on Apr. 16, 2010, provisional application No. 62/304,839, filed on Mar. 7, 2016, provisional application No. 61/055,395, filed on May 22, 2008.

**Publication Classification**

(51) **Int. Cl.**  
*A61N 5/10* (2006.01)  
(52) **U.S. Cl.**  
CPC ..... *A61N 5/1065* (2013.01); *A61N 5/1039*  
(2013.01); *A61N 5/1077* (2013.01); *A61N*  
*2005/1087* (2013.01)

(57) **ABSTRACT**

The invention comprises a method and apparatus for tracking and/or imaging impact of a particle beam treating a tumor using one or more imaging systems positionable about the tumor, such as a positron emission tracking and/or imaging system, where resulting tracking/imaging data: dynamically determines a treatment beam position, tracks a history of treatment beam positions, guides the treatment beam, and/or images a tumor before, during, and/or after treatment with the charged particle beam.






FIG. 1A




FIG. 1B










FIG. 1C







**FIG. 4A**



**FIG. 4B**




FIG. 5




FIG. 6A




FIG. 6B




FIG. 7




FIG. 8




FIG. 9A




FIG. 9B




FIG. 10




FIG. 11




FIG. 12A




FIG. 12B






FIG. 13A








FIG. 16A




FIG. 16C




FIG. 16B




FIG. 17A




FIG. 17B




FIG. 17C




FIG. 17D




FIG. 18




FIG. 19A




FIG. 19B




FIG. 19C




FIG. 20




FIG. 21

**GUIDED CHARGED PARTICLE  
IMAGING/TREATMENT APPARATUS AND  
METHOD OF USE THEREOF**

CROSS-REFERENCES TO RELATED  
APPLICATIONS

**[0001]** This application is a continuation-in-part of U.S. patent application Ser. No. 15/167,617 filed May 27, 2016, which is:

**[0002]** a continuation-in-part of U.S. patent application Ser. No. 15/152,479 filed May 11, 2016, which:

**[0003]** is a continuation-in-part of U.S. patent application Ser. No. 14/216,788 filed Mar. 17, 2014,

**[0004]** which is a continuation-in-part of U.S. patent application Ser. No. 13/087,096 filed Apr. 14, 2011, which claims benefit of U.S. provisional patent application No. 61/324,776 filed Apr. 16, 2010; and

**[0005]** is a continuation-in-part of U.S. patent application Ser. No. 13/788,890 filed Mar. 7, 2013;

**[0006]** is a continuation-in-part of U.S. patent application Ser. No. 14/952,817 filed Nov. 25, 2015, which is a continuation-in-part of U.S. patent application Ser. No. 14/293,861 filed Jun. 2, 2014, which is a continuation-in-part of U.S. patent application Ser. No. 12/985,039 filed Jan. 5, 2011, which claims the benefit of U.S. provisional patent application No. 61/324,776, filed Apr. 16, 2010;

**[0007]** is a continuation-in-part of U.S. patent application Ser. No. 14/860,577 filed Sep. 21, 2015, which is a continuation of U.S. patent application Ser. No. 14/223,289 filed Mar. 24, 2014, which is a continuation-in-part of U.S. patent application Ser. No. 14/216,788 filed Mar. 17, 2014, which is a continuation-in-part of U.S. patent application Ser. No. 12/985,039 filed Jan. 5, 2011, which claims the benefit of U.S. provisional patent application No. 61/324,776, filed Apr. 16, 2010; and

**[0008]** is a continuation-in-part U.S. patent application Ser. No. 15/073,471 filed Mar. 17, 2016, which claims benefit of U.S. provisional patent application No. 62/304,839 filed Mar. 7, 2016,

**[0009]** is a continuation-in-part of U.S. patent application Ser. No. 14/860,577 filed Sep. 21, 2015, which is a continuation of U.S. patent application Ser. No. 14/223,289 filed Mar. 24, 2014, which is a continuation-in-part of U.S. patent application Ser. No. 14/216,788 filed Mar. 17, 2014, which is a continuation-in-part of U.S. patent application Ser. No. 13/572,542 filed Aug. 10, 2012, which is a continuation-in-part of U.S. patent application Ser. No. 12/425,683 filed Apr. 17, 2009, which claims the benefit of U.S. provisional patent application No. 61/055,395 filed May 22, 2008, now U.S. Pat. No. 7,939,809 B2;

**[0010]** all of which are incorporated herein in their entirety by this reference thereto.

BACKGROUND OF THE INVENTION

**[0011]** Field of the Invention

**[0012]** The invention relates generally to imaging and treating a tumor.

**[0013]** Discussion of the Prior Art

Cancer Treatment

**[0014]** Proton therapy works by aiming energetic ionizing particles, such as protons accelerated with a particle accelerator, onto a target tumor. These particles damage the DNA of cells, ultimately causing their death. Cancerous cells, because of their high rate of division and their reduced ability to repair damaged DNA, are particularly vulnerable to attack on their DNA.

**[0015]** Patents related to the current invention are summarized here.

Proton Beam Therapy System

**[0016]** F. Cole, et. al. of Loma Linda University Medical Center “Multi-Station Proton Beam Therapy System”, U.S. Pat. No. 4,870,287 (Sep. 26, 1989) describe a proton beam therapy system for selectively generating and transporting proton beams from a single proton source and accelerator to a selected treatment room of a plurality of patient treatment rooms.

Imaging

**[0017]** P. Adamee, et. al. “Charged Particle Beam Apparatus and Method for Operating the Same”, U.S. Pat. No. 7,274,018 (Sep. 25, 2007) and P. Adamee, et. al. “Charged Particle Beam Apparatus and Method for Operating the Same”, U.S. Pat. No. 7,045,781 (May 16, 2006) describe a charged particle beam apparatus configured for serial and/or parallel imaging of an object.

**[0018]** K. Hiramoto, et. al. “Ion Beam Therapy System and its Couch Positioning System”, U.S. Pat. No. 7,193,227 (Mar. 20, 2007) describe an ion beam therapy system having an X-ray imaging system moving in conjunction with a rotating gantry.

**[0019]** C. Maurer, et. al. “Apparatus and Method for Registration of Images to Physical Space Using a Weighted Combination of Points and Surfaces”, U.S. Pat. No. 6,560,354 (May 6, 2003) described a process of X-ray computed tomography registered to physical measurements taken on the patient’s body, where different body parts are given different weights. Weights are used in an iterative registration process to determine a rigid body transformation process, where the transformation function is used to assist surgical or stereotactic procedures.

**[0020]** M. Blair, et. al. “Proton Beam Digital Imaging System”, U.S. Pat. No. 5,825,845 (Oct. 20, 1998) describe a proton beam digital imaging system having an X-ray source that is movable into a treatment beam line that can produce an X-ray beam through a region of the body. By comparison of the relative positions of the center of the beam in the patient orientation image and the isocentre in the master prescription image with respect to selected monuments, the amount and direction of movement of the patient to make the best beam center correspond to the target isocentre is determined.

**[0021]** S. Nishihara, et. al. “Therapeutic Apparatus”, U.S. Pat. No. 5,039,867 (Aug. 13, 1991) describe a method and apparatus for positioning a therapeutic beam in which a first distance is determined on the basis of a first image, a second distance is determined on the basis of a second image, and the patient is moved to a therapy beam irradiation position on the basis of the first and second distances.

### Problem

[0022] There exists in the art of charged particle cancer therapy a need for confirmation of beam treatment position.

### SUMMARY OF THE INVENTION

[0023] The invention comprises a hybrid imaging/charged particle tumor treatment apparatus and method of use thereof.

### DESCRIPTION OF THE FIGURES

[0024] A more complete understanding of the present invention is derived by referring to the detailed description and claims when considered in connection with the Figures, wherein like reference numbers refer to similar items throughout the Figures.

[0025] FIG. 1A and FIG. 1B illustrate component connections of a charged particle beam therapy system;

[0026] FIG. 1C illustrates a charged particle therapy system;

[0027] FIG. 2A and FIG. 2B illustrate a diode extraction system in standby and functional mode; FIG. 2C and FIG. 2D illustrate a triode in standby and operational mode, respectively;

[0028] FIG. 3 illustrates a method of multi-axis charged particle beam irradiation control;

[0029] FIG. 4A and FIG. 4B illustrate a top view of a beam control tray and a side view of the beam control tray, respectively.

[0030] FIG. 5 illustrates patient specific tray inserts for insertion into the beam control tray;

[0031] FIG. 6A illustrates insertion of the individualized tray assembly into the beam path and FIG. 6B illustrates retraction of the tray assembly into a nozzle of the charged particle cancer therapy system;

[0032] FIG. 7 illustrates a tomography system;

[0033] FIG. 8 illustrates a beam path identification system;

[0034] FIG. 9A illustrates a beam path identification system coupled to a beam transport system and a tomography scintillation detector and FIG. 9B illustrates the scintillation detector rotating with the patient and gantry nozzle;

[0035] FIG. 10 illustrates a treatment delivery control system;

[0036] FIG. 11 illustrates beam state determination systems;

[0037] FIG. 12A and FIG. 12B illustrate control of a patient interface system with a pendant and work-flow control system, respectively;

[0038] FIG. 13A illustrates a two-dimensional-two-dimensional imaging system relative to a cancer treatment beam, FIG. 13B illustrates multiple gantry supported imaging systems, and FIG. 13C illustrates a rotatable cone beam

[0039] FIG. 14A illustrates a scintillation material coupled to a detector array, FIG. 14B illustrates a fiber optic array in a tomography system; FIG. 14C and FIG. 14D illustrate end views of the fiber optic array; and FIG. 14E illustrates a micro-optic array coupled to the scintillation material;

[0040] FIG. 15 illustrates use of multiple layers of scintillation materials;

[0041] FIG. 16A illustrates an array of scintillation optics; FIG. 16B illustrates a scintillating fiber optic; and FIG. 16C illustrates an x-, y-, z-axes array of scintillation optics or scintillation materials;

[0042] FIG. 17A illustrates a scintillation material; FIG. 17B illustrates detector arrays orthogonally coupled to the scintillation material; and FIG. 17C and FIG. 17D illustrate multiple detector arrays coupled to the scintillation material;

[0043] FIG. 18 illustrates subsystems of an imaging system;

[0044] FIG. 19A illustrates a hybrid gantry-imaging system; FIG. 19B illustrates a secondary rotation system, of the gantry, used for imaging; and FIG. 19C illustrates a linearly translatable imaging system of the gantry;

[0045] FIG. 20 illustrates a dynamic charged particle beam positioning system; and

[0046] FIG. 21 illustrates a treatment beam depth of penetration tracking system.

[0047] Elements and steps in the figures are illustrated for simplicity and clarity and have not necessarily been rendered according to any particular sequence. For example, steps that are performed concurrently or in different order are illustrated in the figures to help improve understanding of embodiments of the present invention.

### DETAILED DESCRIPTION OF THE INVENTION

[0048] The invention relates generally to imaging a tumor using one or more imaging systems positionable about the tumor, where resulting image data guides a charged particle beam treatment of the tumor.

[0049] In one embodiment, an imaging system, such as a positron emission tracking system is used to: dynamically determine a treatment beam position, track a history of treatment beam positions, guide the treatment beam, and/or image a tumor before, during, and/or after treatment with the charged particle beam.

[0050] In another embodiment, an imaging system translating on a linear path past a patient operates alternatingly with and/or during a gantry rotating a treatment beam around the patient. More particularly, a method for both imaging a tumor and treating the tumor of a patient using positively charged particles includes the steps of: (1) rotating a gantry support and/or gantry, connected to at least a portion of a beam transport system configured to pass a charged particle treatment beam, circumferentially about the patient and a gantry rotation axis; (2) translating a translatable imaging system past the patient on a path parallel to an axis perpendicular to the gantry rotation axis; (3) imaging the tumor using the translatable imaging system; and (4) treating the tumor using the treatment beam.

[0051] In another embodiment a method for imaging and treating a tumor of a patient with positively charged particles, comprises the steps of: (1) using a rotatable gantry support to support and rotate a section of a positively charged particle beam transport line about a rotation axis and a tumor of a patient; (2) using a rotatable and optionally extendable secondary support to support, circumferentially position, and laterally position a primary and optional secondary imaging system about the tumor; (3) image the tumor using the primary and optional secondary imaging system as a function of rotation and/or translation of the secondary support; and (4) treat, optionally concurrently, the tumor using the positively charged particles as a function of circumferential position of the section of the charged particle beam about the tumor.

[0052] In another embodiment a method and apparatus for imaging a tumor of a patient using positively charged

particles and X-rays, comprises the steps of: (1) transporting the positively charged particles from an accelerator to a patient position using a beam transport line, where the beam transport line comprises a positively charged particle beam path and an X-ray beam path; (2) detecting scintillation induced by the positively charged particles using a scintillation detector system; (3) detecting X-rays using an X-ray detector system; (4) positioning a mounting rail through linear extension/retraction to: at a first time and at a first extension position of the mounting rail, position the scintillation detector system opposite the patient position from the exit nozzle and at a second time and at a second extension position of the mounting rail, position the X-ray detector system opposite the patient position from the exit nozzle; (5) generating an image of the tumor using output of the scintillation detector system and the X-ray detector system; and (6) alternating between the step of detecting scintillation and treating the tumor via irradiation of the tumor using the positively charged particles.

**[0053]** In another embodiment, a method or apparatus for tomographically imaging a sample, such as a tumor of a patient, using positively charged particles is described. Position, energy, and/or vectors of the positively charged particles are determined using a plurality of scintillators, such as layers of chemically distinct scintillators where each chemically distinct scintillator emits photons of differing wavelengths upon energy transfer from the positively charged particles. Knowledge of position of a given scintillator type and a color of the emitted photon from the scintillator type allows a determination of residual energy of the charged particle energy in a scintillator detector. Optionally, a two-dimensional detector array additionally yields x/y-plane information, coupled with the z-axis energy information, about state of the positively charged particles. State of the positively charged particles as a function of relative sample/particle beam rotation is used in tomographic reconstruction of an image of the sample or the tumor.

**[0054]** In another example, a method or apparatus for tomographic imaging of a tumor of a patient using positively charged particles respectively positions a plurality of two-dimensional detector arrays on multiple surfaces of a scintillation material or scintillator. For instance, a first two-dimensional detector array is optically coupled to a first side or surface of a scintillation material, a second two-dimensional detector array is optically coupled to a second side of the scintillation material, and a third two-dimensional detector array is optically coupled to a third side of the scintillation material. Secondary photons emitted from the scintillation material, resultant from energy transfer from the positively charged particles, are detected by the plurality of two-dimensional detector arrays, where each detector array images the scintillation material. Combining signals from the plurality of two-dimensional detector arrays, the path, position, energy, and/or state of the positively charged particle beam as a function of time and/or rotation of the patient relative to the positively charged particle beam is determined and used in tomographic reconstruction of an image of the tumor in the patient or a sample. Particularly, a probabilistic pathway of the positively charged particles through the sample, which is altered by sample constituents, is constrained, which yields a higher resolution, a more accurate and/or a more precise image.

**[0055]** In another example, a scintillation material is longitudinally packaged in a circumferentially surrounding

sheath, where the sheath has a lower index of refraction than the scintillation material. The scintillation material yields emitted secondary photons upon passage of a charged particle beam, such as a positively charged residual particle beam having transmitted through a sample. The internally generated secondary photons within the sheath are guided to a detector element by the difference in index of refraction between the sheath and the scintillation material, similar to a light pipe or fiber optic. The coated scintillation material or fiber is referred to herein as a scintillation optic. Multiple scintillation optics are assembled to form a two-dimensional scintillation array. The scintillation array is optionally and preferably coupled to a detector or two-dimensional detector array, such as via a coupling optic, an array of focusing optics, and/or a color filter array.

**[0056]** In another embodiment, an ion source is coupled to the apparatus. The ion source extraction system facilitates on demand extraction of charged particles at relatively low voltage levels and from a stable ion source. For example, a triode extraction system allows extraction of charged particles, such as protons, from a maintained temperature plasma source, which reduces emittance of the extracted particles and allows use of lower, more maintainable downstream potentials to control an ion beam path of the extracted ions. The reduced emittance facilitates ion beam precision in applications, such as in imaging, tumor imaging, tomographic imaging, and/or cancer treatment.

**[0057]** In another embodiment, a state of a charged particle beam is monitored and/or checked, such as against a previously established radiation plan, in a position just prior to the beam entering the patient. In one example, the charged particle beam state is measured after a final manipulation of intensity, energy, shape, and/or position, such as via use of an insert, a range filter, a collimator, an aperture, and/or a compensator. In one case, one or more beam crossing elements, sheets, coatings, or layers, configured to emit photons upon passage therethrough by the charged particle beam, are positioned between the final manipulation apparatus, such as the insert, and prior to entry into the patient.

**[0058]** In another embodiment, a patient specific tray insert is inserted into a tray frame to form a beam control tray assembly, the beam control tray assembly is inserted into a slot of a tray receiver assembly, and the tray assembly is positioned relative to a gantry nozzle. Optionally, multiple tray inserts, each used to control a beam state parameter, are inserted into slots of the tray receiver assembly. The beam control tray assembling includes an identifier, such as an electromechanical identifier, of the particular insert type, which is communicated to a main controller, such as via the tray receiver assembly. Optionally and preferably, a hand control pendant is used in loading and/or positioning the tray receiver assembly.

**[0059]** In another embodiment, a gantry positions both: (1) a section of a beam transport system, such as a terminal section, used to transport and direct positively charged particles to a tumor and (2) at least one imaging system. In one case, the imaging system is orientated on a same axis as the positively charged particle, such as at a different time through rotation of the gantry. In another case, the imaging system uses at least two crossing beamlines, each beamline coupled to a respective detector, to yield multiple views of the patient. In another case, one or more imaging subsystem yields a two-dimensional image of the patient, such as for

position confirmation and/or as part of a set of images used to develop a three-dimensional image of the patient.

**[0060]** In still another embodiment, multiple linked control stations are used to control position of elements of a beam transport system, nozzle, and/or patient specific beam shaping element relative to a dynamically controlled patient position and/or an imaging surface, element, or system.

**[0061]** In yet another embodiment, a tomography system is optionally used in combination with a charged particle cancer therapy system. The tomography system uses tomography or tomographic imaging, which refers to imaging by sections or sectioning through the use of a penetrating wave, such as a positively charge particle from an injector and/or accelerator. Optionally and preferably, a common injector, accelerator, and beam transport system is used for both charged particle based tomographic imaging and charged particle cancer therapy. In one case, an output nozzle of the beam transport system is positioned with a gantry system while the gantry system and/or a patient support maintains a scintillation plate of the tomography system on the opposite side of the patient from the output nozzle.

**[0062]** In another example, a charged particle state determination system, of a cancer therapy system or tomographic imaging system, uses one or more coated layers in conjunction with a scintillation material, scintillation detector and/or a tomographic imaging system at time of tumor and surrounding tissue sample mapping and/or at time of tumor treatment, such as to determine an input vector of the charged particle beam into a patient and/or an output vector of the charged particle beam from the patient.

**[0063]** In another example, the charged particle tomography apparatus is used in combination with a charged particle cancer therapy system. For example, tomographic imaging of a cancerous tumor is performed using charged particles generated with an injector, accelerator, and guided with a delivery system. The cancer therapy system uses the same injector, accelerator, and guided delivery system in delivering charged particles to the cancerous tumor. For example, the tomography apparatus and cancer therapy system use a common raster beam method and apparatus for treatment of solid cancers. More particularly, the invention comprises a multi-axis and/or multi-field raster beam charged particle accelerator used in: (1) tomography and (2) cancer therapy. Optionally, the system independently controls patient translation position, patient rotation position, two-dimensional beam trajectory, delivered radiation beam energy, delivered radiation beam intensity, beam velocity, timing of charged particle delivery, and/or distribution of radiation striking healthy tissue. The system operates in conjunction with a negative ion beam source, synchrotron, patient positioning, imaging, and/or targeting method and apparatus to deliver an effective and uniform dose of radiation to a tumor while distributing radiation striking healthy tissue.

**[0064]** In another embodiment, a treatment delivery control system (TDCS) or main controller is used to control multiple aspects of the cancer therapy system, including one or more of: an imaging system, such as a CT or PET; a positioner, such as a couch or patient interface module; an injector or injection system; a radio-frequency quadrupole system; a ring accelerator or synchrotron; an extraction system; an irradiation plan; and a display system. The TDCS is preferably a control system for automated cancer therapy once the patient is positioned. The TDCS integrates output of one or more of the below described cancer therapy system

elements with inputs of one or more of the below described cancer therapy system elements. More generally, the TDCS controls or manages input and/or output of imaging, an irradiation plan, and charged particle delivery.

**[0065]** In yet another embodiment, one or more trays are inserted into the positively charged particle beam path, such as at or near the exit port of a gantry nozzle in close proximity to the patient. Each tray holds an insert, such as a patient specific insert for controlling the energy, focus depth, and/or shape of the charged particle beam. Examples of inserts include a range shifter, a compensator, an aperture, a ridge filter, and a blank. Optionally and preferably, each tray communicates a held and positioned insert to a main controller of the charged particle cancer therapy system. The trays optionally hold one or more of the imaging sheets configured to emit light upon transmission of the charged particle beam through a corresponding localized position of the one or more imaging sheets.

**[0066]** For clarity of presentation and without loss of generality, throughout this document, treatment systems and imaging systems are described relative to a tumor of a patient. However, more generally any sample is imaged with any of the imaging systems described herein and/or any element of the sample is treated with the positively charged particle beam(s) described herein.

#### Charged Particle Beam Therapy

**[0067]** Throughout this document, a charged particle beam therapy system, such as a proton beam, hydrogen ion beam, or carbon ion beam, is described. Herein, the charged particle beam therapy system is described using a proton beam. However, the aspects taught and described in terms of a proton beam are not intended to be limiting to that of a proton beam and are illustrative of a charged particle beam system, a positively charged beam system, and/or a multiply charged particle beam system, such as  $C^{4+}$  or  $C^{6+}$ . Any of the techniques described herein are equally applicable to any charged particle beam system.

**[0068]** Referring now to FIG. 1A, a charged particle beam system 100 is illustrated. The charged particle beam preferably comprises a number of subsystems including any of: a main controller 110; an injection system 120; a synchrotron 130 that typically includes: (1) an accelerator system 131 and (2) an internal or connected extraction system 134; a beam transport system 135; a scanning/targeting/delivery system 140; a patient interface module 150; a display system 160; and/or an imaging system 170.

**[0069]** An exemplary method of use of the charged particle beam system 100 is provided. The main controller 110 controls one or more of the subsystems to accurately and precisely deliver protons to a tumor of a patient. For example, the main controller 110 obtains an image, such as a portion of a body and/or of a tumor, from the imaging system 170. The main controller 110 also obtains position and/or timing information from the patient interface module 150. The main controller 110 optionally controls the injection system 120 to inject a proton into a synchrotron 130. The synchrotron typically contains at least an accelerator system 131 and an extraction system 134. The main controller 110 preferably controls the proton beam within the accelerator system, such as by controlling speed, trajectory, and timing of the proton beam. The main controller then controls extraction of a proton beam from the accelerator through the extraction system 134. For example, the con-

troller controls timing, energy, and/or intensity of the extracted beam. The controller **110** also preferably controls targeting of the proton beam through the scanning/targeting/delivery system **140** to the patient interface module **150**. One or more components of the patient interface module **150**, such as translational and rotational position of the patient, are preferably controlled by the main controller **110**. Further, display elements of the display system **160** are preferably controlled via the main controller **110**. Displays, such as display screens, are typically provided to one or more operators and/or to one or more patients. In one embodiment, the main controller **110** times the delivery of the proton beam from all systems, such that protons are delivered in an optimal therapeutic manner to the tumor of the patient.

[0070] Herein, the main controller **110** refers to a single system controlling the charged particle beam system **100**, to a single controller controlling a plurality of subsystems controlling the charged particle beam system **100**, or to a plurality of individual controllers controlling one or more sub-systems of the charged particle beam system **100**.

#### Example I

##### Charged Particle Cancer Therapy System Control

[0071] Referring now to FIG. 1B, an example of a charged particle cancer therapy system **100** is provided. A main controller receives input from one, two, three, or four of a respiration monitoring and/or controlling controller **180**, a beam controller **185**, a rotation controller **147**, and/or a timing to a time period in a respiration cycle controller **148**. The beam controller **185** preferably includes one or more of a beam energy controller **182**, the beam intensity controller **340**, a beam velocity controller **186**, and/or a horizontal/vertical beam positioning controller **188**. The main controller **110** controls any element of the injection system **120**; the synchrotron **130**; the scanning/targeting/delivery system **140**; the patient interface module **150**; the display system **160**; and/or the imaging system **170**. For example, the respiration monitoring/controlling controller **180** controls any element or method associated with the respiration of the patient; the beam controller **185** controls any of the elements controlling acceleration and/or extraction of the charged particle beam; the rotation controller **147** controls any element associated with rotation of the patient **830** or gantry; and the timing to a period in respiration cycle controller **148** controls any aspects affecting delivery time of the charged particle beam to the patient. As a further example, the beam controller **185** optionally controls any magnetic and/or electric field about any magnet in the charged particle cancer therapy system **100**. One or more beam state sensors **190** sense position, direction, intensity, and/or energy of the charged particles at one or more positions in the charged particle beam path. A tomography system **700**, described infra, is optionally used to monitor intensity and/or position of the charged particle beam.

[0072] Referring now to FIG. 1C, an illustrative exemplary embodiment of one version of the charged particle beam system **100** is provided. The number, position, and described type of components is illustrative and non-limiting in nature. In the illustrated embodiment, the injection system **120** or ion source or charged particle beam source generates protons. The injection system **120** optionally includes one or more of: a negative ion beam source, an ion beam focusing

lens, and a tandem accelerator. The protons are delivered into a vacuum tube that runs into, through, and out of the synchrotron. The generated protons are delivered along an initial path **262**. Optionally, focusing magnets **127**, such as quadrupole magnets or injection quadrupole magnets, are used to focus the proton beam path. A quadrupole magnet is a focusing magnet. An injector bending magnet **128** bends the proton beam toward a plane of the synchrotron **130**. The focused protons having an initial energy are introduced into an injector magnet **129**, which is preferably an injection Lamberson magnet. Typically, the initial beam path **262** is along an axis off of, such as above, a circulating plane of the synchrotron **130**. The injector bending magnet **128** and injector magnet **129** combine to move the protons into the synchrotron **130**. Main bending magnets, dipole magnets, turning magnets, or circulating magnets **132** are used to turn the protons along a circulating beam path **264**. A dipole magnet is a bending magnet. The main bending magnets **132** bend the initial beam path **262** into a circulating beam path **264**. In this example, the main bending magnets **132** or circulating magnets are represented as four sets of four magnets to maintain the circulating beam path **264** into a stable circulating beam path. However, any number of magnets or sets of magnets are optionally used to move the protons around a single orbit in the circulation process. The protons pass through an accelerator **133**. The accelerator accelerates the protons in the circulating beam path **264**. As the protons are accelerated, the fields applied by the magnets are increased. Particularly, the speed of the protons achieved by the accelerator **133** are synchronized with magnetic fields of the main bending magnets **132** or circulating magnets to maintain stable circulation of the protons about a central point or region **136** of the synchrotron. At separate points in time the accelerator **133**/main bending magnet **132** combination is used to accelerate and/or decelerate the circulating protons while maintaining the protons in the circulating path or orbit. An extraction element of an inflector/deflector system is used in combination with a Lamberson extraction magnet **137** to remove protons from their circulating beam path **264** within the synchrotron **130**. One example of a deflector component is a Lamberson magnet. Typically the deflector moves the protons from the circulating plane to an axis off of the circulating plane, such as above the circulating plane. Extracted protons are preferably directed and/or focused using an extraction bending magnet **142** and optional extraction focusing magnets **141**, such as quadrupole magnets, and optional bending magnets along a positively charged particle beam transport path **268** in a beam transport system **135**, such as a beam path or proton beam path, into the scanning/targeting/delivery system **140**. Two components of a scanning system **140** or targeting system typically include a first axis control **143**, such as a vertical control, and a second axis control **144**, such as a horizontal control. In one embodiment, the first axis control **143** allows for about 100 mm of vertical or y-axis scanning of the proton beam **268** and the second axis control **144** allows for about 700 mm of horizontal or x-axis scanning of the proton beam **268**. A nozzle system **146** is used for imaging the proton beam, for defining shape of the proton beam, and/or as a vacuum barrier between the low pressure beam path of the synchrotron and the atmosphere. Protons are delivered with control to the patient interface module **150** and to a tumor of a patient. All of the above listed elements are optional and may be used in various permutations and combinations.

#### Ion Extraction from Ion Source

[0073] A method and apparatus are described for extraction of ions from an ion source. For clarity of presentation and without loss of generality, examples focus on extraction of protons from the ion source. However, more generally cations of any charge are optionally extracted from a corresponding ion source with the techniques described herein. For instance,  $C^{4+}$  or  $C^{6+}$  are optionally extracted using the ion extraction methods and apparatus described herein. Further, by reversing polarity of the system, anions are optionally extracted from an anion source, where the anion is of any charge.

[0074] Herein, for clarity of presentation and without loss of generality, ion extraction is coupled with tumor treatment and/or tumor imaging. However, the ion extraction is optional used in any method or apparatus using a stream or time discrete bunches of ions.

#### Diode Extraction

[0075] Referring now to FIG. 2A and FIG. 2B, a first ion extraction system is illustrated. The first ion extraction system uses a diode extraction system 200, where a first element of the diode extraction system is an ion source 122 or first electrode at a first potential and a second element 202 of the diode extraction system is at a second potential. Generally, the first potential is raised or lowered relative to the second potential to extract ions from the ion source 122 along the z-axis or the second potential is raised or lowered relative to the first potential to extract ions from the ion source 122 along the z-axis, where polarity of the potential difference determines if anions or cations are extracted from the ion source 122.

[0076] Still referring to FIG. 2A and FIG. 2B, an example of ion extraction from the ion source 122 is described. As illustrated in FIG. 2A, in a non-extraction time period, a non-extraction diode potential,  $A_1$ , of the ion source 122 is held at a potential equal to a potential,  $B_1$ , of the second element 202. Referring now to FIG. 2B, during an extraction time period, a diode extraction potential,  $A_2$ , of the ion source 122 is raised, causing a positively charged cation, such as the proton, to be drawn out of the ion chamber toward the lower potential of the second element 202. Similarly, if the diode extraction potential,  $A_2$ , of the ion source is lowered relative a potential,  $B_1$ , then an anion is extracted from the ion source 122 toward a higher potential of the second element 202. In the diode extraction system 200, the voltage of a large mass and corresponding large capacitance of the ion source 122 is raised or lowered, which takes time, has an RC time constant, and results in a range of temperatures of the plasma during the extraction time period, which is typically pulsed on and off with time. Particularly, as the potential of the ion source 122 is cycled with time, the ion source 122 temperature cycles, which results in a range of emittance values, resultant from conservation of momentum, and a corresponding less precise extraction beam. Alternatively, potential of the second element 202 is varied, altered, pulsed, or cycled, which reduces a range of emittance values during the extraction process.

#### Triode Extraction

[0077] Referring now to FIG. 2C and FIG. 2D, a second ion extraction system is illustrated. The second ion extraction system uses a triode extraction system 210. The triode

extraction system 210 uses: (1) an ion source 122, (2) a gating electrode 204 also referred to as a suppression electrode, and (3) an extraction electrode 206. Optionally, a first electrode of the triode extraction system 210 is positioned proximate the ion source 122 and is maintained at a potential as described, infra, using the ion source as the first electrode of the triode extraction system. Generally, potential of the gating electrode 204 is raised and lowered to, as illustrated, stop and start extraction of a positive ion. Varying the potential of the gating electrode 204 has the advantages of altering the potential of a small mass with a correspondingly small capacitance and small RC time constant, which via conservation of momentum, reduces emittance of the extracted ions. Optionally, a first electrode maintained at the first potential of the ion source is used as the first element of the triode extraction system in place of the ion source 122 while also optionally further accelerating and/or focusing the extracted ions or set of ions using the extraction electrode 206. Several example further describe the triode extraction system 210.

#### Example I

[0078] Still referring to FIG. 2C and FIG. 2D, a first example of ion beam extraction using the triode extraction system 210 is provided. Optionally and preferably, the ion source 122 is maintained at a stable temperature. Maintaining the ion source 122 at a stable temperature, such as with a constant applied voltage, results in ions with more uniform energy and thus velocity. Hence, extraction of ions from the stable temperature plasma results in extracted ions with more uniform energy or velocity and smaller emittance, where emittance is a property of a charged particle beam in a particle accelerator. Emittance is a measure for the average spread of particle coordinates in position-and-momentum phase space and has the dimension of length, such as meters, or length times angle, such as meters times radians.

#### Example II

[0079] Still referring to FIG. 2C and FIG. 2D, a second example of ion beam extraction using the triode extraction system 210 is provided illustrating voltages of the triode elements for extraction of cations, such as protons. Optionally and preferably, the extraction electrode 206 is grounded at zero volts or is near ground, which allows downstream elements about an ion beam path of the extracted ions to be held at ground or near ground. The ability to maintain downstream elements about the beam path at ground greatly eases design as the downstream elements are often of high mass with high capacitance, thus requiring large power supplies to maintain at positive or negative potentials. The ion source 122, for proton ion formation and extraction therefrom, is optionally maintained at 10 to 100 kV, more preferably at 20 to 80 kV, and most preferably at 30 kV  $\pm$  less than 1, 5, or 10 kV. The gating electrode 204 is maintained at a non-extraction potential at or above the potential of the ion source 122 and is maintained at an extraction potential of less than the potential of the ion source and/or greater than or equal to the potential of the extraction electrode 206.

#### Example III

[0080] Still referring to FIG. 2C and FIG. 2D, a third example of anion beam extraction using the triode extraction system 210 is provided. Generally, for extraction of anions

the potentials of the second example are inverted and/or multiplied by negative one. For instance, if the extraction electrode **206** is held at ground, then the ion source **122** is maintained with a negative voltage, such as at  $-30$  kV, and the gating electrode cycles between the voltage of the ion source **122** and the potential of the extraction electrode **206** to turn off and on extraction of anions from the ion source **122** along the extraction beamline.

#### Example IV

[0081] Still referring to FIG. 2C and FIG. 2D, a fourth example of extraction suppression is provided. As illustrated in FIG. 2C, in the non-extraction mode the ion source potential,  $A_3$ , is equal to the gating electrode potential,  $C_1$ . However, the gating electrode **204**, which is also referred to as a suppression electrode, is optionally held at a higher potential than the ion source potential so as to provide a suppression barrier or a potential resistance barrier keeping cations in the ion source **122**. For instance, for cation extraction, if the ion source potential is  $+30$  kV, then the gating electrode potential is greater than  $+30$  kV, such as  $+32$  kV $\pm 1$ ,  $1.5$ , or  $2$  kV. In a case of the ion source **122** forming anions, the gating electrode potential,  $C_1$ , is optionally held at a lower potential than the ion source potential,  $A_3$ . Generally, during the non-extraction phase, the gating electrode **204** is optionally maintained at a gating potential close to the ion source potential with a bias in voltage relative to the ion source potential repelling ions back into the ion source **122**.

#### Example V

[0082] Still referring to FIG. 2C and FIG. 2D, a fifth example of using the triode extraction system **210** with varying types of ion sources is provided. The triode extraction system **210** is optionally used with an electron cyclotron resonance (ECR) ion source, a dual plasmatron ion source, an indirectly heated cathode ion source, a Freeman type ion source, or a Bernas type ion source.

#### Example VI

[0083] Herein, for clarity of presentation and without loss of generality, the triode extraction system **210** is integrated with an electron cyclotron resonance source. Generally, the electron resonance source generates an ionized plasma by heating or superimposing a static magnetic field and a high-frequency electromagnetic field at an electron cyclotron resonance frequency, which functions to form a localized plasma, where the heating power is optionally varied to yield differing initial energy levels of the ions. As the electron resonance source: (1) moves ions in an arc in a given direction and (2) is tunable in temperature, described infra, emittance of the electron resonance source is low and has an initial beam in a same mean cycling or arc following direction. The temperature of the electron cyclotron resonance ion source is optionally controlled through an external input, such as a tunable or adjustable microwave power, a controllable and variable gas pressure, and/or a controllable and alterable arc voltage. The external input allows the plasma density in the electron cyclotron resonance source to be controlled.

[0084] In a sixth example, an electron resonance source is the ion source **122** of the triode extraction system **210**. Optionally and preferably, the gating electrode **204** of the

triode extraction system is oscillated, such as from about the ion source potential toward the extraction electrode potential, which is preferably grounded. In this manner, the extracted electron beam along the initial path **262** is bunches of ions that have peak intensities alternating with low or zero intensities, such as in an AC wave as opposed to a continuous beam, such as a DC wave.

#### Example VII

[0085] Still referring to FIG. 2C and FIG. 2D, optionally and preferably geometries of the gating electrode **204** and/or the extraction electrode **206** are used to focus the extracted ions along the initial ion beam path **262**.

#### Example VIII

[0086] Still referring to FIG. 2C and FIG. 2D, the lower emittance of the electron cyclotron resonance triode extraction system is optionally and preferably coupled with a downbeam or downstream radio-frequency quadrupole, used to focus the beam, and/or a synchrotron, used to accelerate the beam.

#### Example IX

[0087] Still referring to FIG. 2C and FIG. 2D, the lower emittance of the electron cyclotron resonance triode extraction system is maintained through the synchrotron **130** and to the tumor of the patient resulting in a more accurate, precise, smaller, and/or tighter treatment voxel of the charged particle beam or charged particle pulse striking the tumor.

#### Example X

[0088] Still referring to FIG. 2C and FIG. 2D, the lower emittance of the electron cyclotron resonance triode extraction system reduces total beam spread through the synchrotron **130** and the tumor to one or more imaging elements, such as an optical imaging sheet or scintillation material emitting photons upon passage of the charged particle beam or striking of the charged particle beam, respectively. The lower emittance of the charged particle beam, optionally and preferably maintained through the accelerator system **134** and beam transport system yields a tighter, more accurate, more precise, and/or smaller particle beam or particle burst diameter at the imaging surfaces and/or imaging elements, which facilitates more accurate and precise tumor imaging, such as for subsequent tumor treatment or to adjust, while the patient waits in a treatment position, the charged particle treatment beam position.

[0089] Any feature or features of any of the above provided examples are optionally and preferably combined with any feature described in other examples provided, supra, or herein.

#### Ion Extraction from Accelerator

[0090] Referring now to FIG. 3, both: (1) an exemplary proton beam extraction system **300** from the synchrotron **130** and (2) a charged particle beam intensity control system **305** are illustrated. For clarity, FIG. 3 removes elements represented in FIG. 1C, such as the turning magnets, which allows for greater clarity of presentation of the proton beam path as a function of time. Generally, protons are extracted from the synchrotron **130** by slowing the protons. As described, supra, the protons were initially accelerated in a circulating path, which is maintained with a plurality of

main bending magnets **132**. The circulating path is referred to herein as an original central beamline **264**. The protons repeatedly cycle around a central point in the synchrotron **136**. The proton path traverses through a radio frequency (RF) cavity system **310**. To initiate extraction, an RF field is applied across a first blade **312** and a second blade **314**, in the RF cavity system **310**. The first blade **312** and second blade **314** are referred to herein as a first pair of blades.

[0091] In the proton extraction process, an RF voltage is applied across the first pair of blades, where the first blade **312** of the first pair of blades is on one side of the circulating proton beam path **264** and the second blade **314** of the first pair of blades is on an opposite side of the circulating proton beam path **264**. The applied RF field applies energy to the circulating charged-particle beam. The applied RF field alters the orbiting or circulating beam path slightly of the protons from the original central beamline **264** to an altered circulating beam path **265**. Upon a second pass of the protons through the RF cavity system, the RF field further moves the protons off of the original proton beamline **264**. For example, if the original beamline is considered as a circular path, then the altered beamline is slightly elliptical. The frequency of the applied RF field is timed to apply outward or inward movement to a given band of protons circulating in the synchrotron accelerator. Orbits of the protons are slightly more off axis compared to the original circulating beam path **264**. Successive passes of the protons through the RF cavity system are forced further and further from the original central beamline **264** by altering the direction and/or intensity of the RF field with each successive pass of the proton beam through the RF field. Timing of application of the RF field and/or frequency of the RF field is related to the circulating charged particles circulation pathlength in the synchrotron **130** and the velocity of the charged particles so that the applied RF field has a period, with a peak-to-peak time period, equal to a period of time of beam circulation in the synchrotron **130** about the center **136** or an integer multiple of the time period of beam circulation about the center **136** of the synchrotron **130**. Alternatively, the time period of beam circulation about the center **136** of the synchrotron **130** is an integer multiple of the RF period time. The RF period is optionally used to calculate the velocity of the charged particles, which relates directly to the energy of the circulating charged particles.

[0092] The RF voltage is frequency modulated at a frequency about equal to the period of one proton cycling around the synchrotron for one revolution or at a frequency that is an integral multiplier of the period of one proton cycling about the synchrotron. The applied RF frequency modulated voltage excites a betatron oscillation. For example, the oscillation is a sine wave motion of the protons. The process of timing the RF field to a given proton beam within the RF cavity system is repeated thousands of times with each successive pass of the protons being moved approximately one micrometer further off of the original central beamline **264**. For clarity, the approximately 1000 changing beam paths with each successive path of a given band of protons through the RF field are illustrated as the altered beam path **265**. The RF time period in this process is known, thus energy of the charged particles at time of hitting the extraction material or material **330**, described infra, is known.

[0093] With a sufficient sine wave betatron amplitude, the altered circulating beam path **265** touches and/or traverses a

material **330**, such as a foil or a sheet of foil. The foil is preferably a lightweight material, such as beryllium, a lithium hydride, a carbon sheet, or a material having low nuclear charge components. Herein, a material of low nuclear charge is a material composed of atoms consisting essentially of atoms having six or fewer protons. The foil is preferably about 10 to 150 microns thick, is more preferably about 30 to 100 microns thick, and is still more preferably about 40 to 60 microns thick. In one example, the foil is beryllium with a thickness of about 50 microns. When the protons traverse through the foil, energy of the protons is lost and the speed of the protons is reduced. Typically, a current is also generated, described infra. Protons moving at the slower speed travel in the synchrotron with a reduced radius of curvature **266** compared to either the original central beamline **264** or the altered circulating path **265**. The reduced radius of curvature **266** path is also referred to herein as a path having a smaller diameter of trajectory or a path having protons with reduced energy. The reduced radius of curvature **266** is typically about two millimeters less than a radius of curvature of the last pass of the protons along the altered proton beam path **265**.

[0094] The thickness of the material **330** is optionally adjusted to create a change in the radius of curvature, such as about  $\frac{1}{2}$ , 1, 2, 3, or 4 mm less than the last pass of the protons **265** or original radius of curvature **264**. The reduction in velocity of the charged particles transmitting through the material **330** is calculable, such as by using the pathlength of the betatron oscillating charged particle beam through the material **330** and/or using the density of the material **330**. Protons moving with the smaller radius of curvature travel between a second pair of blades. In one case, the second pair of blades is physically distinct and/or is separated from the first pair of blades. In a second case, one of the first pair of blades is also a member of the second pair of blades. For example, the second pair of blades is the second blade **314** and a third blade **316** in the RF cavity system **310**. A high voltage DC signal, such as about 1 to 5 kV, is then applied across the second pair of blades, which directs the protons out of the synchrotron through an extraction magnet **137**, such as a Lamberson extraction magnet, into a transport path **268**.

[0095] Control of acceleration of the charged particle beam path in the synchrotron with the accelerator and/or applied fields of the turning magnets in combination with the above described extraction system allows for control of the intensity of the extracted proton beam, where intensity is a proton flux per unit time or the number of protons extracted as a function of time. For example, when a current is measured beyond a threshold, the RF field modulation in the RF cavity system is terminated or reinitiated to establish a subsequent cycle of proton beam extraction. This process is repeated to yield many cycles of proton beam extraction from the synchrotron accelerator.

[0096] In another embodiment, instead of moving the charged particles to the material **330**, the material **330** is mechanically moved to the circulating charged particles. Particularly, the material **330** is mechanically or electromechanically translated into the path of the circulating charged particles to induce the extraction process, described supra. In this case, the velocity or energy of the circulating charged particle beam is calculable using the pathlength of the beam path about the center **136** of the synchrotron **130** and from the force applied by the bending magnets **132**.

[0097] In either case, because the extraction system does not depend on any change in magnetic field properties, it allows the synchrotron to continue to operate in acceleration or deceleration mode during the extraction process. Stated differently, the extraction process does not interfere with synchrotron acceleration. In stark contrast, traditional extraction systems introduce a new magnetic field, such as via a hexapole, during the extraction process. More particularly, traditional synchrotrons have a magnet, such as a hexapole magnet, that is off during an acceleration stage. During the extraction phase, the hexapole magnetic field is introduced to the circulating path of the synchrotron. The introduction of the magnetic field necessitates two distinct modes, an acceleration mode and an extraction mode, which are mutually exclusive in time. The herein described system allows for acceleration and/or deceleration of the proton during the extraction step and tumor treatment without the use of a newly introduced magnetic field, such as by a hexapole magnet.

#### Charged Particle Beam Intensity Control

[0098] Control of applied field, such as a radio-frequency (RF) field, frequency and magnitude in the RF cavity system 310 allows for intensity control of the extracted proton beam, where intensity is extracted proton flux per unit time or the number of protons extracted as a function of time.

[0099] Still referring FIG. 3, the intensity control system 305 is further described. In this example, an intensity control feedback loop is added to the extraction system, described supra. When protons in the proton beam hit the material 330 electrons are given off from the material 330 resulting in a current. The resulting current is converted to a voltage and is used as part of an ion beam intensity monitoring system or as part of an ion beam feedback loop for controlling beam intensity. The voltage is optionally measured and sent to the main controller 110 or to an intensity controller subsystem 340, which is preferably in communication or under the direction of the main controller 110. More particularly, when protons in the charged particle beam path pass through the material 330, some of the protons lose a small fraction of their energy, such as about one-tenth of a percent, which results in a secondary electron. That is, protons in the charged particle beam push some electrons when passing through material 330 giving the electrons enough energy to cause secondary emission. The resulting electron flow results in a current or signal that is proportional to the number of protons going through the target or extraction material 330. The resulting current is preferably converted to voltage and amplified. The resulting signal is referred to as a measured intensity signal.

[0100] The amplified signal or measured intensity signal resulting from the protons passing through the material 330 is optionally used in monitoring the intensity of the extracted protons and is preferably used in controlling the intensity of the extracted protons. For example, the measured intensity signal is compared to a goal signal, which is predetermined in an irradiation of the tumor plan. The difference between the measured intensity signal and the planned for goal signal is calculated. The difference is used as a control to the RF generator. Hence, the measured flow of current resulting from the protons passing through the material 330 is used as a control in the RF generator to increase or decrease the number of protons undergoing betatron oscillation and striking the material 330. Hence, the voltage determined off of

the material 330 is used as a measure of the orbital path and is used as a feedback control to control the RF cavity system.

[0101] In one example, the intensity controller subsystem 340 preferably additionally receives input from: (1) a detector 350, which provides a reading of the actual intensity of the proton beam and/or (2) an irradiation plan 360. The irradiation plan provides the desired intensity of the proton beam for each x, y, energy, and/or rotational position of the patient/tumor as a function of time. Thus, the intensity controller 340 receives the desired intensity from the irradiation plan 350, the actual intensity from the detector 350 and/or a measure of intensity from the material 330, and adjusts the amplitude and/or the duration of application of the applied radio-frequency field in the RF cavity system 310 to yield an intensity of the proton beam that matches the desired intensity from the irradiation plan 360.

[0102] As described, supra, the protons striking the material 330 is a step in the extraction of the protons from the synchrotron 130. Hence, the measured intensity signal is used to change the number of protons per unit time being extracted, which is referred to as intensity of the proton beam. The intensity of the proton beam is thus under algorithm control. Further, the intensity of the proton beam is controlled separately from the velocity of the protons in the synchrotron 130. Hence, intensity of the protons extracted and the energy of the protons extracted are independently variable. Still further, the intensity of the extracted protons is controllably variable while scanning the charged particles beam in the tumor from one voxel to an adjacent voxel as a separate hexapole and separated time period from acceleration and/or treatment is not required, as described supra.

[0103] For example, protons initially move at an equilibrium trajectory in the synchrotron 130. An RF field is used to excite or move the protons into a betatron oscillation. In one case, the frequency of the protons orbit is about 10 MHz. In one example, in about one millisecond or after about 10,000 orbits, the first protons hit an outer edge of the target material 130. The specific frequency is dependent upon the period of the orbit. Upon hitting the material 130, the protons push electrons through the foil to produce a current. The current is converted to voltage and amplified to yield a measured intensity signal. The measured intensity signal is used as a feedback input to control the applied RF magnitude or RF field. An energy beam sensor, described infra, is optionally used as a feedback control to the RF field frequency or RF field of the RF field extraction system 310 to dynamically control, modify, and/or alter the delivered charge particle beam energy, such as in a continuous pencil beam scanning system operating to treat tumor voxels without alternating between an extraction phase and a treatment phase. Preferably, the measured intensity signal is compared to a target signal and a measure of the difference between the measured intensity signal and target signal is used to adjust the applied RF field in the RF cavity system 310 in the extraction system to control the intensity of the protons in the extraction step. Stated again, the signal resulting from the protons striking and/or passing through the material 130 is used as an input in RF field modulation. An increase in the magnitude of the RF modulation results in protons hitting the foil or material 130 sooner. By increasing the RF, more protons are pushed into the foil, which results in an increased intensity, or more protons per unit time, of protons extracted from the synchrotron 130.

[0104] In another example, a detector **350** external to the synchrotron **130** is used to determine the flux of protons extracted from the synchrotron and a signal from the external detector is used to alter the RF field, RF intensity, RF amplitude, and/or RF modulation in the RF cavity system **310**. Here the external detector generates an external signal, which is used in a manner similar to the measured intensity signal, described in the preceding paragraphs. Preferably, an algorithm or irradiation plan **360** is used as an input to the intensity controller **340**, which controls the RF field modulation by directing the RF signal in the betatron oscillation generation in the RF cavity system **310**. The irradiation plan **360** preferably includes the desired intensity of the charged particle beam as a function of time and/or energy of the charged particle beam as a function of time, for each patient rotation position, and/or for each x-, y-position of the charged particle beam.

[0105] In yet another example, when a current from material **330** resulting from protons passing through or hitting material is measured beyond a threshold, the RF field modulation in the RF cavity system is terminated or reinitiated to establish a subsequent cycle of proton beam extraction. This process is repeated to yield many cycles of proton beam extraction from the synchrotron accelerator.

[0106] In still yet another embodiment, intensity modulation of the extracted proton beam is controlled by the main controller **110**. The main controller **110** optionally and/or additionally controls timing of extraction of the charged particle beam and energy of the extracted proton beam.

[0107] The benefits of the system include a multi-dimensional scanning system. Particularly, the system allows independence in: (1) energy of the protons extracted and (2) intensity of the protons extracted. That is, energy of the protons extracted is controlled by an energy control system and an intensity control system controls the intensity of the extracted protons. The energy control system and intensity control system are optionally independently controlled. Preferably, the main controller **110** controls the energy control system and the main controller **110** simultaneously controls the intensity control system to yield an extracted proton beam with controlled energy and controlled intensity where the controlled energy and controlled intensity are independently variable and/or continually available as a separate extraction phase and acceleration phase are not required, as described supra. Thus the irradiation spot hitting the tumor is under independent control of:

[0108] time;

[0109] energy;

[0110] intensity;

[0111] x-axis position, where the x-axis represents horizontal movement of the proton beam relative to the patient, and

[0112] y-axis position, where the y-axis represents vertical movement of the proton beam relative to the patient.

[0113] In addition, the patient is optionally independently translated and/or rotated relative to a translational axis of the proton beam at the same time.

#### Beam Transport

[0114] The beam transport system **135** is used to move the charged particles from the accelerator to the patient, such as via a nozzle in a gantry, described infra.

#### Charged Particle Energy

[0115] The beam transport system **135** optionally includes means for determining an energy of the charged particles in the charged particle beam. For example, an energy of the charged particle beam is determined via calculation, such as via equation 1, using knowledge of a magnet geometry and applied magnetic field to determine mass and/or energy. Referring now to equation 1, for a known magnet geometry, charge,  $q$ , and magnetic field,  $B$ , the Larmor radius,  $\rho_L$ , or magnet bend radius is defined as:

$$\rho_L = \frac{v_{\perp}}{\Omega_c} = \frac{\sqrt{2Em}}{qB} \quad (\text{eq. 1})$$

[0116] where:  $v_{\perp}$  is the ion velocity perpendicular to the magnetic field,  $\Omega_c$  is the cyclotron frequency,  $q$  is the charge of the ion,  $B$  is the magnetic field,  $m$  is the mass of the charge particle, and  $E$  is the charged particle energy. Solving for the charged particle energy yields equation 2.

$$E = \frac{(\rho_L q B)^2}{2m} \quad (\text{eq. 2})$$

[0117] Thus, an energy of the charged particle in the charged particle beam in the beam transport system **135** is calculable from the known magnet geometry, known or measured magnetic field, charged particle mass, charged particle charge, and the known magnet bend radius, which is proportional to and/or equivalent to the Larmor radius.

#### Nozzle

[0118] After extraction from the synchrotron **130** and transport of the charged particle beam along the proton beam path **268** in the beam transport system **135**, the charged particle beam exits through the nozzle system **146**. In one example, the nozzle system includes a nozzle foil covering an end of the nozzle system **146** or a cross-sectional area within the nozzle system forming a vacuum seal. The nozzle system includes a nozzle that expands in x/y-cross-sectional area along the z-axis of the proton beam path **268** to allow the proton beam **268** to be scanned along the x-axis and y-axis by the vertical control element and horizontal control element, respectively. The nozzle foil is preferably mechanically supported by the outer edges of an exit port of the nozzle **146**. An example of a nozzle foil is a sheet of about 0.1 inch thick aluminum foil. Generally, the nozzle foil separates atmosphere pressures on the patient side of the nozzle foil from the low pressure region, such as about  $10^{-5}$  to  $10^{-7}$  torr region, on the synchrotron **130** side of the nozzle foil. The low pressure region is maintained to reduce scattering of the circulating charged particle beam in the synchrotron. Herein, the exit foil of the nozzle is optionally the first sheet **760** of the charged particle beam state determination system **750**, described infra.

#### Charged Particle Control

[0119] Referring now to FIG. 4A, FIG. 4B, FIG. 5, FIG. 6A, and FIG. 6B, a charged particle beam control system is described where one or more patient specific beam control assemblies are removably inserted into the charged particle

beam path proximate the nozzle of the charged particle cancer therapy system **100**, where the patient specific beam control assemblies adjust the beam energy, diameter, cross-sectional shape, focal point, and/or beam state of the charged particle beam to properly couple energy of the charged particle beam to the individual's specific tumor.

#### Beam Control Tray

[0120] Referring now to FIG. 4A and FIG. 4B, a beam control tray assembly **400** is illustrated in a top view and side view, respectively. The beam control tray assembly **400** optionally comprises any of a tray frame **410**, a tray aperture **412**, a tray handle **420**, a tray connector/communicator **430**, and means for holding a patient specific tray insert **510**, described infra. Generally, the beam control tray assembly **400** is used to: (1) hold the patient specific tray insert **510** in a rigid location relative to the beam control tray **400**, (2) electronically identify the held patient specific tray insert **510** to the main controller **110**, and (3) removably insert the patient specific tray insert **510** into an accurate and precise fixed location relative to the charged particle beam, such as the proton beam path **268** at the nozzle of the charged particle cancer therapy system **100**.

[0121] For clarity of presentation and without loss of generality, the means for holding the patient specific tray insert **510** in the tray frame **410** of the beam control tray assembly **400** is illustrated as a set of recessed set screws **415**. However, the means for holding the patient specific tray insert **510** relative to the rest of the beam control tray assembly **400** is optionally any mechanical and/or electro-mechanical positioning element, such as a latch, clamp, fastener, clip, slide, strap, or the like. Generally, the means for holding the patient specific tray insert **510** in the beam control tray **400** fixes the tray insert and tray frame relative to one another even when rotated along and/or around multiple axes, such as when attached to a charged particle cancer therapy system **100** dynamic gantry nozzle **610** or gantry nozzle, which is an optional element of the nozzle system **146**, that moves in three-dimensional space relative to a fixed point in the beamline, proton beam path **268**, and/or a given patient position. As illustrated in FIG. 4A and FIG. 4B, the recessed set screws **415** fix the patient specific tray insert **510** into the aperture **412** of the tray frame **410**. The tray frame **410** is illustrated as circumferentially surrounding the patient specific tray insert **510**, which aids in structural stability of the beam control tray assembly **400**. However, generally the tray frame **410** is of any geometry that forms a stable beam control tray assembly **400**.

[0122] Still referring to FIG. 4A and now referring to FIG. 5 and FIG. 6A, the optional tray handle **420** is used to manually insert/retract the beam control tray assembly **400** into a receiving element of the gantry nozzle or dynamic gantry nozzle **610**. While the beam control tray assembly **400** is optionally inserted into the charged particle beam path **268** at any point after extraction from the synchrotron **130**, the beam control tray assembly **400** is preferably inserted into the positively charged particle beam proximate the dynamic gantry nozzle **610** as control of the beam shape is preferably done with little space for the beam shape to defocus before striking the tumor. Optionally, insertion and/or retraction of the beam control tray assembly **400** is semi-automated, such as in a manner of a digital-video disk player receiving a digital-video disk, with a selected auto load and/or a selected auto unload feature.

#### Patient Specific Tray Insert

[0123] Referring again to FIG. 5, a system of assembling trays **500** is described. The beam control tray assembly **400** optionally and preferably has interchangeable patient specific tray inserts **510**, such as a range shifter insert **511**, a patient specific ridge filter insert **512**, an aperture insert **513**, a compensator insert **514**, or a blank insert **515**. As described, supra, any of the range shifter insert **511**, the patient specific ridge filter insert **512**, the aperture insert **513**, the compensator insert **514**, or the blank insert **515** after insertion into the tray frame **410** are inserted as the beam control tray assembly **400** into the positively charged particle beam path **268**, such as proximate the dynamic gantry nozzle **610**.

[0124] Still referring to FIG. 5, the patient specific tray inserts **510** are further described. The patient specific tray inserts comprise a combination of any of: (1) a standardized beam control insert and (2) a patient specific beam control insert. For example, the range shifter insert or **511** or compensator insert **514** used to control the depth of penetration of the charged particle beam into the patient is optionally: (a) a standard thickness of a beam slowing material, such as a first thickness of Lucite, an acrylic, a clear plastic, and/or a thermoplastic material, (b) one member of a set of members of varying thicknesses and/or densities where each member of the set of members slows the charged particles in the beam path by a known amount, or (c) is a material with a density and thickness designed to slow the charged particles by a customized amount for the individual patient being treated, based on the depth of the individual's tumor in the tissue, the thickness of intervening tissue, and/or the density of intervening bone/tissue. Similarly, the ridge filter insert **512** used to change the focal point or shape of the beam as a function of depth is optionally: (1) selected from a set of ridge filters where different members of the set of ridge filters yield different focal depths or (2) customized for treatment of the individual's tumor based on position of the tumor in the tissue of the individual. Similarly, the aperture insert is: (1) optionally selected from a set of aperture shapes or (2) is a customized individual aperture insert **513** designed for the specific shape of the individual's tumor. The blank insert **515** is an open slot, but serves the purpose of identifying slot occupancy, as described infra.

#### Slot Occupancy/Identification

[0125] Referring again to FIG. 4A, FIG. 4B, and FIG. 5, occupancy and identification of the particular patient specific tray insert **510** into the beam control tray assembly **400** is described. Generally, the beam control tray assembly **400** optionally contains means for identifying, to the main controller **110** and/or a treatment delivery control system described infra, the specific patient tray insert **510** and its location in the charged particle beam path **268**. First, the particular tray insert is optionally labeled and/or communicated to the beam control tray assembly **400** or directly to the main controller **110**. Second, the beam control tray assembly **400** optionally communicates the tray type and/or tray insert to the main controller **110**. In various embodiments, communication of the particular tray insert to the main controller **110** is performed: (1) directly from the tray insert, (2) from the tray insert **510** to the tray assembly **400** and subsequently to the main controller **110**, and/or (3) directly from the tray assembly **400**. Generally, communication is performed wire-

lessly and/or via an established electromechanical link. Identification is optionally performed using a radio-frequency identification label, use of a barcode, or the like, and/or via operator input. Examples are provided to further clarify identification of the patient specific tray insert **510** in a given beam control tray assembly **400** to the main controller.

[0126] In a first example, one or more of the patient specific tray inserts **510**, such as the range shifter insert **511**, the patient specific ridge filter insert **512**, the aperture insert **513**, the compensator insert **514**, or the blank insert **515** include an identifier **520** and/or a first electromechanical identifier plug **530**. The identifier **520** is optionally a label, a radio-frequency identification tag, a barcode, a 2-dimensional bar-code, a matrix-code, or the like. The first electromechanical identifier plug **530** optionally includes memory programmed with the particular patient specific tray insert information and a connector used to communicate the information to the beam control tray assembly **400** and/or to the main controller **110**. As illustrated in FIG. 5, the first electromechanical identifier plug **530** affixed to the patient specific tray insert **510** plugs into a second electromechanical identifier plug, such as the tray connector/communicator **430**, of the beam control tray assembly **400**, which is described infra.

[0127] In a second example, the beam control tray assembly **400** uses the second electromechanical identifier plug to send occupancy, position, and/or identification information related to the type of tray insert or the patient specific tray insert **510** associated with the beam control tray assembly to the main controller **110**. For example, a first tray assembly is configured with a first tray insert and a second tray assembly is configured with a second tray insert. The first tray assembly sends information to the main controller **110** that the first tray assembly holds the first tray insert, such as a range shifter, and the second tray assembly sends information to the main controller **110** that the second tray assembly holds the second tray insert, such as an aperture. The second electromechanical identifier plug optionally contains programmable memory for the operator to input the specific tray insert type, a selection switch for the operator to select the tray insert type, and/or an electromechanical connection to the main controller. The second electromechanical identifier plug associated with the beam control tray assembly **400** is optionally used without use of the first electromechanical identifier plug **530** associated with the tray insert **510**.

[0128] In a third example, one type of tray connector/communicator **430** is used for each type of patient specific tray insert **510**. For example, a first connector/communicator type is used for holding a range shifter insert **511**, while a second, third, fourth, and fifth connector/communicator type is used for trays respectively holding a patient specific ridge filter insert **512**, an aperture insert **513**, a compensator insert **514**, or a blank insert **515**. In one case, the tray communicates tray type with the main controller. In a second case, the tray communicates patient specific tray insert information with the main controller, such as an aperture identifier custom built for the individual patient being treated.

#### Tray Insertion/Coupling

[0129] Referring now to FIG. 6A and FIG. 6B a beam control insertion process **600** is described. The beam control insertion process **600** comprises: (1) insertion of the beam

control tray assembly **400** and the associated patient specific tray insert **510** into the charged particle beam path **268** and/or dynamic gantry nozzle **610**, such as into a tray assembly receiver **620** and (2) an optional partial or total retraction of beam of the tray assembly receiver **620** into the dynamic gantry nozzle **610**.

[0130] Referring now to FIG. 6A, insertion of one or more of the beam control tray assemblies **400** and the associated patient specific tray inserts **510** into the dynamic gantry nozzle **610** is further described. In FIG. 6A, three beam control tray assemblies, of a possible n tray assemblies, are illustrated, a first tray assembly **402**, a second tray assembly **404**, and a third tray assembly **406**, where n is a positive integer of 1, 2, 3, 4, 5 or more. As illustrated, the first tray assembly **402** slides into a first receiving slot **403**, the second tray assembly **404** slides into a second receiving slot **405**, and the third tray assembly **406** slides into a third receiving slot **407**. Generally, any tray optionally inserts into any slot or tray types are limited to particular slots through use of a mechanical, physical, positional, and/or steric constraints, such as a first tray type configured for a first insert type having a first size and a second tray type configured for a second insert type having a second distinct size at least ten percent different from the first size.

[0131] Still referring to FIG. 6A, identification of individual tray inserts inserted into individual receiving slots is further described. As illustrated, sliding the first tray assembly **402** into the first receiving slot **403** connects the associated electromechanical connector/communicator **430** of the first tray assembly **402** to a first receptor **626**. The electromechanical connector/communicator **430** of the first tray assembly communicates tray insert information of the first beam control tray assembly to the main controller **110** via the first receptor **626**. Similarly, sliding the second tray assembly **404** into the second receiving slot **405** connects the associated electromechanical connector/communicator **430** of the second tray assembly **404** into a second receptor **627**, which links communication of the associated electromechanical connector/communicator **430** with the main controller **110** via the second receptor **627**, while a third receptor **628** connects to the electromechanical connector placed into the third slot **407**. The non-wireless/direct connection is preferred due to the high radiation levels within the treatment room and the high shielding of the treatment room, which both hinder wireless communication. The connection of the communicator and the receptor is optionally of any configuration and/or orientation.

#### Tray Receiver Assembly Retraction

[0132] Referring again to FIG. 6A and FIG. 6B, retraction of the tray receiver assembly **620** relative to a nozzle end **612** of the dynamic gantry nozzle **610** is described. The tray receiver assembly **620** comprises a framework to hold one or more of the beam control tray assemblies **400** in one or more slots, such as through use of a first tray receiver assembly side **622** through which the beam control tray assemblies **400** are inserted and/or through use of a second tray receiver assembly side **624** used as a backstop, as illustrated holding the plugin receptors configured to receive associated tray connector/communicators **430**, such as the first, second, and third receptors **626**, **627**, **628**. Optionally, the tray receiver assembly **620** retracts partially or completely into the dynamic gantry nozzle **610** using a retraction mechanism **660** configured to alternately retract and extend the tray

receiver assembly 620 relative to a nozzle end 612 of the gantry nozzle 610, such as along a first retraction track 662 and a second retraction track 664 using one or more motors and computer control. Optionally the tray receiver assembly 620 is partially or fully retracted when moving the gantry, nozzle, and/or gantry nozzle 610 to avoid physical constraints of movement, such as potential collision with another object in the patient treatment room.

[0133] For clarity of presentation and without loss of generality, several examples of loading patient specific tray inserts into tray assemblies with subsequent insertion into an positively charged particle beam path proximate a gantry nozzle 610 are provided.

[0134] In a first example, a single beam control tray assembly 400 is used to control the charged particle beam 268 in the charged particle cancer therapy system 100. In this example, a patient specific range shifter insert 511, which is custom fabricated for a patient, is loaded into a patient specific tray insert 510 to form a first tray assembly 402, where the first tray assembly 402 is loaded into the third receptor 628, which is fully retracted into the gantry nozzle 610.

[0135] In a second example, two beam control assemblies 400 are used to control the charged particle beam 268 in the charged particle cancer therapy system 100. In this example, a patient specific ridge filter 512 is loaded into a first tray assembly 402, which is loaded into the second receptor 627 and a patient specific aperture 513 is loaded into a second tray assembly 404, which is loaded into the first receptor 626 and the two associated tray connector/communicators 430 using the first receptor 626 and second receptor 627 communicate to the main controller 110 the patient specific tray inserts 510. The tray receiver assembly 620 is subsequently retracted one slot so that the patient specific ridge filter 512 and the patient specific aperture reside outside of and at the nozzle end 612 of the gantry nozzle 610.

[0136] In a third example, three beam control tray assemblies 400 are used, such as a range shifter 511 in a first tray inserted into the first receiving slot 403, a compensator in a second tray inserted into the second receiving slot 405, and an aperture in a third tray inserted into the third receiving slot 407.

[0137] Generally, any patient specific tray insert 510 is inserted into a tray frame 410 to form a beam control tray assembly 400 inserted into any slot of the tray receiver assembly 620 and the tray assembly is not retracted or retracted any distance into the gantry nozzle 610.

#### Tomography/Beam State

[0138] In one embodiment, the charged particle tomography apparatus is used to image a tumor in a patient. As current beam position determination/verification is used in both tomography and cancer therapy treatment, for clarity of presentation and without limitation beam state determination is also addressed in this section. However, beam state determination is optionally used separately and without tomography.

[0139] In another example, the charged particle tomography apparatus is used in combination with a charged particle cancer therapy system using common elements. For example, tomographic imaging of a cancerous tumor is performed using charged particles generated with an injector, accelerator, and guided with a delivery system that are part of the cancer therapy system, described supra.

[0140] In various examples, the tomography imaging system is optionally simultaneously operational with a charged particle cancer therapy system using common elements, allows tomographic imaging with rotation of the patient, is operational on a patient in an upright, semi-upright, and/or horizontal position, is simultaneously operational with X-ray imaging, and/or allows use of adaptive charged particle cancer therapy. Further, the common tomography and cancer therapy apparatus elements are optionally operational in a multi-axis and/or multi-field raster beam mode.

[0141] In conventional medical X-ray tomography, a sectional image through a body is made by moving one or both of an X-ray source and the X-ray film in opposite directions during the exposure. By modifying the direction and extent of the movement, operators can select different focal planes, which contain the structures of interest. More modern variations of tomography involve gathering projection data from multiple directions by moving the X-ray source and feeding the data into a tomographic reconstruction software algorithm processed by a computer. Herein, in stark contrast to known methods, the radiation source is a charged particle, such as a proton ion beam or a carbon ion beam. A proton beam is used herein to describe the tomography system, but the description applies to a heavier ion beam, such as a carbon ion beam. Further, in stark contrast to known techniques, herein the radiation source is preferably stationary while the patient is rotated.

[0142] Referring now to FIG. 7, an example of a tomography apparatus is described and an example of a beam state determination is described. In this example, the tomography system 700 uses elements in common with the charged particle beam system 100, including elements of one or more of the injection system 120, accelerator 130, targeting/delivery system 140, patient interface module 150, display system 160, and/or imaging system 170, such as the X-ray imaging system. The scintillation material is optionally one or more scintillation plates, such as a scintillating plastic, used to measure energy, intensity, and/or position of the charged particle beam. For instance, a scintillation material 710 or scintillation plate is positioned behind the patient 730 relative to the targeting/delivery system 140 elements, which is optionally used to measure intensity and/or position of the charged particle beam after transmitting through the patient. Optionally, a second scintillation plate or a charged particle induced photon emitting sheet, described infra, is positioned prior to the patient 730 relative to the targeting/delivery system 140 elements, which is optionally used to measure incident intensity and/or position of the charged particle beam prior to transmitting through the patient. The charged particle beam system 100 as described has proven operation at up to and including 330 MeV, which is sufficient to send protons through the body and into contact with the scintillation material. Particularly, 250 MeV to 330 MeV are used to pass the beam through a standard sized patient with a standard sized pathlength, such as through the chest. The intensity or count of protons hitting the plate as a function of position is used to create an image. The velocity or energy of the proton hitting the scintillation plate is also used in creation of an image of the tumor 720 and/or an image of the patient 730. The patient 730 is rotated about the y-axis and a new image is collected. Preferably, a new image is collected with about every one degree of rotation of the patient resulting in about 360 images that are combined into a tomogram using tomographic reconstruction software. The

tomographic reconstruction software uses overlapping rotationally varied images in the reconstruction. Optionally, a new image is collected at about every 2, 3, 4, 5, 10, 15, 30, or 45 degrees of rotation of the patient.

[0143] Herein, the scintillation material **710** or scintillator is any material that emits a photon when struck by a positively charged particle or when a positively charged particle transfers energy to the scintillation material sufficient to cause emission of light. Optionally, the scintillation material emits the photon after a delay, such as in fluorescence or phosphorescence. However, preferably, the scintillator has a fast fifty percent quench time, such as less than 0.0001, 0.001, 0.01, 0.1, 1, 10, 100, or 1,000 milliseconds, so that the light emission goes dark, falls off, or terminates quickly. Preferred scintillation materials include sodium iodide, potassium iodide, cesium iodide, an iodide salt, and/or a doped iodide salt. Additional examples of the scintillation materials include, but are not limited to: an organic crystal, a plastic, a glass, an organic liquid, a luminophor, and/or an inorganic material or inorganic crystal, such as barium fluoride,  $\text{BaF}_2$ ; calcium fluoride,  $\text{CaF}_2$ , doped calcium fluoride, sodium iodide, NaI; doped sodium iodide, sodium iodide doped with thallium, NaI(Tl); cadmium tungstate,  $\text{CdWO}_4$ ; bismuth germanate; cadmium tungstate,  $\text{CdWO}_4$ ; calcium tungstate,  $\text{CaWO}_4$ ; cesium iodide, CsI; doped cesium iodide; cesium iodide doped with thallium, CsI(Tl); cesium iodide doped with sodium CsI (Na); potassium iodide, KI; doped potassium iodide, gadolinium oxysulfide,  $\text{Gd}_2\text{O}_2\text{S}$ ; lanthanum bromide doped with cerium,  $\text{LaBr}_3(\text{Ce})$ ; lanthanum chloride,  $\text{LaCl}_3$ ; cesium doped lanthanum chloride,  $\text{LaCl}_3(\text{Ce})$ ; lead tungstate,  $\text{PbWO}_4$ ; LSO or lutetium oxyorthosilicate ( $\text{Lu}_2\text{SiO}_5$ ); LYSO,  $\text{Lu}_{1.8}\text{Y}_{0.2}\text{SiO}_5(\text{Ce})$ ; yttrium aluminum garnet, YAG (Ce); zinc sulfide,  $\text{ZnS}(\text{Ag})$ ; and zinc tungstate,  $\text{ZnWO}_4$ .

[0144] In one embodiment, a tomogram or an individual tomogram section image is collected at about the same time as cancer therapy occurs using the charged particle beam system **100**. For example, a tomogram is collected and cancer therapy is subsequently performed: without the patient moving from the positioning systems, such as in a semi-vertical partial immobilization system, a sitting partial immobilization system, or the a laying position. In a second example, an individual tomogram slice is collected using a first cycle of the accelerator **130** and using a following cycle of the accelerator **130**, the tumor **720** is irradiated, such as within about 1, 2, 5, 10, 15 or 30 seconds. In a third case, about 2, 3, 4, or 5 tomogram slices are collected using 1, 2, 3, 4, or more rotation positions of the patient **730** within about 5, 10, 15, 30, or 60 seconds of subsequent tumor irradiation therapy.

[0145] In another embodiment, the independent control of the tomographic imaging process and X-ray collection process allows simultaneous single and/or multi-field collection of X-ray images and tomographic images easing interpretation of multiple images. Indeed, the X-ray and tomographic images are optionally overlaid to form a hybrid X-ray/proton beam tomographic image as the patient **730** is optionally in the same position for each image.

[0146] In still another embodiment, the tomogram is collected with the patient **730** in the about the same position as when the patient's tumor is treated using subsequent irradiation therapy. For some tumors, the patient being positioned in the same upright or semi-upright position allows the tumor **720** to be separated from surrounding organs or

tissue of the patient **730** better than in a laying position. Positioning of the scintillation material **710** behind the patient **730** allows the tomographic imaging to occur while the patient is in the same upright or semi-upright position.

[0147] The use of common elements in the tomographic imaging and in the charged particle cancer therapy allows benefits of the cancer therapy, described supra, to optionally be used with the tomographic imaging, such as proton beam x-axis control, proton beam y-axis control, control of proton beam energy, control of proton beam intensity, timing control of beam delivery to the patient, rotation control of the patient, and control of patient translation all in a raster beam mode of proton energy delivery. The use of a single proton or cation beamline for both imaging and treatment facilitates eases patient setup, reduces alignment uncertainties, reduces beam state uncertainties, and eases quality assurance.

[0148] In yet still another embodiment, initially a three-dimensional tomographic proton based reference image is collected, such as with hundreds of individual rotation images of the tumor **720** and patient **730**. Subsequently, just prior to proton treatment of the cancer, just a few 2-dimensional control tomographic images of the patient are collected, such as with a stationary patient or at just a few rotation positions, such as an image straight on to the patient, with the patient rotated about 45 degrees each way, and/or the patient rotated about 90 degrees each way about the y-axis. The individual control images are compared with the 3-dimensional reference image. An adaptive proton therapy is subsequently performed where: (1) the proton cancer therapy is not used for a given position based on the differences between the 3-dimensional reference image and one or more of the 2-dimensional control images and/or (2) the proton cancer therapy is modified in real time based on the differences between the 3-dimensional reference image and one or more of the 2-dimensional control images.

#### Charged Particle State Determination/Verification/Photonic Monitoring

[0149] Still referring to FIG. 7, the tomography system **700** is optionally used with a charged particle beam state determination system **750**, optionally used as a charged particle verification system. The charged particle state determination system **750** optionally measures, determines, and/or verifies one of more of: (1) position of the charged particle beam, such as the treatment beam **269**, (2) direction of the treatment beam **269**, (3) intensity of the treatment beam **269**, (4) energy of the treatment beam **269**, (5) position, direction, intensity, and/or energy of the charged particle beam, such as a residual charged particle beam **267** after passing through a sample or the patient **730**, and (6) a history of the charged particle beam.

[0150] For clarity of presentation and without loss of generality, a description of the charged particle beam state determination system **750** is described and illustrated separately in FIG. 8 and FIG. 9A; however, as described herein elements of the charged particle beam state determination system **750** are optionally and preferably integrated into the nozzle system **146** and/or the tomography system **700** of the charged particle treatment system **100**. More particularly, any element of the charged particle beam state determination system **750** is integrated into the nozzle system **146**, the dynamic gantry nozzle **610**, and/or tomography system **700**, such as a surface of the scintillation material **710** or a surface of a scintillation detector, plate, or system. The nozzle

system **146** or the dynamic gantry nozzle **610** provides an outlet of the charged particle beam from the vacuum tube initiating at the injection system **120** and passing through the synchrotron **130** and beam transport system **135**. Any plate, sheet, fluorophore, or detector of the charged particle beam state determination system is optionally integrated into the nozzle system **146**. For example, an exit foil of the nozzle **610** is optionally a first sheet **760** of the charged particle beam state determination system **750** and a first coating **762** is optionally coated onto the exit foil, as illustrated in FIG. 7. Similarly, optionally a surface of the scintillation material **710** is a support surface for a fourth coating **792**, as illustrated in FIG. 7. The charged particle beam state determination system **750** is further described, infra.

[0151] Referring now to FIG. 7, FIG. 8, and FIG. 9A, four sheets, a first sheet **760**, a second sheet **770**, a third sheet **780**, and a fourth sheet **790** are used to illustrate detection sheets and/or photon emitting sheets upon transmittance of a charged particle beam. Each sheet is optionally coated with a photon emitter, such as a fluorophore, such as the first sheet **760** is optionally coated with a first coating **762**. Without loss of generality and for clarity of presentation, the four sheets are each illustrated as units, where the light emitting layer is not illustrated. Thus, for example, the second sheet **770** optionally refers to a support sheet, a light emitting sheet, and/or a support sheet coated by a light emitting element. The four sheets are representative of  $n$  sheets, where  $n$  is a positive integer.

[0152] Referring now to FIG. 7 and FIG. 8, the charged particle beam state verification system **750** is a system that allows for monitoring of the actual charged particle beam position in real-time without destruction of the charged particle beam. The charged particle beam state verification system **750** preferably includes a first position element or first beam verification layer, which is also referred to herein as a coating, luminescent, fluorescent, phosphorescent, radiance, or viewing layer. The first position element optionally and preferably includes a coating or thin layer substantially in contact with a sheet, such as an inside surface of the nozzle foil, where the inside surface is on the synchrotron side of the nozzle foil. Less preferably, the verification layer or coating layer is substantially in contact with an outer surface of the nozzle foil, where the outer surface is on the patient treatment side of the nozzle foil. Preferably, the nozzle foil provides a substrate surface for coating by the coating layer. Optionally, a binding layer is located between the coating layer and the nozzle foil, substrate, or support sheet. Optionally, the position element is placed anywhere in the charged particle beam path. Optionally, more than one position element on more than one sheet, respectively, is used in the charged particle beam path and is used to determine a state property of the charged particle beam, as described infra.

[0153] Still referring to FIG. 7 and FIG. 8, the coating, referred to as a fluorophore, yields a measurable spectroscopic response, spatially viewable by a detector or camera, as a result of transmission by the proton beam. The coating is preferably a phosphor, but is optionally any material that is viewable or imaged by a detector where the material changes spectroscopically as a result of the charged particle beam hitting or transmitting through the coating or coating layer. A detector or camera views secondary photons emitted from the coating layer and determines a position of a treatment beam **269**, which is also referred to as a current

position of the charged particle beam or final treatment vector of the charged particle beam, by the spectroscopic differences resulting from protons and/or charged particle beam passing through the coating layer. For example, the camera views a surface of the coating surface as the proton beam or positively charged cation beam is being scanned by the first axis control **143**, vertical control, and the second axis control **144**, horizontal control, beam position control elements during treatment of the tumor **720**. The camera views the current position of the charged particle beam or treatment beam **269** as measured by spectroscopic response. The coating layer is preferably a phosphor or luminescent material that glows and/or emits photons for a short period of time, such as less than 5 seconds for a 50% intensity, as a result of excitation by the charged particle beam. The detector observes the temperature change and/or observe photons emitted from the charged particle beam traversed spot. Optionally, a plurality of cameras or detectors are used, where each detector views all or a portion of the coating layer. For example, two detectors are used where a first detector views a first half of the coating layer and the second detector views a second half of the coating layer. Preferably, at least a portion of the detector is mounted into the nozzle system to view the proton beam position after passing through the first axis and second axis controllers **143**, **144**. Preferably, the coating layer is positioned in the proton beam path **268** in a position prior to the protons striking the patient **730**.

[0154] Referring now to FIG. 1 and FIG. 7, the main controller **110**, connected to the camera or detector output, optionally and preferably compares the final proton beam position or position of the treatment beam **269** with the planned proton beam position and/or a calibration reference to determine if the actual proton beam position or position of the treatment beam **269** is within tolerance. The charged particle beam state determination system **750** preferably is used in one or more phases, such as a calibration phase, a mapping phase, a beam position verification phase, a treatment phase, and a treatment plan modification phase. The calibration phase is used to correlate, as a function of  $x$ -,  $y$ -position of the glowing response the actual  $x$ -,  $y$ -position of the proton beam at the patient interface. During the treatment phase, the charged particle beam position is monitored and compared to the calibration and/or treatment plan to verify accurate proton delivery to the tumor **720** and/or as a charged particle beam shutoff safety indicator. Referring now to FIG. 10, the position verification system **179** and/or the treatment delivery control system **112**, upon determination of a tumor shift, an unpredicted tumor distortion upon treatment, and/or a treatment anomaly optionally generates and or provides a recommended treatment change **1070**. The treatment change **1070** is optionally sent out while the patient **730** is still in the treatment position, such as to a proximate physician or over the internet to a remote physician, for physician approval **1072**, receipt of which allows continuation of the now modified and approved treatment plan.

#### Example I

[0155] Referring now to FIG. 7, a first example of the charged particle beam state determination system **750** is illustrated using two cation induced signal generation surfaces, referred to herein as the first sheet **760** and a third sheet **780**. Each sheet is described below.

[0156] Still referring to FIG. 7, in the first example, the optional first sheet 760, located in the charged particle beam path prior to the patient 730, is coated with a first fluorophore coating 762, wherein a cation, such as in the charged particle beam, transmitting through the first sheet 760 excites localized fluorophores of the first fluorophore coating 762 with resultant emission of one or more photons. In this example, a first detector 812 images the first fluorophore coating 762 and the main controller 110 determines a current position of the charged particle beam using the image of the fluorophore coating 762 and the detected photon(s). The intensity of the detected photons emitted from the first fluorophore coating 762 is optionally used to determine the intensity of the charged particle beam used in treatment of the tumor 720 or detected by the tomography system 700 in generation of a tomogram and/or tomographic image of the tumor 720 of the patient 730. Thus, a first position and/or a first intensity of the charged particle beam is determined using the position and/or intensity of the emitted photons, respectively.

[0157] Still referring to FIG. 7, in the first example, the optional third sheet 780, positioned posterior to the patient 730, is optionally a cation induced photon emitting sheet as described in the previous paragraph. However, as illustrated, the third sheet 780 is a solid state beam detection surface, such as a detector array. For instance, the detector array is optionally a charge coupled device, a charge induced device, CMOS, or camera detector where elements of the detector array are read directly, as does a commercial camera, without the secondary emission of photons. Similar to the detection described for the first sheet, the third sheet 780 is used to determine a position of the charged particle beam and/or an intensity of the charged particle beam using signal position and/or signal intensity from the detector array, respectively.

[0158] Still referring to FIG. 7, in the first example, signals from the first sheet 760 and third sheet 780 yield a position before and after the patient 730 allowing a more accurate determination of the charged particle beam through the patient 730 therebetween. Optionally, knowledge of the charged particle beam path in the targeting/delivery system 740, such as determined via a first magnetic field strength across the first axis control 143 or a second magnetic field strength across the second axis control 144 is combined with signal derived from the first sheet 760 to yield a first vector of the charged particles prior to entering the patient 730 and/or an input point of the charged particle beam into the patient 730, which also aids in: (1) controlling, monitoring, and/or recording tumor treatment and/or (2) tomography development/interpretation. Optionally, signal derived from use of the third sheet 780, posterior to the patient 730, is combined with signal derived from tomography system 700, such as the scintillation material 710, to yield a second vector of the charged particles posterior to the patient 730 and/or an output point of the charged particle beam from the patient 730, which also aids in: (1) controlling, monitoring, deciphering, and/or (2) interpreting a tomogram or a tomographic image.

[0159] For clarity of presentation and without loss of generality, detection of photons emitted from sheets is used to further describe the charged particle beam state determination system 750. However, any of the cation induced photon emission sheets described herein are alternatively detector arrays. Further, any number of cation induced

photon emission sheets are used prior to the patient 730 and/or posterior to the patient 730, such as 1, 2, 3, 4, 6, 8, 10, or more. Still further, any of the cation induced photon emission sheets are placed anywhere in the charged particle beam, such as in the synchrotron 130, in the beam transport system 135, in the targeting/delivery system 140, the nozzle 146, in the gantry room, and/or in the tomography system 700. Any of the cation induced photon emission sheets are used in generation of a beam state signal as a function of time, which is optionally recorded, such as for an accurate history of treatment of the tumor 720 of the patient 730 and/or for aiding generation of a tomographic image.

#### Example II

[0160] Referring now to FIG. 8, a second example of the charged particle beam state determination system 750 is illustrated using three cation induced signal generation surfaces, referred to herein as the second sheet 770, the third sheet 780, and the fourth sheet 790. Any of the second sheet 770, the third sheet 780, and the fourth sheet 790 contain any of the features of the sheets described supra.

[0161] Still referring to FIG. 8, in the second example, the second sheet 770, positioned prior to the patient 730, is optionally integrated into the nozzle 146, but is illustrated as a separate sheet. Signal derived from the second sheet 770, such as at point A, is optionally combined with signal from the first sheet 760 and/or state of the targeting/delivery system 140 to yield a first vector,  $v_{1a}$ , from point A to point B of the charged particle beam prior to the sample or patient 730 at a first time,  $t_1$ , and a second vector,  $v_{2a}$ , from point F to point G of the charged particle beam prior to the sample at a second time,  $t_2$ .

[0162] Still referring to FIG. 8, in the second example, the third sheet 780 and the fourth sheet 790, positioned posterior to the patient 730, are optionally integrated into the tomography system 700, but are illustrated as separate sheets. Signal derived from the third sheet 780, such as at point D, is optionally combined with signal from the fourth sheet 790 and/or signal from the tomography system 700 to yield a first vector,  $v_{1b}$ , from point C<sub>2</sub> to point D and/or from point D to point E of the charged particle beam posterior to the patient 730 at the first time,  $t_1$ , and a second vector,  $v_{2a}$ , such as from point H to point I of the charged particle beam posterior to the sample at a second time,  $t_2$ . Signal derived from the third sheet 780 and/or from the fourth sheet 790 and the corresponding first vector at the second time,  $t_2$ , is used to determine an output point, C<sub>2</sub>, which may and often does differ from an extension of the first vector,  $v_{1a}$ , from point A to point B through the patient to a non-scattered beam path of point C<sub>1</sub>. The difference between point C<sub>1</sub> and point C<sub>2</sub> and/or an angle,  $\alpha$ , between the first vector at the first time,  $v_{1a}$ , and the first vector at the second time,  $v_{1b}$ , is used to determine/map/identify, such as via tomographic analysis, internal structure of the patient 730, sample, and/or the tumor 720, especially when combined with scanning the charged particle beam in the x/y-plane as a function of time, such as illustrated by the second vector at the first time,  $v_{2a}$ , and the second vector at the second time,  $v_{2b}$ , forming angle  $\beta$  and/or with rotation of the patient 730, such as about the y-axis, as a function of time.

[0163] Still referring to FIG. 8, multiple detectors/detector arrays are illustrated for detection of signals from multiple sheets, respectively. However, a single detector/detector array is optionally used to detect signals from multiple

sheets, as further described infra. As illustrated, a set of detectors **810** is illustrated, including a second detector **814** imaging the second sheet **770**, a third detector **816** imaging the third sheet **780**, and a fourth detector **818** imaging the fourth sheet **790**. Any of the detectors described herein are optionally detector arrays, are optionally coupled with any optical filter, and/or optionally use one or more intervening optics to image any of the four sheets **760**, **770**, **780**, **790**. Further, two or more detectors optionally image a single sheet, such as a region of the sheet, to aid optical coupling, such as F-number optical coupling.

[0164] Still referring to FIG. **8**, a vector of the charged particle beam is determined. Particularly, in the illustrated example, the third detector **816**, determines, via detection of secondary emitted photons, that the charged particle beam transmitted through point D and the fourth detector **818** determines that the charged particle beam transmitted through point E, where points D and E are used to determine the first vector at the second time,  $v_{1b}$ , as described supra. To increase accuracy and precision of a determined vector of the charged particle beam, a first determined beam position and a second determined beam position are optionally and preferably separated by a distance,  $d_1$ , such as greater than 0.1, 0.5, 1, 2, 3, 5, 10, or more centimeters. A support element **752** is illustrated that optionally connects any two or more elements of the charged particle beam state determination system **750** to each other and/or to any element of the charged particle beam system **100**, such as a rotating platform **756** used to co-rotate the patient **730** and any element of the tomography system **700**.

### Example III

[0165] Still referring to FIG. **9A**, a third example of the charged particle beam state determination system **750** is illustrated in an integrated tomography-cancer therapy system **900**.

[0166] Referring to FIG. **9A**, multiple sheets and multiple detectors are illustrated determining a charged particle beam state prior to the patient **730**. As illustrated, a first camera **812** spatially images photons emitted from the first sheet **760** at point A, resultant from energy transfer from the passing charged particle beam, to yield a first signal and a second camera **814** spatially images photons emitted from the second sheet **770** at point B, resultant from energy transfer from the passing charged particle beam, to yield a second signal. The first and second signals allow calculation of the first vector,  $v_{1a}$ , with a subsequent determination of an entry point **732** of the charged particle beam into the patient **730**. Determination of the first vector,  $v_{1a}$ , is optionally supplemented with information derived from states of the magnetic fields about the first axis control **143**, the vertical control, and the second axis control **144**, the horizontal axis control, as described supra.

[0167] Still referring to FIG. **9A**, the charged particle beam state determination system is illustrated with multiple resolvable wavelengths of light emitted as a result of the charged particle beam transmitting through more than one molecule type, light emission center, and/or fluorophore type. For clarity of presentation and without loss of generality a first fluorophore in the third sheet **780** is illustrated as emitting blue light, b, and a second fluorophore in the fourth sheet **790** is illustrated as emitting red light, r, that are both detected by the third detector **816**. The third detector is optionally coupled with any wavelength separation device,

such as an optical filter, grating, or Fourier transform device. For clarity of presentation, the system is described with the red light passing through a red transmission filter blocking blue light and the blue light passing through a blue transmission filter blocking red light. Wavelength separation, using any means, allows one detector to detect a position of the charged particle beam resultant in a first secondary emission at a first wavelength, such as at point C, and a second secondary emission at a second wavelength, such as at point D. By extension, with appropriate optics, one camera is optionally used to image multiple sheets and/or sheets both prior to and posterior to the sample. Spatial determination of origin of the red light and the blue light allow calculation of the first vector at the second time,  $V_{1b}$ , and an actual exit point **736** from the patient **730** as compared to a non-scattered exit point **734** from the patient **730** as determined from the first vector at the first time,  $V_{1a}$ .

[0168] Still referring to FIG. **9A** and referring now to FIG. **9B**, the integrated tomography-cancer therapy system **900** is illustrated with an optional configuration of elements of the charged particle beam state determination system **750** being co-rotatable with the nozzle **146** of the cancer therapy system **100**. More particularly, in one case sheets of the charged particle beam state determination system **750** positioned prior to, posterior to, or on both sides of the patient **730** co-rotate with the scintillation material **710** about any axis, such as illustrated with rotation about the y-axis. Further, any element of the charged particle beam state determination system **750**, such as a detector, two-dimensional detector, multiple two-dimensional detectors, and/or light coupling optic move as the gantry moves, such as along a common arc of movement of the nozzle **146** and/or at a fixed distance to the common arc. For instance, as the gantry moves, a monitoring camera positioned on the opposite side of the tumor **720** or patient **730** from the nozzle **146** maintains a position on the opposite side of the tumor **720** or patient **730**. In various cases, co-rotation is achieved by co-rotation of the gantry of the charged particle beam system and a support of the patient, such as the rotatable platform **756**, which is also referred to herein as a movable or dynamically positionable patient platform, patient chair, or patient couch. Mechanical elements, such as the support element **752** affix the various elements of the charged particle beam state determination system **750** relative to each other, relative to the nozzle **146**, and/or relative to the patient **730**. For example, the support elements **752** maintain a second distance,  $d_2$ , between a position of the tumor **720** and the third sheet **780** and/or maintain a third distance,  $d_3$ , between a position of the third sheet **780** and the scintillation material **710**. More generally, support elements **752** optionally dynamically position any element about the patient **730** relative to one another or in x,y,z-space in a patient diagnostic/treatment room, such as via computer control.

[0169] Referring now to FIG. **9B**, positioning the nozzle **146** of a gantry **960** on an opposite side of the patient **730** from a detection surface, such as the scintillation material **710**, in a gantry movement system **950** is described. Generally, in the gantry movement system **950**, as the gantry **960** rotates about an axis the nozzle **146** and/or one or more magnets of the beam transport system **135** are repositioned. As illustrated, the nozzle **146** is positioned by the gantry **960** in a first position at a first time,  $t_1$ , and in a second position at a second time,  $t_2$ , where n positions are optionally possible. An electromechanical system, such as a patient

table, patient couch, patient couch, patient rotation device, and/or a scintillation plate holder maintains the patient **730** between the nozzle **146** and the scintillation material **710** of the tomography system **700**. Similarly, not illustrated for clarity of presentation, the electromechanical system maintains a position of the third sheet **780** and/or a position of the fourth sheet **790** on a posterior or opposite side of the patient **730** from the nozzle **146** as the gantry **960** rotates or moves the nozzle **146**. Similarly, the electromechanical system maintains a position of the first sheet **760** or first screen and/or a position of the second sheet **770** or second screen on a same or prior side of the patient **730** from the nozzle **146** as the gantry **960** rotates or moves the nozzle **146**. As illustrated, the electromechanical system optionally positions the first sheet **760** in the positively charged particle path at the first time,  $t_1$ , and rotates, pivots, and/or slides the first sheet **760** out of the positively charged particle path at the second time,  $t_2$ . The electromechanical system is optionally and preferably connected to the main controller **110** and/or the treatment delivery control system **112**. The electromechanical system optionally maintains a fixed distance between: (1) the patient and the nozzle **146** or the nozzle end **612**, (2) the patient **730** or tumor **720** and the scintillation material **710**, and/or (3) the nozzle **146** and the scintillation material **710** at a first treatment time with the gantry **960** in a first position and at a second treatment time with the gantry **960** in a second position. Use of a common charged particle beam path for both imaging and cancer treatment and/or maintaining known or fixed distances between beam transport/guide elements and treatment and/or detection surface enhances precision and/or accuracy of a resultant image and/or tumor treatment, such as described supra.

#### System Integration

[0170] Any of the systems and/or elements described herein are optionally integrated together and/or are optionally integrated with known systems.

#### Treatment Delivery Control System

[0171] Referring now to FIG. 10, a centralized charged particle treatment system **1000** is illustrated. Generally, once a charged particle therapy plan is devised, a central control system or treatment delivery control system **112** is used to control sub-systems while reducing and/or eliminating direct communication between major subsystems. Generally, the treatment delivery control system **112** is used to directly control multiple subsystems of the cancer therapy system without direct communication between selected subsystems, which enhances safety, simplifies quality assurance and quality control, and facilitates programming. For example, the treatment delivery control system **112** directly controls one or more of: an imaging system, a positioning system, an injection system, a radio-frequency quadrupole system, a linear accelerator, a ring accelerator or synchrotron, an extraction system, a beam line, an irradiation nozzle, a gantry, a display system, a targeting system, and a verification system. Generally, the control system integrates subsystems and/or integrates output of one or more of the above described cancer therapy system elements with inputs of one or more of the above described cancer therapy system elements.

[0172] Still referring to FIG. 10, an example of the centralized charged particle treatment system **1000** is provided.

Initially, a doctor, such as an oncologist, prescribes **1010** or recommends tumor therapy using charged particles. Subsequently, treatment planning **1020** is initiated and output of the treatment planning step **1020** is sent to an oncology information system **1030** and/or is directly sent to the treatment delivery system **112**, which is an example of the main controller **110**.

[0173] Still referring to FIG. 10, the treatment planning step **1020** is further described. Generally, radiation treatment planning is a process where a team of oncologist, radiation therapists, medical physicists, and/or medical dosimetrists plan appropriate charged particle treatment of a cancer in a patient. Typically, one or more imaging systems **170** are used to image the tumor and/or the patient, described infra. Planning is optionally: (1) forward planning and/or (2) inverse planning. Cancer therapy plans are optionally assessed with the aid of a dose-volume histogram, which allows the clinician to evaluate the uniformity of the dose to the tumor and surrounding healthy structures. Typically, treatment planning is almost entirely computer based using patient computed tomography data sets using multimodality image matching, image coregistration, or fusion.

#### Forward Planning

[0174] In forward planning, a treatment oncologist places beams into a radiotherapy treatment planning system including: how many radiation beams to use and which angles to deliver each of the beams from. This type of planning is used for relatively simple cases where the tumor has a simple shape and is not near any critical organs.

#### Inverse Planning

[0175] In inverse planning, a radiation oncologist defines a patient's critical organs and tumor and gives target doses and importance factors for each. Subsequently, an optimization program is run to find the treatment plan which best matches all of the input criteria.

#### Oncology Information System

[0176] Still referring to FIG. 10, the oncology information system **1030** is further described. Generally, the oncology information system **1030** is one or more of: (1) an oncology-specific electronic medical record, which manages clinical, financial, and administrative processes in medical, radiation, and surgical oncology departments; (2) a comprehensive information and image management system; and (3) a complete patient information management system that centralizes patient data; and (4) a treatment plan provided to the charged particle beam system **100**, main controller **110**, and/or the treatment delivery control system **112**. Generally, the oncology information system **1030** interfaces with commercial charged particle treatment systems.

#### Safety System/Treatment Delivery Control System

[0177] Still referring to FIG. 10, the treatment delivery control system **112** is further described. Generally, the treatment delivery control system **112** receives treatment input, such as a charged particle cancer treatment plan from the treatment planning step **1020** and/or from the oncology information system **1030** and uses the treatment input and/or treatment plan to control one or more subsystems of the charged particle beam system **100**. The treatment delivery control system **112** is an example of the main controller **110**,

where the treatment delivery control system receives subsystem input from a first subsystem of the charged particle beam system **100** and provides to a second subsystem of the charged particle beam system **100**: (1) the received subsystem input directly, (2) a processed version of the received subsystem input, and/or (3) a command, such as used to fulfill requisites of the treatment planning step **1020** or direction of the oncology information system **1030**. Generally, most or all of the communication between subsystems of the charged particle beam system **100** go to and from the treatment delivery control system **112** and not directly to another subsystem of the charged particle beam system **100**. Use of a logically centralized treatment delivery control system has many benefits, including: (1) a single centralized code to maintain, debug, secure, update, and to perform checks on, such as quality assurance and quality control checks; (2) a controlled logical flow of information between subsystems; (3) an ability to replace a subsystem with only one interfacing code revision; (4) room security; (5) software access control; (6) a single centralized control for safety monitoring; and (7) that the centralized code results in an integrated safety system **1040** encompassing a majority or all of the subsystems of the charged particle beam system **100**. Examples of subsystems of the charged particle cancer therapy system **100** include: a radio frequency quadrupole **1050**, a radio frequency quadrupole linear accelerator, the injection system **120**, the synchrotron **130**, the accelerator system **131**, the extraction system **134**, any controllable or monitorable element of the beam line **268**, the targeting/delivery system **140**, the nozzle **146**, a gantry **1060** or an element of the gantry **1060**, the patient interface module **150**, a patient positioner **152**, the display system **160**, the imaging system **170**, a patient position verification system **179**, any element described supra, and/or any subsystem element. A treatment change **1070** at time of treatment is optionally computer generated with or without the aid of a technician or physician and approved while the patient is still in the treatment room, in the treatment chair, and/or in a treatment position.

#### Safety

[0178] Referring now to FIG. **11**, a redundant safety system **1100** is described. In one optional and preferred embodiment, the charged particle beam system **100** includes redundant systems for determination of one or more of: (1) beam position, (2) beam direction, (3) beam intensity, (4) beam energy, and (5) beam shape. The redundant safety system **1000** is further described herein.

#### Beam Position

[0179] A beam positioning system **1110** or beam position determination/verification system is linked to the main controller **100** or treatment delivery control system **112**. The beam positioning system **1110** includes any electromechanical system, optical system, and/or calculation for determining a current position of the charged particle beam. In a first case, after calibration, the scanning/targeting/delivery system **140** uses x/y-positioning magnets, such as in the first axis control **143** and the second axis control **144**, to position the charged particle beam. In a second case, a photonic emission position system **1114** is used to measure a position of the charged particle beam, where the photonic emission system **1114** uses a secondary emission of a photon upon

passage of the charged particle beam, such as described supra for the first sheet **760**, the second sheet **770**, the third sheet **780**, and the fourth sheet **790**. In a third case, a scintillation positioning system **1116**, such as via use of a detector element in the tomography system **700**, is used to measure a position of the charged particle beam. Any permutation or combination of the three cases described herein yield multiple or redundant measures of the charged particle beam position and therefrom one or more measures of a charged particle beam vector during a period of time.

#### Beam Intensity

[0180] A beam intensity system **1120** or beam intensity determination/verification system is linked to the main controller **100** or treatment delivery control system **112**. Herein, intensity is a number of positively charged particles passing a point or plane as a function of time. The beam intensity system **1110** includes any electromechanical system, optical system, and/or calculation for determining a current intensity of the charged particle beam. In a first case, the extraction system **134** uses an electron emission system **1122**, such as a secondary emission of electrons upon passage of the charged particle beam through the extraction material **330**, to determine an intensity of the charged particle beam. In a second case, the duration of the applied RF-field and/or a magnitude of the RF-field applied in the RF-cavity system **310** is used to calculate the intensity of the charged particle beam, as described supra. In a third case, a photon emission system **1124**, such as a magnitude of a signal representing the emitted photons from the photonic emission system **1114**, is used to measure the intensity of the charged particle beam. In a fourth case, a scintillation intensity determination system **1126** measures the intensity of the charged particle beam, such as with a detector of the tomography system **700**.

#### Beam Energy

[0181] A beam energy system **1130** or beam energy determination/verification system is linked to the main controller **100** or treatment delivery control system **112**. Herein, energy is optionally referred to as a velocity of the positively charged particles passing a point, where energy is dependent upon mass of the charged particles. The beam energy system **1110** includes any electromechanical system, optical system, and/or calculation for determining a current energy of the charged particle beam. In a first case, an RF-cavity energy system **1132** calculates an energy of the charged particles in the charged particle beam, such as via relating a period of an applied RF-field in the RF-cavity system **310** to energy, such as described supra. In a second case, an in-line energy system **1134** is used to measure a value related to beam energy, such as described above in equations 1 and 2. In a third case, a scintillation energy system **1136** is used to measure an energy of the charged particle beam, such as via use of a detector in the tomography system **700**.

[0182] Optionally and preferably, two or more measures/determination/calculations of a beam state property, such as position, direction, shape, intensity, and/or energy yield a redundant measure of the measured state for use in a beam safety system and/or an emergency beam shut-off system. Optionally and preferably, the two or more measures of a beam state property are used to enhance precision and/or accuracy of determination of the beam state property

through statistical means. Optionally and preferably, any of the beam state properties are recorded and/or used to predict a future state, such as position, intensity, and/or energy of the charged particle beam, such as in a neighboring voxel in the tumor 720 adjacent to a currently treated voxel in the tumor 720 of the patient 730.

#### Motion Control System

[0183] Referring now to FIG. 12A, a motion control system 1200 is illustrated. Generally, the motion control system controls, as a function of time: (1) the charged particle beam state, such as direction, shape, intensity, and/or energy; (2) a patient position; and/or (3) an imaging system. The motion control system 1200 is further described herein.

[0184] The motion control system 1200 optionally uses one or more patient interface controllers 1210, such as an external motion control system 1212, an internal motion control system 1214, an external pendant 1216, and an internal pendant 1218. As illustrated, the patient 730 is in a treatment room 1222 separated from a control room 1224 by a radiation shielded wall 1226 and a window 1228 or view port. The external motion control system 1212, internal motion control system 1214, external pendant 1216, and the internal pendant 1218 optionally and preferably control the same elements, allowing one or more operators control of the motion control system. Any of the patient interface controllers 1210 are optionally linked to each other or to the main controller 110 via wireless means; however, interconnections of the patient interface controllers 1210 to each other and/or to the main controller 110 are preferably hard-wired due to high radiation levels in the treatment room 1222. For example, the external pendant 1216 is linked via a first communication bundle 1217 to the external motion control system 1212, the internal pendant 1218 is linked via a second communication bundle 1219 to the internal motion system controller 1214, and/or the internal and external motion control system 1212, 1214 are hardwired to each other and/or to the main controller 110. The first communication bundle 1217 and the second communication bundle 1219 optionally provide power to the external pendant 1216 and the internal pendant 1218, respectively. The second communication bundle 1219 is optionally attached and/or linked to the nozzle system 146 and/or an element of the beam transport system 135 to keep the second communication bundle: (1) accessible to the operator, (2) out of the way of the charged particle beam, and/or (3) out of the way of motion of the patient 730/patient interface module 150. Optionally, a patient specific treatment module 1290 is replaceably plugged into and/or attached to the one or more patient interface controllers 1210, such as the internal pendant 1218. The patient treatment module 1290 optionally contains one or more of: image information about the individual being treated and/or preprogrammed treatment steps for the individual being treated, where some controls of the charged particle beam system 100, such as related to charged particle beam aiming and/or patient positioning are optionally limited by the preprogrammed treatment steps of any information/hardware of the patient treatment module. Optionally, the internal pendant 1218 replaceably mounts to a bracket, hook, slot, or the like mounted on the nozzle system 146 or the beam transport system 135 to maintain close access for the operator when not in use. The operator optionally and preferably uses, at times, a mobile control

pendant, such as the external pendant 1216 or the internal pendant 1218. The operator optionally has access via a direct doorway 1229 between treatment room 1222 and the control room 1224. Use of multiple patient interface controllers 1210 gives flexibility to an operator of the motion control system 1200, as further described infra.

#### Example I

[0185] In a first example, the operator of the motion control system 1200 is optionally seated or standing by a fixed position controller, such as by a desktop or wall mounted version of the external motion control system 1212. Similarly, the internal motion control system 1214 is optionally and preferably in a fixed position, such as at a desktop system or wall mounted system.

#### Example II

[0186] In a second example, the operator optionally and preferably uses, at times, the external pendant 1216, which allows the operator to view the patient 730, the beam transport system 135, beam path housing 139, the patient interface module 150, and/or the imaging system 170 through the safety of the window 1228. Optionally and preferably, the beam transport system 135 is configured with one or more mechanical stops to not allow the charged particle beam to aim at the window 1228, thereby providing a continuously safe zone for the operator. Direct viewing and control of the charged particle beam system 100, imaging system 170, and/or tomography system 700 relative to the current position of the patient 730 allows backup security in terms of unexpected aim of a treatment beam and/or movement of the patient 730. Controlled elements and/or processes of the charged particle beam system 100 via the pendants is further described, infra.

#### Example III

[0187] In a third example, the operator optionally and preferably uses, at times, the internal pendant 1218, which allows the operator both direct access and view of: (1) the patient 730, (2) the beam transport system 135, (3) the patient interface module 150, and/or (4) the imaging system 170, which has multiple benefits. In a first case, the operator can adjust any element of the patient interface module 150, such as a patient positioning device and/or patient motion constraint device. In a second case, the operator has access to load/unload: (1) the patient specific tray insert 510 into the beam control tray assembly 400; (2) the beam control tray assembly 400 into the nozzle system 146, as described supra; and/or (3) any imaging material, such as an X-ray film.

#### Example IV

[0188] In a fourth example, the gantry comprises at least two imaging devices, where each imaging device moves with rotation of the gantry and where the two imaging devices view the patient 730 along two axes forming an angle of ninety degrees, in the range of eighty-five to ninety-five degrees, and/or in the range of seventy-five to one hundred five degrees.

## Pendant

[0189] Referring still to FIG. 12A and referring now to FIG. 12B, a pendant system 1250, such as a system using the external pendant 1216 and/or internal pendant 1218 is described. In a first case, the external pendant 1216 and internal pendant 1218 have identical controls. In a second case, controls and/or functions of the external pendant 1216 intersect with controls and/or function of the internal pendant 1218. Particular processes and functions of the internal pendant 1218 are provided below, without loss of generality, to facilitate description of the external and internal pendants 1216, 1218. The internal pendant 1218 optionally comprises any number of input buttons, screens, tabs, switches, or the like. The pendant system 1250 is further described, infra.

## Example I

[0190] Referring now to FIG. 12B, a first example of the internal pendant 1218 is provided. In this example, in place of and/or in conjunction with a particular button, such as a first button 1270 and/or a second button 1280, moving or selecting a particular element, processes are optionally described, displayed, and/or selected within a flow process control unit 1260 of the internal pendant 1218. For example, one or more display screens and/or printed elements describe a set of processes, such as a first process 1261, a second process 1263, a third process 1265, and a fourth process 1267 and are selected through a touch screen selection process or via a selection button, such as a corresponding first selector 1262, second selector 1264, third selector 1266, and fourth selector 1268. Optionally, a next button a-priori or previously scheduled in treatment planning to select a next process is lit up on the pendant.

## Example II

[0191] Referring still to FIG. 12B, a second example of the internal pendant 1218 is provided. In this example, one or more buttons or the like, such as the first button 1270, and/or one or more of the processes, such as the first process 1261, are customizable, such as to an often repeated set of steps and/or to steps particular to treatment of a given patient 730. The customizable element, such as the first button 1270, is optionally further setup, programmed, controlled, and/or limited via information received from the patient treatment module 1290. In this example, a button, or the like, operates as an emergency all stop button, which at the minimum shuts down the accelerator, redirects the charged particle beam to a beam stop separate from a path through the patient, or stops moving the patient 730.

## Example III

[0192] In place of and/or in conjunction with a particular button, such as the first button 1270 and/or the second button 1280, moving or selecting a particular element, processes are optionally described, displayed, and/or selected within a flow process control unit 1260 of the internal pendant 1218. For example, one or more display screens and/or printed elements describe a set of processes, such as a first process 1261, a second process 1263, a third process 1265, and a fourth process 1267 and are selected through a touch screen selection process or via a selection button, such as a corresponding first selector 1262, second selector 1264, third selector 1266, and fourth selector 1268.

[0193] Referring still to FIG. 12B, as illustrated for clarity and without loss or generalization, the first process 1261 and/or a display screen thereof operable by the first selector 1262 selects, initiates, and/or processes a set of steps related to the beam control tray assembly 400. For instance, the first selector 1262, functioning as a tray button: (1) confirms presence a requested patient specific tray insert 510 in a requested tray assembly; (2) confirms presence of a request patient specific tray insert in a receiving slot of the control tray assembly; (3) retracts the beam control tray assembly 400 into the nozzle system 146; (4) confirms information using the electromechanical identifier plug, such as the first electromechanical identifier plug 530; (5) confirms information using the patient treatment module 1290; and/or (6) performs a set of commands and/or movements identified with the first selector 1262 and/or identified with the first process 1261. Similarly, the second process 1263, corresponding to a second process display screen and/or the second selector 1264; the third process 1265, corresponding to a third process display screen and/or the third selector 1266; and the fourth process 1267, corresponding to a fourth process display screen and/or the fourth selector 1268 control and/or activate a set of actions, movements, and/or commands related to positioning the patient 730, imaging the patient 730, and treating the patient 730, respectively.

## Integrated Cancer Treatment-Imaging System

[0194] One or more imaging systems 170 are optionally used in a fixed position in a cancer treatment room and/or are moved with a gantry system, such as a gantry system supporting: a portion of the beam transport system 135, the targeting/delivery control system 140, and/or moving or rotating around a patient positioning system, such as in the patient interface module. Without loss of generality and to facilitate description of the invention, examples follow of an integrated cancer treatment-imaging system. In each system, the beam transport system 135 and/or the nozzle system 146 indicates a positively charged beam path, such as from the synchrotron, for tumor treatment and/or for tomography, as described supra.

## Example I

[0195] Referring now to FIG. 13A, a first example of an integrated cancer treatment-imaging system 1300 is illustrated. In this example, the charged particle beam system 100 is illustrated with a treatment beam 269 directed to the tumor 720 of the patient 730 along the z-axis. Also illustrated is a set of imaging sources 1310, imaging system elements, and/or paths therefrom and a set of detectors 1320 corresponding to a respective element of the set of imaging sources 1310. Herein, the set of imaging sources 1310 are referred to as sources, but are optionally any point or element of the beam train prior to the tumor or a center point about which the gantry rotates. Hence, a given imaging source is optionally a dispersion element used to for cone beam. As illustrated, a first imaging source 1312 yields a first beam path 1332 and a second imaging source 1314 yields a second beam path 1334, where each path passes at least into the tumor 720 and optionally and preferably to a first detector array 1322 and a second detector array 1324, respectively, of the set of detectors 1320. Herein, the first beam path 1332 and the second beam path 1334 are illustrated as forming a ninety degree angle, which yields

complementary images of the tumor **720** and/or the patient **730**. However, the formed angle is optionally any angle from ten to three hundred fifty degrees. Herein, for clarity of presentation, the first beam path **1332** and the second beam path **1334** are illustrated as single lines, which optionally is an expanding, uniform diameter, or focusing beam. Herein, the first beam path **1332** and the second beam path **1334** are illustrated in transmission mode with their respective sources and detectors on opposite sides of the patient **730**. However, a beam path from a source to a detector is optionally a scattered path and/or a diffuse reflectance path. Optionally, one or more detectors of the set of detectors **1320** are a single detector element, a line of detector elements, or preferably a two-dimensional detector array. Use of two two-dimensional detector arrays is referred to herein as a two-dimensional-two-dimensional imaging system or a 2D-2D imaging system.

[0196] Still referring to FIG. **13A**, the first imaging source **1312** and the second imaging source **1314** are illustrated at a first position and a second position, respectively. Each of the first imaging source **1312** and the second imaging source **1322** optionally: (1) maintain a fixed position; (2) provide the first beam path **1332** and the second beam path **1334**, respectively, through the gantry **960**, such as through a set of one or more holes or slits; (3) provide the first beam path **1332** and the second beam path **1334**, respectively, off axis to a plane of movement of the nozzle system **760**; (4) move with the gantry **960** as the gantry **960** rotates about at least a first axis; (5) move with a secondary imaging system independent of movement of the gantry, as described supra; and/or (5) represent a narrow cross-diameter section of an expanding cone beam path.

[0197] Still referring to FIG. **13A**, the set of detectors **1320** are illustrated as coupling with respective elements of the set of sources **1310**. Each member of the set of detectors **1320** optionally and preferably co-moves/and/or co-rotates with a respective member of the set of sources **1310**. Thus, if the first imaging source **1312** is statically positioned, then the first detector **1322** is optionally and preferably statically positioned. Similarly, to facilitate imaging, if the first imaging source **1312** moves along a first arc as the gantry **960** moves, then the first detector **1322** optionally and preferably moves along the first arc or a second arc as the gantry **960** moves, where relative positions of the first imaging source **1312** on the first arc, a point that the gantry **960** moves about, and relative positions of the first detector **1322** along the second arc are constant. To facilitate the process, the detectors are optionally mechanically linked, such as with a mechanical support to the gantry **960** in a manner that when the gantry **960** moves, the gantry moves both the source and the corresponding detector. Optionally, the source moves and a series of detectors, such as along the second arc, capture a set of images. As illustrated in FIG. **13A**, the first imaging source **1312**, the first detector array **1322**, the second imaging source **1314**, and the second detector array **1324** are coupled to a rotatable imaging system support **1812**, which optionally rotates independently of the gantry **960** as further described infra. As illustrated in FIG. **13B**, the first imaging source **1312**, the first detector array **1322**, the second imaging source **1314**, and the second detector array **1324** are coupled to the gantry **960**, which in this case is a rotatable gantry.

[0198] Still referring to FIG. **13A**, optionally and preferably, elements of the set of sources **1310** combined with

elements of the set of detectors **1320** are used to collect a series of responses, such as one source and one detector yielding a detected intensity and rotatable imaging system support **1812** preferably a set of detected intensities to form an image. For instance, the first imaging source **1312**, such as a first X-ray source or first cone beam X-ray source, and the first detector **1322**, such as an X-ray film, digital X-ray detector, or two-dimensional detector, yield a first X-ray image of the patient at a first time and a second X-ray image of the patient at a second time, such as to confirm a maintained location of a tumor or after movement of the gantry **760** or rotation of the patient **730**. A set of  $n$  images using the first imaging source **1312** and the first detector **1322** collected as a function of movement of the gantry **760** and/or as a function of movement and/or rotation of the patient **730** are optionally and preferably combined to yield a three-dimensional image of the patient **730**, such as a three-dimensional X-ray image of the patient **730**, where  $n$  is a positive integer, such as greater than 1, 2, 3, 4, 5, 10, 15, 25, 50, or 100. The set of  $n$  images is optionally gathered as described in combination with images gathered using the second imaging source **1314**, such as a second X-ray source or second cone beam X-ray source, and the second detector **1324**, such as a second X-ray detector, where the use of two, or multiple, source/detector combinations are combined to yield images where the patient **730** has not moved between images as the two, or the multiple, images are optionally and preferably collected at the same time, such as with a difference in time of less than 0.01, 0.1, 1, or 5 seconds. Longer time differences are optionally used. Preferably the  $n$  two-dimensional images are collected as a function of rotation of the gantry **960** about the tumor and/or the patient and/or as a function of rotation of the patient **730** and the two-dimensional images of the X-ray cone beam are mathematically combined to form a three-dimensional image of the tumor **720** and/or the patient **730**. Optionally, the first X-ray source and/or the second X-ray source is the source of X-rays that are divergent forming a cone through the tumor. A set of images collected as a function of rotation of the divergent X-ray cone around the tumor with a two-dimensional detector that detects the divergent X-rays transmitted through the tumor is used to form a three-dimensional X-ray of the tumor and of a portion of the patient, such as in X-ray computed tomography.

[0199] Still referring to FIG. **13A**, use of two imaging sources and two detectors set at ninety degrees to one another allows the gantry **960** or the patient **730** to rotate through half an angle required using only one imaging source and detector combination. A third imaging source/detector combination allows the three imaging source/detector combination to be set at sixty degree intervals allowing the imaging time to be cut to that of one-third that gantry **960** or patient **730** rotation required using a single imaging source-detector combination. Generally,  $n$  source-detector combinations reduces the time and/or the rotation requirements to  $1/n$ . Further reduction is possible if the patient **730** and the gantry **960** rotate in opposite directions. Generally, the used of multiple source-detector combination of a given technology allow for a gantry that need not rotate through as large of an angle, with dramatic engineering benefits.

[0200] Still referring to FIG. **13A**, the set of sources **1310** and set of detectors **1320** optionally use more than one imaging technology. For example, a first imaging technology uses X-rays, a second used fluoroscopy, a third detects

fluorescence, a fourth uses cone beam computed tomography or cone beam CT, and a fifth uses other electromagnetic waves. Optionally, the set of sources **1310** and the set of detectors **1320** use two or more sources and/or two or more detectors of a given imaging technology, such as described supra with two X-ray sources to  $n$  X-ray sources.

[0201] Still referring to FIG. **13A**, use of one or more of the set of sources **1310** and use of one or more of the set of detectors **1320** is optionally coupled with use of the positively charged particle tomography system described supra. As illustrated in FIG. **13A**, the positively charged particle tomography system uses a second mechanical support **1343** to co-rotate the scintillation material **710** with the gantry **960**, as well as to co-rotate an optional sheet, such as the first sheet **760** and/or the fourth sheet **790**.

#### Example II

[0202] Referring now to FIG. **13B**, a second example of the integrated cancer treatment-imaging system **1300** is illustrated using greater than three imagers.

[0203] Still referring to FIG. **13B**, two pairs of imaging systems are illustrated. Particularly, the first and second imaging source **1312**, **1314** coupled to the first and second detectors **1322**, **1324** are as described supra. For clarity of presentation and without loss of generality, the first and second imaging systems are referred to as a first X-ray imaging system and a second X-ray imaging system. The second pair of imaging systems uses a third imaging source **1316** coupled to a third detector **1326** and a fourth imaging source **1318** coupled to a fourth detector **1328** in a manner similar to the first and second imaging systems described in the previous example. Here, the second pair of imaging systems optionally and preferably uses a second imaging technology, such as fluoroscopy. Optionally, the second pair of imaging systems is a single unit, such as the third imaging source **1316** coupled to the third detector **1326**, and not a pair of units. Optionally, one or more of the set of imaging sources **1310** are statically positioned while one or more of the set of imaging sources **1310** co-rotate with the gantry **960**. Pairs of imaging sources/detector optionally have common and distinct distances, such as a first distance,  $d_1$ , such as for a first source-detector pair and a second distance,  $d_2$ , such as for a second source-detector or second source-detector pair. As illustrated, the tomography detector or the scintillation material **710** is at a third distance,  $d_3$ . The distinct differences allow the source-detector elements to rotate on a separate rotation system at a rate different from rotation of the gantry **960**, which allows collection of a full three-dimensional image while tumor treatment is proceeding with the positively charged particles.

#### Example III

[0204] For clarity of presentation, referring now to FIG. **13C**, any of the beams or beam paths described herein is optionally a cone beam **1390** as illustrated. The patient support **152** is an mechanical and/or electromechanical device used to position, rotate, and/or constrain any portion of the tumor **720** and/or the patient **730** relative to any axis.

#### Tomography Detector System

[0205] A tomography system optically couples the scintillation material to a detector. As described, supra, the tomography system optionally and preferably uses one or

more detection sheets, beam tracking elements, and/or tracking detectors to determine/monitor the charged particle beam position, shape, and/or direction in the beam path prior to and/or posterior to the sample, imaged element, patient, or tumor. Herein, without loss of generality, the detector is described as a detector array or two-dimensional detector array positioned next to the scintillation material; however, the detector array is optionally optically coupled to the scintillation material using one or more optics. Optionally and preferably, the detector array is a component of an imaging system that images the scintillation material **710**, where the imaging system resolves an origin volume or origin position on a viewing plane of the secondary photon emitted resultant from passage of the residual charged particle beam **267**. As described, infra, more than one detector array is optionally used to image the scintillation material **710** from more than one direction, which aids in a three-dimensional reconstruction of the photonic point(s) of origin, positively charged particle beam path, and/or tomographic image.

#### Detector Array

[0206] Referring now to FIG. **14A**, in a tomography system **1400**, a detector array **1410** is optically coupled to the scintillation material **710**. For clarity of presentation and without loss of generality, the detector array **1410**, which is preferably a two-dimensional detector array, is illustrated with a detection side directly coupled to the scintillation material **710**, such as through physical contact or through an intervening layer of an optical coupling material or optical coupling fluid with an index of refraction between that of the scintillation material **710** and front side of the detector array **1410**. However, the detector array **1410** is optionally remotely located from the scintillation material **710** and coupled using light coupling optics. As illustrated, secondary photons emitted from the scintillation material **710**, resultant from passage of the residual charge particle beam **267**, strike a range of detector elements according to a probability distribution function. Generally, the positively charged particles from the accelerator after passing through the sample strike the scintillation material resultant in emitted electrons and photons, the photons are detected, and the path of the charged particles and/or the energy of the charged particles after passing through the sample is back calculated using the detection position(s) of the photons in the detector array.

[0207] Referring now to FIG. **14B**, the tomography system **1400** is illustrated with an optical array between the scintillation material **710** and the detector array **1410**. For clarity of presentation and without loss of generality, the optical array is referred to herein as a fiber optic array **1420**, which is preferably a two-dimensional fiber optic array. The individual elements of the optical array are optionally of any geometry, such as a square or rectangular cross-section in place of a round cross-section of a fiber optic. Generally, the scintillation material **710** is optically coupled to the fiber optic array **1420** and the fiber optic array **1420** is optically coupled to the detector array **1410**, which may be mass produced. In one case, elements of the fiber optic array **1420** couple 1:1 with elements of the detector array **1410**. In a second preferable case, the intermediate fiber optic array **1420** is primarily used to determine position of detected photons and many detector elements of the detector array couple to a single fiber optic element of the fiber optic array

**1420** or vice-versa. In the second case, signals from detector elements not aligned with a given fiber core, but instead aligned with a cladding or buffer material about the fiber are removed in post-processing.

[0208] Referring now to FIG. 14C and FIG. 14D, the fiber optic array **1420** is illustrated with a fiber array configuration that is close-packed **1422** and orderly **1424**, respectively. The close-packed **1422** system captures a higher percentage of photons while the orderly **1424** system couples readily with an array of detector elements in the detector array **1424**. Since post-processing is optionally and preferably used to determine which detector element signals to use, the packing structure of the fiber optic array **1420** is optionally of any geometry.

[0209] Referring now to FIG. 14E, the tomography system **1400** is illustrated with an optional micro-optic array **1412** coupling and focusing photons from the scintillation material **710** to the detector array **1410**. Generally, the array of micro-optics couples more light to the detector elements of the detector array **1410**, which increases the signal-to-noise ratio of the detected signals.

#### Multiplexed Scintillation

[0210] Referring now to FIG. 15, a multiplexed scintillation system **1500** is illustrated. In one case of the multiplexed scintillation system **1500**, multiple frequencies of light are detected where the detected frequency wavelength, wavelength range, or color is representative of energy, or residual energy after passing through the sample, of the residual charged particle beam **267**. In another case, changing distributions of secondary photons, resultant from passage of the residual charged particle beam **267**, are detected and used to determine state of the residual charged particle beam **267**, such as position, direction, intensity, and/or energy. In still another case, a set of different scintillation materials are used to determine state of the residual charged particle beam **267**. To clarify and without loss of generality, several examples of multiplexed scintillation follow.

#### Example I

[0211] In a first example, the scintillation material **710** results in emission of photons at different wavelengths dependent upon the energy of the residual charged particle beam **267**, which is the treatment beam **269** after passing through a sample, such as the tumor **720** of the patient **730**. For instance, as the residual charged particle beam **267** slows in the scintillation material, the wavelength of secondary photons increases resultant in a color shift as a function of position along the path or vector of the residual charged particle beam **267**. Hence, use of wavelengths of the photons detected by detector elements in the detector array **1410**, or as described infra multiple detector arrays, viewing varying depths of the scintillation material **710** are used to back calculate state of the residual charged particle beam **267**.

#### Example II

[0212] In a second example, the scintillation material **710** results in emission of differing numbers of photons as a function of the energy of the residual charged particle beam **267**, which changes as a function of depth of penetration into the scintillation material **710**. For instance, as the residual charged particle beam **267** slows in the scintillation material

**710**, the intensity of secondary photons changes as a function of position along the path or vector of the residual charged particle beam **267**. Hence, use of the intensity of the signals of detector element of the detector array **1410**, or as described infra multiple detector arrays, viewing varying depths of the scintillation material **710** are used to back calculate state of the residual charged particle beam **267** as a function of depth in the scintillation material **710**.

#### Example III

[0213] In a third example, the scintillation material **710** is a set of  $n$  scintillation materials having differing secondary photon emission properties as a function of incident or transiting positively charged particles, where  $n$  is a positive integer such as greater than 1, 2, 3, 4, 5, or 10. For clarity of presentation and without loss of generality, cases of using a set of scintillation materials are described herein.

[0214] Referring now to FIG. 15, in a first case, three scintillation materials are used in a scintillation block, section, or volume of the multiplexed scintillation system **1500**. Particularly, a first scintillation material **711**, a second scintillation material **712**, and a third scintillation material **713** are used at a first, second, and third depth along a path of the residual charge particle beam **267** or  $z$ -axis. Further, as illustrated, the first scintillation material **711**, the second scintillation material **712**, and the third scintillation material **713** emit light at three separate wavelengths, such as from three distinct chemical compositions of the three scintillation materials **711**, **712**, **713**. For clarity of presentation, the three wavelengths are denoted blue (B), green (G), and red (R); however, any wavelength, range of wavelength, or ranges of wavelengths from **200** to **2500** is optionally used. As illustrated, when the residual charged particle beam **267** has only enough energy to penetrate into the first scintillation material **711**, then only blue light is emitted. Further, when the residual charged particle beam **267** has sufficient energy to penetrate into only the second scintillation material **712**, then only blue light and green light is emitted. In this case, the colors of the emitted light yields additional information on the path of the positively charged particles, which provides a useful constraint on back calculation of the state of the residual charged particle beam **267**. Still further, when the residual charged particle beam **267** has a large enough energy to penetrate into the third scintillation material **713**, then blue, green, and red light is emitted; again adding useful information on the state of the residual charged particle beam **267** and useful constraints on back calculation of the residual charged particle beam state.

[0215] In a second case the set of scintillation materials comprise different thicknesses, such as  $n$  thicknesses, where  $n$  is a positive integer. Still referring to FIG. 15, for clarity of presentation and without loss of generality, three thicknesses of scintillation materials are illustrated along a longitudinal  $z$ -axis of the residual charged particle beam **267**. Particularly, the first scintillation material **711** is illustrated with a first pathlength,  $b_1$ ; the second scintillation material **712** is illustrated with a second pathlength,  $b_2$ ; and the third scintillation material **713** is illustrated with a third pathlength,  $b_3$ . By using thinner layers, relative to a homogeneous scintillation material, of a given light emitting color, identification, post-processing, and/or back calculation of the points of origin of secondary emission of photons, resultant from passage of the residual charged particle beam **267**, are constrained and thus the path of the residual charged

particle beam **267** and corresponding treatment beam **269** through the tumor **720** is identified with more accuracy and/or precision. The layers of scintillation material optionally emit  $n$  wavelengths or bands of light. Further, the use of one material emitting a first color at a first layer is optionally used again for another non-adjacent layer. Similarly, a pattern of colors from corresponding layers is optionally repeated as a function of position along the residual charged particle beam **267**, such as B, G, R, B, G, R, . . . , B, G, R.

#### Example IV

[0216] In a fourth example, a color filter array **1414** is optically coupled to the detector array **1410**, where the color filter array **1414** is in a secondary photon path between the scintillation material **710** and the detector array **1410**. Similarly and preferably, a two-dimensional color filter array is optically coupled to a two-dimensional detector array in the secondary photon path. Using the color filter array **1414** as a portion of an imaging system, a point of origin of the secondary photon is determined, which yields information on path of the residual charged particle beam **267**. For clarity of presentation, the color filter array **1414** is described as a Bayer matrix; a cyan, yellow, green, magenta filter, which is a CYGM filter; a red, green, blue, emerald filter, which is a RGBE filter, and/or a two color filter array. Generally any repeating array of color filters or even non-repeating pattern of optical filters is used in the color filter array **1414**.

#### Example VI

[0217] Generally, components of the tomography system, described supra, are combined in any combination and/or permutation. For instance, still referring to FIG. **15**, a sixth example is provided using: (1) the first scintillation material **711** with the first pathlength,  $b_1$ ; (2) the second scintillation material **712** with the second pathlength,  $b_2$ ; (3) the third scintillation material **713** with the third pathlength,  $b_3$ ; (4) the color filter array **1414**; (5) the micro-optics array **1412**; and (6) the detector array **1410**, all in two-dimensional configurations as part of an imaging system imaging the scintillation materials and secondary photons emitted therefrom, resultant from passage, transit, energy transfer from, interaction with, or termination of the residual charged particles in the residual charged particle beam **267**. Calculation of position and direction of the residual charged particle beam **267**, with or without use of an imaging sheet, allows a more accurate determination of an exit point of the treatment beam **269** or start of the residual energy beam **269** from the patient **730** and a corresponding path of the charged particle beam from the prior side of the patient **730**, through the patient **730**, and to the posterior exit point of the patient **730**.

#### Scintillation Array

[0218] Referring now to FIG. **16A**, the scintillation material **710** is optionally configured as an array of scintillation materials and/or as an array of scintillation sections **1610** in a multiplexed scintillation detector **1600**, where elements of the array of scintillation sections **1610** are optionally physically separated. For clarity of presentation and without loss of generality examples follow that described and/or illustrate the array of scintillation sections **1610** as an element of the tomography system.

#### Example I

[0219] In a first example, referring still to FIG. **16A**, the scintillation material **710** described above is illustrated in a configuration of an array of scintillation sections **1610** or an array of scintillation optics. As illustrated, elements of the array of scintillation sections **1610** having a first index of refraction are separated by a separation material or cladding **1422** having a second index of refraction that is less than the first index of refraction. For example, the first index of refraction is greater than 1.4, in a range of 1.3 to 1.7, and/or in a range of 1.4 to 1.6 and the second index of refraction is in a range of 1.0 to 1.3 or 1.4. The difference in index of refraction forms a light-pipe similar to a fiber optic, for the photons at or above a total internal reflectance angle threshold. Referring now to FIG. **16B**, the core scintillation material **710** and the surrounding cladding **1422** is further illustrated within a buffer material **1424**. While the light-pipe in FIG. **16B** is illustrated with a circular cross-sectional shape, generally the light pipe cross-sectional shape is of any geometry, such as a rounded corner polygon, square, or rectangle. Referring again to FIG. **16A**, the residual charged particle beam **267** is illustrated as inducing emission of two photons, illustrated as dashed lines. The first photon passes straight to a first detector element **1415** of the detector array **1410**. The second photon reflects off of the surrounding cladding **1422** into the first detector element. As illustrated with the dotted line, without the surrounding cladding **1422**, having a lower index of refraction than the scintillation material **710**, the second photon would have struck a second detector element **1416** of the detector array **1410**. Hence, by restricting,  $x$ - and/or  $y$ -axis movement of the photon, as limited by the respective index of refractions, detected and determined resolution of the path of the residual charged particle is enhanced and a corresponding enhancement of the tomographic image is achieved, as described supra.

#### Example II

[0220] In a second example, still referring to FIG. **16B**, individual elements of the array of scintillation sections **1610** or scintillation optics are comprised of individual scintillation materials, such as the first scintillation material **711**, the second scintillation material **712**, and the third scintillation material **713**. Optionally, the surrounding cladding **1422** is only used between a repeating set of the scintillation materials, in this case between every three longitudinal elements of scintillation materials.

#### Example III

[0221] In a third example, referring now to FIG. **16C**, as in the first example the scintillation material **710** described above is illustrated as an array of scintillation sections **1610**, where individual longitudinal paths of the scintillation sections **1610** along the  $z$ -axis are separated by the cladding with the second lower index of refraction compared indices of refraction of a set of scintillation materials. However, in this example, the longitudinal paths of a given scintillation section comprises  $n$  sections of scintillation materials, where  $n$  is a positive integer of 2, 3, 4, 5, or more. As illustrated, longitudinal sections comprise the second scintillation material **712** between the first scintillation material **711** and the third scintillation material **713**. Further, as illustrated at a first time,  $t_1$ , the residual charged energy beam **267** strikes the first scintillation material generating a blue photon, B,

detected at a third detector element **1417**, where the blue photon is maintained in a resolved x/y-range by the surrounding cladding **1422**. Similarly, at a second time,  $t_2$ , and third time,  $t_3$ , respectively, residual charged energy beams generate a green photon, G, and a red photon, R, respectively, which are detected with a fourth detector element **1418** and a fifth detector element **1419**, respectively. Again, the surrounding cladding **1422** limits x/y-plane translation of the green photon and the red photon. As: (1) the color of the photon, B, G, R, is indicative of the z-axis energy of the residual charged particle beam **267** in the longitudinally segmented sections of the elements of the fiber optic array **1410** and (2) the x/y-plane position of the residual charged particle beam **267** is restricted by the cladding **1422** between the axially separated scintillation sections **1610** of the scintillation optic array, x, y, and z information or spatial position and energy information about the residual charged particle beam **267** is obtained as a function of time, which is used in a back calculation of the path of: (1) the treatment beam **269** or imaging beam and (2) presence and structure of constituents of the patient **730**, such as the tumor **720**, blood, bone, muscle, connective tissue, collagen, elastin, and/or fat.

#### Example IV

[0222] In another example, one or more imaging optic, such as a light directing optic and/or a focusing optic, used to image the scintillation material comprises the scintillation material **710**.

#### Enhanced Multi-Directional Scintillation Detection

[0223] Photons emitted from the scintillation material, resultant from energy transfer from a passing residual charged particle beam **267**, emit in many directions. Hence, detection and/or imaging of the photons in many planes or directions provides an opportunity for enhanced signal-to-noise, resolution, accuracy, and/or precision of determination of state of the residual charged particle beam **267** and from that enhanced resolution, accuracy, and precision of the imaged sample, such as the tumor **720** of the patient **730**.

[0224] Referring now to FIG. 17A, herein the scintillation material **710**, in the form of a block or as segmented sections has a prior surface **714** or front surface, a posterior surface **715** or back surface, a dexter surface **716** or viewer's left surface, a sinister surface **717** or viewer's right surface, a top surface **718**, and a bottom surface **719**.

[0225] Generally, the detector array **1410** and/or any of the accessories thereof, such as the micro-optics array **1412**, color filter array **1414**, axially separated sections, and/or longitudinally separated sections, is optionally used on any surface of the scintillation material **710**. Further, referring now to FIG. 17B, the detector array **1410** is optionally a set of detector arrays **1700**, such as n detector arrays where n is a positive integer. In FIG. 17B, the set of detector arrays **1700** includes: (1) a second detector array **1702** optically coupled to the posterior surface **715** of the scintillation material **710**; (2) a fourth detector array **1704** optically coupled to the sinister surface **716** of the scintillation material **710**; and (3) a fifth detector array **1705** optically coupled to the top surface **718** of the scintillation material **710**. The use of multiple detector arrays, each configured to image the scintillation material **710**, enhances accuracy and precision of knowledge of path of the residual charged particle beam **267** through enhanced accuracy, precision,

and resolution of points of origin of the resultant emitted photons and as discussed above the resulting accuracy, precision, and resolution of the imaged object. As illustrated, use of three detector arrays set at orthogonal angles allows imaging of the scintillation material in three dimensions, which aids in determination of the path of the residual charged particle beam **267**. Optionally, each of the set of detector arrays **1700** is set at any orientation in the x-, y-, z-axes space.

[0226] Referring now to FIG. 17B, FIG. 17C, and FIG. 17D, the set of detector arrays **1700** is illustrated with six detector arrays: (1) a first detector array **1701** optically coupled to the prior surface **714** of the scintillation material **710**; (2) a second detector array **1702** optically coupled to the posterior surface **715** of the scintillation material **710**; (3) a third detector array **1703** optically coupled to the dexter surface **716** of the scintillation material **710**; (4) a fourth detector array **1704** optically coupled to the sinister surface **717** of the scintillation material **710**; (5) a fifth detector array **1705** optically coupled to the top surface **718** of the scintillation material **710**; and (6) a sixth detector array **1706** optically coupled to the bottom surface **719** of the scintillation material **710**. Use of a detector array on each surface of the scintillation material **710** allows detection of secondary photons, resultant from the residual charged particle beam **267**, with a corresponding increase and/or maximum percentage of detection of the emitted photons. The larger number of detected photons, with the multiple detector arrays, yields a larger number of data points to more accurately and precisely determine state of the residual charged particle beam with a corresponding enhancement of the tomographic image, as described supra.

[0227] Still referring to FIG. 17C, optionally, the prior surface **714** of the scintillation material **710** comprises an aperture **1710** through which the residual charged particle beam **267** passes. Optionally, no aperture is used on the prior surface **714** of the scintillation material **710** and the densities and pathlengths of the first detector array **1701** are used in a calculation of an energy of the residual charged particle beam **267**.

#### Imaging

[0228] Generally, medical imaging is performed using an imaging apparatus to generate a visual and/or a symbolic representation of an interior constituent of the body for diagnosis, treatment, and/or as a record of state of the body. Typically, one or more imaging systems are used to image the tumor and/or the patient. For example, the X-ray imaging system and/or the positively charged particle imaging system, described supra, are optionally used individually, together, and/or with any additional imaging system, such as use of X-ray radiography, magnetic resonance imaging, medical ultrasonography, thermography, medical photography, positron emission tomography (PET) system, single-photon emission computed tomography (SPECT), and/or another nuclear/charged particle imaging technique.

[0229] Referring now to FIG. 18, the imaging system **170** is further described. As described supra, the imaging system **170** optionally uses:

[0230] a positive ion beam tumor irradiation system **171**;

[0231] two or more imaging systems **172**, where the individual imaging systems generate data for a composite image of the sample;

- [0232] a concurrent treatment imaging system 173, where imaging occurs during treatment of the tumor 720 with the positively charged particle or in-between treatment of voxels of the tumor 720;
- [0233] an intermittent or periodic imaging system 174, where one or more update images, confirmation images, and/or adjustment images are collected to update a previous image, alter a treatment plan, and/or stop a current treatment of the tumor 720 with the treatment beam 269;
- [0234] a tomography beam imaging system 175 comprising generating tomograms from any radiology technology;
- [0235] a dynamic feedback system 176, such as use of a positron emission tomography signal to dynamically control state and/or movement of a positive ion tumor treatment beam;
- [0236] a relative rotational motion system 177 between the patient and an imaging beam; and/or
- [0237] a relative linear motion system 178 between the patient and a radiography imaging beam.
- [0238] To clarify the imaging system and without loss of generality several examples are provided.

#### Example I

[0239] In a first example, a positron emission tomography system is used to monitor, as a function of time, a precise and accurate location of the treatment beam 269 relative to the tumor 720. Signal from the positron emission tomography system is optionally: (1) recorded to provide a reviewable history of treatment of the tumor 720 with the positively charged particle beam or treatment beam 269 and/or (2) used to dynamically monitor the position of the treatment beam 269 and to function as a feedback control signal to dynamically adjust position of the treatment beam 269 as a function of time while scanning through treatment voxels of the tumor 720.

#### Example II

[0240] In a second example, an imaging system images the tumor 720 as a function of imaging system paths, which is movement of at least a portion of the imaging system beam along a first path relative to the tumor 720, while the charged particle beam system 100 treats a series of voxels of the tumor 720 along a set of treatment beam paths. In various cases: (1) the imaging system paths and treatment beam paths are essentially parallel paths, such as the two paths forming an angle with the tumor of less than 10, 5, 2, or 1 degrees; (2) the imaging system paths and treatment beam paths are essentially perpendicular to one another, such as forming an angle with the tumor 720 of greater than 70, 80, 85, 88, or 89 degrees and less than 91, 92, 95, 100, or 110 degrees; (3) as the treatment beam 269 and gantry nozzle 610, of the particle beam system 100, rotates around the tumor 720 with rotation of the gantry 960 at a first rotational rate, the imaging system path rotates around the tumor 720 at a second rotational rate; and (4) as the treatment beam 269 and gantry nozzle 610, of the particle beam system 100, relatively rotates around the tumor 720, the imaging system paths translate along a vector, such as while the tumor 720 is treated along a set of rotated lines joined at the tumor, the imaging system paths form a set of essentially parallel lines,

such as a set of vectors along a plane and/or a set of vectors passing through a first or prior side of the tumor.

[0241] Referring now to FIGS. 19(A-C), a hybrid cancer treatment-imaging system 1800 is illustrated. Generally, the gantry 960, which optionally and preferably supports the gantry nozzle 610, rotates around the tumor 720, as illustrated in FIG. 19B, and/or an isocentre 263, as illustrated in FIG. 19A, of the charged particle beam. As illustrated, the gantry 960 rotates about a gantry rotation axis 1811, such as using a rotatable gantry support 1810. In one case, the gantry 960 is supported on a first end 962 by a first buttress, wall, or support, not illustrated, and on a second end 964 by a second buttress, wall, or support, not illustrated. A first optional rotation track 1813 and a second optional rotation track coupling the rotatable gantry support and the gantry 960 are illustrated, where the rotation tracks are any mechanical connection. Further, as illustrated, for clarity of presentation, only a portion of the gantry 960 is illustrated to provide visualization of the supported beam transport system 135 or a section of the beamline between the synchrotron 130 and the patient 730. To further clarify, the gantry 960 is illustrated, at one moment in time, supporting the gantry nozzle 610 of the beam transport system 135 in an orientation resulting in a vertical vector of the treatment beam 269. As the rotatable gantry support 1810 rotates, the gantry 960, the beam transport line 135, the gantry nozzle 610 and the treatment beam 269 rotate about the gantry rotation axis 1811, illustrated as the x-axis, forming a set of treatment beam vectors originating at circumferential positions about tumor 720 or isocentre 263 and passing through the tumor 720. Optionally, an X-ray beam path 1801, from an X-ray source, runs through and moves with the dynamic gantry nozzle 610 parallel to the treatment beam 269. Prior to, concurrently with, intermittently with, and/or after the tumor 720 is treated with the set of treatment beam vectors, one or more elements of the imaging system 170 image the tumor 720 of the patient 730.

[0242] Still referring to FIG. 19A, the hybrid cancer treatment-imaging system 1800 is illustrated with an optional set of rails 1820 and an optional rotatable imaging system support 1812 that rotates the set of rails 1820, where the set of rails 1820 optionally includes n rails where n is a positive integer. Elements of the set of rails 1820 support elements of the imaging system 170, the patient 730, and/or a patient positioning system. The rotatable imaging system support 1812 is optionally concentric with the rotatable gantry support 1810. The rotatable gantry support 1810 and the rotatable imaging system support 1812 optionally: co-rotate, rotate at the same rotation rate, rotate at different rates, or rotate independently. A reference point 1815 is used to illustrate the case of the rotatable gantry support 1810 remaining in a fixed position, such as a treatment position at a third time,  $t_3$ , and a fourth time,  $t_4$ , while the rotatable imaging system support 1812 rotates the set of rails 1820.

[0243] Still referring to FIG. 19A, any rail of the set of rails optionally rotates circumferentially around the x-axis, as further described infra. For instance, the first rail 1822 is optionally rotated as a function of time with the gantry 960, such as on an opposite side of the gantry nozzle 610 from the tumor 720 of the patient 730.

[0244] Still referring to FIG. 19A, a first rail 1822 of the set of rails 1820 is illustrated in a first retracted position at a first time,  $t_1$ , and at a second extended position at a second time,  $t_2$ . The first rail 1822 is illustrated with a set of n

detector types **1830**, such as a first detector **1832** or first detector array at a first extension position of the first rail **1822** and a second detector **1834** or second detector array at a second extension position of the first rail **1822**, where  $n$  is a positive integer, such as 1, 2, 3, 4, 5, or more. The first detector **1832** and the second detector **1834** are optionally and preferably two detector array types, such as an X-ray detector and a scintillation detector. In use, the scintillation detector is positioned, at the second extended position of the first rail **1822**, opposite the tumor **730** from the gantry nozzle **610** when detecting scintillation, resultant from passage of the residual charged particle beam **267** into the scintillation material **710**, such as for generating tomograms, tomographic images, and/or a three-dimensional tomographic reconstruction of the tumor **720**. In use, the first rail **1822** is positioned at a third extended position, not illustrated, which places the second detector or X-ray detector opposite the tumor **720** from the gantry nozzle **610**, such as for generating an X-ray image of the tumor **720**. Optionally, the first rail **1822** is attached to the rotatable gantry support **1810** and rotates with the first gantry support **1810**. The first rail **1822** is optionally retracted, such as illustrated at the first time,  $t_1$ , such as for some patient positions about the isocentre **263**.

[0245] Still referring to FIG. 19A and referring again to FIG. 19B and FIG. 19C, a second rail **1824** and a third rail **1826** of the set of rails **1820** are illustrated at a retracted position at a first time,  $t_1$ , and an extended position at a second time,  $t_2$ . Generally, the second rail **1824** and the third rail **1826** are positioned on opposite sides of the patient **730**, such as a sinister side and a dexter side of the patient **730**. Generally, the second rail **1824**, also referred to as a source side rail, positions an imaging source system element and the third rail **1826**, also referred to as a detector side rail, positions an imaging detector system element on opposite sides of the patient **730**. Optionally and preferably, the second rail **1824** and the third rail **1826** extend away from the first buttress **962** and retract toward the first buttress **962** together, which keeps a source element mounted, directly or indirectly, on the second rail **1824** opposite the patient **730** from a detector element mounted, directly or indirectly, on the third rail **1826**. Optionally, the second rail **1824** and the third rail **1826** translate, such as linearly, on opposite sides of an axis perpendicular to the gantry rotation axis **1811**, as further described infra. Optionally, the second rail **1824** and the third rail **1826** position PET detectors for monitoring emissions from the tumor **720** and/or the patient **730**, as further described infra.

[0246] Still referring to FIG. 19B, a rotational imaging system **1840** is described. For example, the second rail **1824** is illustrated with: (1) a first source system element **1841** of a first imaging system, or first imaging system type, at a first extension position of the second rail **1824**, which is optically coupled with a first detector system element **1851** of the first imaging system on the third rail **1826** and (2) a second source system element **1843** of a second imaging system, or second imaging system type, at a second extension position of the second rail **1824**, which is optically coupled with a second detector system element **1853** of the second imaging system on the third rail **1826**, which allows the first imaging system to image the patient **730** in a treatment position and, after translation of the first rail **1824** and the second rail **1826**, the second imaging system to image the patient in the patient's treatment position. Optionally the first imaging system or primary imaging system and the second imaging

system or secondary imaging system are supplemented with a tertiary imaging system, which uses any imaging technology. Optionally, first signals from the first imaging system are fused with second signals from the second imaging system to: (1) form a hybrid image; (2) correct an image; and/or (3) form a first image using the first signals and modified using the second signals or vice-versa.

[0247] Still referring to FIG. 19B, the second rail **1824** and third rail **1826** are optionally alternately translated inward and outward relative to the patient, such as away from the first buttress and toward the first buttress. In a first case, the second rail **1824** and the third rail **1826** extend outward on either side of the patient, as illustrated in FIG. 19B. Further, in the first case the patient **730** is optionally maintained in a treatment position, such as in a constrained laying position that is not changed between imaging and treatment with the treatment beam **269**. In a second case, the patient **730** is translated toward the first buttress **962** to a position between the second rail **1824** and the third rail **1826**, as illustrated in FIG. 19B. In the second case, the patient is optionally imaged out of the treatment beam path **269**, as illustrated in FIG. 19B. Further, in the second case the patient **730** is optionally maintained in a treatment position, such as in a constrained laying position that is not changed until after the patient is translated back into a treatment position and treated. In a third case, the second rail **1824** and the third rail **1826** are translated away from the first buttress **962** and the patient **730** is translated toward the first buttress **962** to yield movement of the patient **730** relative to one or more elements of the first imaging system type or second imaging system type. Optionally, images using at least one imaging system type, such as the first imaging system type, are collected as a function of the described relative movement of the patient **730**, such as along the x-axis and/or as a function of rotation of the first imaging system type and the second imaging system type around the x-axis, where the first imaging type and second imaging system type use differing types of sources, use differing types of detectors, are generally thought of as distinct by those skilled in the art, and/or have differing units of measure. Optionally, the source is an emissions from the body, such as a radioactive emission, decay, and/or gamma ray emission, and the second rail **1824** and the third rail **1826** position and/or translate one or more emission detectors, such as a first positron emission detector on a first side of the tumor **720** and a second positron emission detector on an opposite side of the tumor **730**.

[0248] Still referring to FIG. 19B, a hybrid cancer treatment-rotational imaging system **1804** is illustrated. In one example of the hybrid cancer treatment-rotational imaging system **1804**, the second rail **1824** and third rail **1826** are optionally circumferentially rotated around the patient **730**, such as after relative translation of the second rail **1824** and third rail **1826** to opposite sides of the patient **730**. As illustrated, the second rail **1824** and third rail **1826** are affixed to the rotatable imaging system support **1812**, which optionally rotates independently of the rotatable gantry support **1810**. As illustrated, the first source system element **1841** of the first imaging system, such as a two-dimensional X-ray imaging system, affixed to the second rail **1824** and the first detector system element **1851** collect a series of preferably digital images, preferably two-dimensional images, as a function of co-rotation of the second rail **1824** and the third rail **1826** around the tumor **720** of the patient, which is positioned along the gantry rotation axis **1811**

and/or about the isocentre **263** of the charged particle beam line in a treatment room. As a function of rotation of the rotatable imaging system support **1812** about the gantry rotation axis **1811** and/or a rotation axis of the rotatable imaging system support **1812**, two-dimensional images are generated, which are combined to form a three-dimensional image, such as in tomographic imaging. Optionally, collection of the two-dimensional images for subsequent tomographic reconstruction are collected: (1) with the patient in a constrained treatment position, (2) while the charged particle beam system **100** is treating the tumor **720** of the patient **730** with the treatment beam **269**, (3) during positive charged particle beam tomographic imaging, and/or (4) along an imaging set of angles rotationally offset from a set of treatment angles during rotation of the gantry **960** and/or rotation of the patient **730**, such as on a patient positioning element of a patient positioning system.

[0249] Referring now to FIG. **19C**, a hybrid tumor treatment-vertical imaging system **1806** is illustrated, such as with a translatable imaging system **1860** is described. In one example of the hybrid tumor treatment vertical imaging system **1806**, the second rail **1824** and the third rail **1826** are used to acquire a set of images with linear translation of the second rail **1824** and the third rail **1826** past the tumor **720** of the patient **730**, such as with movement along an axis as a function of time, such as, as illustrated, along a vertical axis at the fifth time,  $t_5$ , and a sixth time,  $t_6$ . As illustrated, the second rail **1824** and the third rail co-translate along a rail support **1864**, where the rail support **1864** is optionally positioned inside the rotatable gantry support **1810** and/or the rotatable imaging system support **1812**. Optionally and preferably, source elements and detector elements moving past the tumor **720** of the patient **730** on the second rail **1824** and third rail **1826**, respectively, are used to collect a scanning set of images, such as PET images, of the tumor as a function of translation along the rail support **1864**. In the hybrid tumor treatment-vertical imaging system **1806**, the second rail **1824** and elements supported thereon and the third rail and elements supported thereon optionally extend and/or retract, as described supra. Further, in the hybrid tumor treatment-vertical imaging system **1806**, the second rail **1824** and elements supported thereon and the third rail and elements supported thereon optionally rotate about the isocentre, such as with rotation of the rotatable gantry support **1810** and/or the rotatable imaging system support **1812**. Optionally, any member of the set of rails **1820** extends/retracts, rotates, and/or translates past the tumor **720** of the patient **730** at the same time.

[0250] Optionally, the vertical imaging system **1806** moves a PET system detector system element, such as a detector or coupling device, to a position corresponding to a depth of penetration of the treatment beam **269** into the tumor **720** of the patient **730**. For clarity of presentation and without loss of generality, an example is provided where the treatment beam **269** is vertical and passes through the gantry nozzle **610** directly above the tumor **720**. The treatment beam **269** is of a known energy at a known time, where the known energy is intentionally varied to yield a corresponding varied depth of penetration of the treatment beam **269** into the tumor, such as described by the peak of the Bragg peak. A detector system element of the positron emission tomography system, supported on the vertical imaging system **1806**, it optionally translated vertically to observe the depth of penetration of the treatment beam **269**. For

instance, as the treatment beam energy is decreased, the depth of penetration of the treatment beam **269** into the tumor **720** of the patient **730** decreases and the detector system element of the positron emission tomography system is raised vertically. Similarly, as the treatment beam energy is increased, the depth of penetration of the treatment beam **269** into the tumor **720** of the patient **730** increases and the detector system element of the positron emission tomography system is lowered vertically. Optionally, as the gantry **960** rotates, the vertical imaging system rotates.

[0251] Still referring to FIG. **19C**, a reference point **1816** is used to illustrate the case of the rotatable gantry support **1810** rotating between a fifth time,  $t_5$ , and a sixth time,  $t_6$ , while the translatable imaging system **1860** moves the second rail **1824** and the third rail **1826** along a linear axis, illustrated as the z-axis or treatment beam axis.

[0252] Optionally, one or more of the imaging systems described herein monitor treatment of the tumor **720** and/or are used as feedback to control the treatment of the tumor by the treatment beam **269**.

[0253] Referring now to FIG. **20**, a dynamic treatment beam guiding system **2000** is described. As the treatment beam **269** irradiates the tumor **2010** radioactive nuclei or isotopes are formed **2020** where the treatment beam **269** strikes the tumor **720** of the patient **730**. The radioactive nuclei emits a positron **2030** that rapidly undergoes electron-positron annihilation **2040**, which results in a gamma ray emission **2050**. Thus, monitoring location of one or more gamma ray emissions is a measure of the current location of the treatment beam **269**. Also, dependent upon the half-life of the formed radioactive nuclei, monitoring location of the gamma ray emissions provides a measure of where the treatment beam **269** has interacted with the tumor **720** and/or the patient **730**, which yields information on treatment coverage and/or provides a history of treatment. By monitoring new voxels or positions of gamma ray emission and/or by monitoring intensity drop off of gamma-ray emission over each of multiple recently treated voxels, a current treatment position of the treatment beam **269** is determined. The main controller **110** and/or a dynamic positioning system **2060** is optionally used to dynamically correct and/or alter the current position of the treatment beam **269**, such as through control of one or more of the extraction energy, beam guiding magnets, or beam shaping elements.

[0254] Referring now to FIG. **21**, treatment position determination system **2100** is illustrated. Herein, for clarity of presentation and without loss of generality, a Bragg peak is used to describe a treatment position. However, the techniques described herein additionally apply to the tail of the Bragg peak. As illustrated, at a first time,  $t_1$ , the treatment beam **269** having a first energy has a first treatment position **2111** at a first depth,  $d_1$ , into the patient **730** where the first treatment position **2111** is illustrated at a depth corresponding to the Bragg peak. Through the process illustrated in FIG. **20**, the treatment beam **269** yields gamma rays **2121** that are detected using a first gamma ray detector **2131** and a second gamma ray detector **2132**, such as respectively mounted on first support **2141** on a first side of the tumor **720** and a second support **2142** on a second side of the tumor **720**, which is preferably the opposite side of the patient to capture paired gamma ray signals. Optionally and preferably, a common support is used to mount the first and second supports **2141**, **2142**. As illustrated, at a second time,  $t_2$ , the treatment beam **269** having a second energy has a second

treatment position **2112** at a second depth,  $d_2$ , where the first and second gamma ray detectors **2131**, **2132** are translated, such as via the first and second support **2141**, **2142** to positions on opposite sides of the tumor **720**, to a second position corresponding to the second energy and the second depth of penetration into the patient **730**. Similarly, at a third time,  $t_3$ , of  $n$  times, where  $n$  is a positive integer, the treatment beam **269** having a third energy, of  $n$  energies corresponding to a set of treatment depths **2110**, treats a third treatment position **2113** at a third depth,  $d_3$ , where the first and second gamma ray detectors **2131**, **2132** are translated to positions on opposite sides of the third depth of penetration. In this manner, the first and second gamma ray detectors **2131**, **2132** are aligned with the expected depth of penetration corresponding with the current energy of the treatment beam. Generally, as energy of the treatment beam **269** increases, the gamma ray detectors are positioned on opposite sides of a large depth of penetration and as the energy of the treatment beam decreases **269**, the detectors are positioned on opposite sides of a relatively shallower depth of penetration. If the first and second gamma ray detectors **2131**, **2132** yield low signals versus an expected signal, then the expected treatment position is not met and the position of the first and second gamma ray detectors **2131**, **2132** is scanned or dithered, such as along the z-axis as illustrated, to find the maximum gamma ray signal corresponding to an actual depth of penetration. The inventor notes that previously treated positions yield decaying intensities of emitted gamma rays based upon the initial intensity of the treatment beam **269**, historical trail positions of the treatment beam **269** overlapping the monitored position, cross-sectional target of the target atom, such as carbon, nitrogen, oxygen, or any atom, and concentration of the target atom, all of which are calculable and are optionally and preferably used to: monitor total treatment of a voxel, determine a current treatment position of the treatment beam **269**, dynamically control subsequent positions of the treatment beam **269**, and/or record a history of actual treatment. Any deviation between planned treatment and actual treatment is noted in the treatment record, such as for a subsequent treatment, and/or is used in dynamic control of the charged particle beam. Optionally, an array of gamma ray detectors is used, such as in place of a moveable gamma ray detector. Similarly, optionally a pair of gamma ray detectors is used in place of the illustrated translatable first and second gamma ray detectors **2131**, **2132**.

[0255] Still referring to FIG. 21, optionally two or more off-axis gamma ray detector are used to determine treatment of a current voxel and/or treatment of a previously treated voxel. For example, if a first pair of gamma ray detectors are used to determine a z-axis position, a second pair of gamma ray detectors are used to determine an x-axis position or a y-axis position. Similarly, a first pair of gamma ray detector arrays is optionally combined with a second pair of gamma ray detector arrays to increase accuracy and/or precision of treatment of a given voxel. Similarly, gamma ray detectors are optionally positioned and/or moved in a manner similar to placement and/or movement of any of the X-ray sources and/or X-ray detectors described above, such as supported and rotated about the tumor **720** of the patient **730** using the gantry **960**, the rotatable gantry support **1810**, and/or the rotatable imaging system support **1812**.

[0256] The inventor notes that the positron emission system here uses radioactive nuclei or isotopes formed in-vivo,

in stark contrast to positron emission tomography systems that generate isotopes externally that are subsequently injected into the body. As the in-vivo radioactive nuclei are formed in the tumor **720**, the gamma rays are emitted from the tumor **720** and not the tumor and all of the surrounding tissue. This aids a signal-to-noise ratio of the acquired gamma ray signal as the background noise of additional non-tumor body tissues emitting gamma rays is substantially reduced and completely removed if only considering the Bragg peak resultant in the peak signal. Further, as the radioactive nuclei is formed in-vivo, time is not required to recover the radioactive nuclei, move the radioactive nuclei to the patient, inject the radioactive nuclei in the patient, and let the radioactive nuclei disperse in the patient, which take ten minutes or more. In stark contrast, the radioactive nuclei is formed in-situ, such as in the tumor **720**, which allows analysis of any radioactive nuclei with a short half-life, such as less than 10, 5, 2, 1, 0.5, 0.1, or 0.05 minutes. Still further, as position of the treatment beam **269** is monitored or determined as a function of time, tomograms of the tumor are generated. The individual tomograms are optionally combined to form an image of the treated tumor or of the tumor as a function of treatment time allowing a video of the collapse of the tumor to be generated and analyzed for real-time modification of the tumor treatment with the treatment beam, to enhance protocols for future tumor treatments of others, and or to monitor sections of the tumor requiring a second treatment.

#### Multiple Beam Energies

[0257] Optionally, the sample, patient, and/or tumor is imaged using two or more energies of the treatment beam **269**. In analysis, resulting images or responses using a first beam energy and a second beam energy, of the two or more energies, are used in an analysis that removes at least one background signal or error from one or more voxels and/or pixels of the obtained images, such as by: taking a ratio of the two signals, calculating a difference between the two signals, by normalizing the images, and/or by comparing the images. By comparing images, tomograms, values, and/or signals obtained with at least two incident beam energies of the treatment beam **269**, background interference is reduced and/or removed. In the case of imaging a tumor, the process of comparing signals with differing incident beam energies reduces and/or removes interference related to skin, collagen, elastic, protein, albumin, globulin, water, urea, glucose, hemoglobin, lactic acid, cholesterol, fat, blood, interstitial fluid, extracellular fluid, intracellular fluid, a sample constituent, temperature, and/or movement of the sample so that the intended element for imaging, such as the tumor, is enhanced in terms of at least one of resolution, accuracy, precision, identification, and spatial boundary.

[0258] Still yet another embodiment includes any combination and/or permutation of any of the elements described herein.

[0259] The main controller, a localized communication apparatus, and/or a system for communication of information optionally comprises one or more subsystems stored on a client. The client is a computing platform configured to act as a client device or other computing device, such as a computer, personal computer, a digital media device, and/or a personal digital assistant. The client comprises a processor that is optionally coupled to one or more internal or external input device, such as a mouse, a keyboard, a display device,

a voice recognition system, a motion recognition system, or the like. The processor is also communicatively coupled to an output device, such as a display screen or data link to display or send data and/or processed information, respectively. In one embodiment, the communication apparatus is the processor. In another embodiment, the communication apparatus is a set of instructions stored in memory that is carried out by the processor.

**[0260]** The client includes a computer-readable storage medium, such as memory. The memory includes, but is not limited to, an electronic, optical, magnetic, or another storage or transmission data storage medium capable of coupling to a processor, such as a processor in communication with a touch-sensitive input device linked to computer-readable instructions. Other examples of suitable media include, for example, a flash drive, a CD-ROM, read only memory (ROM), random access memory (RAM), an application-specific integrated circuit (ASIC), a DVD, magnetic disk, an optical disk, and/or a memory chip. The processor executes a set of computer-executable program code instructions stored in the memory. The instructions may comprise code from any computer-programming language, including, for example, C originally of Bell Laboratories, C++, C#, Visual Basic® (Microsoft, Redmond, Wash.), Matlab® (MathWorks, Natick, Mass.), Java® (Oracle Corporation, Redwood City, Calif.), and JavaScript® (Oracle Corporation, Redwood City, Calif.).

**[0261]** Herein, any number, such as 1, 2, 3, 4, 5, is optionally more than the number, less than the number, or within 1, 2, 5, 10, 20, or 50 percent of the number.

**[0262]** The particular implementations shown and described are illustrative of the invention and its best mode and are not intended to otherwise limit the scope of the present invention in any way. Indeed, for the sake of brevity, conventional manufacturing, connection, preparation, and other functional aspects of the system may not be described in detail. Furthermore, the connecting lines shown in the various figures are intended to represent exemplary functional relationships and/or physical couplings between the various elements. Many alternative or additional functional relationships or physical connections may be present in a practical system.

**[0263]** In the foregoing description, the invention has been described with reference to specific exemplary embodiments; however, it will be appreciated that various modifications and changes may be made without departing from the scope of the present invention as set forth herein. The description and figures are to be regarded in an illustrative manner, rather than a restrictive one and all such modifications are intended to be included within the scope of the present invention. Accordingly, the scope of the invention should be determined by the generic embodiments described herein and their legal equivalents rather than by merely the specific examples described above. For example, the steps recited in any method or process embodiment may be executed in any order and are not limited to the explicit order presented in the specific examples. Additionally, the components and/or elements recited in any apparatus embodiment may be assembled or otherwise operationally configured in a variety of permutations to produce substantially the same result as the present invention and are accordingly not limited to the specific configuration recited in the specific examples.

**[0264]** Benefits, other advantages and solutions to problems have been described above with regard to particular embodiments; however, any benefit, advantage, solution to problems or any element that may cause any particular benefit, advantage or solution to occur or to become more pronounced are not to be construed as critical, required or essential features or components.

**[0265]** As used herein, the terms “comprises”, “comprising”, or any variation thereof, are intended to reference a non-exclusive inclusion, such that a process, method, article, composition or apparatus that comprises a list of elements does not include only those elements recited, but may also include other elements not expressly listed or inherent to such process, method, article, composition or apparatus. Other combinations and/or modifications of the above-described structures, arrangements, applications, proportions, elements, materials or components used in the practice of the present invention, in addition to those not specifically recited, may be varied or otherwise particularly adapted to specific environments, manufacturing specifications, design parameters or other operating requirements without departing from the general principles of the same.

**[0266]** Although the invention has been described herein with reference to certain preferred embodiments, one skilled in the art will readily appreciate that other applications may be substituted for those set forth herein without departing from the spirit and scope of the present invention. Accordingly, the invention should only be limited by the Claims included below.

1. A method for treating a tumor of a patient using a treatment beam, comprising the steps of:

transporting positively charged particles, of the treatment beam, from an accelerator to the tumor using a beam transport system, the charged particles sequentially resultant in formation of a radioactive nuclei in the tumor, positron emission from the radioactive nuclei, electron positron annihilation, and gamma ray emission;

providing a dynamic treatment beam positioning system comprising a set of detectors and a controller;

detecting the gamma ray emission using said set of detectors;

said controller, using output of said set of detectors:

determining a corresponding source voxel of the gamma ray emission;

comparing position of said source voxel to a target voxel;

generating a feedback signal to said beam transport system; and

dynamically adjusting state of subsequent positively charged particles of the treatment beam using the feedback signal.

2. The method of claim 1, further comprising the step of: positioning a first detector and a second detector, of said set of detectors, on opposite sides of the patient.

3. The method of claim 2, further comprising the steps of: moving said set of detectors toward an exit nozzle of said beam transport system to monitor a first depth of penetration of the treatment beam into the tumor; and moving said set of detectors away from said exit nozzle to monitor a second depth of penetration of the treatment beam into the tumor, the second depth of penetration larger than the first depth of penetration.

4. The method of claim 3, further comprising the step of: said controller positioning said set of detectors as a function of energy of the treatment beam.
5. The method of claim 2, further comprising the step of: translating said first detector and said second detector past the patient; and imaging the tumor during said step of translating.
6. The method of claim 5, said step of imaging further comprising the steps of:  
subsequent to a treatment session of the tumor using the treatment beam, generating a first positron emission tomography image of the tumor using ongoing gamma ray emission resultant from ongoing decay of the radioactive nuclei.
7. The method of claim 6, said step of imaging further comprising the steps of:  
subsequent to said step of generating the first positron emission tomography image and prior to the patient departing from a treatment position used in the treatment session, generating a second positron emission tomography image of the tumor using ongoing gamma ray emission resultant from ongoing decay of the radioactive nuclei; and comparing the second positron emission tomography image to the first positron emission tomography image.
8. The method of claim 6, further comprising the step of: generating at least one two-dimensional X-ray image of the patient while the patient remains in the treatment position used in the treatment session.
9. The method of claim 8, further comprising the step of: forming a hybrid positron emission tomography-X-ray tomography image of an original tumor position of the tumor.
10. The method of claim 2, said step of detecting further comprising the step of:  
detecting gamma ray emission resultant from an atom irradiated in said step of transporting, the radioactive decay of the atom comprising a half-life of less than two minutes.
11. The method of claim 10, the radioactive decay of the atom comprising a half-life of less than thirty seconds.
12. The method of claim 2, further comprising the step of: determining a current treatment voxel through detection of gamma rays from a voxel not previously emitting gamma rays in a current treatment session of the tumor.
13. The method of claim 12, said step of determining a current treatment voxel further comprising the step of:

- determining the current treatment voxel on a first axis with a first pair of gamma ray detectors of said set of detectors; and  
determining the current treatment voxel on a second axis with a second pair of gamma ray detectors of said set of detectors, the first axis and the second axis forming an angle of greater than forty degrees.
14. An apparatus for treating a tumor of a patient using a treatment beam, comprising:  
a beam transport system configured to transport positively charged particles, of the treatment beam, from an accelerator to a patient positioning system, the charged particles sequentially resultant in formation of radioactive nuclei in the tumor, positron emission from the isotope, electron positron annihilation, and gamma ray emission; and  
a dynamic treatment beam positioning system, comprising:  
a set of detectors configured to detect the gamma ray emission;  
a controller, said controller:  
configured to determine a corresponding source voxel of the gamma ray emission;  
compare position of the source voxel to a target voxel; and  
generate a feedback signal to said beam transport system, said feedback signal used to dynamically adjust state of subsequent positively charged particles of the treatment beam.
15. The apparatus of claim 14, said dynamic treatment beam positioning system further comprising:  
a first mount mounting a first detector of said set of detectors on a first side of the tumor during use; and  
a second mount mounting a second detector of said set of detectors on a second side of the tumor during use, the first side and the second side on opposite sides of the tumor.
16. The apparatus of claim 15, said set of detectors further comprising:  
a first array of detectors positioned on a first side of a beam path of the treatment beam during use; and  
a second array of detectors positioned on a second side of the beam path during use, the beam path between the first side and the second side.
17. The apparatus of claim 14, said set of detectors further comprising:  
an arc of gamma ray emission detectors circumferentially positioned about the patient positioning system.

\* \* \* \* \*