



US007859473B2

(12) **United States Patent**
Gibson

(10) **Patent No.:** **US 7,859,473 B2**
(45) **Date of Patent:** **Dec. 28, 2010**

(54) **SYSTEMS AND METHODS FOR MANAGING INFORMATION RELATING TO MEDICAL FLUIDS AND CONTAINERS THEREFOR**

(75) Inventor: **Chad M. Gibson**, Cincinnati, OH (US)

(73) Assignee: **Mallinckrodt Inc.**, Hazelwood, MO (US)

(*) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 686 days.

4,401,108 A	8/1983	Galkin et al.
4,526,404 A	7/1985	Vazquez
4,604,847 A	8/1986	Moulding, Jr. et al.
4,690,676 A	9/1987	Moulding, Jr. et al.
4,781,696 A	11/1988	Moulding, Jr. et al.
4,853,521 A	8/1989	Claeys et al.
4,854,324 A	8/1989	Hirschman et al.
4,897,789 A	1/1990	King et al.

(Continued)

(21) Appl. No.: **11/617,987**

(22) Filed: **Dec. 29, 2006**

(65) **Prior Publication Data**

US 2007/0229266 A1 Oct. 4, 2007

Related U.S. Application Data

(62) Division of application No. PCT/US2006/012620, filed on Apr. 4, 2006.

(60) Provisional application No. 60/668,647, filed on Apr. 6, 2005, provisional application No. 60/716,166, filed on Sep. 12, 2005, provisional application No. 60/668,681, filed on Apr. 6, 2005, provisional application No. 60/681,252, filed on May 16, 2005, provisional application No. 60/718,545, filed on Sep. 19, 2005.

(51) **Int. Cl.**
H01Q 11/12 (2006.01)
H01Q 21/00 (2006.01)
G08B 13/14 (2006.01)

(52) **U.S. Cl.** **343/742; 343/867; 340/572.1**

(58) **Field of Classification Search** **343/741, 343/742, 867; 340/572.1**

See application file for complete search history.

(56) **References Cited**

U.S. PATENT DOCUMENTS

2,195,976 A	4/1940	Schlitt
4,296,334 A	10/1981	Wong

FOREIGN PATENT DOCUMENTS

EP	1 433 456 A1	6/2004
----	--------------	--------

(Continued)

OTHER PUBLICATIONS

European Patent Office, European Search Report, Application No. EP07021642.9-2310, Feb. 28, 2005 (5 pages).


(Continued)

Primary Examiner—Shih-Chao Chen
(74) *Attorney, Agent, or Firm*—Wood, Herron & Evans, LLP

(57) **ABSTRACT**

The present invention relates to management of information relating to medical fluids, containers therefore, and medical fluid administration devices for administering such medical fluids to patients. Data tags (e.g., RFID tags) are generally associated with containers of the invention and may be electromagnetically read from and/or written to using an electromagnetic device, for example, that may be associated with a medical fluid administration device of the invention.

19 Claims, 25 Drawing Sheets



U.S. PATENT DOCUMENTS							
			6,958,053	B1	10/2005	Reilly	
4,954,969	A	9/1990	6,961,000	B2*	11/2005	Chung	340/572.1
4,978,335	A	12/1990	6,962,601	B2	11/2005	Becker et al.	
5,047,014	A	9/1991	6,980,419	B2	12/2005	Smith et al.	
5,078,683	A	1/1992	6,995,641	B2	2/2006	Hasson et al.	
5,108,372	A	4/1992	6,995,673	B1	2/2006	Osredkar	
5,195,976	A	3/1993	7,012,529	B2	3/2006	Sajkowsky	
5,274,239	A	12/1993	7,034,683	B2	4/2006	Ghazarian	
5,306,252	A	4/1994	7,158,030	B2*	1/2007	Chung	340/572.1
5,317,506	A	5/1994	7,171,426	B2	1/2007	Farmer et al.	
5,322,511	A	6/1994	7,175,081	B2	2/2007	Andreasson	
5,378,231	A	1/1995	7,258,276	B2	8/2007	Linton et al.	
5,383,858	A	1/1995	7,264,148	B2	9/2007	Tachibana	
5,445,621	A	8/1995	7,304,913	B2	12/2007	Niemiec	
5,568,810	A	10/1996	2001/0028308	A1	10/2001	De La Huerga	
5,651,775	A	7/1997	2001/0034506	A1	10/2001	Hirschman et al.	
5,662,612	A	9/1997	2002/0038392	A1	3/2002	De La Huerga	
5,681,285	A	10/1997	2002/0044059	A1	4/2002	Reeder	
5,699,803	A	12/1997	2002/0076429	A1	6/2002	Wironen	
5,729,236	A	3/1998	2002/0106411	A1	8/2002	Wironen	
5,739,508	A	4/1998	2002/0111584	A1	8/2002	Walker	
5,741,232	A	4/1998	2002/0118111	A1	8/2002	Brown	
5,840,026	A	11/1998	2002/0143320	A1	10/2002	Levin	
5,865,805	A	2/1999	2002/0176893	A1	11/2002	Wironen	
5,882,338	A	3/1999	2002/0180606	A1	12/2002	Kitaguchi	
5,920,054	A	7/1999	2002/0196150	A1	12/2002	Wildman	
5,920,554	A	7/1999	2003/0011476	A1	1/2003	Godfrey	
5,927,351	A	7/1999	2003/0040700	A1	2/2003	Hickle et al.	
5,928,197	A	7/1999	2003/0052788	A1	3/2003	Chung	
5,954,700	A	9/1999	2003/0055685	A1	3/2003	Cobb	
5,980,501	A	11/1999	2003/0060754	A1	3/2003	Reilly	
5,997,502	A	12/1999	2003/0060761	A1	3/2003	Evans	
6,019,745	A	2/2000	2003/0074223	A1	4/2003	Hickle	
6,036,458	A	3/2000	2003/0125665	A1	7/2003	Rosenman	
6,069,853	A	5/2000	2003/0135388	A1	7/2003	Kowalik	
6,090,064	A	7/2000	2003/0139640	A1	7/2003	Whittacre et al.	
6,110,152	A	8/2000	2003/0160698	A1	8/2003	Andreasson et al.	
6,161,031	A	12/2000	2003/0174046	A1	9/2003	Abrams	
6,167,394	A	12/2000	2003/0183226	A1	10/2003	Brand et al.	
6,221,051	B1	4/2001	2003/0183683	A1	10/2003	Stewart	
6,269,340	B1	7/2001	2003/0216688	A1	11/2003	Huybregts	
6,269,648	B1	8/2001	2003/0233069	A1	12/2003	Gillespie, Jr. et al.	
6,344,030	B1	2/2002	2004/0008123	A1	1/2004	Carrender et al.	
6,351,215	B2	2/2002	2004/0019464	A1	1/2004	Martucci et al.	
6,366,206	B1	4/2002	2004/0024361	A1	2/2004	Fago	
6,385,483	B1	5/2002	2004/0027244	A9	2/2004	Menard	
6,387,092	B1	5/2002	2004/0051368	A1	3/2004	Caputo et al.	
6,397,098	B1	5/2002	2004/0054325	A1	3/2004	Ginsburg	
6,397,334	B1	5/2002	2004/0100376	A1	5/2004	Lye	
6,402,717	B1	6/2002	2004/0104271	A1	6/2004	Martucci et al.	
6,440,158	B1	8/2002	2004/0176984	A1	9/2004	White	
6,475,192	B1	11/2002	2004/0193449	A1	9/2004	Wildman	
6,575,906	B1	6/2003	2004/0199075	A1	10/2004	Evans, III	
RE38,189	E	7/2003	2004/0199076	A1	10/2004	Nemoto	
6,626,862	B1	9/2003	2004/0199115	A1	10/2004	Rosenman	
6,633,513	B1	10/2003	2004/0220824	A1	11/2004	Imai	
6,641,052	B2	11/2003	2005/0029277	A1	2/2005	Tachibana	
6,685,678	B2	2/2004	2005/0035862	A1	2/2005	Wildman	
6,704,592	B1	3/2004	2005/0038386	A1	2/2005	Fago	
6,714,121	B1	3/2004	2005/0104363	A1	5/2005	Wright	
6,733,478	B2	5/2004	2005/0107698	A1	5/2005	Powers et al.	
6,733,495	B1	5/2004	2005/0182358	A1	8/2005	Veit et al.	
6,743,202	B2	6/2004	2005/0198800	A1	9/2005	Reich	
6,751,615	B2	6/2004	2005/0199698	A1	9/2005	Glynn	
6,808,513	B2	10/2004	2005/0277833	A1	12/2005	Williams, Jr.	
6,828,577	B2	12/2004	2006/0030773	A1	2/2006	Uber, III et al.	
6,846,298	B1	1/2005	2006/0072030	A1	4/2006	Silverbrook	
6,861,954	B2	3/2005	2006/0109105	A1	5/2006	Varner	
6,861,993	B2	3/2005	2006/0152367	A1	7/2006	Narayanaswamy	
6,901,283	B2	5/2005	2006/0183912	A1	8/2006	Assaf	
6,903,656	B1	6/2005	2006/0183913	A1	8/2006	Assaf	
6,920,088	B2	7/2005	2006/0187050	A1	8/2006	Wagner	
			2007/0003593	A1	1/2007	Wironen	

2007/0021382 A1 1/2007 Assaf
2007/0163583 A1 7/2007 Brand
2007/0167919 A1 7/2007 Nemoto
2007/0189026 A1 8/2007 Chemel
2007/0191690 A1 8/2007 Hasse
2007/0197974 A1 8/2007 Gibson
2007/0198297 A1 8/2007 Perkins
2007/0208308 A1 9/2007 Gibson
2007/0208445 A1 9/2007 Gibson
2007/0225672 A1 9/2007 Wagner
2007/0272746 A1 11/2007 Ortiz
2008/0195249 A1 8/2008 Rouso

FOREIGN PATENT DOCUMENTS

EP 1433456 A1 6/2004
EP 1723977 A1 11/2006
WO 03026724 A1 4/2003
WO 2005010796 A2 7/2004
WO 2004086997 A1 10/2004
WO 2004088567 A2 10/2004
WO 2004095379 A1 11/2004

OTHER PUBLICATIONS

Office Action from Canadian Patent Office, dated Feb. 3, 2009, for Canadian Application No. 2,622,407 (two pages).

* cited by examiner




FIG. 1A




FIG. 1B




FIG. 1C




FIG. 2A




FIG. 2B




FIG. 2C




FIG. 2D




FIG. 3A




FIG. 3B




FIG. 3C




FIG. 4




FIG. 5A




FIG. 5B




FIG. 6




FIG. 7




FIG. 8




FIG. 9




FIG. 10




FIG. 11




FIG. 12




FIG. 13A




FIG. 13B




FIG. 13C




FIG. 13D




FIG. 14




FIG. 15




FIG. 16




FIG. 17




FIG. 18




FIG. 19




FIG. 20




FIG. 21




FIG. 22




FIG. 22A

**SYSTEMS AND METHODS FOR MANAGING
INFORMATION RELATING TO MEDICAL
FLUIDS AND CONTAINERS THEREFOR**

CROSS REFERENCE TO RELATED
APPLICATIONS

This application is a divisional of and claims priority to PCT Application PCT/US2006/012620, entitled SYSTEMS AND METHODS FOR MANAGING INFORMATION RELATING TO MEDICAL FLUIDS AND CONTAINERS THEREFOR, filed on Apr. 4, 2006, the disclosure of which is hereby incorporated by reference in its entirety, which claims the benefit of the following U.S. Provisional Applications that are hereby incorporated in their entireties by reference herein:

U.S. Provisional Application Ser. No. 60/668,647, filed 6 Apr. 2005 and entitled SYSTEM AND METHOD FOR TRACKING INFORMATION RELATING TO A PHARMACEUTICAL CONTAINER AND/OR PHARMACEUTICAL DISPOSED THEREIN;

U.S. Provisional Application Ser. No. 60/716,166, filed 12 Sep. 2005 and entitled SYSTEM AND METHOD OF TRACKING INFORMATION RELATING TO A PHARMACEUTICAL CONTAINER IN A CT SCANNING SUITE;

U.S. Provisional Application Ser. No. 60/668,681, filed 6 Apr. 2005 and entitled APPARATUS AND METHOD FOR LABELING RADIOPHARMACEUTICALS;

U.S. Provisional Application Ser. No. 60/681,252, filed 16 May 2005 and entitled APPARATUS AND METHOD FOR LABELING RADIOPHARMACEUTICALS; and

U.S. Provisional Application Ser. No. 60/718,545, filed 19 Sep. 2005 and entitled ANTENNA SYSTEM AND METHOD OF READING A DATA TAG ON A CONTRAST MEDIA CONTAINER.

FIELD OF THE INVENTION

The present invention relates generally to medical fluids (e.g., radiopharmaceuticals, contrast media) and, more particularly, to tracking and/or managing information relating to medical fluids, containers therefore, and/or medical fluid administration devices used to administer such medical fluids.

BACKGROUND

Proper administration of pharmaceuticals (e.g., contrast media, radiopharmaceuticals) is dependent on human reliability to insure the correct drug is administered properly. In the case of injectable pharmaceuticals, the consequences of mistakes can be severe. Statistically the accuracy of the health care system in providing correct injections is excellent. However, with millions of injections per year, there is a continuing effort to further reduce mistakes, a great majority of which are the result of human error.

Of particular interest is the packaging, distribution and use of contrast media or a contrast agent. As used herein, a contrast media or agent is a substance that is introduced into, on, or around a physiological structure (e.g., tissue, vasculature, cell); and because of the differences in absorption by the contrast media and the surrounding tissues, the contrast media allows a radiographic visualization of the structure. Contrast media is used in x-ray computed tomography (CT), magnetic resonance imaging (MR), ultrasound imaging, angiographic imaging, and other procedures. Often, a con-

tainer, for example, a syringe, is filled with a desired quantity of the contrast media by an independent supplier; and the filled syringes of contrast media are sold or otherwise provided to a hospital, imaging service provider or other health care facility.

Over the useful life of the contrast media and its associated syringe, there are three principal areas of interest for tracking purposes: 1) the location where the contrast media is packaged in a container (e.g., a syringe); 2) the distribution and storage of the filled syringe; and 3) the use and disposal of the syringe. The filling of a syringe with contrast media can occur at a supplier's facility separate from a health care facility; or in some circumstances, within a pharmacy of the health care facility. Contrast media comes in many types and concentrations and can be filled in syringes of different sizes that also vary with the type of injector to be used. Further, the contrast media has a limited shelf life and a more limited life when open to atmosphere or when heated in preparation for injection. Thus, in order to properly fill a syringe with contrast media, knowledge of the contrast media's use, the injector and sometimes an identity of a patient are required. In addition, proper use of the contrast media requires knowledge of its age and other information relating to when the syringe was filled.

Currently, all this information is manually collected by pharmacists and X-ray technologists. The technologist then uses this information to manually set up the injection; and currently, this information must be manually transposed onto various records. Known systems for managing pharmaceuticals provide filled syringes with bar codes having SKUs and other indicia relating to various filled sizes and concentrations of contrast media. But this system is limited in use and does not provide an efficient management of all of the parameters needed in a medical environment and particularly in connection with the use of contrast media. There is a need for a more automated system for entering information relating to contrast media upon filling a syringe. There is a further need to automatically track a particular syringe through a distribution system whether from a supplier external to a health care facility and/or from a pharmacy within the facility.

A typical X-ray department has an X-ray contrast warming device or box. This device is used to raise the temperature of the contrast media to body temperature before it is manually injected or installed in an injector. Additionally, it is considered normal for X-ray departments to store more than the day's requirement of contrast media in the warmer box. This creates a complex situation for an X-ray technologist responsible for manually keeping track of sometimes dozens of contrast media syringes. The syringes have to be tracked by quantity, type and time in the warmer box; and the contrast media syringes should be used on a first-in first-out basis. As a result, a situation may result where there is too much of one type or not enough of another type. This manual tracking of contrast media syringes may also result in some syringes staying in the warmer box too long, and others being mistakenly removed before they have been properly warmed. Therefore, there is a need for a more automated system for tracking contrast media syringes in a warmer box.

Power injectors are frequently used to inject X-ray contrast media into patients receiving X-ray imaging procedures. X-ray technologists may encounter distractions in the course of executing an X-ray procedure thus leading to the possibility of injecting a patient using an empty syringe. An empty syringe injection often occurs when a technologist retracts a plunger of a syringe with the power injector after an injection but inadvertently does not replace the empty syringe with a new full syringe—when the next patient is prepared for imag-

ing the technologist fails to recognize the empty syringe loaded in the power injector because the fully retracted empty syringe looks like a full syringe with contrast media. To reduce the risk of using an empty syringe, power injectors often prompt the technologist with a message asking the technologist to confirm that air has been purged out of the syringe and tubing. However, a technologist may answer “yes” to the prompt without carefully checking the syringe and tubing with the result that air is injected into a patient. Therefore, there is a need for a more automated system for preventing use of an empty syringe.

It is possible to refill almost any empty syringe with contrast media. Some syringes are intended to be refilled, whereas others are not. However, some engage in a practice of refilling syringes that are not intended to be refilled and/or refill a syringe improperly with a risk of trapping air within the syringe. Therefore, there is a need for an automated system for tracking the use of a syringe and preventing its subsequent unauthorized re-use.

The installed base of power injectors in the world is very large due to their reliability and long useful life. Throughout the life of a power injector, the diameter and length of syringes used in that injector may vary due to tooling, material or process changes over time, or even normal variations from batch-to-batch. Known power injectors have fixed programming for syringe sizes and are not setup to automatically make adjustments for minor variations in the diameter and length of a syringe. By assuming a diameter and length for a syringe, the volume delivery accuracy of a power injector is limited. For example, variations in syringe size result in a typical volume accuracy specification for a power injector of about ± 2 milliliters (“ml”) per injection, even though the electronics and mechanical transmission are capable of much better. Therefore, there is a need for an automated system for determining variations in syringe size, so that better volume delivery accuracy can be achieved.

When a power injector fails to operate correctly, a service engineer must be called. In analyzing a power injector experiencing operating problems, the injector is operated in a “service” mode, which is often achieved by installing electrical jumpers in an injector control. The service mode makes testing and troubleshooting the power injector easier, but the service mode often disables some safety features of the injector. Use of a jumper is simple technology; and it is relatively easy for a customer to invoke service mode without authorization, for example, to avoid the inconvenience of various safety checks when using the injector. Furthermore, service mode may also be accidentally left enabled. Since a jumper is located on rear connection panels, it is not readily visible; and it is possible for the jumper to be mistakenly left in the power injector, in which case the injector is left in service mode. If the service mode is used for a medical procedure, either deliberately or mistakenly, the injector may not perform in a safe manner. Therefore, there is a need for a better system for placing a power injector in a service mode and preventing normal use of the power injector while it is in the service mode.

Sometimes, when a power injector is not operating properly, the improper operation cannot be repeated, is intermittent or just cannot be solved by the service engineer. In such cases the power injector is temporarily replaced and returned to the factory for a more thorough examination. Upon the power injector being returned, factory personnel sometimes do not receive sufficient information about the power injector’s defective operation to effectively resolve the problem.

Therefore, there is a need for a better system of communicating defective operating conditions to factory personnel for service purposes.

Often power injector manufacturers embed all possible features into the injector’s software, even though some customers do not want particular features. Manufacturers do this to reduce the development cost and the complexity of installations. However, when the manufacturer has a very high value feature, the manufacturer must find a cost-effective and reliable method of activating that feature for only those customers who have paid for it. Therefore, there is a need for a better system that permits a manufacturer to embed all operating features but automatically activate only those features that a particular customer has purchased.

There is also a need for an automated system that tracks syringes from the time they are filled with a contrast media, through their distribution to a health care facility and/or an imaging suite, through the injection of the contrast media from the syringe and then the disposal or authorized refilling of the syringe. There is a further need for such an automated system to communicate information regarding the injection of contrast media to patient records.

Similar problems and needs also exist with respect to the manufacture, storage and use of other pharmaceuticals such as radioactive pharmaceuticals or radiopharmaceuticals. Radiopharmaceuticals, are often prepared at a radiopharmacy in which a syringe or vial may be filled with a desired quantity of the radiopharmaceutical. The syringe or vial may then be placed into a container called a “pig” that generally includes lead and/or other radiation shielding material to protect handlers from exposure to radiation from the radiopharmaceutical. After delivery, the pig may be opened; the syringe or vial may be removed; and the radiopharmaceutical may be administered to a patient. The used syringe or vial may then be put back in the pig, and pig and syringe or vial may be returned to the radiopharmacy for disposal of the syringe and reuse or disposal of the pig. For purposes of this document, the term “container” means a structure for holding a radiopharmaceutical and from which the radiopharmaceutical may be dispensed, for example, a syringe, vial, etc.

Some radiopharmacies have nuclear medicine tracking systems that use bar code readers to read bar codes on prescription labels to facilitate shipment and receipt of the radiopharmaceutical pig and syringe or vial. Therefore, a person in a receiving nuclear medicine department can scan the prescription label on the pig to enter data into a procedural data system. While this known use of bar codes has improved the reliability of passing prescription information through a distribution channel, bar codes have a significant disadvantage. Bar codes store only a limited amount of information, are “read only” devices and therefore, do not permit coded information to be changed or updated or new data to be added to the prescription labels. Further, a bar code must be in a “line of sight” of a reader to be useful.

While a syringe or vial may be disposed of after use, the radiopharmaceutical pig is cleaned and reconditioned for reuse. Therefore, instead of using adhesives to attach a pharmaceutical label to a pig, it is known to attach the label to the pig with elastic bands, resilient clear plastic sleeves, etc. While such techniques make a pig easier to clean for reuse, they do have a disadvantage in that reliably maintaining a label and pig together may require substantial human effort in initially applying the label and then checking and double checking the correctness of the label and pig combination over the life of the prescription.

The proper handling and use of radiopharmaceuticals may be said to require highly disciplined processes—and while

the occurrence of mistakes is statistically small, errors still occur in the handling and delivery of radiopharmaceuticals. Thus, there is a need to provide a prescription label for a radiopharmaceutical that addresses the disadvantages described above.

SUMMARY

The present invention is generally directed to managing information relating to a medical fluid, a container therefore, and/or a medical fluid administration device. Containers of the invention typically have a data tag associated therewith to enable information to be read from and/or written to the data tag of the container. This allows information regarding the container and/or the medical fluid associated therewith to be ascertained, and optionally updated, for example, during and/or between various stages of manufacture, transport, storage, use, and/or disposal.

As used herein, a “medical fluid” generally refers to a fluid that is designed to be administered (e.g., intravenously) to a medical patient as a part of a medical procedure (e.g., diagnostic procedure, therapeutic procedure). Examples of medical fluids include, but are not limited to, contrast media, radiopharmaceuticals, and saline. A “container” of the invention generally refers to any container designed to have a medical fluid disposed therein. Examples of containers of the invention include, but are not limited to, syringes, IV bags, and bulk contrast media containers. An “administration device” of the invention refers to any electronic device designed to at least assist in transferring medical fluid from a container to a patient. Examples of medical fluid administration devices of the invention include, but are not limited to, infusion pumps and power injectors.

A first aspect of the invention is directed to a syringe having a medical fluid disposed therein. The syringe includes a data tag for storing data, such as data relating to a software update for a powered fluid injector, a product promotion, and/or an electronic coupon code for sales of further products. Incidentally, a “data tag” herein refers to any device capable of having data electromagnetically read therefrom and/or written thereto (e.g., RFID tag).

A second aspect of the invention is directed to a medical fluid administration device capable of at least assisting in delivering a medical fluid from a container to a patient in a medical procedure. The container includes a data tag for storing data, and the administration device includes an electromagnetic device. Herein, an “electromagnetic device” refers to any device capable of electromagnetically reading data from and/or writing data to a data tag. The data read from the data tag may relate to configuration information for the administration device, a software update for the administration device, a product promotion, and/or an electronic coupon code for purchases of further products. In the case of the data tag including data relating to configuration information, and upon the data being read from the data tag by the electromagnetic device, the configuration information may be used by the administration device to execute a self-configuration cycle.

A third aspect of the invention is directed to a system for use in association with a medical fluid administration device. The system includes a service data tag (e.g., as a component of a badge or card) that may be used by service personnel, and an electromagnetic device associated with the administration device. This electromagnetic device is operable to read data from and/or write data to the service data tag (e.g., to provide

data relating to an identity of the service person and/or configuration information for that particular administration device).

With regard to this third aspect of the invention, the administration device of some embodiments may enable a service mode upon the electromagnetic device detecting data from the service data tag. In some embodiments, the electromagnetic device may write data to the service data tag that relates to service activity information, administration device configuration information and/or administration device use information (e.g., fluid administration protocol statistics, container identifications, medical fluid use information).

A fourth aspect of the invention is directed to a warmer for warming a container having a medical fluid disposed therein. The container has a data tag for storing data associated therewith. The warmer includes both a heating element for elevating the temperature of the medical fluid and an electromagnetic device operable to read data from and/or write data to the data tag associated with the container. The data tag may contain data (which may be read by the electromagnetic device) relating to the amount of medical fluid in the container, the concentration of the medical fluid, manufacturing information regarding the medical fluid and/or the container, the container capacity, the container dimensions, a use code for the medical fluid, and configuration information for a medical fluid administration device to be used in administering the medical fluid to a patient.

With regard to this fourth aspect of the invention, some embodiments may include a user interface (e.g., touch screen) for facilitating user selection of a container in the warmer. In some embodiments, the electromagnetic device may be used to write data relating to use of the medical fluid to the data tag. For example, the electromagnetic device may be used to write data to (and/or read data from) the data tag that relates to a date the container was placed in the warmer, an expiration date for contrast media in the container, and/or administration information for an administration device to be used in administering the medical fluid in the container.

Still a fifth aspect of the invention is directed to a container having a medical fluid disposed therein and a data tag associated therewith. In the case that the medical fluid is a radiopharmaceutical, the data on the data tag of some embodiments may relate to an identity of the radiopharmaceutical, a radioactivity level of the radiopharmaceutical, manufacturing information for the radiopharmaceutical, a use code for the radiopharmaceutical (e.g., identifying whether a radiopharmaceutical container has previously been used in a radiopharmaceutical administration procedure), and/or configuration information for an administration device to be utilized in administering the radiopharmaceutical (e.g., a code that is required by the administration device prior to use of the container, a software update for the administration device, a product promotion, references to information).

Yet a sixth aspect of the invention is directed to a radiopharmaceutical administration device for use in administering a radiopharmaceutical to a patient. This administration device is designed to at least assist in delivering a radiopharmaceutical from a container to a patient. The container has a data tag associated therewith, and the administration device includes an electromagnetic device for reading data from and/or writing data to the data tag. In some embodiments, the data included on the data tag identifies the amount and/or identity of radiopharmaceutical in the container, manufacturing information for the radiopharmaceutical in the container, the radioactivity level of the radiopharmaceutical in the container, a use code for the radiopharmaceutical in the container, configuration information for the administration device to be

used in administering the radiopharmaceutical from the container, and/or particular data regarding a radiopharmaceutical container previously used with the administration device. In some embodiments, the data tag may store data indicative of configuration information for the administration device that includes a code required by the administration device prior to use of the radiopharmaceutical container (e.g., data used by the administration device in self-configuration upon reading of the data tag), a software update for the administration device, a product promotion, and/or references to information. For instance, in some embodiments, the administration device may utilize an electronic coupon code included in the data tag in purchases of further products.

A seventh aspect of the invention is directed to a system for use in a medical procedure with respect to a patient. The system includes a hospital information system, a container having a medical fluid disposed therein, and an administration device for administering the medical fluid to a patient. Associated with the container is a data tag that is readable by electromagnetic signals and that stores signals representing product promotions, coupons, Internet links of the supplier, and/or recommended software updates for administration devices with which the container is intended for use. The system also includes an electromagnetic device for reading data from and/or writing data to the data tag associated with the container. This electromagnetic device may be mounted on the administration device and is preferably in electrical communication with both the hospital information system and the administration device (e.g., the control thereof).

Still further, the system includes an imaging apparatus (e.g., CT scanner) that includes an imaging control, which is preferably in electrical communication with the hospital information system, the control of the administration device, and the electromagnetic device. Incidentally, "electrical communication" or the like herein refers to objects that are directly and/or indirectly connected in a manner such that electricity (e.g., data in the form of electronic signals) can be conveyed between them. Data associated with administration (e.g., injection, infusion) of the medical fluid may be transferred between the hospital information system, the data tag, the control of the administration device, and the imaging control. Some embodiments of this seventh aspect may include a printer in electrical communication with the administration device (e.g., the control thereof).

An eighth aspect of the invention is directed to an administration device for use with a container having medical fluid disposed. In some embodiments, the medical fluid is metallic and/or diamagnetic. The container has a data tag that is readable by electromagnetic signals associated therewith, and the administration device includes an electromagnetic device adapted to read data from and/or write data to the data tag. In some embodiments, this electromagnetic device includes first and second antenna loops, each of which forms one side of a V-shape and is tuned to a radio frequency. Each of the first and second antenna loops may include a signal lead and a ground lead.

Still referring to the eighth aspect of the invention, the electromagnetic device of some embodiments may include first and second tuning circuits that correspond with the first and second antenna loops. These tuning circuits may each include an input and an output. The output of the first tuning circuit may be connected to the signal lead of the first antenna loop and may function to tune the first antenna loop to a radio frequency. Similarly, the output of the second tuning circuit may be connected to the signal lead of the second antenna loop and may function to tune the second antenna loop to a radio frequency (e.g., the same radio frequency as the first

antenna loop). The second antenna loop of the electromagnetic device may be nonparallel (e.g., form an angle of less than 180 degrees) with the first antenna loop.

Some embodiments of this eighth aspect may include additional antenna loops beyond the first and second antenna loops. For instance, some embodiments may include a third antenna loop having both a signal lead and a ground lead, and a third tuning circuit that includes an input and an output. As with the outputs of the first and second tuning circuits, the output of the third tuning circuit may be connected to the signal lead of the third antenna loop and may function to tune the third antenna loop to a radio frequency (e.g., the same radio frequency as the first and/or second antenna loop).

In some embodiments of the eighth aspect of the invention, the administration device may be utilized to support the container. For instance, in some embodiments, the administration device is an electronic fluid injector, and the electromagnetic device is mounted in association with the injector. The administration device may include both a first printed circuit board that supports the first antenna loop and the first tuning circuit, and a second printed circuit board that supports the second antenna loop and the second tuning circuit. The first printed circuit board may be oriented in any of a number of appropriate orientations relative to the second circuit board. For instance, in some embodiments, the first circuit board forms an angle of less than about 180 degrees with the second printed circuit board. The first printed circuit board may support a driver circuit electrically connectable to the first antenna loop, the second antenna loop, the first tuning circuit, and/or the second tuning circuit. This driver circuit may include a power terminal and a ground terminal.

In some embodiments of this eighth aspect, the input of the first tuning circuit is connected to the power terminal, and the ground lead of the first antenna loop is connected to the ground terminal. In addition, the input of the second tuning circuit is not connected to the power terminal or the ground terminal, and the ground lead of the second antenna loop is connected to the ground terminal.

In other embodiments of the eighth aspect, the input of the first tuning circuit is not connected to the power terminal or the ground terminal, and the ground lead of the first antenna loop is connected to the ground terminal. In addition, the input of the second tuning circuit is connected to the power terminal, and the ground lead of the second antenna loop is connected to the ground terminal.

In still other embodiments of the eighth aspect, the input of the first tuning circuit is connected to the power terminal, and the ground lead of the first antenna loop is connected to the ground terminal. In addition, the input of the second tuning circuit is connected to the ground terminal, and the ground lead of the second antenna loop is connected to the ground terminal.

In yet other embodiments of the eighth aspect, the input of the first tuning circuit is connected to the ground terminal, and the ground lead of the first antenna loop is connected to the ground terminal. In addition, the input of the second tuning circuit is connected to the power terminal, and the ground lead of the second antenna loop is connected to the ground terminal.

Some embodiments of the eighth aspect may be equipped with a switching circuit including first and second switches. The first switch may include a first contact connected to the input of the first tuning circuit, a second contact connected to the ground terminal, a third contact connected to the power terminal, and a fourth contact not connected to the ground terminal or the power terminal. This first switch is preferably operable to electrically connect the first contact with at least

one of the second contact, the third contact and the fourth contact. Similarly, the second switch may include a fifth contact connected to the input of the second tuning circuit, a sixth contact connected to the ground terminal, a seventh contact connected to the power terminal, and an eighth contact not connected to the ground terminal or the power terminal. This second switch is preferably operable to electrically connect the fifth contact with at least one of the sixth contact, the seventh contact and the eighth contact.

In a ninth aspect, the invention is directed to a method of using a medical fluid administration device that includes an electromagnetic device operable to read data from and/or write data to a data tag. This data tag is associated with a container that has medical fluid disposed therein. In this method, first and second antenna loops of the electromagnetic device are electrically connected in a first circuit configuration and are tuned to a substantially identical radio frequency. These first and second antenna loops may be oriented in a nonparallel relationship relative to one another. An electromagnetic (e.g., RF) communication may be attempted between the electromagnetic device and the data tag, at least in part, by providing electromagnetic power to the first circuit configuration. A determination may be made as to whether or not electromagnetic communication is or was established between the electromagnetic device and the data tag. If it is determined that electromagnetic communication is/was not made, the first and second antenna loops may be electrically reconnected in a further (e.g., second) circuit configuration different from the first circuit configuration. Then, another electromagnetic communication between the electromagnetic device and the data tag may be attempted, at least in part, by providing electromagnetic power to the further circuit configuration. The process of determining whether or not an electrical communication exists, electrically reconnecting the first and second antenna loops, and attempting another electromagnetic communication may be repeated as desired (e.g., until determining that a successful electromagnetic communication has been established between the electromagnetic device and the data tag).

A tenth aspect of the invention is directed to a method of using a medical fluid administration device that includes an electromagnetic device operable to read data from and/or write data to a data tag. In this method, a data tag is disposed near an antenna system of the electromagnetic device, and a material that interferes with electromagnetic signals (e.g., metallic material, diamagnetic material) is disposed between the data tag and the antenna system. Even though the material is disposed between the data tag and the antenna system, data may still be electromagnetically read from and/or written to the data tag using the electromagnetic device and the antenna system thereof.

In some embodiments of this tenth aspect, the data tag is a component of a container that has medical fluid (which, in this case, is or includes the material) disposed therein. In such embodiments, the medical fluid may be, for example, water, saline, contrast media, or a combination thereof. In such embodiments, the container may be placed near (e.g., in contact with) the administration device in a manner such that the data tag of the container is located near the antenna system and such that the material in the container is located between the data tag and the antenna system. While not always the case, the electromagnetic device and the antenna system thereof may be components of the administration device.

Some embodiments of the antenna system of this tenth aspect may include first and second antenna loops. In these embodiments, the first and second antenna loops may be electrically connected in a first antenna configuration, and

electromagnetic signals from this first antenna configuration may be emitted to at least attempt to electromagnetically read data from and/or electromagnetically write data to the data tag. In response to a failure to electromagnetically read data from and/or electromagnetically write data to the data tag when the first and second antennas are in the first configuration, the first and second antenna loops may be electrically reconnected in another (e.g., second) antenna configuration, and electromagnetic signals from the new antenna configuration may be emitted to again at least attempt to electromagnetically read data from and/or electromagnetically write data to the data tag.

In an eleventh aspect, the invention is directed to a container assembly that includes a medical fluid container that is enclosable inside an enclosure. Associated with the container are both a data tag that includes a data store and an antenna system that is electrically connectable to the data tag. The construction of the enclosure of this eleventh aspect is such that a frequency of electromagnetic signal necessary to read data from and/or write data to the data tag is substantially prevented from passing through the material of the enclosure. The antenna system of this eleventh aspect is designed so that an antenna thereof is located outside the enclosure while the container and the data store of the data tag are enclosed in the enclosure. This antenna system permits data to be read from and/or written to the data store while the container and the data store of the data tag are enclosed within the enclosure.

Still a twelfth aspect of the invention is directed to a radiopharmaceutical assembly that includes a radiopharmaceutical container (e.g., a syringe having a radiopharmaceutical disposed therein) and a radiopharmaceutical pig that is enclosable about the container to fully surround and support the container. In addition, this twelfth aspect includes a data tag that includes a data store and that is attached to the radiopharmaceutical container. An antenna system is electrically connectable to the data tag upon the radiopharmaceutical container (and the data tag attached thereto) being placed in the radiopharmaceutical pig. This antenna system permits data to be read from and/or written to the data store of the data tag while the radiopharmaceutical pig is closed around the radiopharmaceutical container and the data tag.

In some embodiments of this twelfth aspect, the radiopharmaceutical pig may be characterized as having both a first pig component (e.g., a base) adapted to support the radiopharmaceutical container with the data tag and a second pig component (e.g., a cap) that is attachable to the first pig component and adapted to fully enclose the radiopharmaceutical container with the data tag within the radiopharmaceutical pig. In such embodiments, the antenna system may be adapted to be electrically connectable to the data tag upon the radiopharmaceutical container being placed in the first pig component of the radiopharmaceutical pig. The antenna system of these embodiments permits data to be read from and/or written to the data store of the data tag while the first pig component is attached to the second pig component and while the radiopharmaceutical container and the data tag are enclosed inside the radiopharmaceutical pig. In some of these embodiments the antenna system may include an antenna electrically connected to the data tag, an inner antenna adjacent an inner surface of one of the first pig component and the second pig component, an outer antenna adjacent an outer surface of one of the first pig component and the second pig component, and a conductive lead electrically connecting the inner antenna with the outer antenna. The antenna of some embodiments of the twelfth aspect may be attached to (e.g., fixed to) the radiopharmaceutical container.

11

Still referring to the twelfth aspect of the invention, some embodiments of the antenna system may be characterized as having an antenna locatable outside the radiopharmaceutical pig, and a conductive lead that has one end connected to the data tag within the radiopharmaceutical pig and an opposite end connected to the antenna located outside the radiopharmaceutical pig.

In yet a thirteenth aspect, the invention is directed to a power injector capable of supporting a syringe that has a medical fluid disposed therein. Particularly, the medical fluid is located between a plunger and a discharge tip of the syringe. The syringe includes a data tag for storing data that is electromagnetically readable from the data tag. The injector of this thirteenth aspect includes a powerhead having a plunger drive adapted to interface with (e.g., be connected to) the plunger of the syringe. An injector control of the injector is operatively connected to the powerhead. Further, an electromagnetic device of the injector is mounted on the powerhead and is in electrical communication with the injector control. This electromagnetic device includes a plurality of antennas operative to transmit electromagnetic signals to and receive electromagnetic signals from the data tag (e.g., to read data stored in the data tag).

In some embodiments of this thirteenth aspect, the electromagnetic device may include a plurality of tuning circuits electrically connected to respective antennas for tuning the respective antennas to a desired frequency(ies). For instance, in some embodiments, the tuning circuits may be utilized to tune the respective antennas to a frequency of about 13.56 Megahertz. A driver circuit of the electromagnetic device may be electrically connectable to the tuning circuits and the injector control. This driver circuit may function to provide drive signals to the tuning circuits causing the respective antennas to transmit electromagnetic signals to and receive electromagnetic signals from the data tag (e.g., to read data stored in the data tag). Some embodiments may include a switching circuit electrically connected between the driver circuit and the tuning circuits. This switching circuit may be utilized to connect the antennas in different circuit configurations. In some embodiments, at least one of the switching circuit and the driver circuit are located in the powerhead of the injector.

Still referring to the thirteenth aspect of the invention, some embodiments of the powerhead may include a forward end adapted to receive and support the syringe. In some embodiments, this forward end may include or be characterized as a mount of sorts adapted to accommodate (e.g., receive and support) the syringe. In some embodiments, the mount may include a pressure jacket for supporting the syringe. In such embodiments, the antennas may be mounted on the pressure jacket. Some embodiments of the mount may not include a pressure jacket. Some embodiments of the mount may include what may be referred to as a cradle to support the syringe. In such embodiments, the antennas may be supported by and/or located within the cradle.

Some embodiments of the thirteenth aspect may have a pressure jacket that includes an inner sleeve and an outer sleeve disposed about the inner sleeve. One or more antennas may be located between the inner sleeve and the outer sleeve of the pressure jacket. For instance, in some embodiments, a plurality of antennas may be disposed between the inner and outer sleeves and equally spaced about a circumference of the pressure jacket. In some embodiments, one or more tuning circuits may be located between the inner sleeve and the outer sleeve.

Some embodiments of the injector of the thirteen aspect of the invention may include a heater (e.g., for heating the medi-

12

cal fluid disposed in the syringe). For example, in some embodiments, the heater may be attached to or a component of a pressure jacket of the injector. As another example, in some embodiments, the heater may be attached to or a component of a cradle of the injector. In embodiments equipped with a heater, the heater may be electrically connected to the injector control.

The syringe employed in this thirteen aspect of the invention may exhibit any of a number of appropriate structural designs/configurations. For instance, in some embodiments, the plunger of the syringe is substantially wholly contained within a barrel of the syringe. Further, the syringe employed in this thirteen aspect of the invention may exhibit any of a number of appropriate sizes (e.g., volume capacities). As an example, the syringe of some embodiments exhibits a volumetric capacity capable of accommodating a volume of fluid in excess of about 90 ml.

A fourteenth aspect of the invention is directed to a system for managing data relating to a container and/or a medical fluid disposed therein. The container includes a data tag operable to have data written thereto and read therefrom. A filling station of the system may be utilized to place the medical fluid in the container. This filling station includes an electromagnetic device operable to at least write data (e.g., relating to the fluid in the container) to the data tag. Further, a disposal station of the system may be utilized in disposing of and/or preparing for disposal of the container (which may or may not still have medical fluid therein). This disposal station also includes an electromagnetic device operable to write data (e.g., relating to disposal of the container) to the data tag. The system may also include a hospital information system in electrical communications with one or more electromagnetic devices of the system.

In some embodiments of this fourteenth aspect, the system may include a warmer that may be utilized to heat the fluid in the container. This warmer is generally equipped with an electromagnetic device operable to write data (e.g., relating to placing the container in and/or removing the container from the warmer) to the data tag.

Some embodiments of the fourteenth aspect may include a medical fluid administration device. For example, in some embodiments, the administration device is a power injector for use with a syringe. The power injector generally includes both a control and an electromagnetic device that is electrically connected to the control and operable to write data (e.g., relating to administration of the medical fluid into the patient) to the data tag.

In some embodiments, the system of the fourteenth aspect may include a packaging station that may be used in placement of the container into a package. This packaging station may include an electromagnetic device operable to write data (e.g., relating to the package, the fluid and/or the container) to the data tag.

Some embodiments of the system may include a storage area for storing the container (which may or may not already have the medical fluid disposed therein). This storage area generally includes an electromagnetic device operable to write data (e.g., relating to placing the syringe in and/or removing the syringe from the storage area) to the data tag.

In some embodiments of the fourteenth aspect, the medical fluid that is in or is to be placed in the container is a radiopharmaceutical. In such embodiments, a packaging station (e.g., radiopharmacy) of the system may be used during placement of the container into a radiopharmaceutical pig. Further, the packaging station may be utilized when placing the radiopharmaceutical pig in a package (e.g., a transport package). This packaging station may include an electromag-

netic device operable to write data (e.g., relating to the radiopharmaceutical, the container, the pig, and/or the package) to the data tag.

Some embodiments of the system may include a calibration station that includes an electromagnetic device operable to write data (e.g., relating to radioactivity level of the radiopharmaceutical in the container) to the data tag. Some embodiments of the system may include a treatment room where the radiopharmaceutical pig may be received and the container having the radiopharmaceutical disposed therein is removed for administration of the radiopharmaceutical to a patient. This treatment room may include an electromagnetic device operable to write data (e.g., relating to administration of the radiopharmaceutical to the patient) to the data tag. A storage area of the system may include an electromagnetic device operable to write data (e.g., relating to placing the pig into and/or removing the pig from the storage area) to the data tag.

Various refinements exist of the features noted in relation to the above-mentioned aspects of the present invention. Further features may also be incorporated in the above-mentioned aspects of the present invention as well. These refinements and additional features may exist individually or in any combination. For instance, various features discussed below in relation to any of the exemplary embodiments of the present invention may be incorporated into any of the aspects of the present invention alone or in any combination.

BRIEF DESCRIPTION OF THE FIGURES

The accompanying figures, which are incorporated herein and constitute a part of this specification, illustrate exemplary embodiments of the invention and, together with a general description of aspects of the invention given above, and the detailed description of various exemplary embodiments given below, serve to explain various principles of the invention.

FIG. 1A is a schematic drawing of a system for tracking a syringe filled with contrast media over a syringe life cycle.

FIG. 1B is a schematic drawing of a system for tracking a container filled with a radiopharmaceutical over a container life cycle.

FIG. 1C is a schematic drawing of a system for tracking an IV bag filled with a medical fluid over an IV bag life cycle.

FIGS. 2A-2D are perspective views of a syringe that illustrate different manners of applying a tracking device to a syringe filled with contrast media in the system shown in FIG. 1A.

FIG. 3A is a schematic block diagram of components associated with the system illustrated in FIG. 1A.

FIG. 3B is a schematic block diagram of components associated with the system illustrated in FIG. 1B.

FIG. 3C is a schematic block diagram of components associated with the system illustrated in FIG. 1C.

FIG. 4 is a schematic drawing illustrating activities and operations associated with use and disposal of a container of contrast media in an imaging suite.

FIG. 5A is a perspective view of one embodiment of an injector that may be used in the system of FIG. 1A.

FIG. 5B is a perspective view of an embodiment of an injector and a field engineer identification card that may be used in the system of FIG. 1A.

FIG. 6 is a flowchart of an exemplary method of manufacturing and distributing a syringe or other container as shown in FIGS. 1A and 1B.

FIG. 7 is a flowchart of an exemplary method of stocking and preparing for use of a syringe or other container as shown in FIGS. 1A and 1B.

FIG. 8 is a flowchart of an exemplary method of using a syringe or other container as shown in FIGS. 1A and 1B.

FIG. 9 is a flowchart of an exemplary method of a field maintenance process for a syringe filled with contrast media as shown in FIG. 1A.

FIG. 10 is a schematic drawing illustrating a variation in RF signal strength in coupling a transmitting antenna with a receiving antenna angled with respect to the transmitting antenna.

FIG. 11 is perspective view of a contrast media power injector having an RF data tag on a syringe mounted in a power injector.

FIG. 12 is a perspective view of an exemplary embodiment illustrating a syringe positioned above a faceplate of a contrast media power injector having multiple, nonparallel antenna loops for a read/write device in accordance with the principles of the present invention.

FIGS. 13A-13D are schematic drawings of four different circuit configurations for the multiple, nonparallel antenna loops of FIG. 12.

FIG. 14 is a schematic drawing of the multiple, nonparallel antenna loops of FIG. 11 with switches for connecting the antenna loops in the four different circuit configurations of FIGS. 13A-13D.

FIG. 15 is schematic drawing of a flowchart illustrating a communications cycle utilizing the multiple, nonparallel antenna loops of FIG. 12.

FIG. 16 is a cross-sectional drawing of a pressure jacket for a contrast media power injector as shown in FIG. 11, which is equipped with a multiple loop, nonparallel antenna system for the contrast media power injector similar to that illustrated in FIG. 12.

FIG. 17 is a schematic drawing of an electromagnetic radio frequency R/W device utilizing the multiple loop, nonparallel antenna system of FIG. 16.

FIG. 18 illustrates different manners of applying a tracking device to a radiopharmaceutical container and respective pig in the system shown in FIG. 1.

FIG. 19 is a flowchart of an exemplary method of post-processing a radiopharmaceutical container and associated pig.

FIG. 20 is a perspective view of an exemplary embodiment of an RF tag and antenna system that is applicable to a radiopharmaceutical syringe and associated radiopharmaceutical pig in accordance with the principles of the present invention.

FIG. 21 is a perspective view of another exemplary embodiment of an RF tag and antenna system that is applicable to a radiopharmaceutical syringe and associated radiopharmaceutical pig in accordance with the principles of the present invention.

FIG. 22 is a perspective view of a further exemplary embodiment of an RF tag and antenna system that is applicable to a radiopharmaceutical syringe and associated radiopharmaceutical pig in accordance with the principles of the present invention.

FIG. 22A is an exploded view showing a path of an antenna lead in the further embodiment of the radiopharmaceutical syringe and associated radiopharmaceutical pig shown in FIG. 22.

DETAILED DESCRIPTION OF EXEMPLARY EMBODIMENTS

Referring to FIG. 1A, an exemplary embodiment of a container life cycle 18a relates to medical fluid containers, for example, a syringe 20 suitable for storing contrast media. The syringes 20 may be manufactured at a supplier facility 24 that

is remote from a facility 42 in which a syringe 20 is to be used. Within the supplier facility 24, the syringe 20 is first filled with a contrast media at a filling station 28, and thereafter, labels 30 may be applied to respective syringes 20 at a labeling station 32. The syringes 20 may then be packaged either singularly or as a batch in an appropriate shipping carton 34 at a packaging station and the shipping cartons 34 may be temporarily queued or stored in a shipping/receiving department 38.

Orders for the syringes 20 can be received from various sources, for example, a purchasing office 25 within a health care facility 42, or a doctor's office 27 that may be part of, or independent from, the health care facility 42. Further, the orders may or may not be associated with a particular patient.

Based on the orders, the shipping cartons 34 may enter a distribution channel 40 by which they may be delivered to various facilities 42, for example, hospitals, image service providers, and/or other health care facilities. In the example of FIG. 1A, the facility 42 is a hospital that has a shipping/receiving area 44 for receiving the cartons 34 of prefilled syringes 20. Incidentally, "prefilled" herein describes a container that is designed to be sold and/or delivered to a user with at least some medical fluid already disposed in the container. Often, the cartons 34 are temporarily stored in a room 46 that may or may not be associated with a pharmacy within the hospital 42. As desired, the cartons 34 may be transferred to a preparation room 48 at which the syringes 20 may be unpacked and placed in a warming oven 36 to raise the temperature of the contrast media up to about body temperature (e.g., between about 97° F. and about 100° F.). At appropriate times, one or more syringes 20 may be removed from the warming oven 36, carried to the imaging suite 26a and loaded into a powered fluid injector 50. The injector 50 operates to inject the contrast fluid into an examination subject or patient 52. After use, the spent syringe 20 may be processed for an authorized refilling or disposed of (e.g., in a disposal area 112) in a known manner. For purposes herein, the term "prefilled syringe" means a syringe 20 prefilled with a medical fluid (e.g., contrast media) at a location remote from the preparation room 48 and imaging suite 26a.

As with any substance to be injected into an animal, there are a great many regulated practices as well as unregulated common practices that are desirable to be followed in the filling, distribution, preparation and use of a prefilled syringe. Further, the regulated and common practices may differ depending on the type of contrast media being used. Consequently, it is generally desirable to generate and provide a substantial amount of data relating to the handling of the syringe 20 throughout its life cycle, for example, at substantially every step from its filling to its disposal. Further, it is generally preferred that the data be transferable from one location, for example, the respective filling and labeling stations 28, 32, to another location, for example, the respective preparation and imaging rooms 48, 26a. Today, such data has been known to be recorded and transferred utilizing typed and/or hand-written information located on the syringes 20 and/or cartons 34 as well as typed and/or hand-written records associated therewith. However, during the life of a syringe 20, the data is desired to be utilized in computer systems that may, most often, not be integrated and sometimes, in databases that may not be compatible.

In order to provide a common data acquisition and storage system for each syringe 20, which can be utilized during any portion, and at every stage, of the container life cycle 18a, a system of radio frequency identification device ("RFID") tags and readers is used.

The object of an RFID-based system is to carry data in transponders, generally known as tags, and to retrieve data, by machine-readable means, at a suitable time and place to satisfy a particular application need. Thus, a tag or transponder may typically include an RF driver circuit and associated antenna. The RF driver circuit often utilizes an integrated circuit chip having a programmable processor and associated memory, which are capable of storing the data and performing necessary demodulation and, if applicable, modulation functions. Data within a tag may provide any manner of information relating to a prefilled syringe that is useful over the life of the syringe. It is generally preferred that an RFID system include a means for reading data from, and in some applications, writing data to, the tags, as well as a means for communicating the data to a computer or information management system. Thus, an RFID system preferably has the versatility to permit data to be written into, and read from, a tag at different times and at different locations.

Wireless communication is most often used to transfer data between a tag and a reader. Such communication is often based upon propagating electromagnetic waves, for example, radio frequency waves, by antenna structures present in both tags and readers. It is known to use either a common antenna or different antennas with an RFID tag to read data from, and write data to, the tag; closed loop, open loop, stripline, dipole and/or other antennas may be used. Further, RFID tags may be passive, that is, without an independent power supply, or active, that is, with a power supply such as a battery. In applications described herein, the choice of a particular antenna configuration and whether to use an active or passive RFID tag may or may not be application dependent.

An exemplary embodiment of a syringe manufacturing process implemented at a supplier facility 24 is illustrated in FIG. 6. First, at 502, a syringe 20 is filled with contrast media 22 at a filling station 28. Thereafter, at 504, a label 30 containing human readable and/or machine-readable indicia is applied to the syringe 20 at the labeling station 32. As part of the labeling process, an RFID tag 60 is applied to the syringe 20. The RFID tag 60 incorporates an RFID chip and associated antenna in a known manner, for example, as shown in FIG. 5A by the RFID chip 212 and antenna 210; and the RFID tag 60 may be a part of or separate from the label 30. As shown in FIGS. 2A-2D, the RFID tag can be applied at any suitable location on the syringe 20. For example, as shown in FIG. 2A, the RFID tag 60 can be applied to a rear surface 55 of a syringe flange 56; and as shown in FIG. 2B, the RFID tag 60 can be applied to an outer cylindrical surface 57 of the syringe. In another embodiment shown in FIG. 2C, prior to the syringe 20 being loaded into a power head of an injector, the RFID tag 60 can be peeled off of the syringe 20 and applied to the injector. Upon removing the syringe 20 from the injector power head, the RFID tag may be reapplied to the syringe 20. In a still further embodiment shown in FIG. 2D, the RFID tag 60 can be applied to a rear surface 58 of a plunger 59. The plunger 59 may have a core 61 covered by a molded material 63, and an RFID tag can be applied to or integrated into the plunger structure at various locations 65a, 65b, 65c, etc. As shown in FIG. 2D, an RFID tag may be applied as shown at 60' on the discharge extension (e.g., nozzle) extending from the distal end of the syringe 20, or as shown at 60", an RFID tag can be applied to a front wall (e.g., tapering front wall) of the syringe 20.

Within the supplier facility 24 of FIG. 1A, a read/write ("R/W") device 62 is connected to a labeling computer 64 and, at 506 (FIG. 6), is operative to write data in the RFID tag 60 relating to contrast media or other pharmaceutical and its

associated prefilled syringe or other container **20**. Data that can be written to the RFID tag **60** includes, but is not limited to, the following:

- A unique container identification number.
- A security code that limits access to the RFID tag to those R/W devices that are able to provide the security code.
- A volume of the pharmaceutical filled in the container.
- A total available volume and/or physical dimensions of the available volume in the container.
- An identity, or type, of the pharmaceutical in the container.
- A concentration of the pharmaceutical.
- A formula of the pharmaceutical.
- A manufacturing date.
- An identity of a factory, production line, filling station machine, and/or batch number associated with the container.
- A date and time at which the container is filled.
- An expiration time and/or date and/or a shelf life of the pharmaceutical.
- NDC codes.
- One or more vendor specific inventory codes, for example, an SKU code.
- An identity of the country in which the container was filled.
- An identity of the container and/or container packaging.
- Product promotions and/or coupons and/or Internet links of the supplier.
- Recommended software updates for power injectors in which the container is intended for use.

Thereafter, at **508**, the syringe **20** is loaded into a shipping carton **34**; and, at **510**, the cartons **34** are stocked as inventory in a shipping/receiving department **38**. Based on orders received, as indicated at **512**, the cartons **24** may be further combined or palletized into a case or batch **67** for shipment to a customer; and a label **66** can be optionally applied to an individual shipping carton **34** or a unified case or batch **67** of cartons. The label **66** can include human readable, machine-readable indicia and/or be an RFID tag. Such indicia or RFID tag data may include but is not limited to an identification of the supplier and the product, the product expiration date and the packaging. The packaging code identifies whether the package is a single syringe, a carton of syringes or a case of syringes. In preparing one or a batch of cartons **34** for shipment, an R/W device **68** connected to a shipping computer **70** may be used to read data from, and write data to, the RFID tags **60** on the syringes **20** within the cartons **34**. In addition, if applicable, the R/W device **68** may be used to read data from, and write data to, RFID tags associated with the labels **66**. Thus, the shipping computer **70** is able to identify parameters, for example, type of syringe, type of contrast media, contrast media concentration, etc., and confirm that those parameters meet the specifications of a particular order. Thus, the R/W device **68** can be used to write into either the RFID tags **60** on the syringes **20**, and/or the RFID tags on labels **66**, data including, but not limited to, the following:

- An identity of the customer.
- Purchase invoice and tracking numbers.
- Purchase and/or shipment dates.
- Customer specific marketing data.
- Customer specific software updates for power injectors owned by the customer.

The cartons **34** then enter the distribution channel **40** and are received by a receiving department **44** of an imaging facility such as the hospital **42**. An example of a syringe stocking and preparation process is illustrated in FIG. 7. Upon receiving the cartons **34**, a R/W device **72** connected to a shipping/receiving computer **74** reads, at **602**, the syringe RFID tags **60** and/or the shipping carton RFID tags **66**. As shown in FIG. 3A, the shipping/receiving computer **74** stores

the read data in an inventory database **76**. The shipping/receiving computer **74** is connected via a communications link, for example, an Ethernet LAN, etc. to a hospital administration computer **78** and other computers; and one or more versions of the inventory database **76** can be maintained in any of those computers. Thus, the receiving computer **76**, or another computer, is able to confirm that the delivered syringes conform to hospital purchase orders and, if applicable, automatically authorize payment of invoices therefore. Further, via the shipping/receiving computer **74**, the syringe RFID tags **60** within the cartons **34** can, at **604**, be updated with other data including, but not limited to:

- A time and date that the container was received.
- A hospital SKU code.
- Doctor related information.
- Patient related information.
- An identity of a stock room or other storage area.
- An identity of a particular preparation room and/or imaging suite in which the pharmaceutical is to be used.
- An identity of a particular power injector, which is to be used.

Thereafter, at **606**, cartons are delivered to a room **46**. As seen in FIGS. 3A and 1A, within the room **46**, a R/W device **77** connected to a computer **79** can be used to read the syringe RFID tags **60** and update a database within the computer **79**. Further, or alternatively, as shown in FIG. 3A, the computer **79**, via the communications link **80**, can be used to update the inventory database **76** within administration computer **78**, thereby confirming delivery of the syringes to the room **46** from the shipping/receiving area **44**.

The communications link **80** may be implemented by an Ethernet, USB, RS-232, RS-422, or other interface that uses a standard PC-based communications protocol, for example, BLUETOOTH, parallel, IrDA, ZigBee, 802.11b/g, or other comparable wired or wireless connection.

Subsequently, instructions are provided to move a shipping carton **34** from the room **46** to a preparation room **48**. The R/W device **77** is used to read the RFID tags, at **606**, and find the cartons **34** containing the desired syringes. Further, reading the RFID tags permits an identification of the oldest inventory. (Since contrast media has a shelf life, it may be appropriate to follow a first-in/first-out inventory procedure.) Thereafter, at **608**, an identified shipping carton **34** is delivered to the preparation room **48**.

In the preparation room **48**, the syringes **20** are removed from a carton **34** and placed in the warmer **36** to bring the contrast media up to about body temperature. As shown in FIGS. 1A, 3A and 4, an RAN device **81** is connected to a warmer control **82** having a user interface **86**. The warmer control **82** is electrically connected to an imaging information system **87** that, in turn, is connected to the communications link **80**, and hence, to the other computers in the hospital **42**. Upon placing a syringe in the warmer **36**, the R/W device **81** reads, at **610**, a respective RFID tag **60** and transmits data with respect to the syringe **20** to a work-in-process database **84** in the imaging information system **87** as illustrated in FIG. 3A. Further, or alternatively, the imaging information system **87**, via the communications link **80**, can be used to update the inventory database **76**, thereby allowing other computers to track information written to and read from the syringe RFID tags **60** in the warmer **36**. R/W device **81** may also write to each RFID tag **60** the time and date each respective syringe **20** is placed in the warmer **36**. Further, upon a technologist requesting, via the user interface **86**, a particular contrast media, the warmer control **82** can, via the user interface **86**, identify to the technologist a particular syringe inside the warmer **36**, such as the syringe that has been in the warmer for

the longest period of time. (Not only does contrast media have a limited shelf life, but the time spent in the warmer **36** should also be limited. Thus, inventory in the warmer **36** may also be handled on a first-in/first-out basis.) Upon removing a syringe **20** from the warmer, at **612**, the R/W device **81** writes the removal time and date to a respective RFID tag **60** and reads data identifying the syringe being removed. The work-in-process database **84** and other databases are appropriately updated; and the warmer control **82** via the user interface **86** confirms to the technologist that the correct syringe has been removed.

Referring to FIGS. **1A**, **3A**, **4** and **5A**, one or more syringes **20a**, **20b** are then carried into an imaging suite **26a** and loaded into respectively one or both of the mounts or faceplates **88a**, **88b** that are attachable on a powerhead **90** of a powered fluid injector **50** in a known manner. An exemplary injector is shown and described in U.S. patent application Ser. No. 10/964,003, the entirety of which is hereby incorporated by reference. Although the powerhead **90** discussed herein is a dual head injector, embodiments of the present invention explicitly contemplate single head injectors as well. A suitable single-head injector is shown in U.S. Pat. No. 5,300,031, the entirety of which is hereby incorporated by reference.

In the illustrated application, in which the injector receives multiple syringes, a user-filled syringe having a volume of about 200 ml is mountable in a pressure jacket **250** of faceplate **88a**. Further, a pre-filled syringe having a volume in excess of about 90 ml or more may also be mountable in faceplate **88b**. The injector powerhead **90** includes hand-operated knobs **92a** and **92b** that are operative via an injector control circuit to control motors within respective plunger drives **95a**, **95b**. The plunger drives **95a**, **95b** are operable to move plungers within the respective syringes **20a**, **20b** in a known manner. Exemplary operations of a powerhead **90** and injector control **93** are shown and described in U.S. patent application Ser. No. 10/964,002, the entirety of which is hereby incorporated herein by reference. Additional exemplary operations are described in U.S. Pat. Nos. 5,662,612, 5,681,286 and 6,780,170, the entirety of which are hereby incorporated by reference. As seen in FIG. **3A**, the injector control **93** is electrically connected to the hospital information system **78** via the communications link **80**, and/or may be otherwise electrically connected to the imaging information system **87** by a communications link that uses a technology such as those noted above with reference to the communications link **80**.

The injector powerhead **90** has a user interface **94**, for example, a touch screen, for displaying current status and operating parameters of the injector **50**. Powerhead **90** is often mounted to a wheeled stand **100**, which permits easy positioning of the powerhead **90** in the vicinity of the examination subject **52**. The injector **50** also has a remotely located console **96** with remote user interface **97**, for example, a touch screen, a power supply **98** and other switches and components (not shown). The console **96** may be used by an operator to enter programs and control the operation of the injector **50** from a remote location in a known manner. It will be appreciated that elements of the injector control **93** may be incorporated into the powerhead **90** or may be incorporated in other elements of the injector such as the power supply **98** or console **96**, or may be distributed among these elements.

The faceplate **88b** has an outward extending cradle **99** that supports a heater **106** mounted on a printed circuit ("PC") board **102**. The heater **106** is electrically connected to the injector control via a cable or connector and is operable by the injector control **93** to heat the syringe **20b** in a known manner. The PC board **102** further supports a R/W device **104b** and an

associated antenna system **229b**. The R/W device **104b** is also electrically connected to the injector control **93** and console **96**. Further, the R/W device **104b** may be activated by the injector control **93** to read data from an RFID tag **60b** on a respective syringe **20b**. Data may be written to, and/or read from, the RFID tag **60b** at any specified time when a syringe **20b** is in proximity of a respective faceplate **88**. Thus, the system has the ability to determine when syringes **20a**, **20b** are mounted in the respective faceplates **88a**, **88b**. The data may be encrypted, and the data and data transfer may comply with 21 CFR 11, JCAHO, and HIPAA requirements.

One example of a process for utilizing the syringe **20b** within the imaging suite **26a** is shown in FIG. **8**. This example is described principally with respect to the syringe **20b** loaded in faceplate **88b**; however the description is equally applicable to the syringe **20a** loaded in faceplate **88a**. The description is further applicable to an injection process in which media is dispensed from both syringes **20a**, **20b**, either sequentially or simultaneously. Simultaneous dispensing from both syringes may be done at controlled and selected flow rates to achieve any desired concentration of the resulting mixture of media and/or media and saline in the two syringes.

Referring to the process of FIG. **8**, first, at **702**, the R/W device **104b** is activated to read data stored in the RFID tag **60b** relating to contrast media or other pharmaceutical and its associated prefilled syringe or other container **20b**. As shown at **704**, that information includes, but is not limited to:

A container identification and/or serial number that is checked against a database of previously used containers to block, if appropriate, a potential reuse of the container.

A container security code, which may be matched with the security code of the injector being used.

Information relating to container volume and volume delivery to assist the technologist in setting up the injector.

Container volume and/or dimension information in order to provide a more precise real time dispensing control of volume.

Pharmaceutical type and concentration data to confirm it is correct for a selected protocol.

ID, batch and lot numbers that can be used to test the container and/or pharmaceutical against recall data.

Shelf life data and fill date, which is compared to a current date to determine whether a recommended shelf life has been exceeded.

The R/W device **104b** also writes the current time and date to the RFID device **60b** to permit tracking of open-to-atmosphere time for the syringe **20b**, which is also limited. During the contrast media injection process, the displacement of the syringe plunger is precisely controlled in accordance with data read from the RFID tag **60b** relating to available syringe volume and/or dimensions thereof. Further, plunger feed is tracked, so that the contrast media remaining in the syringe can be continuously determined.

The faceplates **88a**, **88b** have a bidirectional communications link with the injector control **93**, which may be used to transfer any of the above information between the syringes **20a**, **20b** and the injector control **93**. Thus, the injector control **93** may have syringe and drug information that may facilitate a procedure setup and result in reduced time and error. In addition, the injector control **93** may read or write other information to and from the faceplates **88a**, **88b**, which is not directly pertinent to syringe information. Examples of this may include, but are not limited to:

Enabling or disabling of the faceplate electronics.

Heating of the faceplate for contrast media warming.

21

In step 706 of FIG. 8, the media is used in connection with a procedure. As seen in FIG. 4, before, during and after injection of the contrast media, a technologist operates a CT scanner control 101 that is effective to cause a CT scanner 103 to scan a patient 105 shown in phantom. The injector control 93 may have one or more interfaces to a CAN communications bus 111, which is a known interface for the CT scanner control 101. The protocol is defined by the scanner manufacturers. Data and data transfer between the injector and scanner comply with 21 CFR 11, JCAHO, and HIPAA requirements.

Returning to FIG. 8, as shown at 706, data transfer between the injector control 93 and CT scanner control 101 may be bidirectional and may relate to the contrast media or other pharmaceutical and its associated prefilled syringe or other container 20b. Such data includes, but is not limited to, the following:

- Pharmaceutical brand name, concentration, lot number.
- Pharmaceutical expiration date, volume.
- Injected volume, flow rate (achieved, target).
- Injection time.
- Patient name, weight, age, ID number, for example, SS no., hospital ID, etc.
- Injector serial number, firmware version.
- Procedure number and/or name.
- Technologist name and/or identification number.
- Hospital name and/or identification number.
- Used or unused status of container.
- CT scanner setup and procedure information.
- CT scanner ID and/or serial no.
- CT images.
- Hospital information system data.
- Injector functional control.
- CT scanner functional control.

Upon the injector control 93 determining that the desired volume of contrast media has been delivered, the injection process is stopped. At the end of the injection process, as shown in FIG. 8 at 708, the injector control 93 is operative to determine an exact volume of contrast media injected; and the injector control writes to the RFID tag 60b and/or updates the imaging information system 87 with data and information that includes, but is not limited to the following:

- Time and date that the injection process was finished.
- Injected volume, flow rate (achieved, target).
- Volume of pharmaceutical remaining in the container.
- Injection time.
- Patient name, weight, age, ID number, for example, SS no., hospital ID, etc.
- Injector serial number, firmware version.
- Procedure number and/or name.
- Technologist name and/or identification number.
- Hospital name and/or identification number.
- Used or unused status of syringe.
- CT Scanner Information.

As illustrated in FIG. 4, the injector control 93 has an interface providing a communications link 107 to a hard-copy printer 109. The printer 109 may be, but is not limited to, a thermal, ink-jet, or laser based printer. The printer 109 may be used to print pages and/or labels of various sizes and colors at specified times upon requests of a user, the CT scanner control 101, the hospital information system 78, or the injector control 93. The labels may be made part of patient records, requisition sheets, or other forms. Data output and data transfer may comply with 21 CFR 11, JCAHO, and HIPAA requirements.

Returning to FIG. 8, as shown at 710, a label or page may be printed to provide information relating to the contrast

22

media or other pharmaceutical, its associated prefilled syringe or other container 20b, and the use thereof. Such information includes, but is not limited to, the following:

- Pharmaceutical brand name, concentration, lot number.
- Pharmaceutical expiration date, volume.
- Injected volume, pressure, flow rate (achieved, target).
- Injection time.
- Patient name, weight, age, ID number, for example, SS no., hospital ID, etc.
- Injector serial number, firmware version.
- Procedure number and/or name.
- Technologist name and/or identification number.
- Hospital name and/or identification number.
- Used or unused status of syringe.
- Graphs or charts, for example, pressure, flow rate, etc.
- CT scanner information.
- CT scan information.
- Open (white) space or blanks for tech initials, drawings, etc.

Thus, any of the above information can be exchanged between the injector control 93 and hospital information system 78. Potential uses for this capability include but are not limited to:

- Electronic inclusion of volume of contrast media injected and other procedure information in patient record.
- Electronic re-ordering of supplies.
- Automated billing.
- Automated scheduling.

After the injection process, the injector control 93 can write to the RFID tag 60b to set a syringe-used flag that will help to prevent a reuse of the syringe 20b. The syringe 20b is then removed from the faceplate 88b; and if the procedure was aborted and the syringe was not used, it can be placed back into the warmer 36. In that process, information is read from, and written to, the RFID tag 60b as previously described. Further, the image information system 87 is also able to track the open-to-atmosphere time of the syringe and warn the technologists when an open-to-atmosphere time is exceeded.

If the syringe 20b removed from the faceplate 88b is empty, the syringe is typically transported to a disposal area 112 (FIGS. 1A, 3A and 4); and prior to disposal, another R/W device 114 connected to one of the other computers 75 reads the RFID tag 60b. The inventory database 76 can thus track the identity of the syringe 20 being destroyed. Further, the syringe disposal information can be communicated to a supplier computer 116 via a communications link 118 as seen in FIG. 3A, for example, via the Internet 83, a telephonic connection, or other comparable wired or wireless connection.

In an alternative embodiment, empty syringes, instead of being destroyed, are returned to the supplier 24 for further processing, for example, disposal or refilling. In the latter example, the syringes 20 pass through the hospital shipping/receiving area 44 and the RFID tags are again read to identify the syringes leaving the hospital; and the inventory database 76 is updated accordingly. Upon entering the supplier shipping/receiving area 38, the RFID tags 60b are again read to update a supplier inventory database 120 tracking syringes within the supplier's facilities. The RFID tags 60b on the syringes 20 are updated or replaced depending on whether the syringe is destroyed or reconditioned and refilled by the supplier.

In the system shown and described herein, the injector control 93 facilitates information collection and transfer throughout a CT procedure. The RFID-enabled syringes provide quicker and more accurate data recording, as well as an automated transfer of drug information. The printer allows for a hard copy of selected information to be incorporated into

the patient or hospital record. The CT interface via CAN, facilitates information flow and collection at a single point, either the CT scanner system or the injector. The hospital information system interface improves this information flow a step further, potentially creating an all-electronic system with minimal user intervention; this provides the opportunity for reduced error and efficiency in the CT scanning suite.

With respect to another exemplary embodiment, on occasion, field engineers make service calls to a power injector, e.g. for routine maintenance or to diagnose failed operation. During such service calls, the field engineer is able to operate the injector in a "service" mode without having to install electrical jumpers in the injector control. Instead, referring to FIG. 5B, the service mode function is initiated by a field engineer using an intelligent identification ("ID") card **122**. Such an ID card **122** has an RFID tag **124** that incorporates an RFID chip and associated antenna in a known manner.

An exemplary process for using the ID card **122** for injector maintenance is shown in FIG. 9. As indicated at **802**, the RFID tag **124** is loaded at the supplier facility **24** with data including, but not limited to, the following:

- An identification of the field engineer.
- Latest updates and software information.
- Specific software revisions.

To initiate service of a power injector, the field engineer places the ID card **122** on an empty faceplate **88b**, thereby allowing the R/W device **104b** to read and write to the RFID tag **124**. As indicated at **804** of FIG. 9, upon reading an appropriate identification and security code from the RFID tag **124**, a field engineer identification and service time and date are stored in the injector control **93**. Thereafter, the injector user interfaces **94, 97** (see FIG. 5A) are effective to switch the injector **50** into a service mode, thereby disabling several operational checks and features that are used in a normal injection cycle but which inhibit operating the injector **50** for service purposes. The R/W device **104** continues to periodically read the identification and security codes from the RFID tag **124**. Upon failure to successfully read the RFID tag **124**, for example, because the ID card **122** has been removed from the faceplate **88b**, the injector control **93** automatically switches the injector **50** out of the service mode. Thus, the previously disabled operational checks and features are re-enabled, and the injector is ready to operate in a normal injection cycle. Further, at **804**, the injector control **93** is operative to read from the RFID tag **124** information and data relating to factory updates to the injector components and software.

In the process of servicing the injector **50**, as indicated at **806**, the field engineer initiates uploads of software upgrades from the RFID tag **124** to the injector control **93**. In addition, mechanical components are serviced, mechanical upgrades are installed and their operation is verified. As a final step of the service operation as indicated at **808**, the injector control **93** writes to the RFID tag **124** on the ID card **122** data including, but not limited to, the following:

- The latest software revision installed.
- A confirmation that mechanical and software upgrades have been installed.
- The date of service and serial number of the injector.
- Protocol, statistics or details relating to the injector operation since the last service.

Upon the field engineer returning to the supplier facility **24**, the RFID tag **124** is read; and the service information is stored in a history file associated with the particular injector that was serviced.

The use of an RF communications system between an RFID tag **60** on a container **20** and a power injector control **93**

provides for further exemplary embodiments of the RF communications system. Known RFID systems use electromagnetic (EM) fields to communicate between an R/W device that includes a tuned antenna and one or more RFID tags or transponders. In one exemplary embodiment, the R/W device sends out data using EM fields at a specific frequency; and with passive RFID tags, this EM energy powers the tag, which in turn enables processing of this received data. Following receipt of the data, the RFID tag may transmit data that is received and processed by the R/W device.

An RFID is difficult to implement around metallic or diamagnetic materials, for example, water, saline or a medical fluid in a container such as a contrast media in a syringe. These materials absorb and/or reflect RF energy, making successful read-write RFID operations difficult, especially with the low power regulations for RF frequencies. In addition, the angle between a plane of the RFID tag antenna and a plane of the R/W device antenna is critical. For optimum performance, the plane of the RFID tag antenna should be substantially parallel to the plane of the R/W device antenna. As shown in FIG. 10, for single plane antennas, as an acute angle **200** between an RFID tag antenna plane **202** and an R/W device antenna plane **204** increases, a signal strength coupling the antennas in the two planes **200, 204** decreases. In other words, as the angle **200** increases, the RF signal strength transferable from the R/W device antenna to the RFID tag antenna decreases. Similarly, the signal strength transferable from the RFID tag antenna back to the R/W device antenna also diminishes. Further, that signal strength is substantially equal to the output signal strength of the R/W device antenna minus any attenuation from metallic and diamagnetic materials divided by the cosine of the angle **200**.

Referring back to FIG. 5A, orientation of the syringe **20b** places the RFID tag antenna **210** relatively close to the R/W device **104b**; and therefore, coupling RF signals therebetween to facilitate reading data from, and/or writing data to, the RFID tag **60b**. However, with the syringe **20b** oriented as shown in FIG. 11, contrast media in the syringe **20b** is between the RFID tag antenna **210** and the R/W device **104b**. The contrast media attenuates the RF field strength from the antenna of the R/W device **104b** and interferes with its RF coupling with the RFID tag antenna **210**.

In one exemplary embodiment of the invention, referring to FIG. 12, a syringe **20b** having a label **30b** with an antenna **210** and RF driver **212** is positioned above faceplate **88b**, ready to be loaded therein. A first PC board **102** and a second PC board **103** are mounted in faceplate **88b**, so as to be nonparallel. The PC boards **102, 103** form sides of a V-shape and thus, form an angle of less than 180 degrees therebetween. PC board **102** supports a first antenna loop **220** and its associated tuning circuit **226**, and PC board **103** supports a second antenna loop **222** and its associated tuning circuit **228**. The first and second antenna loops **220, 222** and respective tuning circuits **226, 228** are connected to an R/W RF driver circuit **224b** through a switching circuit **241b** to collectively form the electromagnetic R/W device **104b**. In an alternative embodiment, the R/W RF driver circuit **224b** and switching circuit **241b** may be mounted on a separate PC board **102b** (shown in phantom), which is located beneath, and electrically connected to, the PC board **102**. In other embodiments, the R/W RF driver circuit **224b** and/or the switching circuit **241b** may be mounted in the power head **90** in association with the injector control **93**.

Further, as shown in FIGS. 13A-13D, an antenna system **229b** comprising the antenna loops **220, 222**, respective tuning circuits **226, 228** and switching circuit **241b** is connect-

able in different electrical configurations to achieve an optimum RF coupling between the R/W device **104b** and the RFID tag **60b**.

Referring to FIG. **13A**, power from the R/W RF driver circuit **224b** is applied to the input **230** of a tuning circuit **226** that is connected to a signal lead **231** of the primary antenna loop **220** on PC board **102**. Further, input **234** of the tuning circuit **228**, which is connected to a signal lead **235** of the secondary antenna loop **222** on PC board **103**, is left open or floating. A primary antenna loop ground lead **232** is connected to ground with the secondary antenna loop ground lead **236**. In this configuration, the powered primary antenna loop **220** on PC board **102** is tuned to a frequency indicated by a protocol of the RFID tag **60b**, for example, about 13.56 Megahertz, which permits propagation of the RF signal into the surrounding area. An RF signal from the primary antenna loop **220** is coupled with the secondary antenna loop **222** on PC board **103**, because the secondary antenna loop **222** is also tuned to resonate at about 13.56 Megahertz.

The angled, V-shape orientation of the PC boards **102**, **103** and respective areas of antenna loops **220**, **222** provide an expanded or increased total antenna area for the R/W device **104b**. Thus, with the antenna configuration of FIG. **13A**, as shown in FIG. **12**, an effective antenna area extends circumferentially around a substantially greater area of a syringe **20b** than is possible with the single PC board **102** shown in FIG. **5A**. Further, the antenna power provided by the RF driver circuit **224b** is also spread over a larger area represented by the combined areas of antenna loops **220**, **222**. Upon the syringe **20b** being loaded onto the faceplate **88b**, with some orientations of the syringe **20b**, the larger antenna area shown in FIG. **13A** improves the RF coupling with the antenna **210** of the RFID tag **60b**.

As shown in FIG. **13B**, antenna loop **222** on PC board **103** can be made the primary loop by disconnecting or opening an input **230** of the tuning circuit **226** and connecting the tuning circuit input **234** of the antenna loop **222** to the power output of the R/W RF driver circuit **224b**. First antenna loop ground lead **232** and second antenna loop ground lead **236** continue to be connected to ground. Again, both antenna loops **220**, **222** are tuned to resonate at the RFID tag frequency, that is, about 13.56 Megahertz. The antenna configuration of FIG. **13B** may provide better RF coupling with the antenna **210** of the RFID tag **60b** depending on the orientation of the syringe **20b** and thus, the circumferential location of the RFID tag **60b**.

Another configuration of the antenna loops **220**, **222** is shown in FIG. **13C** wherein the tuning circuit input **230** of the first antenna loop **220** is connected to the power output of the R/W RF driver circuit **224b**; and first antenna loop ground lead **232** is connected to ground. The tuning circuit input **234** and ground lead **236** of antenna loop **222** are connected to ground, which prevents the second antenna loop **222** from resonating at the RFID tag frequency, which, in this application, is 13.56 MHz. This effectively reduces the area of the antenna system **229b** to the area of the primary antenna loop **220**, and all of the power from the R/W RF driver circuit **224b** is applied across the area of the primary antenna loop **220**, which is tuned to resonate at the RFID tag frequency, that is, about 13.56 Megahertz. Upon the syringe **20b** being loaded onto the faceplate **88b**, depending on the orientation of the syringe **20b** and the RFID tag antenna **210**, the smaller antenna area of the circuit in FIG. **13C** may improve the RF coupling with the antenna **210** of the RFID tag **60b**.

Referring to FIG. **13D**, alternatively to FIG. **13C**, the tuning circuit input **234** of the second antenna loop **222** on PC board **103** is connected to the power output of the R/W RF driver circuit **224b**; and tuning circuit input **230** of the first

antenna loop **220** is connected to ground along with antenna loop ground leads **232** and **236**. Thus, the first antenna loop **220** does not resonate at the RFID tag frequency of 13.56 MHz; and only the second antenna loop **222** is tuned to resonate at that frequency. With some orientations of the syringe **20b**, this antenna configuration provides the best RF coupling with the antenna **210** of the RFID tag **60b**.

In some applications, a user may be instructed to load the syringe **20b** in the faceplate **88b** so that the label **30b** is always in the same orientation. Or, in other applications, the RFID tag **60b** may be removable from the syringe and mountable at a fixed location on the injector **50**. In those applications, an R/W antenna can be designed and placed in a fixed location to have optimum RF coupling with an RFID tag. However, in still further applications, a user may have no limitations on where the RFID tag **60b** is located on the syringe **20b** or how the RFID tag **60b** is oriented when the syringe **20b** is mounted on a faceplate **88b**. In those applications, the RFID tag **60b** may have any circumferential location around a barrel of the syringe **20b** or within the faceplate **88b**. Further, in such applications, it is difficult to precisely predict which of the antenna configurations in FIGS. **13A-13D** will provide the best RF coupling with an RFID tag having an unknown orientation with respect to R/W device **104b**. This is due, in part, to the complex and somewhat unpredictable EM fields formed around materials that reflect and/or absorb such fields. Therefore, in another exemplary embodiment of the invention, all of the antenna configurations of FIGS. **13A-13B** may be utilized.

Referring to FIG. **14**, switches **238**, **240** on PC board **102** comprise the switching circuit **241b**, which is used to selectively connect respective tuning circuit inputs **230**, **234** to either a power output or terminal **242** from R/W RF driver circuit **224b**, a ground terminal **244** or an open state represented by contacts **246**. The ground leads **232**, **236** of respective antenna loops **220**, **222** are always connected to the ground **244**. The contacts of switches **238**, **240** have notations to FIGS. **13A-13D** indicating the switch states corresponding to the antenna configurations of FIGS. **13A-13D**.

In use, referring to FIGS. **12** and **15**, a communications cycle is initiated either automatically by the injector control **93** detecting a syringe **20b** being loaded into the faceplate **88b** (such as by the movement of a mounting arm of the faceplate **88b**, causing a magnet in the mounting arm to move into confronting relationship with a magnetic sensor in the injector), or manually by an operator providing an input to the injector control **93**. In either event, the injector control, at **900**, operates the switches **238**, **240** to connect the antenna loops **220**, **222** in a first of the four circuit configurations, for example, the circuit configuration shown in FIG. **13A**. Thereafter, the injector control **93** initiates, at **902**, a communications protocol between the R/W RF driver circuit **224b** and the RF driver circuit **212** of the RFID tag **60b**. Initiating a communications protocol is a known process by which the R/W RF driver circuit **224b** causes the R/W antenna system **229b** to emit an electromagnetic signal in order to establish a reliable RF coupling with the tag antenna **210** and thus, establish an RF communications with the RFID tag **60b**. Upon establishing an RF communications, the R/W device **104b** can read data from and/or write data to the RFID tag **60b**.

If, at **904**, the injector control **93** determines that the communications protocol and hence, the RF communications link, has been established, the injector control **93** commands, at **906**, the R/W drive **104b** to proceed with the reading of data from, and/or the writing of data to, the RFID tag **60b**. However, if, at **904**, the injector control **93** determines that the communications protocol failed, and a successful RF com-

communications between the R/W device **104b** and the RFID tag **60b** is not made, the injector control **93** determines, at **908**, whether all antenna loop configurations have been tried. If not, the injector control **93** operates, at **910**, the switches **238**, **240** to connect the antenna loops **220**, **222** into another one of the four circuit configurations shown in FIGS. **13A-13B**. Thereafter, the injector control **93** automatically iterates through the process steps **902-908** to reconnect the antenna loops **220-222** in different circuit configurations in an attempt to establish a successful RF communications protocol or link. If, at **908**, the injector control **93** has tried all of the antenna loop configurations without success, it sets, at **912**, a protocol failure flag or error message.

FIGS. **11-14** illustrate different embodiments of an antenna system **229b** that may be employed with an electromagnetic R/W device **104b** to read a data tag **60b** applied to a syringe **20b** mounted in an open faceplate **88b**. In a further embodiment, referring to FIG. **5A**, a syringe **20a**, that often is a user-filled disposable syringe, is mounted within a translucent or transparent pressure jacket **250** of faceplate **88a**. The syringe **20a** is secured in the pressure jacket **250** by a cap **252** in a known manner. A data tag **60a** is integrated into a label **30a** applied to the syringe **20a**, and the structure and operation of data tag **60a** is substantially identical to the data tag **60b** previously described. When utilizing the pressure jacket **250** of faceplate **88a**, it is desirable that the data tag **60a** be readable regardless of its orientation inside the pressure jacket **250**.

Referring to FIGS. **5A** and **16**, in a further exemplary embodiment of an RFID communications system, to enhance readability of a data tag **60a**, the pressure jacket **250** may be equipped with an antenna system **229a**, which includes of an array of antenna loops **254**, **256**, **258** spaced about a circumference of the syringe **20a**. While equal spacing of the antenna loops is shown, other spacing may be used. The pressure jacket **250** has inner and outer cylindrical sleeves **260**, **262**, respectively. As illustrated, the antenna loops **254**, **256**, **258** may be molded between the inner and outer sleeves **260**, **262**. Referring to FIG. **17**, the antenna loops **254**, **256**, **258** have respective tuning circuits **264**, **266**, **268**, which may be molded between the inner and outer cylindrical sleeves **260**, **262**. Tuning circuit input leads **270**, **272**, **274** and a ground lead **276** may be bundled into a cable **278** that extends from the face plate **88a** to a switching circuit **241a** located in the power head **90**. The switching circuit **241a** may operate in any appropriate manner, such as in a manner like that previously described with respect to the switching circuit **241b** of FIG. **14**. The switching circuit **241a** may be controlled by an R/W driver circuit **224a** that may be located in the power head **90**. To exchange data with the data tag **60a**, the R/W driver circuit **224a** may execute a communications cycle utilizing the antenna loops **254**, **256**, **258** in a manner similar to that described with respect to FIG. **15**. Thus, in initiating communications with the data tag **60a**, the R/W RF driver circuit **224a** may connect the antenna loops **254**, **256**, **258** in different circuit configurations in order to find a circuit configuration providing the most reliable communications with the data tag **60a**. By using more than two antenna loops, less power may be required to initiate a communications cycle with the data tag **60a**. In additional exemplary embodiments, while the antenna system **229a** is shown as including three antenna loops, other embodiments may include other appropriate quantities and/or arrangements of antenna loops. Further, while the antenna system **229a** is shown as a component of the pressure jacket **250**, other embodiments may include an antenna system having a plurality of antenna loops that is not associated with a pressure jacket.

In its various embodiments, the antenna systems **229a**, **229b** may advantageously incorporate one or more antenna loops that can be powered individually, or mutually coupled together, to produce several tuned antenna and EM field configurations. In some environments, the antenna systems **229a**, **229b** may be characterized as providing an effective low power system for reading data from and/or writing data to a data tag that may be disposed at any location on a contrast media syringe. Moreover, that contrast media syringe may exhibit virtually any orientation relative to a faceplate of a power injector **50** with which it may be associated. Thus, the antenna systems **229a**, **229b** may positively address various challenges relating to use of an RF communications system around metallic or diamagnetic materials, e.g., water, saline, contrast media, or other fluids, and/or in a regulated environment that may mandate use of a relatively low power RF signal.

The exemplary embodiments described with respect to FIG. **1A** relate generally to a life cycle of a container **20** such as a syringe filled with a pharmaceutical such as a contrast media. However, referring to FIG. **1B**, a container life cycle **18b** may relate to other types of containers **20c** that are used to store radiopharmaceuticals. While much of the container life cycle **18b** of FIG. **1B** is generally similar to container life cycle **18a** of FIG. **1A**, radiopharmaceuticals require different handling and storage. The container **20c** is schematically shown as a syringe, but the container **20c** may be a vial or other container suitable for use with a radiopharmaceutical. Within the supplier facility **24**, after the container **20c** is filled with a radiopharmaceutical at a drawing-up or filling station **28**, a quality control check of the radiopharmaceutical may be performed at quality control station **31**. Thereafter, the container **20c** is placed or loaded into a pig **33**, which generally includes lead and/or other radiation shielding material to protect handlers from exposure to radiation from the radiopharmaceutical.

In a manner similar to that described with respect to container **20** of FIG. **1A**, as shown in FIG. **1B**, the loaded pig **33** may then be packaged either singularly or as a batch in an appropriate shipping carton **34** and shipped to a customer or user. Often, the cartons **34** are stored in a nuclear medicine department **29** within the hospital **42**, which generally includes a radiopharmacy **48** and treatment room **26b**. As required, a radiopharmaceutical container may be removed from a pig and placed in a calibration tool **49** to calibrate an activity level of the radiopharmaceutical to a desired level prior to its use. The radiopharmaceutical container may then be placed back into the pig; and at an appropriate time, the pig may be carried to a treatment room **26b**. The radiopharmaceutical container may again be removed from the pig, and the radiopharmaceutical may be injected into a patient **52** either manually or using a powered injector such as that shown and described herein. In various embodiments, different manual or powered injectors may utilize various principles of the invention, and are thus, included within the scope of this disclosure.

After use, the radiopharmaceutical container may be placed in the pig and returned to the supplier facility **24**; and at a post processing station **51**, the radiopharmaceutical container may be disposed of and the pig may be cleaned for reuse.

An exemplary embodiment of a radiopharmaceutical container draw-up and packaging process implemented at a supplier facility **24** is illustrated in FIG. **6**. A radiopharmaceutical container **20c** is filled, at **502**, with a radiopharmaceutical at a draw-up station **28**. Thereafter, at **504**, a label **30** and/or RFID tag **60** are applied to the radiopharmaceutical container **20c** at

the labeling station 32. The RFID tag 60 can be integrated with, or separate from, the label, and the RFID tag 60 incorporates an RFID chip and associated antenna in a known manner.

As shown in FIG. 18, the RFID tag 60 can be applied at any suitable location on a radiopharmaceutical container. For example, the RFID tag 60 can be part of a label 30 that is applied to a radiopharmaceutical syringe 20d or a radiopharmaceutical vial 20e. In the example of the radiopharmaceutical syringe 20d, an RFID tag can be applied to, or integrated into, the syringe structure at different locations as previously described with respect to FIGS. 2A-2D. In a further embodiment, the syringe label 30 may be removable; and immediately prior to the syringe 20d being loaded into a power injector, a portion of the label 30 including the RFID tag can be peeled off and applied to the injector or an associated reader. Upon removing the radiopharmaceutical syringe 20d from the injector, the RFID tag 30 is reapplied to the radiopharmaceutical container 20d. An identical or different label 30 can also or alternatively be applied to a radiopharmaceutical syringe pig 33a or a radiopharmaceutical vial pig 33b. Further, a label 30 with an RFID tag 60 can be applied to a carton 34, for example, a satchel, designed to transport a plurality of pigs.

Within the supplier facility 24 of FIG. 1B, a read/write (“R/W”) device 62 is connected to a label computer 64 and, at 506 (FIG. 6), is operative to read data from and/or write data to the RFID tag 60 for a particular radiopharmaceutical container 20c. As shown in FIG. 3B, the draw-up station 28 may include a draw-up station computer 41 in electrical communications with an R/W device 43; and depending on the application, either or both of the R/W devices 43, 62 can be used to write data to the RFID tag 60, which data includes but is not limited to the data previously described with respect to step 506. With a radiopharmaceutical, the data may also include all of the dose and prescription information that is currently being printed on a prescription label and/or encoded into a bar code, measured radioactivity levels, for example, Tc-99 and Mo-99, and time when measured, an identity of radioactive elements used, for example, Tc-99 and Mo-99, their respective sources, and other suitable data.

Returning to FIG. 6, processes shown in phantom at 507 and 509 are performed that are unique to the radiopharmaceutical containers 20c. First, at 507, quality control checks may be performed (e.g., at a quality control station 31) to determine, for example, a purity of the radiopharmaceutical, the correctness of information on the label, dosage information, etc. As shown in FIG. 3B, the quality control station 31 may include a quality control computer 45 and an associated R/W device 47 that may be used to read data from and/or write data to the RFID tag 60 depending on the quality control checks performed and/or other system specifications.

The container 20c may then, at 509, be inserted into a pig 33 for handling, storage and transportation. A label 65 can optionally be applied to the pig 33. The label 65 can include human readable indicia, machine readable indicia and/or an RFID tag as described with respect to the label 30. As part of the process of inserting the container 20c into the pig, either the R/W device 62 or another R/W device can be used to read data from and/or write data to the RFID tag 65. Data that can be written to the RFID tag 65 may include data written to the RFID tag 60 on the container 20c as well as data that includes, but is not limited to, the following:

A unique identification number for the pig.

An identity of a factory, production line, and/or batch number associated with the pig.

A date and time at which the container was inserted into the pig.

Any other data associated with the order, the radiopharmaceutical, its container 20c and associated pig 33.

At 508 in FIG. 6 (in a manner similar to that previously described with respect to FIG. 1A), one or more pigs 33 may be loaded into a shipping carton 34 (see FIG. 1B). At 510, the cartons 34 may be stocked as inventory in a shipping/receiving department 38. Based on orders received, as indicated at 512, the cartons 24 may be further combined or palletized into a case or batch 67 for shipment to a customer; and a label 66 can be optionally applied to an individual shipping carton 34 or a unified case or batch 67 of cartons.

Referring to FIGS. 1B and 7, the cartons 34 may then enter the distribution channel 40 and may be received by a receiving department 44 of a treatment facility such as the hospital 42. A stocking and preparation process may be executed in process steps 602 and 604, which are similar to those previously described. Also in step 606, cartons may be delivered to a hospital radiopharmacy 48 (or nuclear medicine department of a healthcare facility or other appropriate location), and within the radiopharmacy 48, an R/W device 77 connected to a computer 79 can be used to read data from and/or write data to the pig RFID tags 65. As shown in FIG. 3B, the computer 79, via the communications link 80, can also be used to update the medicine tracking database 76 within the hospital administration computer 78.

Processes unique to radiopharmaceutical containers are shown in phantom at 607 and 609 in FIG. 7. Specifically, within the radiopharmacy 48, a calibration tool 49 is often used, at 607, to check or validate a radioactivity level of the dosage of the radiopharmaceutical within a container. This check/validation can be performed using any appropriate process and/or calibration tool. As shown in FIG. 3B, the calibration tool 49 may have a calibration computer 85 connected to an R/W device 89 that, during the check/validation process, can be used to read data from and/or write check/validation data to the container RFID tags 30 and/or the pig RFID tags 65. This check/validation data may include but is not limited to

A check/validation time and date.

The decay factor or half life of the radiopharmaceutical.

The prescribed activity level (curie level of radiation) at injection time.

The activity level at another time, for example, the draw-up time.

A measured radioactivity level.

A desired radioactivity level at time of treatment.

An identity of the radioactive element injected.

An identity of the calibration tool and operator, etc.

Continuing in FIG. 7, at the appropriate time, at 609, a pig 33 may be delivered to a treatment room for use. The radiopharmaceutical can be administered manually or using a power injector. In most, but not all cases, a syringe 20d or vial 20e containing the radiopharmaceutical is removed from a respective pig 33 for manual administration; but in other applications, a power injector and process as previously shown and described with respect to FIG. 8 may be used. With a radiopharmaceutical, the R/W device 104 associated with the injector control 93 (see FIG. 3B) may write the current time and date to the RFID tag 60 to permit tracking of out-of-pig time (e.g., the duration of time that a syringe or vial is not housed within the pig), if desired. During the radiopharmaceutical injection process, the displacement of the radiopharmaceutical container plunger may be precisely controlled, and plunger feed may be tracked (e.g., recorded and written to a tag associated with syringe and/or pig).

It should be noted that labeling systems described herein have potential for eliminating a need for the calibration tool **49**. For example, the R/W device **104** of FIG. **3B** can read a radioactivity level and time and date of measurement written into the RFID tag by the quality control station **31** (FIG. **1B**). Injector control **93** can then calculate the time elapsed between the measured radioactivity level and the scheduled treatment time and date. The injector control **93** can further calculate the decay in radioactivity level over the elapsed time; and then, being programmed with the prescribed radiopharmaceutical dose, the injector control can calculate the correct unit dose volume to be injected. Thus, a calibration tool **49** may not be required. If the radiopharmaceutical is to be injected manually, the computer **79** and associated R/W device **77** can be used by a clinician or other appropriate personnel in a similar fashion to provide a display of the computed current unit dosage without using a calibration tool.

After the injection process, referring to FIGS. **1B**, **5A** and **19**, the radiopharmaceutical container **20c** may be removed from the faceplate **88b** and placed back into a respective pig **33** as indicated at **802** in FIG. **19**. The pig **33** may then be placed in the same or a different carton and, at **804**, returned to the shipping department **44** and, at **806**, returned to the supplier facility **24**. As shown in **807**, the label associated with the radiopharmaceutical container may be read just prior to disposal to assist in determining how long the container will have to be stored in a radiation-shielding disposal and/or storage container before substantially all of its radioactivity has decayed. For instance, the initial radioactivity of the radiopharmaceutical may be written to the tag at the time of filling the container. Subsequent to that initial fill time, the radioactivity of that radiopharmaceutical decays. Since the rate of decay is generally known, one may utilize the rate of decay and the duration of time that has passed from the initial fill time to determine how much storage time may be needed to sufficiently ensure that the spent container no longer has a significant amount of radioactivity associated therewith. This calculation of storage time may be accomplished manually and/or electronically (e.g., using an appropriate computer interconnected with the reader utilized to read the tag just prior to disposal).

At post processing station **51** within the supplier facility **24** (FIG. **1B**), at **808**, the used radiopharmaceutical container may undergo suitable processing for disposal and, at **810**, the associated pig may be cleaned for reuse. During post processing, any of the computers previously described can be used to read data from and/or write data to the RFID tags on the container **20c**, pig **33**, carton **34** and/or pallet **67**. Such activity may be application dependent to fulfill the needs of a particular supplier, customer, doctor and/or hospital. As shown in FIG. **3B**, a post processing computer **53** may be connected to an R/W device **55** that can be used to read data from and/or write data to the RFID tags **60** on one or both the radiopharmaceutical container or the pig. The post processing computer **53** may be able (via a communications link **57**) to update a supplier inventory database **120** tracking radiopharmaceutical containers and pigs within the supplier's facilities. The RFID tags **60** on the radiopharmaceutical pigs **33** may be updated or replaced. Further, if desired, data relating to the radiopharmaceutical containers and pigs can be communicated from a supplier computer **116** to computer **79** within the hospital **42** via a communications link **118**, for example, an Internet connection, a telephonic connection, or other suitable link.

In methods as contemplated herein, RF tags **60** may be applied to a radioactive pharmaceutical container **20c** that is

subsequently placed in a lead lined pig **33**. In such a circumstance, the pig limits the usability of the RF tags **60** and may prevent use thereof unless the container **20c** is removed from the pig **33**. Therefore, it would be highly desirable to be able to read data from, and write data to, the RF tag **60** on the radiopharmaceutical container **20c** when it is stored inside the pig **33**. Such is achieved by an exemplary embodiment of a pig-mounted antenna system shown in FIGS. **20-22**.

Referring to FIG. **20**, in a first embodiment, a radiopharmaceutical pig **33b** has an elongated base **322** and an elongated cap **324**. The base **322** and cap **324** can be formed in any of a wide variety of shapes and sizes, however, a substantially cylindrical shape is illustrated. The cap **324** is joined to the base **322** by a threaded interconnection **325** in a known manner. A cap shielding element **326** within the cap **324** and a base shielding element **328** within the base **322** are used to block radiation that may be emitted from the radiopharmaceutical within a syringe **20c**. The shielding elements **326**, **328** can be formed from any material that is effective to block radiation, for example, lead, tungsten, a filled polymer composite material, etc. The cap shielding element **326** forms a protrusion **329** that overlaps the base shielding element **328** when the cap **324** is mounted on the base **322**. This overlap of the shields **326**, **328** facilitates a blockage of radiation through a discontinuity in the shields caused by the cap **324** being separable from the base **322**.

The cap **324** further has a cap shell **330** comprised of an outer shell portion **332** and an inner shell portion **334**. Similarly, the base **322** has a cap shell **336** comprised of an outer shell portion **338** and an inner shell portion **340**. The base and cap shells **328**, **330** are made from a plastic material, for example, a polycarbonate resin, etc.

A label **30** is affixed to the radiopharmaceutical syringe **22c** by known means, for example, an adhesive, tape, elastic bands, etc. Indeed, the label **30** may be affixed to the radiopharmaceutical syringe **20c** in any appropriate manner (e.g., so that it is not easily removable). The label **30** contains indicia **346** that is in human readable and/or machine readable form. The label **30** further has an RFID tag **60** that comprises an RFID integrated circuit chip **212** and at least one radio frequency antenna **210**. The radiopharmaceutical syringe **20c** is often manufactured at a facility independent of the healthcare facility where it is to be used. Therefore, data relating to the radiopharmaceutical syringe **20c** is often collected at the point of its manufacture. Further, additional data is often collected at different points in a distribution channel at which the radiopharmaceutical pig **33b** containing the radiopharmaceutical syringe **20c** is handled. Data is also collected upon the radiopharmaceutical syringe **20c** being used and thereafter, upon its disposal or cleaning for an authorized reuse. Thus, over the life of the radiopharmaceutical syringe **20c** and associated radiopharmaceutical pig **33b**, data that can be written into the RF ID tag **60** at different times in the life cycle of the syringe **20c** has been previously described. Such data includes but is not limited to the decay factor for a radiopharmaceutical (e.g., half life of pharmaceutical), its prescribed activity level (curie level of radiation) at injection time, the activity level at another time (such as filling time), and/or the time at which the preparing physician or radiopharmacist assumed the radiopharmaceutical would be injected. The activity level is a function of time due to the short half life of most radiopharmaceuticals, so the activity level is designed for a specific injection time.

In order to obtain a maximum benefit from the data stored within the RFID tag **60**, it is necessary to be able to read the tag when the radiopharmaceutical syringe **20c** is housed within the radiopharmaceutical pig **33b**. In the embodiment

of FIG. 20, at least one radio frequency inner antenna 358 is applied over an inner surface of the inner base shell 340; and at least one radio frequency outer antenna 364 is applied over an outer surface of the outer base shell 338. A hole 360 extends through the inner base shell 340, the base shield 328, and the outer base shell 338. At least one connecting lead 362, for example, a copper wire lead, extends through the hole 360 and has one end connected to the inner antenna 358 and an opposite end connected to the outer antenna 364.

The inner antenna 358 is designed to couple with the RFID antenna 210 connected to the RFID chip 212. The outer antenna 364 is designed to electromagnetically couple with a read/write (“R/W”) device 366 in the same way that the RFID antenna 210 would couple with the R/W device 366. The R/W device 366 is connected to a computer 368 in a known manner. The R/W device 366 electromagnetically couples with the RFID antenna 210 via the inner and outer antennas 358, 364 respectively. Therefore, any time the radiopharmaceutical pig 33b is handled in its life cycle, the R/W device 366 can be used to read information from, and/or write information to, the RFID chip 212 of the RFID tag 60 on the radiopharmaceutical syringe 20c via an RFID antenna system comprising the antennas 210, 358, 362, 364. It should be noted that the antenna may simply comprise leads of a sufficient length to be used as an RFID antenna, in which case there may not be a coiled antenna section 364.

Another exemplary embodiment of a radiopharmaceutical pig 33b and radiopharmaceutical syringe 20c utilizing the RFID tag 60 is shown in FIG. 21. In this embodiment, inner and outer antennas 358, 364 are located on respective inner and outer surfaces 370, 372 of a top of the cap 324. The antennas 358, 364 are electrically connected by at least one lead 362 extending through a hole 374 in the top of the cap 324. The R/W device 366 is able to electromagnetically couple with the RFID antenna 210 via the inner and outer antennas 358, 364 respectively. Therefore, at any time the radiopharmaceutical pig 33b is handled in its life cycle, the R/W device 366 can be used to read information from, and/or write information to, the RFID chip 212 of the RFID tag 60 on the radiopharmaceutical syringe 20c via an RFID antenna system comprising the antennas 210, 358, 364.

Placing the antennas 358, 362 in the top of the cap 324 has some advantages. First, the top of the cap 324 often experiences less radiation exposure than the base shell 336. Further, the cap outer surface 372 often experiences less physical contact than the base outer shell 338 during the handling of the radiopharmaceutical pig 33b; and hence, the outer antenna 362 on the cap outer surface 372 is less subject to physical damage.

A further exemplary embodiment of a radiopharmaceutical pig 33b and radiopharmaceutical syringe 20c utilizing an RFID tag 60 is shown in FIGS. 22 and 22A. In this embodiment, the RFID tag 60 has an RFID chip 212 on a first portion of a label 30c that is attached to the radiopharmaceutical syringe 20c in a manner described earlier with respect to FIG. 20. A second portion of the label 30d is located outside of the radiopharmaceutical pig 33b and has at least one RFID antenna 210 thereon. The RFID chip 212 on the first label portion 30c is electrically connected to the antenna 210 by at least one electrically conductive lead 376 integral with a tether 378. The conductive lead 376 and tether 378 may be formed from any materials that provide the desired electrical and mechanical properties, for example, an insulated or un-insulated copper wire, a copper trace laminated on a substrate, etc. The threaded connector 325 is designed to provide a clearance for the conductive lead 376 and tether 378, so that the cap 324 can be attached and removed from the base 322

without damaging the conductive lead 376 and tether 378. The R/W device 366 is able to electromagnetically couple with the RFID antenna 210, and the RFID antenna 210 communicates data to and from the RFID chip 212 via the conductive lead 376. Therefore, at any time the radiopharmaceutical pig 33b is handled in its life cycle, the R/W device 366 can be used to read information from, and/or write information to, the RFID chip 212 of the RFID tag 60 on the radiopharmaceutical syringe 20c via an RFID antenna system comprising the antenna 210 and conductive lead 376.

In use, upon receiving an order for a radiopharmaceutical, a label 30 having an RFID chip 212 and associated antenna 210 is applied to the radiopharmaceutical syringe 20c, and the radiopharmaceutical syringe 20c can be placed in a radiopharmaceutical pig 33b. At that time, data including but not limited to the identity of the syringe and pig can be written to the RFID tag 60 in a manner previously described with respect to FIGS. 1A and 1B. The radiopharmaceutical syringe 20c and pig 33b are then transported to a location where the syringe 20c is filled with a desired radiopharmaceutical. This location may be at a radiopharmaceutical supplier or a location of a user of the radiopharmaceutical syringe 20c. In either event, regardless of where the radiopharmaceutical syringe 20c is filled, as previously described, data can be entered into the RFID tag 60 relating to the filling process, the radiopharmaceutical being filled, and the how the radiopharmaceutical is to be used. After being filled, the pig 33b holding the syringe 20c filled with the radiopharmaceutical may be transported and stored several times before it is delivered for use in a preparation and/or imaging room. During use, the syringe 20c is removed from the pig 33b, and the radiopharmaceutical is injected into an examination subject or patient. After use, the empty syringe 20c is placed back in the pig 33b and returned to the pharmaceutical supplier or other location for proper disposal of the radiopharmaceutical syringe 20c and reconditioning of the radiopharmaceutical pig 33b for reuse.

Every time the radiopharmaceutical pig 33b and/or radiopharmaceutical syringe 20c is handled over their respective life cycles, in a manner as previously described, an R/W device 366 can be used to read data from, and/or write data to, the RFID tag 60, thereby providing complete chronological history of the radiopharmaceutical pig 33b and syringe radiopharmaceutical 20c over the respective life cycles. The systems illustrated in FIGS. 1A, 3A, 1B, 3B have an advantage in that almost any information is able to be transferred between all entities involved in a life cycle of a syringe 20, which is any entity that can communicate with the communication link 80. Therefore, data available from a website on the internet 83 can be utilized during the life cycle of the syringe 20. Such internet communications capabilities permits remote service of a power injector 50, downloading of an injection protocol, communication with a remotely located physician, media supplier or other entity of interest and other functions.

While the various principles of the invention have been illustrated by way of describing various exemplary embodiments, and while such embodiments have been described in considerable detail, there is no intention to restrict, or in any way limit, the scope of the appended claims to such detail. Additional advantages and modifications will readily appear to those skilled in the art. For example, in the described embodiments of FIGS. 20-22, an RFID chip 212 may be positioned inside the pig. In some embodiments, the chip 212 may be located outside the pig along with an associated antenna, and the chip may be physically attached to the syringe 20c by a string or other attachment so that the radiopharmaceutical syringe 20c and RFID information therein remain associated. Alternatively, the pig 33b may carry an

RFID tag and antenna with no mechanical attachment to the syringe, but it may simply be known that the data therein relates to the syringe that is in the pig.

Further, in the exemplary embodiments shown and described herein, the antenna systems **229a**, **229b** use one, two and three antenna loops; however, in alternative embodiments, any number of antenna loops may be used. The antenna loops may be configured in any shape and be in the same plane or in different planes. Further, the antenna loops may or may not be overlapping. In may, however, be preferable that the antenna loops be individually tuned to resonate at a specific frequency used by the RFID protocol. Further, in the described embodiment, a switching circuit **241b** is located on the same PC board **102** as an RF driver circuit **224b**; however, in alternative embodiments, a switching circuit may be located on the second PC board **103**, be split between the two PC boards **102**, **103** or located elsewhere, for example, with the power injector as shown in FIG. **17**.

In addition, in the described embodiments, the R/W antenna systems **229a**, **229b** are applied to a pharmaceutical injection assembly; however, in alternative embodiments, the R/W antenna systems **229a**, **229b** utilizing multiple nonparallel antennas may be applied to any devices that support a medical fluid container. Such devices include but are not limited to a warmer oven or warming box, a container filling station, a pig or other nuclear medicine container, a dose calibration station, a handheld powered medical fluid dispenser, a syringe disposal station, or other device.

The systems of the described embodiments relate to containers of medical fluids. Two examples described in detail relate to contrast media and respective syringes and radiopharmaceuticals and respective containers. In alternative embodiments, referring to FIG. **1C**, the container may be an IV bag **130** filled with a medical fluid. Tubing **132** from the IV bag **130** may interface with an infusion pump **134** so that a flow of medical fluid from the IV bag **130** may be regulated via use of the pump **134**. While one end of the tubing **132** is generally associated with the IV bag **130**, the other end of the tubing **132** may be connected to a patient in a known manner. The IV bag **130** may have a label **30** with a data tag **60** as previously described herein, for example, an RFID tag. Further, the infusion pump **134** may be in electrical communication with an electromagnetic device capable of reading data from and/or writing data to the data tag **60** of the IV bag **130**. For example, the electromagnetic device may be attached to and/or located within the infusion pump **134**. As shown in FIG. **3C**, the infusion pump **134** may have a control **136** connected to the communications link **80** in a manner similar to that described with respect to the injector control **93** shown in FIGS. **1A** and **1B**. Thus, the systems of FIGS. **1C** and **3C** may permit activity relating to the IV bag **130**, the medical fluid therein, and/or the infusion pump **134** to be tracked and recorded (e.g., over a life cycle of the IV bag **130**).

There are many known structures for mounting a syringe to a power injector, and the faceplates shown and described herein are only two such structures. Other mounting structures may not permit removal from the power head. The inventions claimed herein can be applied to power heads having any type of structure for mounting a syringe thereto. In the shown and described embodiment, a heater **106** is mounted on the PC boards **102**, **103**; however, in alternative embodiments, the heater **106** may not be used and therefore, deleted from PC boards **102**, **103**.

When introducing elements of the present invention or various embodiments thereof, the articles “a”, “an”, “the”, and “said” are intended to mean that there are one or more of the elements. The terms “comprising”, “including”, and

“having” are intended to be inclusive and mean that there may be additional elements other than the listed elements. Moreover, the use of “top” and “bottom”, “front” and “rear”, “above” and “below” and variations of these and other terms of orientation is made for convenience, but does not require any particular orientation of the components. Therefore, the invention, in its broadest aspects, is not limited to the specific details shown and described herein. Consequently, departures may be made from the details described herein without departing from the spirit and scope of the claims, which follow.

What is claimed is:

1. A medical fluid injection assembly for use with a container having medical fluid disposed therein, the container having a data tag readable by electromagnetic signals, and the assembly comprising:

an electromagnetic device adapted to read data from and/or write data to the data tag while the container is supported on the medical fluid injection assembly, the electromagnetic device comprising:

a first antenna loop forming one side of a V-shape and being tuned to a radio frequency, and

a second antenna loop forming an opposing side of the V-shape and being tuned to the radio frequency; and

an electronic medical fluid administration device supporting the container at a substantially fixed location relative to said antenna loops and operable to inject the medical fluid into a patient.

2. A medical fluid injection assembly for use with a container having medical fluid disposed therein, the container having a data tag readable by electromagnetic signals, and the assembly comprising:

an electromagnetic device adapted to read data from and/or write data to the data tag while the container is supported on the medical fluid injection assembly, the electromagnetic device comprising

a first antenna loop comprising a signal lead and a ground lead,

a first tuning circuit comprising an input and an output, the output of the first tuning circuit connected to the signal lead of the first antenna loop, and the first tuning circuit tuning the first antenna loop to a radio frequency,

a second antenna loop comprising a signal lead and a ground lead, and

a second tuning circuit comprising an input and an output, the output of the second tuning circuit connected to the signal lead of the second antenna loop, and the second tuning circuit tuning the second antenna loop to the radio frequency; and

an electronic medical fluid administration device supporting the container at a substantially fixed location relative to said antenna loops, and operable to inject the medical fluid into a patient.

3. The assembly of claim **2** wherein the second antenna loop is nonparallel with the first antenna loop and forms an angle of less than 180 degrees with the first antenna loop.

4. The assembly of claim **2** further comprising:

a third antenna loop comprising a signal lead and a ground lead, and

a third tuning circuit comprising an input and an output, the output of the third tuning circuit connected to the signal lead of the third antenna loop, and the third tuning circuit tuning the third antenna loop to the radio frequency.

5. The assembly of claim **2** wherein the device is mounted in association with the administration device.

6. The assembly of claim 5 wherein the administration device comprises a mount comprising:

a first printed circuit board supporting the first antenna loop and the first tuning circuit; and

a second printed circuit board supporting the second antenna loop and the second tuning circuit.

7. The assembly of claim 6 wherein the first printed circuit board forms an angle of less than about 180 degrees with the second printed circuit board.

8. The assembly of claim 7 wherein the first printed circuit board supports a driver circuit electrically connectable to the first antenna loop, the second antenna loop, the first tuning circuit, and the second tuning circuit.

9. The assembly of claim 8 wherein the driver circuit comprises a power terminal and a ground terminal.

10. The assembly of claim 9 wherein the input of the first tuning circuit is connected to the power terminal, the ground lead of the first antenna loop is connected to the ground terminal, the input of the second tuning circuit is not connected to the power terminal or the ground terminal, and the ground lead of the second antenna loop is connected to the ground terminal.

11. The assembly of claim 9 wherein the input of the first tuning circuit is not connected to the power terminal or the ground terminal, the ground lead of the first antenna loop is connected to the ground terminal, the input of the second tuning circuit is connected to the power terminal, and the ground lead of the second antenna loop is connected to the ground terminal.

12. The assembly of claim 9 wherein the input of the first tuning circuit is connected to the power terminal, the ground lead of the first antenna loop is connected to the ground terminal, the input of the second tuning circuit is connected to the ground terminal, and the ground lead of the second antenna loop is connected to the ground terminal.

13. The assembly of claim 9 wherein the input of the first tuning circuit is connected to the ground terminal, the ground lead of the first antenna loop is connected to the ground terminal, the input of the second tuning circuit is connected to the power terminal, and the ground lead of the second antenna loop is connected to the ground terminal.

14. The assembly of claim 9 further comprising a switching circuit comprising:

a first switch comprising

a first contact connected to the input of the first tuning circuit,

a second contact connected to the ground terminal,

a third contact connected to the power terminal, and

a fourth contact not connected to the ground terminal or the power terminal, the first switch being operable to electrically connect the first contact with one of the second contact, the third contact and the fourth contact; and

a second switch comprising

a fifth contact connected to the input of the second tuning circuit,

a sixth contact connected to the ground terminal,

a seventh contact connected to the power terminal, and

an eighth contact not connected to the ground terminal or the power terminal, the second switch being operable to

electrically connect the fifth contact with one of the sixth contact, the seventh contact and the eighth contact.

15. A medical fluid administration device for use with a medical fluid container containing at least one of a metallic and a diamagnetic fluid material, the medical fluid container having a data tag for storing data relating to the fluid material, the administration device comprising:

an electromagnetic device adapted to read data from, and/or write data to, the data tag, the electromagnetic device while the container is supported on the medical fluid administration device, comprising

a first antenna loop forming one side of a V-shape and being tuned to the radio frequency, and

a second antenna loop forming an opposing side of the V-shape and being tuned to the radio frequency; and

the medical fluid administration device supporting the container at a substantially fixed location relative to said antenna loops and operable to inject the at least one of the metallic and the diamagnetic fluid into a patient.

16. A method of using a medical fluid injection assembly comprising an electromagnetic device operable to read data from and/or write data to a data tag, the data tag being readable by electromagnetic signals and being associated with a container having medical fluid disposed therein, the method comprising:

electrically connecting first and second antenna loops of the device in a first circuit configuration, wherein the first and second antenna loops are tuned to a substantially identical radio frequency, and the first antenna loop is oriented in a nonparallel relationship with the second antenna loop; and

attempting an RF communication between the device and the data tag, wherein the attempting comprises providing RF power to the first circuit configuration;

during the attempting, supporting the container on an electronic medical fluid administration device at a substantially fixed location relative to said antenna loops; and injecting the medical fluid into a patient by the electronic medical fluid administration device.

17. The method of claim 16 further comprising: determining whether or not RF communication is or was established between the device and the data tag;

if determined unsuccessful, electrically reconnecting the first and second antenna loops in a further circuit configuration different from the first circuit configuration; and

attempting another RF communication between the device and the data tag, wherein the attempting another RF communication comprises providing RF power to the further circuit configuration.

18. The method of claim 17 further comprising iterating the determining, the electrically reconnecting, and the attempting another RF communication until determining that a successful RF communication has been established between the device and the data tag.

19. The method of claim 18 wherein the first and second antenna loops, together, exhibit a V-shape.